

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BP024486.D Time Analyzed: 10:37

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	147752	7.711	596585	10.48	373130	14.35
UPPER LIMIT	295504	8.211	1193170	10.981	746260	14.846
LOWER LIMIT	73876	7.211	298293	9.981	186565	13.846
EPA SAMPLE NO.						
01 WC-13-S-202504	207548	7.71	829684	10.49	528997	14.35
02 PB167805TB	147117	7.71	562398	10.48	333420	14.35
03 WC-12-S-202504	159742	7.71	610644	10.48	349744	14.35
04 WC-12-A-202504	147765	7.71	578416	10.48	346607	14.34

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.