

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO02

Lab Code: CHEM

SAS No.: Q1913 SDG NO.: Q1913

Lab File ID: BF142249.D

DFTPP Injection Date: 05/01/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.4
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	31.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	42.4
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142250.D	05/01/2025	10:17
PB167810BL	PB167810BL	BF142254.D	05/01/2025	12:10
PB167810BS	PB167810BS	BF142255.D	05/01/2025	12:39
PB167774TB	PB167774TB	BF142256.D	05/01/2025	13:07
WC-13-A-202504	Q1913-04	BF142260.D	05/01/2025	15:06