

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO02

Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG No.: Q1913

Instrument ID: MSVOA_X Calibration Date/Time: 04/30/2025 10:01

Lab File ID: VX045985.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.703	0.712		1.28	20
1,1-Dichloroethene	0.600	0.646		7.67	20
2-Butanone	0.550	0.628		14.18	20
Carbon Tetrachloride	0.518	0.632		22.01	20
Chloroform	1.306	1.474		12.86	20
Benzene	1.463	1.629		11.35	20
1,2-Dichloroethane	0.604	0.716		18.54	20
Trichloroethene	0.348	0.381		9.48	20
Tetrachloroethene	0.353	0.398		12.75	20
Chlorobenzene	1.068	1.246	0.3	16.67	20
1,2-Dichloroethane-d4	0.914	0.976		6.78	20
Dibromofluoromethane	0.355	0.405		14.09	20
Toluene-d8	1.238	1.287		3.96	20
4-Bromofluorobenzene	0.451	0.514		13.97	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.