



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050225\
Data File : VX046012.D
Acq On : 02 May 2025 08:56
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.4	6961	PASS
75	95	30	60	57.6	17110	PASS
95	95	100	100	100.0	29717	PASS
96	95	5	9	6.7	2004	PASS
173	174	0.00	2	0.8	173	PASS
174	95	50	100	68.5	20357	PASS
175	174	5	9	7.6	1553	PASS
176	174	95	101	95.8	19511	PASS
177	176	5	9	6.6	1287	PASS