

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: SW8260-Low

Datafile : VX045988.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0430WBS01	Vinyl chloride	20	17.3	ug/L	86			65	117	
	1,1-Dichloroethene	20	18.5	ug/L	93			74	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	21.8	ug/L	109			77	113	
	Chloroform	20	21.6	ug/L	108			79	113	
	Benzene	20	20.2	ug/L	101			82	109	
	1,2-Dichloroethane	20	22.8	ug/L	114			80	115	
	Trichloroethene	20	19.9	ug/L	100			77	113	
	Tetrachloroethene	20	19.0	ug/L	95			67	123	
	Chlorobenzene	20	21.3	ug/L	106			82	109	