

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1913

Client: AECOM

Analytical Method: SW8260D

Datafile : VX046016.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0502WBS01	Vinyl chloride	20	15.3	ug/L	77			65	117	
	1,1-Dichloroethene	20	16.6	ug/L	83			74	110	
	2-Butanone	100	96.6	ug/L	97			65	122	
	Carbon Tetrachloride	20	19.2	ug/L	96			77	113	
	Chloroform	20	19.5	ug/L	98			79	113	
	Benzene	20	18.1	ug/L	91			82	109	
	1,2-Dichloroethane	20	19.7	ug/L	99			80	115	
	Trichloroethene	20	17.9	ug/L	90			77	113	
	Tetrachloroethene	20	18.6	ug/L	93			67	123	
	Chlorobenzene	20	19.2	ug/L	96			82	109	