

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO02
 Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913
 Lab File ID: VX046013.D Date Analyzed: 05/02/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	88170	5.54	150204	6.75	131792	10.05
UPPER LIMIT	176340	6.044	300408	7.251	263584	10.549
LOWER LIMIT	44085	5.044	75102	6.251	65896	9.549
EPA SAMPLE NO.						
WC-13-A-202504RE	65405	5.54	123884	6.76	113310	10.05
VX0502WBL01	66865	5.54	129422	6.76	121808	10.05
VX0502WBS01	77025	5.55	135229	6.76	118203	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.