

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO02
 Lab Code: CHEM Case No.: Q1913 SAS No.: Q1913 SDG NO.: Q1913
 Lab File ID: VX046013.D Date Analyzed: 05/02/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63914	12.018				
UPPER LIMIT	127828	12.518				
LOWER LIMIT	31957	11.518				
EPA SAMPLE NO.						
WC-13-A-202504RE	47494	12.02				
VX0502WBL01	52200	12.02				
VX0502WBS01	56323	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.