

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1914

Client: CDM Smith

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1914-11MSD	Client Sample ID:	SS-8MSD					DataFile:	BM050084.D		
Naphthalene	1800	2000	3300	ug/Kg	72		7		72	110	20
Acenaphthylene	1800	0	1400	ug/Kg	78	*	8		79	118	20
Acenaphthene	1800	190	1500	ug/Kg	73		9		70	121	20
Fluorene	1800	0	1700	ug/Kg	94		5		68	116	20
Phenanthrene	1800	1400	2500	ug/Kg	61		0		52	128	20
Anthracene	1800	220	1700	ug/Kg	82		6		62	124	20
Fluoranthene	1800	2000	3100	ug/Kg	61		0		44	125	20
Pyrene	1800	1800	2700	ug/Kg	50		13		26	142	20
Benzo(a)anthracene	1800	930	2300	ug/Kg	76		7		71	114	20
Chrysene	1800	860	2200	ug/Kg	74		7		57	121	20
Benzo(b)fluoranthene	1800	980	2200	ug/Kg	68		0		67	121	20
Benzo(k)fluoranthene	1800	330	1700	ug/Kg	76		0		57	134	20
Benzo(a)pyrene	1800	770	2100	ug/Kg	74		0		70	142	20
Indeno(1,2,3-cd)pyrene	1800	420	1900	ug/Kg	82		0		40	129	20
Dibenz(a,h)anthracene	1800	140	1700	ug/Kg	87		7		43	123	20
Benzo(g,h,i)perylene	1800	490	2000	ug/Kg	84		0		24	125	20