

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1914

Client: CDM Smith

Analytical Method: 8270E

DataFile: BF142253.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Limits		
							Qual	Qual	Low	High	RPD
PB167798BSD	Naphthalene	50	39.8	ug/L	80	1			64	107	20
	Acenaphthylene	50	40.8	ug/L	82	1			79	103	20
	Acenaphthene	50	44.0	ug/L	88	0			59	113	20
	Fluorene	50	40.7	ug/L	81	1			64	107	20
	Phenanthrene	50	40.7	ug/L	81	2			62	109	20
	Anthracene	50	41.3	ug/L	83	1			65	110	20
	Fluoranthene	50	41.8	ug/L	84	0			64	110	20
	Pyrene	50	43.0	ug/L	86	0			71	103	20
	Benzo(a)anthracene	50	41.3	ug/L	83	0			62	107	20
	Chrysene	50	43.0	ug/L	86	1			61	108	20
	Benzo(b)fluoranthene	50	39.3	ug/L	79	12			77	113	20
	Benzo(k)fluoranthene	50	44.5	ug/L	89	11			77	105	20
	Benzo(a)pyrene	50	42.6	ug/L	85	0			72	131	20
	Indeno(1,2,3-cd)pyrene	50	42.8	ug/L	86	1			72	105	20
	Dibenz(a,h)anthracene	50	43.0	ug/L	86	2			78	115	20
	Benzo(g,h,i)perylene	50	43.1	ug/L	86	3			75	118	20