

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BM050062.D Time Analyzed: 13:25

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	329367	7.757	1090200	10.55	636680	14.40
UPPER LIMIT	658734	8.257	2180400	11.045	1273360	14.898
LOWER LIMIT	164684	7.257	545100	10.045	318340	13.898
EPA SAMPLE NO.						
01 SS-4	266849	7.75	987063	10.54	682416	14.39
02 SS-1	239935	7.75	828165	10.54	540888	14.39
03 SS-91	241297	7.75	826545	10.54	542720	14.39
04 SS-2	275696	7.75	997778	10.54	688876	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.