

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q1914

SDG NO.: Q1914

Lab File ID: BF142238.D

DFTPP Injection Date: 04/30/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.6
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	27.3
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	37.1
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.3
275	10.0 - 60.0% of mass 198	22.8
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.9 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142239.D	04/30/2025	11:24
SSTDICC005	SSTDICC005	BF142240.D	04/30/2025	11:52
SSTDICC010	SSTDICC010	BF142241.D	04/30/2025	12:20
SSTDICC020	SSTDICC020	BF142242.D	04/30/2025	12:49
SSTDICCC040	SSTDICCC040	BF142243.D	04/30/2025	13:17
SSTDICC050	SSTDICC050	BF142244.D	04/30/2025	13:46
SSTDICC060	SSTDICC060	BF142245.D	04/30/2025	14:15
SSTDICC080	SSTDICC080	BF142246.D	04/30/2025	14:43