

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q1914 SDG NO.: Q1914

Lab File ID: BM050078.D

DFTPP Injection Date: 05/02/2025

Instrument ID: BNA_M

DFTPP Injection Time: 09:54

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	20.9
68	Less than 2.0% of mass 69	0.4 (1.5) 1
69	Mass 69 relative abundance	25.2
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	32.5
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	26.7
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	12.2
442	Greater than 50% of mass 198	77.4
443	15.0 - 24.0% of mass 442	15.1 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050079.D	05/02/2025	10:33
PB167803BL	PB167803BL	BM050080.D	05/02/2025	11:13
PB167803BS	PB167803BS	BM050081.D	05/02/2025	11:52
SS-8	Q1914-09	BM050082.D	05/02/2025	12:41
SS-8MS	Q1914-10MS	BM050083.D	05/02/2025	13:20
SS-8MSD	Q1914-11MSD	BM050084.D	05/02/2025	14:00
SS-7	Q1914-08	BM050089.D	05/02/2025	17:17