

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM050225

Review By	Rahul	Review On	5/5/2025 11:31:05 AM
Supervise By	Jagrut	Supervise On	5/5/2025 11:51:36 AM
SubDirectory	BM050225	HP Acquire Method	BNA_M
		HP Processing Method	BM042825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050078.D	02 May 2025 09:54		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM050079.D	02 May 2025 10:33		RC/JU	Ok
3	PB167803BL	PB167803BL	BM050080.D	02 May 2025 11:13		RC/JU	Ok
4	PB167803BS	PB167803BS	BM050081.D	02 May 2025 11:52		RC/JU	Ok
5	Q1914-09	SS-8	BM050082.D	02 May 2025 12:41		RC/JU	Ok,M
6	Q1914-10MS	SS-8MS	BM050083.D	02 May 2025 13:20		RC/JU	Ok,M
7	Q1914-11MSD	SS-8MSD	BM050084.D	02 May 2025 14:00		RC/JU	Ok,M
8	Q1914-12	SS-9	BM050085.D	02 May 2025 14:39	Internal Standard Fail	RC/JU	Not Ok
9	Q1914-06	SS-5	BM050086.D	02 May 2025 15:19	Internal Standard Fail	RC/JU	Not Ok
10	Q1914-04	SS-3	BM050087.D	02 May 2025 15:58	Internal Standard Fail	RC/JU	Not Ok
11	Q1907-01DL	CO-008R-WCDL	BM050088.D	02 May 2025 16:38	Internal Standard Fail	RC/JU	Not Ok
12	Q1914-08	SS-7	BM050089.D	02 May 2025 17:17		RC/JU	Ok,M
13	Q1922-01	MH-R	BM050090.D	02 May 2025 17:57	Internal Standard Fail	RC/JU	ReRun
14	Q1906-09	WC-6	BM050091.D	02 May 2025 18:36	Internal Standard Fail	RC/JU	ReRun
15	Q1935-05	MH-MM	BM050092.D	02 May 2025 19:15	Internal standard fail	RC/JU	ReRun
16	Q1923-01	TR-05-04302025	BM050093.D	02 May 2025 19:55		RC/JU	Ok,NS
17	Q1933-01	OK-01-050125	BM050094.D	02 May 2025 20:34	Internal standard fail	RC/JU	ReRun

M : Manual Integration