

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG No.: Q1914

Instrument ID: BNA_F Calibration Date/Time: 05/01/2025 10:17

Lab File ID: BF142250.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.252	1.250		-0.2	
Phenol-d6	1.512	1.575		4.2	
Nitrobenzene-d5	0.304	0.333		9.5	
Naphthalene	0.997	1.030		3.3	
2-Fluorobiphenyl	1.418	1.409		-0.6	
Acenaphthylene	1.952	1.996		2.3	
Acenaphthene	1.140	1.185		3.9	20.0
Fluorene	1.299	1.322		1.8	
2,4,6-Tribromophenol	0.182	0.196		7.7	
Phenanthrene	1.080	1.116		3.3	
Anthracene	1.105	1.142		3.3	
Fluoranthene	1.068	1.147		7.4	20.0
Pyrene	1.845	1.938		5.0	
Terphenyl-d14	1.375	1.422		3.4	
Benzo (a) anthracene	1.343	1.378		2.6	
Chrysene	1.222	1.297		6.1	
Benzo (k) fluoranthene	1.136	1.240		9.2	
Benzo (a) pyrene	1.137	1.193		4.9	20.0
Indeno (1,2,3-cd) pyrene	1.465	1.533		4.6	
Dibenzo (a,h) anthracene	1.189	1.258		5.8	
Benzo (g,h,i) perylene	1.195	1.250		4.6	

All other compounds must meet a minimum RRF of 0.010.