

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1914

Client: CDM Smith

Analytical Method: 8270E

DataFile: BF142252.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167798BS	Naphthalene	50	40.4	ug/L	81				64	107	
	Acenaphthylene	50	41.2	ug/L	82				79	103	
	Acenaphthene	50	43.9	ug/L	88				59	113	
	Fluorene	50	41.1	ug/L	82				64	107	
	Phenanthrene	50	40.0	ug/L	80				62	109	
	Anthracene	50	40.9	ug/L	82				65	110	
	Fluoranthene	50	41.9	ug/L	84				64	110	
	Pyrene	50	42.9	ug/L	86				71	103	
	Benzo(a)anthracene	50	41.2	ug/L	82				62	107	
	Chrysene	50	43.6	ug/L	87				61	108	
	Benzo(k)fluoranthene	50	39.7	ug/L	79				77	105	
	Benzo(a)pyrene	50	42.5	ug/L	85				72	131	
	Indeno(1,2,3-cd)pyrene	50	42.2	ug/L	84				72	105	
	Dibenz(a,h)anthracene	50	42.3	ug/L	85				78	115	
	Benzo(g,h,i)perylene	50	42.0	ug/L	84				75	118	