

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG No.: Q1914

Instrument ID: MSVOA_X Calibration Date/Time: 04/30/2025 10:01

Lab File ID: VX045985.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.463	1.629		11.35	20
Toluene	0.885	1.002		13.22	20
Ethyl Benzene	1.914	2.259		18.02	20
m/p-Xylenes	0.696	0.828		18.97	20
o-Xylene	0.686	0.796		16.03	20
Isopropylbenzene	4.005	4.466		11.51	20
n-propylbenzene	4.613	5.277		14.39	20
1,3,5-Trimethylbenzene	3.312	3.775		13.98	20
tert-Butylbenzene	3.279	3.730		13.75	20
1,2,4-Trimethylbenzene	3.324	3.798		14.26	20
sec-Butylbenzene	4.027	4.708		16.91	20
p-Isopropyltoluene	3.320	3.912		17.83	20
n-Butylbenzene	2.879	3.538		22.89	20
1,2-Dichloroethane-d4	0.914	0.976		6.78	20
Dibromofluoromethane	0.355	0.405		14.09	20
Toluene-d8	1.238	1.287		3.96	20
4-Bromofluorobenzene	0.451	0.514		13.97	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.