

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN050525

Review By	John Carlone	Review On	5/6/2025 10:01:09 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:40:44 PM
SubDirectory	VN050525	HP Acquire Method	HP Processing Method 82N041525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133808
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133809,VP133810

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086477.D	05 May 2025 10:23		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086478.D	05 May 2025 11:32	pH#Lot#V12668	JC\MD	Ok,M
3	VN0505MBL01	VN0505MBL01	VN086479.D	05 May 2025 12:13		JC\MD	Ok
4	VN0505WBL01	VN0505WBL01	VN086480.D	05 May 2025 12:38		JC\MD	Ok
5	VN0505WBS01	VN0505WBS01	VN086481.D	05 May 2025 13:02		JC\MD	Ok,M
6	VN0505WBSD01	VN0505WBSD01	VN086482.D	05 May 2025 13:36		JC\MD	Ok,M
7	IBLK	IBLK	VN086483.D	05 May 2025 14:00		JC\MD	Ok
8	Q1942-01	STORAGE-BLANK-SO	VN086484.D	05 May 2025 14:24	vial A pH<2	JC\MD	Ok
9	Q1942-02	STORAGE-BLANK-WA	VN086485.D	05 May 2025 14:48	vial A pH<2	JC\MD	Ok
10	Q1942-03	STORAGE-BLANK-WA	VN086486.D	05 May 2025 15:12	vial A pH<2	JC\MD	Ok
11	Q1942-04	STORAGE-BLANK-SAI	VN086487.D	05 May 2025 15:36	vial A pH<2	JC\MD	Ok
12	Q1956-07	FB-05022025	VN086488.D	05 May 2025 16:00	vial A pH<2 FB	JC\MD	Ok
13	Q1929-05RE	WC-A1-03-GRE	VN086489.D	05 May 2025 16:25	vial B pH#5.0 Surrogate Fail	JC\MD	Confirms
14	Q1929-09RE	WC-A1-04-GRE	VN086490.D	05 May 2025 16:49	vial B pH#5.0 Surrogate Fail	JC\MD	Confirms
15	VN0505MBS01	VN0505MBS01	VN086491.D	05 May 2025 17:13		JC\MD	Ok,M
16	VN0505MBSD01	VN0505MBSD01	VN086492.D	05 May 2025 17:37		JC\MD	Ok,M
17	Q1914-06ME	SS-5ME	VN086493.D	05 May 2025 18:01	run straight meoh	JC\MD	Not Ok
18	Q1914-07MEDL	SS-6MEDL	VN086494.D	05 May 2025 18:25		JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN050525

Review By	John Carlone	Review On	5/6/2025 10:01:09 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:40:44 PM
SubDirectory	VN050525	HP Acquire Method	HP Processing Method 82N041525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133808		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133809,VP133810		

19	Q1914-08DL	SS-7DL	VN086495.D	05 May 2025 18:49		JC\MD	Ok
20	Q1914-09MEDL	SS-8MEDL	VN086496.D	05 May 2025 19:14		JC\MD	Ok
21	Q1914-12	SS-9	VN086497.D	05 May 2025 19:38	run soil	JC\MD	Not Ok
22	IBLK	IBLK	VN086498.D	05 May 2025 20:02		JC\MD	Ok
23	Q1939-03	GB2	VN086499.D	05 May 2025 20:26	Need Straight Run	JC\MD	Not Ok
24	Q1939-06	GB3B	VN086500.D	05 May 2025 20:50	vial C	JC\MD	Ok
25	IBLK	IBLK	VN086501.D	05 May 2025 21:14		JC\MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VN086502.D	05 May 2025 21:38		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN050625

Review By	John Carlone	Review On	5/7/2025 9:53:45 AM
Supervise By	Maresh Dadoda	Supervise On	5/7/2025 1:29:39 PM
SubDirectory	VN050625	HP Acquire Method	HP Processing Method 82N041525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133826
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133827,VP133828

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086503.D	06 May 2025 09:13		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086504.D	06 May 2025 10:35	CCC Failed Low For Com.#17,31	JC\MD	Ok,M
3	VN0506MBL01	VN0506MBL01	VN086505.D	06 May 2025 11:11		JC\MD	Ok
4	VN0506WBL01	VN0506WBL01	VN086506.D	06 May 2025 11:35		JC\MD	Ok
5	VN0506WBS01	VN0506WBS01	VN086507.D	06 May 2025 11:59		JC\MD	Ok,M
6	VN0506WBSD01	VN0506WBSD01	VN086508.D	06 May 2025 12:34		JC\MD	Ok,M
7	Q1947-04	MH-P	VN086509.D	06 May 2025 12:58	vial A pH#5.0	JC\MD	Ok
8	Q1947-08	MH-O	VN086510.D	06 May 2025 13:22	vial A pH#5.0	JC\MD	Ok
9	Q1951-04	MH-KK	VN086511.D	06 May 2025 13:46	vial A pH#5.0	JC\MD	Ok,M
10	Q1956-05	COMP1	VN086512.D	06 May 2025 14:11	vial A pH#5.0	JC\MD	Ok,M
11	Q1959-02	SOIL-PILE	VN086513.D	06 May 2025 14:35	vial A pH#5.0	JC\MD	Ok,M
12	Q1960-03	72-11933	VN086514.D	06 May 2025 14:59	vial A pH#5.0	JC\MD	Ok,M
13	VN0506MBS01	VN0506MBS01	VN086515.D	06 May 2025 15:23		JC\MD	Ok,M
14	VN0506MBSD01	VN0506MBSD01	VN086516.D	06 May 2025 15:47		JC\MD	Ok,M
15	Q1939-03	GB2	VN086517.D	06 May 2025 16:11	vial C, CCC Failed Low	JC\MD	ReRun
16	Q1914-06ME	SS-5ME	VN086518.D	06 May 2025 16:35		JC\MD	Ok
17	Q1914-07ME	SS-6ME	VN086519.D	06 May 2025 16:59	Need 20X	JC\MD	Dilution

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN050625

Review By	John Carlone	Review On	5/7/2025 9:53:45 AM
Supervise By	Maresh Dadoda	Supervise On	5/7/2025 1:29:39 PM
SubDirectory	VN050625	HP Acquire Method	HP Processing Method 82N041525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133826		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133827,VP133828		

18	Q1914-08	SS-7	VN086520.D	06 May 2025 17:23	Need 20X	JC\MD	Dilution
19	Q1914-09ME	SS-8ME	VN086521.D	06 May 2025 17:48	Need 20X	JC\MD	Dilution
20	Q1966-03	SILICA-GEL	VN086522.D	06 May 2025 18:12	CCC Failed Low,Run@ Lower Dilution	JC\MD	ReRun
21	IBLK	IBLK	VN086523.D	06 May 2025 18:36		JC\MD	Ok
22	IBLK	IBLK	VN086524.D	06 May 2025 19:00		JC\MD	Ok
23	IBLK	IBLK	VN086525.D	06 May 2025 19:24		JC\MD	Ok
24	VSTDCCC050	VSTDCCC050EC	VN086526.D	06 May 2025 19:48		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133559,VP133563
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574
CCC	VP133564,VP133565
Internal Standard/PEM	
ICV/I.BLK	VP133575
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045524.D	01 Apr 2025 16:15		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX045525.D	01 Apr 2025 17:06	%D failed for Comp. #53 in 01PPB	JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX045526.D	01 Apr 2025 17:29		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX045527.D	01 Apr 2025 17:52		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX045528.D	01 Apr 2025 18:15		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX045529.D	01 Apr 2025 18:38		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX045530.D	01 Apr 2025 19:02		JC/MD	Ok,M
8	IBLK	IBLK	VX045531.D	01 Apr 2025 19:25		JC/MD	Ok
9	VSTDICV050	ICVVX040225	VX045532.D	01 Apr 2025 19:48		JC/MD	Ok,M
10	BFB	BFB	VX045533.D	02 Apr 2025 09:30		JC/MD	Ok
11	VSTDIC050	VSTDIC050	VX045534.D	02 Apr 2025 10:02	pH#Lot#V12668	JC/MD	Ok,M
12	VX0402MBL01	VX0402MBL01	VX045535.D	02 Apr 2025 10:30		JC/MD	Ok
13	VX0402WBL01	VX0402WBL01	VX045536.D	02 Apr 2025 10:53		JC/MD	Ok
14	VX0402WBS01	VX0402WBS01	VX045537.D	02 Apr 2025 11:16		JC/MD	Ok,M
15	VX0402WBSD01	VX0402WBSD01	VX045538.D	02 Apr 2025 11:44		JC/MD	Ok,M
16	Q1697-01	MW-19B-72-040125	VX045539.D	02 Apr 2025 12:07	vial A pH<2 Need 200X	JC/MD	Dilution
17	Q1697-02	IW-01-55-040125	VX045540.D	02 Apr 2025 12:30	vial A pH<2 Need 10X	JC/MD	Dilution

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	IBLK	IBLK	VX045541.D	02 Apr 2025 12:54		JC/MD	Ok
19	Q1697-01DL	MW-19B-72-040125DL	VX045542.D	02 Apr 2025 13:21	vial B pH<2	JC/MD	Ok
20	Q1697-02DL	IW-01-55-040125DL	VX045543.D	02 Apr 2025 13:44	vial B pH<2	JC/MD	Ok
21	Q1697-03	IW-02-55-040125	VX045544.D	02 Apr 2025 14:07	Need lower dilution	JC/MD	Not Ok
22	Q1697-04	IW-02-55-040125-FD	VX045545.D	02 Apr 2025 14:31	Need lower dilution	JC/MD	Not Ok
23	Q1697-05	IW-03-55-040125	VX045546.D	02 Apr 2025 14:54	Need lower dilution	JC/MD	Not Ok
24	Q1697-07	MW-19B-72-040125	VX045547.D	02 Apr 2025 15:18	Not On Login	JC/MD	Not Ok
25	IBLK	IBLK	VX045548.D	02 Apr 2025 15:41		JC/MD	Ok
26	Q1697-03	IW-02-55-040125	VX045549.D	02 Apr 2025 16:04	vial A pH<2	JC/MD	Ok
27	Q1697-04	IW-02-55-040125-FD	VX045550.D	02 Apr 2025 16:27	vial A pH<2	JC/MD	Ok
28	Q1697-05	IW-03-55-040125	VX045551.D	02 Apr 2025 16:51	vial A pH<2	JC/MD	Ok
29	Q1697-06	TB-01-040125	VX045552.D	02 Apr 2025 17:14	vial A pH<2 TB;Hit of comp.#44	JC/MD	ReRun
30	Q1623-01	Storage-Blank-SOIL-RE	VX045553.D	02 Apr 2025 17:37	vial A pH<2	JC/MD	Ok
31	Q1623-02	Storage-Blank-WATER	VX045554.D	02 Apr 2025 18:01	vial A pH<2	JC/MD	Ok
32	Q1623-03	Storage-Blank-WATER	VX045555.D	02 Apr 2025 18:24	vial A pH<2	JC/MD	Ok
33	Q1623-04	Storage-Blank-SAMPLE	VX045556.D	02 Apr 2025 18:47	vial A pH<2	JC/MD	Ok
34	VSTDCCC050	VSTDCCC050EC	VX045557.D	02 Apr 2025 19:10		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX043025

Review By	John Carlone	Review On	5/2/2025 10:12:27 AM
Supervise By	Semsettin Yesilyurt	Supervise On	5/5/2025 4:29:00 AM
SubDirectory	VX043025	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133790		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133791,VP133792		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045984.D	30 Apr 2025 09:32		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045985.D	30 Apr 2025 10:01	pH#Lot#V12668	JC/MD	Ok,M
3	VX0430MBL01	VX0430MBL01	VX045986.D	30 Apr 2025 10:32		JC/MD	Ok
4	VX0430WBL01	VX0430WBL01	VX045987.D	30 Apr 2025 10:55		JC/MD	Ok
5	VX0430WBS01	VX0430WBS01	VX045988.D	30 Apr 2025 11:19		JC/MD	Ok,M
6	VX0430WBSD01	VX0430WBSD01	VX045989.D	30 Apr 2025 11:46		JC/MD	Ok,M
7	IBLK	IBLK	VX045990.D	30 Apr 2025 12:09		JC/MD	Ok
8	Q1914-14	FB04282025	VX045991.D	30 Apr 2025 12:32	vial A pH<2 FB	JC/MD	Ok
9	Q1904-02	VNJ-210	VX045992.D	30 Apr 2025 12:56	vial A pH#5.0	JC/MD	Ok
10	Q1901-09	B-170-SB01	VX045993.D	30 Apr 2025 13:19	vial A pH#5.0	JC/MD	Ok
11	Q1901-10	B-167-SB02	VX045994.D	30 Apr 2025 13:42	vial A pH#5.0	JC/MD	Ok
12	Q1915-01	WC-04282025	VX045995.D	30 Apr 2025 14:06	vial A pH#5.0	JC/MD	Ok
13	Q1901-08	B-167-SB01	VX045996.D	30 Apr 2025 14:29	vial A pH#5.0	JC/MD	Ok
14	Q1905-04	MH-G	VX045997.D	30 Apr 2025 14:53	vial A pH#5.0	JC/MD	Ok
15	Q1905-08	MH-H	VX045998.D	30 Apr 2025 15:16	vial A pH#5.0	JC/MD	Ok
16	Q1906-04	WC-4	VX045999.D	30 Apr 2025 15:40	vial A pH#5.0	JC/MD	Ok
17	Q1906-08	WC-5	VX046000.D	30 Apr 2025 16:03	vial A pH#5.0	JC/MD	Ok
18	Q1906-12	WC-6	VX046001.D	30 Apr 2025 16:27	vial A pH#5.0	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX043025

Review By	John Carlone	Review On	5/2/2025 10:12:27 AM
Supervise By	Semsettin Yesilyurt	Supervise On	5/5/2025 4:29:00 AM
SubDirectory	VX043025	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133790		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133791,VP133792		

19	Q1906-16	WC-7	VX046002.D	30 Apr 2025 16:50	vial A pH#5.0	JC/MD	Ok
20	Q1907-02	CO-008R-WC	VX046003.D	30 Apr 2025 17:13	vial A pH#5.0	JC/MD	Ok,M
21	Q1912-04	MH-E	VX046004.D	30 Apr 2025 17:37	vial A pH#5.0	JC/MD	Ok
22	Q1912-08	MH-F	VX046005.D	30 Apr 2025 18:00	vial A pH#5.0	JC/MD	Ok
23	Q1913-01	WC-12-S-202504	VX046006.D	30 Apr 2025 18:24	vial A pH#5.0	JC/MD	Ok
24	Q1913-03	WC-13-S-202504	VX046007.D	30 Apr 2025 18:47	vial A pH#5.0	JC/MD	Ok
25	Q1913-04	WC-13-A-202504	VX046008.D	30 Apr 2025 19:10	vial A pH#5.0 Surrogate Fail	JC/MD	ReRun
26	Q1913-02	WC-12-A-202504	VX046009.D	30 Apr 2025 19:34	vial A pH#5.0	JC/MD	Ok
27	Q1901-11	B-170-SB02	VX046010.D	30 Apr 2025 19:57	vial A pH#5.0	JC/MD	Ok
28	VSTDCCC050	VSTDCCC050EC	VX046011.D	30 Apr 2025 20:21		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY050125

Review By	Semsettin Yesilyurt	Review On	5/5/2025 2:51:02 PM
Supervise By	Maresh Dadoda	Supervise On	5/5/2025 2:57:16 PM
SubDirectory	VY050125	HP Acquire Method	MSVOA_Y HP Processing Method 82y042225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133796
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133797,VP133798 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022094.D	01 May 2025 08:20		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022095.D	01 May 2025 08:50		SY/MD	Ok,M
3	VY0501SBL01	VY0501SBL01	VY022096.D	01 May 2025 10:11		SY/MD	Ok
4	VY0501SBS01	VY0501SBS01	VY022097.D	01 May 2025 10:42		SY/MD	Ok,M
5	VY0501SBSD01	VY0501SBSD01	VY022098.D	01 May 2025 11:04		SY/MD	Ok,M
6	Q1920-01	291	VY022099.D	01 May 2025 11:54	vial-A	SY/MD	Ok
7	Q1911-01RE	EO-03-04292025RE	VY022100.D	01 May 2025 12:17	vial-B Internal Standard Fail	SY/MD	Confirms
8	Q1916-03	WC-12-VOC	VY022101.D	01 May 2025 12:41	vial-B	SY/MD	Ok
9	Q1923-01	TR-05-04302025	VY022102.D	01 May 2025 13:04	vial-A Internal Standard Fail	SY/MD	ReRun
10	Q1922-03	MH-R-VOC	VY022103.D	01 May 2025 13:28	vial-A	SY/MD	Ok
11	Q1922-07	MH-2-VOC	VY022104.D	01 May 2025 13:51	vial-A	SY/MD	Ok
12	Q1901-01	B-170-SB00	VY022105.D	01 May 2025 14:15	vial-A	SY/MD	Ok
13	Q1901-03	B-170-SB01	VY022106.D	01 May 2025 14:38	vial-A	SY/MD	Ok
14	Q1914-01	SS-1	VY022107.D	01 May 2025 15:02	vial-A	SY/MD	Ok
15	Q1914-02	SS-91	VY022108.D	01 May 2025 15:28	vial-A	SY/MD	Ok
16	Q1914-03	SS-2	VY022109.D	01 May 2025 16:04	vial-A	SY/MD	Ok
17	Q1914-04	SS-3	VY022110.D	01 May 2025 16:27	vial-A	SY/MD	Ok
18	Q1914-05	SS-4	VY022111.D	01 May 2025 16:49	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY050125

Review By	Semsettin Yesilyurt	Review On	5/5/2025 2:51:02 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/5/2025 2:57:16 PM		
SubDirectory	VY050125	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133796			
CCC		VP133797,VP133798			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1914-06	SS-5	VY022112.D	01 May 2025 17:16	vial-A Need MeOH	SY/MD	Dilution
20	Q1914-07	SS-6	VY022113.D	01 May 2025 17:37	vial-A Internal Standard Fail;Need MeOH	SY/MD	Dilution
21	Q1914-09	SS-8	VY022114.D	01 May 2025 18:00	vial-A Internal Standard Fail; Surrogate Fail;Need MeOH	SY/MD	Dilution
22	Q1914-10MS	SS-8MS	VY022115.D	01 May 2025 18:23	vial-A Surrogate fail	SY/MD	Ok,M
23	Q1914-11MSD	SS-8MSD	VY022116.D	01 May 2025 18:46	vial-A Surrogate fail	SY/MD	Ok,M
24	VSTDCCC050	VSTDCCC050EC	VY022117.D	01 May 2025 19:08		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY050625

Review By	Maresh Dadoda	Review On	5/7/2025 1:28:03 PM
Supervise By	Semsettin Yesilyurt	Supervise On	5/7/2025 1:47:17 PM
SubDirectory	VY050625	HP Acquire Method	MSVOA_Y
		HP Processing Method	82y042225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133823
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133824,VP133825 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022169.D	06 May 2025 08:00		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022170.D	06 May 2025 09:14		SY/MD	Ok,M
3	VY0506SBL01	VY0506SBL01	VY022171.D	06 May 2025 09:45		SY/MD	Ok
4	VY0506SBS01	VY0506SBS01	VY022172.D	06 May 2025 10:17		SY/MD	Ok,M
5	VY0506SBSD01	VY0506SBSD01	VY022173.D	06 May 2025 10:39		SY/MD	Ok,M
6	Q1961-01	EO-01-050525	VY022174.D	06 May 2025 11:29	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q1952-01	CLEAN-FILL	VY022175.D	06 May 2025 11:53	vial-B	SY/MD	Ok
8	Q1962-01	HD-01-050525	VY022176.D	06 May 2025 12:16	vial-A Internal Standard Fail	SY/MD	ReRun
9	Q1960-02	72-11933	VY022177.D	06 May 2025 12:40	vial-A	SY/MD	Ok
10	Q1914-12	SS-9	VY022178.D	06 May 2025 13:03	vial-A	SY/MD	Ok
11	Q1964-03	MH-LL-VOC	VY022179.D	06 May 2025 13:26	vial-A	SY/MD	Ok,M
12	VSTDCCC050	VSTDCCC050EC	VY022180.D	06 May 2025 14:42		SY/MD	Ok,M

M : Manual Integration