

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1914

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VN086515.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0506MBS01	Benzene	2000	1800	ug/Kg	90			84	121	
	Toluene	2000	1900	ug/Kg	95			83	122	
	Ethyl Benzene	2000	1900	ug/Kg	95			82	124	
	m/p-Xylenes	4000	3800	ug/Kg	95			83	124	
	o-Xylene	2000	2000	ug/Kg	100			83	123	
	Isopropylbenzene	2000	1900	ug/Kg	95			82	124	
	N-propylbenzene	2000	1800	ug/Kg	90			81	123	
	1,3,5-Trimethylbenzene	2000	1900	ug/Kg	95			82	124	
	tert-Butylbenzene	2000	1900	ug/Kg	95			81	125	
	1,2,4-Trimethylbenzene	2000	1900	ug/Kg	95			81	125	
	Sec-butylbenzene	2000	1900	ug/Kg	95			80	124	
	p-Isopropyltoluene	2000	1900	ug/Kg	95			81	125	
	n-Butylbenzene	2000	1700	ug/Kg	85			78	126	