

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1914

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VY022172.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0506SBS01	Benzene	20	21.3	ug/Kg	106			84	121	
	Toluene	20	21.4	ug/Kg	107			83	122	
	Ethyl Benzene	20	20.3	ug/Kg	102			82	124	
	m/p-Xylenes	40	41.6	ug/Kg	104			83	124	
	o-Xylene	20	20.3	ug/Kg	102			83	123	
	Isopropylbenzene	20	19.7	ug/Kg	99			82	124	
	N-propylbenzene	20	20.6	ug/Kg	103			81	123	
	1,3,5-Trimethylbenzene	20	20.3	ug/Kg	102			82	124	
	tert-Butylbenzene	20	19.5	ug/Kg	98			81	125	
	1,2,4-Trimethylbenzene	20	20.6	ug/Kg	103			81	125	
	Sec-butylbenzene	20	20.5	ug/Kg	103			80	124	
	p-Isopropyltoluene	20	20.2	ug/Kg	101			81	125	
	n-Butylbenzene	20	20.0	ug/Kg	100			78	126	