

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1914

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VY022173.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0506SBSD01	Benzene	20	22.4	ug/Kg	112	6		84	121	20
	Toluene	20	22.4	ug/Kg	112	5		83	122	20
	Ethyl Benzene	20	21.4	ug/Kg	107	5		82	124	20
	m/p-Xylenes	40	43.9	ug/Kg	110	6		83	124	20
	o-Xylene	20	21.1	ug/Kg	106	4		83	123	20
	Isopropylbenzene	20	21.0	ug/Kg	105	6		82	124	20
	N-propylbenzene	20	21.4	ug/Kg	107	4		81	123	20
	1,3,5-Trimethylbenzene	20	21.4	ug/Kg	107	5		82	124	20
	tert-Butylbenzene	20	20.8	ug/Kg	104	6		81	125	20
	1,2,4-Trimethylbenzene	20	21.7	ug/Kg	109	6		81	125	20
	Sec-butylbenzene	20	21.6	ug/Kg	108	5		80	124	20
	p-Isopropyltoluene	20	21.3	ug/Kg	106	5		81	125	20
	n-Butylbenzene	20	21.2	ug/Kg	106	6		78	126	20