

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914
 Lab File ID: VN086504.D Date Analyzed: 05/06/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:35
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	186971	8.22	338775	9.09	314072	11.86
UPPER LIMIT	373942	8.724	677550	9.594	628144	12.359
LOWER LIMIT	93485.5	7.724	169388	8.594	157036	11.359
EPA SAMPLE NO.						
SS-5ME	169551	8.22	319127	9.09	297541	11.87
SS-6ME	172187	8.22	322534	9.09	313788	11.87
SS-7	198517	8.22	370271	9.09	347259	11.86
SS-8ME	202938	8.22	380050	9.09	349216	11.87
VN0506MBL01	194854	8.22	360705	9.10	339328	11.87
VN0506MBS01	186931	8.22	334765	9.10	293306	11.87
VN0506MBSD01	179114	8.22	320919	9.10	294238	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.