

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02  
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914  
 Lab File ID: VX045985.D Date Analyzed: 04/30/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 10:01  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT # | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|------|---------------|--------|
| 12 HOUR STD    | 83019         | 5.55  | 141256        | 6.75 | 122754        | 10.05  |
| UPPER LIMIT    | 166038        | 6.049 | 282512        | 7.25 | 245508        | 10.549 |
| LOWER LIMIT    | 41509.5       | 5.049 | 70628         | 6.25 | 61377         | 9.549  |
| EPA SAMPLE NO. |               |       |               |      |               |        |
| FB04282025     | 59292         | 5.55  | 117758        | 6.76 | 108001        | 10.06  |
| VX0430WBL01    | 62274         | 5.54  | 125239        | 6.76 | 116463        | 10.05  |
| VX0430WBS01    | 79591         | 5.54  | 138531        | 6.76 | 123154        | 10.05  |
| VX0430WBSD01   | 77091         | 5.54  | 132529        | 6.76 | 119517        | 10.05  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.