

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914
 Lab File ID: VY022170.D Date Analyzed: 05/06/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:14
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	239745	7.71	360580	8.62	331226	11.42
UPPER LIMIT	479490	8.213	721160	9.116	662452	11.92
LOWER LIMIT	119873	7.213	180290	8.116	165613	10.92
EPA SAMPLE NO.						
SS-9	197156	7.71	383574	8.62	339845	11.42
VY0506SBL01	251073	7.71	481164	8.62	415589	11.42
VY0506SBS01	219542	7.71	344933	8.62	314858	11.42
VY0506SBSD01	213818	7.71	337521	8.62	309806	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.