

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02  
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914  
 Lab File ID: VN086478.D Date Analyzed: 05/05/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 11:32  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 133852        | 13.788 |  |  |  |  |
| UPPER LIMIT    | 267704        | 14.288 |  |  |  |  |
| LOWER LIMIT    | 66926         | 13.288 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| SS-6MEDL       | 140970        | 13.79  |  |  |  |  |
| SS-7DL         | 128135        | 13.79  |  |  |  |  |
| SS-8MEDL       | 141242        | 13.79  |  |  |  |  |
| VN0505MBL01    | 117072        | 13.79  |  |  |  |  |
| VN0505MBS01    | 137470        | 13.79  |  |  |  |  |
| VN0505MBSD01   | 129515        | 13.79  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.