

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914
 Lab File ID: VN086504.D Date Analyzed: 05/06/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:35
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	155373	13.788				
UPPER LIMIT	310746	14.288				
LOWER LIMIT	77686.5	13.288				
EPA SAMPLE NO.						
SS-5ME	131140	13.79				
SS-6ME	148688	13.79				
SS-7	152767	13.79				
SS-8ME	167366	13.79				
VN0506MBL01	141530	13.79				
VN0506MBS01	131486	13.79				
VN0506MBSD01	133843	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.