

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914
 Lab File ID: VY022095.D Date Analyzed: 05/01/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:50
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	227184	13.347				
UPPER LIMIT	454368	13.847				
LOWER LIMIT	113592	12.847				
EPA SAMPLE NO.						
SS-1	218119	13.35				
SS-91	255221	13.35				
SS-2	258914	13.35				
SS-3	236192	13.35				
SS-4	199649	13.35				
SS-5	237049	13.35				
SS-6	327199	13.35				
SS-8	511573 *	13.35				
SS-8MS	378955	13.33				
SS-8MSD	398456	13.35				
VY0501SBL01	180066	13.35				
VY0501SBS01	225406	13.35				
VY0501SBSD01	219160	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.