

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914
 Lab File ID: VY022170.D Date Analyzed: 05/06/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:14
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	184988	13.346				
UPPER LIMIT	369976	13.846				
LOWER LIMIT	92494	12.846				
EPA SAMPLE NO.						
SS-9	120680	13.35				
VY0506SBL01	146510	13.35				
VY0506SBS01	170277	13.35				
VY0506SBSD01	166836	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.