

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG No.: Q1914
 Instrument ID: MSVOA_N Calibration Date(s): 04/15/2025 04/15/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:29 14:29
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN086277.D		RRF005 = VN086278.D		RRF010 = VN086279.D			
		RRF020 = VN086280.D		RRF050 = VN086281.D		RRF100 = VN086282.D			
COMPOUND	RRF001	RRF005	RRF010	RRF020	RRF050	RRF100	RRF	% RSD	
Benzene	1.648	1.545	1.389	1.508	1.409	1.411	1.485	6.8	
Toluene	1.003	0.963	0.865	0.941	0.892	0.895	0.927	5.6	
Ethyl Benzene	2.210	2.112	1.873	2.028	1.918	1.941	2.014	6.4	
m/p-Xylenes	0.809	0.795	0.705	0.757	0.727	0.734	0.754	5.4	
o-Xylene	0.791	0.792	0.709	0.753	0.713	0.718	0.746	5.2	
Isopropylbenzene	4.666	4.271	3.991	4.181	3.728	3.706	4.090	8.9	
n-propylbenzene	5.366	5.072	4.629	4.862	4.488	4.501	4.820	7.3	
1,3,5-Trimethylbenzene	3.800	3.518	3.178	3.418	3.083	3.088	3.348	8.5	
tert-Butylbenzene	3.382	3.057	2.795	2.886	2.608	2.675	2.901	9.8	
1,2,4-Trimethylbenzene	3.782	3.644	3.262	3.431	3.171	3.156	3.408	7.6	
sec-Butylbenzene	4.451	4.309	3.897	3.991	3.763	3.729	4.023	7.3	
p-Isopropyltoluene	3.603	3.487	3.174	3.290	3.158	3.185	3.316	5.6	
n-Butylbenzene	3.209	3.013	2.746	2.921	2.838	2.910	2.939	5.4	
1,2-Dichloroethane-d4		0.746	0.745	0.707	0.719	0.708	0.725	2.6	
Dibromofluoromethane		0.267	0.255	0.230	0.215	0.194	0.232	12.8	
Toluene-d8		1.247	1.281	1.185	1.251	1.237	1.240	2.8	
4-Bromofluorobenzene		0.459	0.456	0.421	0.459	0.467	0.452	4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.