

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG No.: Q1914

Instrument ID: MSVOA_N Calibration Date/Time: 05/06/2025 10:35

Lab File ID: VN086504.D Init. Calib. Date(s): 04/15/2025 04/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:29 14:29

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.485	1.462		-1.55	20
Toluene	0.927	0.953		2.81	20
Ethyl Benzene	2.014	1.981		-1.64	20
m/p-Xylenes	0.754	0.762		1.06	20
o-Xylene	0.746	0.762		2.14	20
Isopropylbenzene	4.090	3.729		-8.83	20
n-propylbenzene	4.820	4.382		-9.09	20
1,3,5-Trimethylbenzene	3.348	3.125		-6.66	20
tert-Butylbenzene	2.901	2.643		-8.89	20
1,2,4-Trimethylbenzene	3.408	3.236		-5.05	20
sec-Butylbenzene	4.023	3.668		-8.82	20
p-Isopropyltoluene	3.316	3.160		-4.7	20
n-Butylbenzene	2.939	2.699		-8.17	20
1,2-Dichloroethane-d4	0.725	0.698		-3.72	20
Dibromofluoromethane	0.232	0.250		7.76	20
Toluene-d8	1.240	1.178		-5	20
4-Bromofluorobenzene	0.452	0.460		1.77	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.