

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG No.: Q1914

Instrument ID: MSVOA_Y Calibration Date/Time: 05/06/2025 09:14

Lab File ID: VY022170.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.387	1.468		5.84	20
Toluene	0.898	0.976		8.69	20
Ethyl Benzene	1.829	2.031		11.04	20
m/p-Xylenes	0.728	0.817		12.23	20
o-Xylene	0.671	0.744		10.88	20
Isopropylbenzene	3.341	3.536		5.84	20
n-propylbenzene	3.978	4.340		9.1	20
1,3,5-Trimethylbenzene	2.752	2.992		8.72	20
tert-Butylbenzene	2.469	2.618		6.03	20
1,2,4-Trimethylbenzene	2.753	3.014		9.48	20
sec-Butylbenzene	3.612	3.931		8.83	20
p-Isopropyltoluene	3.067	3.352		9.29	20
n-Butylbenzene	2.742	3.028		10.43	20
1,2-Dichloroethane-d4	0.462	0.462		0	20
Dibromofluoromethane	0.330	0.334		1.21	20
Toluene-d8	1.245	1.283		3.05	20
4-Bromofluorobenzene	0.419	0.436		4.06	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.