

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914
 Lab File ID: VN086478.D Date Analyzed: 05/05/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	154484	8.22	297307	9.10	276417	11.87
UPPER LIMIT	308968	8.718	594614	9.6	552834	12.365
LOWER LIMIT	77242	7.718	148654	8.6	138209	11.365
EPA SAMPLE NO.						
SS-6MEDL	166833	8.22	307337	9.10	302781	11.87
SS-7DL	162880	8.22	306914	9.10	283717	11.87
SS-8MEDL	167617	8.22	312635	9.10	300291	11.87
VN0505MBL01	156968	8.22	298822	9.10	277824	11.87
VN0505MBS01	174392	8.22	323563	9.10	294006	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.