

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914
Lab File ID: VN086276.D BFB Injection Date: 04/15/2025
Instrument ID: MSVOA_N BFB Injection Time: 10:47
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.6
75	30.0 - 60.0% of mass 95	50.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.4 (0.6) 1
174	50.0 - 100.0% of mass 95	67.6
175	5.0 - 9.0% of mass 174	5.1 (7.6) 1
176	95.0 - 101.0% of mass 174	64.6 (95.5) 1
177	5.0 - 9.0% of mass 176	4.4 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VN086277.D	04/15/2025	11:29
VSTDIC005	VSTDIC005	VN086278.D	04/15/2025	12:21
VSTDIC010	VSTDIC010	VN086279.D	04/15/2025	12:45
VSTDIC020	VSTDIC020	VN086280.D	04/15/2025	13:09
VSTDIC050	VSTDIC050	VN086281.D	04/15/2025	13:51
VSTDIC100	VSTDIC100	VN086282.D	04/15/2025	14:29