

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG No.: Q1914

Instrument ID: MSVOA_N Calibration Date/Time: 05/05/2025 11:32

Lab File ID: VN086478.D Init. Calib. Date(s): 04/15/2025 04/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:29 14:29

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.485	1.541		3.77	20
Toluene	0.927	0.979		5.61	20
Ethyl Benzene	2.014	1.999		-0.75	20
m/p-Xylenes	0.754	0.772		2.39	20
o-Xylene	0.746	0.767		2.82	20
Isopropylbenzene	4.090	3.762		-8.02	20
n-propylbenzene	4.820	4.463		-7.41	20
1,3,5-Trimethylbenzene	3.348	3.185		-4.87	20
tert-Butylbenzene	2.901	2.643		-8.89	20
1,2,4-Trimethylbenzene	3.408	3.330		-2.29	20
sec-Butylbenzene	4.023	3.690		-8.28	20
p-Isopropyltoluene	3.316	3.176		-4.22	20
n-Butylbenzene	2.939	2.742		-6.7	20
1,2-Dichloroethane-d4	0.725	0.817		12.69	20
Dibromofluoromethane	0.232	0.264		13.79	20
Toluene-d8	1.240	1.280		3.23	20
4-Bromofluorobenzene	0.452	0.478		5.75	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.