

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG No.: Q1914

Instrument ID: MSVOA_Y Calibration Date/Time: 05/01/2025 08:50

Lab File ID: VY022095.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.387	1.370		-1.23	20
Toluene	0.898	0.925		3.01	20
Ethyl Benzene	1.829	1.811		-0.98	20
m/p-Xylenes	0.728	0.745		2.34	20
o-Xylene	0.671	0.676		0.75	20
Isopropylbenzene	3.341	3.204		-4.1	20
n-propylbenzene	3.978	3.849		-3.24	20
1,3,5-Trimethylbenzene	2.752	2.700		-1.89	20
tert-Butylbenzene	2.469	2.372		-3.93	20
1,2,4-Trimethylbenzene	2.753	2.737		-0.58	20
sec-Butylbenzene	3.612	3.524		-2.44	20
p-Isopropyltoluene	3.067	3.035		-1.04	20
n-Butylbenzene	2.742	2.631		-4.05	20
1,2-Dichloroethane-d4	0.462	0.431		-6.71	20
Dibromofluoromethane	0.330	0.335		1.51	20
Toluene-d8	1.245	1.270		2.01	20
4-Bromofluorobenzene	0.419	0.427		1.91	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.