

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG No.: Q1914
 Instrument ID: MSVOA_N Calibration Date(s): 04/15/2025 04/15/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:29 14:29
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF001 = VN086277.D		RRF005 = VN086278.D		RRF010 = VN086279.D			
		RRF020 = VN086280.D		RRF050 = VN086281.D		RRF100 = VN086282.D			
COMPOUND	RRF001	RRF005	RRF010	RRF020	RRF050	RRF100	RRF	% RSD	
Benzene	1.648	1.545	1.389	1.508	1.409	1.411	1.485	6.8	
Toluene	1.003	0.963	0.865	0.941	0.892	0.895	0.927	5.6	
Ethyl Benzene	2.210	2.112	1.873	2.028	1.918	1.941	2.014	6.4	
m/p-Xylenes	0.809	0.795	0.705	0.757	0.727	0.734	0.754	5.4	
o-Xylene	0.791	0.792	0.709	0.753	0.713	0.718	0.746	5.2	
Isopropylbenzene	4.666	4.271	3.991	4.181	3.728	3.706	4.090	8.9	
n-propylbenzene	5.366	5.072	4.629	4.862	4.488	4.501	4.820	7.3	
1,3,5-Trimethylbenzene	3.800	3.518	3.178	3.418	3.083	3.088	3.348	8.5	
tert-Butylbenzene	3.382	3.057	2.795	2.886	2.608	2.675	2.901	9.8	
1,2,4-Trimethylbenzene	3.782	3.644	3.262	3.431	3.171	3.156	3.408	7.6	
sec-Butylbenzene	4.451	4.309	3.897	3.991	3.763	3.729	4.023	7.3	
p-Isopropyltoluene	3.603	3.487	3.174	3.290	3.158	3.185	3.316	5.6	
n-Butylbenzene	3.209	3.013	2.746	2.921	2.838	2.910	2.939	5.4	
1,2-Dichloroethane-d4		0.746	0.745	0.707	0.719	0.708	0.725	2.6	
Dibromofluoromethane		0.267	0.255	0.230	0.215	0.194	0.232	12.8	
Toluene-d8		1.247	1.281	1.185	1.251	1.237	1.240	2.8	
4-Bromofluorobenzene		0.459	0.456	0.421	0.459	0.467	0.452	4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG No.: Q1914
 Instrument ID: MSVOA_X Calibration Date(s): 04/01/2025 04/01/2025
 Heated Purge: (Y/N) N Calibration Time(s): 17:06 19:02
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX045525.D		RRF005 = VX045526.D		RRF020 = VX045527.D			
		RRF050 = VX045528.D		RRF100 = VX045529.D		RRF150 = VX045530.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Benzene	1.414	1.416	1.519	1.481	1.483	1.465	1.463	2.8	
Toluene	0.817	0.866	0.939	0.910	0.905	0.875	0.885	4.8	
Ethyl Benzene	1.608	1.819	2.002	2.007	1.993	2.053	1.914	8.9	
m/p-Xylenes	0.594	0.669	0.732	0.728	0.723	0.729	0.696	7.9	
o-Xylene	0.562	0.676	0.725	0.719	0.716	0.714	0.686	9.2	
Isopropylbenzene	3.581	3.850	4.224	4.181	4.043	4.151	4.005	6.2	
n-propylbenzene	3.734	4.346	5.012	4.896	4.819	4.872	4.613	10.6	
1,3,5-Trimethylbenzene	2.931	3.119	3.571	3.511	3.367	3.371	3.312	7.3	
tert-Butylbenzene	2.894	3.124	3.437	3.427	3.387	3.405	3.279	6.8	
1,2,4-Trimethylbenzene	2.906	3.129	3.529	3.523	3.438	3.420	3.324	7.6	
sec-Butylbenzene	3.195	3.828	4.316	4.300	4.232	4.292	4.027	11.1	
p-Isopropyltoluene	2.734	3.157	3.515	3.524	3.506	3.484	3.320	9.6	
n-Butylbenzene	2.117	2.486	2.974	3.160	3.256	3.283	2.879	16.5	
1,2-Dichloroethane-d4		0.962	0.900	0.868	0.904	0.937	0.914	3.9	
Dibromofluoromethane		0.372	0.342	0.345	0.353	0.362	0.355	3.5	
Toluene-d8		1.257	1.233	1.214	1.230	1.257	1.238	1.5	
4-Bromofluorobenzene		0.413	0.448	0.453	0.481	0.460	0.451	5.5	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG No.: Q1914
 Instrument ID: MSVOA_Y Calibration Date(s): 04/22/2025 04/22/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 13:39 16:15
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY021953.D		RRF010 = VY021954.D		RRF020 = VY021955.D			
		RRF050 = VY021956.D		RRF100 = VY021957.D		RRF150 = VY021958.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Benzene	1.423	1.439	1.351	1.337	1.415	1.358	1.387	3.1	
Toluene	0.838	0.915	0.875	0.894	0.952	0.913	0.898	4.4	
Ethyl Benzene	1.693	1.840	1.718	1.835	1.990	1.898	1.829	6.1	
m/p-Xylenes	0.664	0.720	0.699	0.734	0.797	0.754	0.728	6.3	
o-Xylene	0.599	0.639	0.635	0.689	0.747	0.716	0.671	8.3	
Isopropylbenzene	3.061	3.316	3.102	3.325	3.639	3.599	3.341	7.2	
n-propylbenzene	3.708	3.996	3.778	3.977	4.258	4.152	3.978	5.3	
1,3,5-Trimethylbenzene	2.467	2.757	2.614	2.800	2.978	2.898	2.752	6.8	
tert-Butylbenzene	2.246	2.382	2.263	2.494	2.725	2.701	2.469	8.5	
1,2,4-Trimethylbenzene	2.415	2.717	2.631	2.784	3.027	2.945	2.753	8	
sec-Butylbenzene	3.254	3.724	3.409	3.602	3.898	3.786	3.612	6.7	
p-Isopropyltoluene	2.717	3.016	2.869	3.094	3.386	3.321	3.067	8.4	
n-Butylbenzene	2.477	2.790	2.545	2.742	2.976	2.919	2.742	7.2	
1,2-Dichloroethane-d4	0.525	0.477	0.459	0.466	0.418	0.427	0.462	8.3	
Dibromofluoromethane	0.335	0.341	0.330	0.330	0.319	0.327	0.330	2.3	
Toluene-d8	1.220	1.260	1.211	1.276	1.244	1.261	1.245	2	
4-Bromofluorobenzene	0.408	0.429	0.401	0.426	0.423	0.428	0.419	2.8	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.