

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02  
 Lab Code: CHEM Case No.: Q1914 SAS No.: Q1914 SDG NO.: Q1914  
 Lab File ID: VN086478.D Date Analyzed: 05/05/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 11:32  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	133852	13.788				
UPPER LIMIT	267704	14.288				
LOWER LIMIT	66926	13.288				
EPA SAMPLE NO.						
SS-6MEDL	140970	13.79				
SS-7DL	128135	13.79				
SS-8MEDL	141242	13.79				
VN0505MBL01	117072	13.79				
VN0505MBS01	137470	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.