

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN041525

Review By	John Carlone	Review On	4/16/2025 9:07:40 AM
Supervise By	Semsettin Yesilyurt	Supervise On	4/16/2025 3:44:21 PM
SubDirectory	VN041525	HP Acquire Method	MSVOA_N HP Processing Method 82N041525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133665 VP133667,VP133668,VP133669,VP133670,VP133671,VP133672
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131746 VP133673

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086276.D	15 Apr 2025 10:47		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN086277.D	15 Apr 2025 11:29		JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN086278.D	15 Apr 2025 12:21	Comp.#13,16 is on Linear Regression	JC\MD	Ok,M
4	VSTDICC010	VSTDICC010	VN086279.D	15 Apr 2025 12:45	Comp.#43 is on Quadratic Regression	JC\MD	Ok,M
5	VSTDICC020	VSTDICC020	VN086280.D	15 Apr 2025 13:09		JC\MD	Ok,M
6	VSTDICCC050	VSTDICCC050	VN086281.D	15 Apr 2025 13:51		JC\MD	Ok,M
7	VSTDICC100	VSTDICC100	VN086282.D	15 Apr 2025 14:29		JC\MD	Ok,M
8	VIBLK	VIBLK	VN086283.D	15 Apr 2025 15:06		JC\MD	Ok
9	VSTDICV050	ICVVN041525	VN086284.D	15 Apr 2025 15:30		JC\MD	Ok,M

M : Manual Integration