

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY050625

Review By	Mahesh Dadoda	Review On	5/7/2025 1:28:03 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/7/2025 1:47:17 PM		
SubDirectory	VY050625	HP Acquire Method	MSVOA_Y	HP Processing Method	82y042225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133823
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133824,VP133825 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022169.D	06 May 2025 08:00		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022170.D	06 May 2025 09:14		SY/MD	Ok,M
3	VY0506SBL01	VY0506SBL01	VY022171.D	06 May 2025 09:45		SY/MD	Ok
4	VY0506SBS01	VY0506SBS01	VY022172.D	06 May 2025 10:17		SY/MD	Ok,M
5	VY0506SBSD01	VY0506SBSD01	VY022173.D	06 May 2025 10:39		SY/MD	Ok,M
6	Q1961-01	EO-01-050525	VY022174.D	06 May 2025 11:29	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q1952-01	CLEAN-FILL	VY022175.D	06 May 2025 11:53	vial-B	SY/MD	Ok
8	Q1962-01	HD-01-050525	VY022176.D	06 May 2025 12:16	vial-A Internal Standard Fail	SY/MD	ReRun
9	Q1960-02	72-11933	VY022177.D	06 May 2025 12:40	vial-A	SY/MD	Ok
10	Q1914-12	SS-9	VY022178.D	06 May 2025 13:03	vial-A	SY/MD	Ok
11	Q1964-03	MH-LL-VOC	VY022179.D	06 May 2025 13:26	vial-A	SY/MD	Ok,M
12	VSTDCCC050	VSTDCCC050EC	VY022180.D	06 May 2025 14:42		SY/MD	Ok,M

M : Manual Integration