

DATA PACKAGE

VOLATILE ORGANICS
GENERAL CHEMISTRY
METALS
SEMI-VOLATILE ORGANICS

PROJECT NAME : CON ED UTEN MOUNT VERNON, NY

CDM SMITH

110 Fieldcrest Ave

Raritan Center

Edison, NJ - 08837

Phone No: 732-225-7000

ORDER ID : Q1915

ATTENTION : Marcie Ann Encinas



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) TCLP VOA- Case Narrative	4
2.2) TCLP BNA- Case Narrative	6
2.3) Metals-TCLP- Case Narrative	8
2.4) Genchem- Case Narrative	9
3) Qualifier Page	10
4) QA Checklist	12
5) TCLP VOA Data	13
6) TCLP BNA Data	34
7) Metals-TCLP Data	69
8) Genchem Data	120
9) Shipping Document	131
9.1) CHAIN OF CUSTODY	132
9.2) Lab Certificate	133

1
2
3
4
5
6
7
8
9

Cover Page

Order ID : Q1915

Project ID : Con Ed UTEN Mount Vernon, NY

Client : CDM Smith

Lab Sample Number

Q1915-01

Client Sample Number

WC-04282025

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:46 am, May 08, 2025

Date: 5/8/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

CDM Smith

Project Name: Con Ed UTEN Mount Vernon, NY

Project # N/A

Chemtech Project # Q1915

Test Name: TCLP VOA

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 04/29/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Ignitability, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP VOA and TCLP ZHE Extraction. This data package contains results for TCLP VOA.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of TCLP VOA was based on method 8260D and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the Requirements.

The Continuous Calibration File ID VX045985.D met the requirements except for Carbon Tetrachloride is failing high but no positive hit in associate sample therefore no corrective action taken.

The Tuning criteria met requirements.

E. Additional Comments:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:46 am, May 08, 2025

Signature _____

CASE NARRATIVE

CDM Smith

Project Name: Con Ed UTEN Mount Vernon, NY

Project # N/A

Chemtech Project # Q1915

Test Name: TCLP BNA

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 04/29/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Ignitability, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP VOA and TCLP ZHE Extraction. This data package contains results for TCLP BNA.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df The analysis of TCLP BNA was based on method 8270E and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds .

The MSD recoveries met the acceptable requirements .

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 20% in the Initial Calibration (8270-BF043025.M) for 2,4-Dinitrotoluene, this compound is passing on Linear Regression.

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:46 am, May 08, 2025

Signature_____



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

CDM Smith

Project Name: Con Ed UTEN Mount Vernon, NY

Project # N/A

Chemtech Project # Q1915

Test Name: TCLP Mercury, TCLP ICP Metals

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 04/29/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Ignitability, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP VOA and TCLP ZHE Extraction. This data package contains results for TCLP Mercury, TCLP ICP Metals.

C. Analytical Techniques:

The analysis of TCLP ICP Metals was based on method 6010D, digestion based on method 3010 (waters). The analysis and digestion of TCLP Mercury was based on method 7470A and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:47 am, May 08, 2025



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

CDM Smith

Project Name: Con Ed UTEN Mount Vernon, NY

Project # N/A

Chemtech Project # Q1915

Test Name: Corrosivity, Ignitability, Reactive Cyanide, Reactive Sulfide

A. Number of Samples and Date of Receipt:

1 Solid sample was received on 04/29/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Ignitability, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP VOA and TCLP ZHE Extraction. This data package contains results for Corrosivity, Ignitability, Reactive Cyanide, Reactive Sulfide.

C. Analytical Techniques:

The analysis of Ignitability was based on method 1030, The analysis of Reactive Cyanide was based on method 9012B, The analysis of Reactive Sulfide was based on method 9034 and The analysis of Corrosivity was based on method 9045D.

D. QA/ QC Samples:

The Holding Times were met for all samples except for WC-04282025 of Corrosivity as sample was received out of holding time.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:47 am, May 08, 2025

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1915

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 05/08/2025

LAB CHRONICLE

OrderID:	Q1915	OrderDate:	4/29/2025 2:29:00 PM
Client:	CDM Smith	Project:	Con Ed UTEN Mount Vernon, NY
Contact:	Marcie Ann Encinas	Location:	L41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1915-01	WC-04282025	TCLP	TCLP VOA	8260D	04/28/25		04/30/25	04/29/25

Hit Summary Sheet
SW-846

SDG No.: Q1915
Client: CDM Smith

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:				0				
Total Voc :								
Total Concentration:								



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	04/28/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/29/25
Client Sample ID:	WC-04282025	SDG No.:	Q1915
Lab Sample ID:	Q1915-01	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045995.D	1		04/30/25 14:06	VX043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	5.00	U	0.26	5.00	ug/L
75-35-4	1,1-Dichloroethene	5.00	U	0.23	5.00	ug/L
78-93-3	2-Butanone	25.0	U	0.98	25.0	ug/L
56-23-5	Carbon Tetrachloride	5.00	U	0.25	5.00	ug/L
67-66-3	Chloroform	5.00	U	0.25	5.00	ug/L
71-43-2	Benzene	5.00	U	0.15	5.00	ug/L
107-06-2	1,2-Dichloroethane	5.00	U	0.22	5.00	ug/L
79-01-6	Trichloroethene	5.00	U	0.090	5.00	ug/L
127-18-4	Tetrachloroethene	5.00	U	0.23	5.00	ug/L
108-90-7	Chlorobenzene	5.00	U	0.12	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	56.5		74 - 125	113%	SPK: 50
1868-53-7	Dibromofluoromethane	54.9		75 - 124	110%	SPK: 50
2037-26-5	Toluene-d8	53.3		86 - 113	107%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		77 - 121	106%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	56600	5.55			
540-36-3	1,4-Difluorobenzene	111000	6.757			
3114-55-4	Chlorobenzene-d5	108000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	44600	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: Q1915

Client: CDM Smith

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1915-01	WC-04282025	1,2-Dichloroethane-d4	50	56.5	113	74	125
		Dibromofluoromethane	50	54.9	110	75	124
		Toluene-d8	50	53.3	107	86	113
		4-Bromofluorobenzene	50	52.8	106	77	121

A

B

C

D

E

F

G

Surrogate Summary

SDG No.: Q1915

Client: CDM Smith

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX0430WBL01	VX0430WBL01	1,2-Dichloroethane-d4	50	56.3	113	74	125
		Dibromofluoromethane	50	52.5	105	75	124
		Toluene-d8	50	50.4	101	86	113
		4-Bromofluorobenzene	50	51.9	104	77	121
VX0430WBS01	VX0430WBS01	1,2-Dichloroethane-d4	50	54.6	109	74	125
		Dibromofluoromethane	50	56.5	113	75	124
		Toluene-d8	50	52.1	104	86	113
		4-Bromofluorobenzene	50	55.7	111	77	121
VX0430WBSD0	VX0430WBSD01	1,2-Dichloroethane-d4	50	54.3	109	74	125
		Dibromofluoromethane	50	56.5	113	75	124
		Toluene-d8	50	52.0	104	86	113
		4-Bromofluorobenzene	50	56.5	113	77	121

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1915
Client: CDM Smith
Analytical Method: SW8260-Low **Datafile :** VX045988.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0430WBS01	Vinyl chloride	20	17.3	ug/L	86			65	117	
	1,1-Dichloroethene	20	18.5	ug/L	93			74	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	21.8	ug/L	109			77	113	
	Chloroform	20	21.6	ug/L	108			79	113	
	Benzene	20	20.2	ug/L	101			82	109	
	1,2-Dichloroethane	20	22.8	ug/L	114			80	115	
	Trichloroethene	20	19.9	ug/L	100			77	113	
	Tetrachloroethene	20	19.0	ug/L	95			67	123	
	Chlorobenzene	20	21.3	ug/L	106			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: SW8260-Low

Datafile : VX045989.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0430WBSD01	Vinyl chloride	20	16.6	ug/L	83	4		65	117	20
	1,1-Dichloroethene	20	18.5	ug/L	93	0		74	110	20
	2-Butanone	100	120	ug/L	120	9		65	122	20
	Carbon Tetrachloride	20	21.8	ug/L	109	0		77	113	20
	Chloroform	20	21.0	ug/L	105	3		79	113	20
	Benzene	20	20.1	ug/L	101	0		82	109	20
	1,2-Dichloroethane	20	22.6	ug/L	113	1		80	115	20
	Trichloroethene	20	20.2	ug/L	101	1		77	113	20
	Tetrachloroethene	20	20.1	ug/L	101	6		67	123	20
	Chlorobenzene	20	20.8	ug/L	104	2		82	109	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0430WBL01

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q1915

SAS No.: Q1915 SDG NO.: Q1915

Lab File ID: VX045987.D

Lab Sample ID: VX0430WBL01

Date Analyzed: 04/30/2025

Time Analyzed: 10:55

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0430WBS01	VX0430WBS01	VX045988.D	04/30/2025
VX0430WBSD01	VX0430WBSD01	VX045989.D	04/30/2025
WC-04282025	Q1915-01	VX045995.D	04/30/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915
Lab File ID: VX045524.D BFB Injection Date: 04/01/2025
Instrument ID: MSVOA_X BFB Injection Time: 16:15
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.6
75	30.0 - 60.0% of mass 95	58.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	66.8
175	5.0 - 9.0% of mass 174	5.2 (7.8) 1
176	95.0 - 101.0% of mass 174	65.3 (97.8) 1
177	5.0 - 9.0% of mass 176	4.4 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX045525.D	04/01/2025	17:06
VSTDIC005	VSTDIC005	VX045526.D	04/01/2025	17:29
VSTDIC020	VSTDIC020	VX045527.D	04/01/2025	17:52
VSTDIC050	VSTDIC050	VX045528.D	04/01/2025	18:15
VSTDIC100	VSTDIC100	VX045529.D	04/01/2025	18:38
VSTDIC150	VSTDIC150	VX045530.D	04/01/2025	19:02

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915
Lab File ID: VX045984.D BFB Injection Date: 04/30/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:32
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.5
75	30.0 - 60.0% of mass 95	57.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1 (1.4) 1
174	50.0 - 100.0% of mass 95	69.5
175	5.0 - 9.0% of mass 174	5.3 (7.7) 1
176	95.0 - 101.0% of mass 174	67.1 (96.5) 1
177	5.0 - 9.0% of mass 176	4.5 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045985.D	04/30/2025	10:01
VX0430WBL01	VX0430WBL01	VX045987.D	04/30/2025	10:55
VX0430WBS01	VX0430WBS01	VX045988.D	04/30/2025	11:19
VX0430WBSD01	VX0430WBSD01	VX045989.D	04/30/2025	11:46
WC-04282025	Q1915-01	VX045995.D	04/30/2025	14:06

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915
Lab File ID: VX045985.D Date Analyzed: 04/30/2025
Instrument ID: MSVOA_X Time Analyzed: 10:01
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	83019	5.55	141256	6.75	122754	10.05
UPPER LIMIT	166038	6.049	282512	7.25	245508	10.549
LOWER LIMIT	41509.5	5.049	70628	6.25	61377	9.549
EPA SAMPLE NO.						
WC-04282025	56553	5.55	111268	6.76	107559	10.06
VX0430WBL01	62274	5.54	125239	6.76	116463	10.05
VX0430WBS01	79591	5.54	138531	6.76	123154	10.05
VX0430WBSD01	77091	5.54	132529	6.76	119517	10.05

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915
Lab File ID: VX045985.D Date Analyzed: 04/30/2025
Instrument ID: MSVOA_X Time Analyzed: 10:01
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	60558	12.018				
UPPER LIMIT	121116	12.518				
LOWER LIMIT	30279	11.518				
EPA SAMPLE NO.						
WC-04282025	44562	12.02				
VX0430WBL01	49781	12.02				
VX0430WBS01	58135	12.02				
VX0430WBSD01	57970	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	VX0430WBL01	SDG No.:	Q1915
Lab Sample ID:	VX0430WBL01	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045987.D	1		04/30/25 10:55	VX043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	1.00	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	1.00	U	0.23	1.00	ug/L
78-93-3	2-Butanone	5.00	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	1.00	U	0.25	1.00	ug/L
67-66-3	Chloroform	1.00	U	0.25	1.00	ug/L
71-43-2	Benzene	1.00	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	1.00	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	1.00	U	0.090	1.00	ug/L
127-18-4	Tetrachloroethene	1.00	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	1.00	U	0.12	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	56.3		74 - 125	113%	SPK: 50
1868-53-7	Dibromofluoromethane	52.5		75 - 124	105%	SPK: 50
2037-26-5	Toluene-d8	50.4		86 - 113	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.9		77 - 121	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	62300	5.544			
540-36-3	1,4-Difluorobenzene	125000	6.757			
3114-55-4	Chlorobenzene-d5	116000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	49800	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	VX0430WBS01	SDG No.:	Q1915
Lab Sample ID:	VX0430WBS01	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045988.D	1		04/30/25 11:19	VX043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	17.3		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.5		0.23	1.00	ug/L
78-93-3	2-Butanone	110		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	21.8		0.25	1.00	ug/L
67-66-3	Chloroform	21.6		0.25	1.00	ug/L
71-43-2	Benzene	20.2		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	22.8		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.9		0.090	1.00	ug/L
127-18-4	Tetrachloroethene	19.0		0.23	1.00	ug/L
108-90-7	Chlorobenzene	21.3		0.12	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.6		74 - 125	109%	SPK: 50
1868-53-7	Dibromofluoromethane	56.5		75 - 124	113%	SPK: 50
2037-26-5	Toluene-d8	52.1		86 - 113	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.7		77 - 121	111%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	79600	5.544			
540-36-3	1,4-Difluorobenzene	139000	6.757			
3114-55-4	Chlorobenzene-d5	123000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	58100	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	VX0430WBSD01	SDG No.:	Q1915
Lab Sample ID:	VX0430WBSD01	Matrix:	TCLP
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045989.D	1		04/30/25 11:46	VX043025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	16.6		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.5		0.23	1.00	ug/L
78-93-3	2-Butanone	120		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	21.8		0.25	1.00	ug/L
67-66-3	Chloroform	21.0		0.25	1.00	ug/L
71-43-2	Benzene	20.1		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	22.6		0.22	1.00	ug/L
79-01-6	Trichloroethene	20.2		0.090	1.00	ug/L
127-18-4	Tetrachloroethene	20.1		0.23	1.00	ug/L
108-90-7	Chlorobenzene	20.8		0.12	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.3		74 - 125	109%	SPK: 50
1868-53-7	Dibromofluoromethane	56.5		75 - 124	113%	SPK: 50
2037-26-5	Toluene-d8	52.0		86 - 113	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.5		77 - 121	113%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	77100	5.544			
540-36-3	1,4-Difluorobenzene	133000	6.757			
3114-55-4	Chlorobenzene-d5	120000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	58000	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG No.: Q1915

Instrument ID: MSVOA_X Calibration Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Calibration Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX045525.D		RRF005 = VX045526.D		RRF020 = VX045527.D		
		RRF050 = VX045528.D		RRF100 = VX045529.D		RRF150 = VX045530.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Vinyl Chloride	0.671	0.662	0.701	0.716	0.738	0.730	0.703	4.4
1,1-Dichloroethene	0.563	0.588	0.612	0.604	0.618	0.616	0.600	3.5
2-Butanone	0.474	0.537	0.582	0.586	0.571	0.548	0.550	7.5
Carbon Tetrachloride	0.450	0.497	0.521	0.536	0.545	0.556	0.518	7.5
Chloroform	1.244	1.309	1.348	1.315	1.309	1.309	1.306	2.6
Benzene	1.414	1.416	1.519	1.481	1.483	1.465	1.463	2.8
1,2-Dichloroethane	0.533	0.585	0.649	0.622	0.620	0.617	0.604	6.7
Trichloroethene	0.349	0.322	0.356	0.351	0.351	0.356	0.348	3.7
Tetrachloroethene	0.346	0.373	0.371	0.347	0.333	0.347	0.353	4.5
Chlorobenzene	0.951	1.054	1.123	1.084	1.086	1.112	1.068	5.8
1,2-Dichloroethane-d4		0.962	0.900	0.868	0.904	0.937	0.914	3.9
Dibromofluoromethane		0.372	0.342	0.345	0.353	0.362	0.355	3.5
Toluene-d8		1.257	1.233	1.214	1.230	1.257	1.238	1.5
4-Bromofluorobenzene		0.413	0.448	0.453	0.481	0.460	0.451	5.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG No.: Q1915

Instrument ID: MSVOA_X Calibration Date/Time: 04/30/2025 10:01

Lab File ID: VX045985.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.703	0.712		1.28	20
1,1-Dichloroethene	0.600	0.646		7.67	20
2-Butanone	0.550	0.628		14.18	20
Carbon Tetrachloride	0.518	0.632		22.01	20
Chloroform	1.306	1.474		12.86	20
Benzene	1.463	1.629		11.35	20
1,2-Dichloroethane	0.604	0.716		18.54	20
Trichloroethene	0.348	0.381		9.48	20
Tetrachloroethene	0.353	0.398		12.75	20
Chlorobenzene	1.068	1.246	0.3	16.67	20
1,2-Dichloroethane-d4	0.914	0.976		6.78	20
Dibromofluoromethane	0.355	0.405		14.09	20
Toluene-d8	1.238	1.287		3.96	20
4-Bromofluorobenzene	0.451	0.514		13.97	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

LAB CHRONICLE

OrderID:	Q1915	OrderDate:	4/29/2025 2:29:00 PM
Client:	CDM Smith	Project:	Con Ed UTEN Mount Vernon, NY
Contact:	Marcie Ann Encinas	Location:	L41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1915-01	WC-04282025	TCLP	TCLP BNA	8270E	04/28/25	04/30/25	05/01/25	04/29/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1915
Client: CDM Smith

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								
				0.000				
			Total Svoc :		0.00			
			Total Concentration:		0.00			



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	04/30/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/30/25
Client Sample ID:	PB167774TB	SDG No.:	Q1915
Lab Sample ID:	PB167774TB	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142256.D	1	04/30/25 13:15	05/01/25 13:07	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	50.0	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	50.0	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	50.0	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	100	U	11.0	100	ug/L
67-72-1	Hexachloroethane	50.0	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	50.0	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	50.0	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	50.0	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	50.0	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	50.0	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	50.0	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	100	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	116		10 - 139	78%	SPK: 150
13127-88-3	Phenol-d6	116		10 - 134	77%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.3		49 - 133	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	72.1		52 - 132	72%	SPK: 100
118-79-6	2,4,6-Tribromophenol	131		44 - 137	87%	SPK: 150
1718-51-0	Terphenyl-d14	67.9		48 - 125	68%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	218000		6.904		
1146-65-2	Naphthalene-d8	840000		8.186		
15067-26-2	Acenaphthene-d10	443000		9.945		
1517-22-2	Phenanthrene-d10	773000		11.427		
1719-03-5	Chrysene-d12	539000		14.068		
1520-96-3	Perylene-d12	394000		15.562		

Report of Analysis

Client:	CDM Smith	Date Collected:	04/30/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/30/25
Client Sample ID:	PB167774TB	SDG No.:	Q1915
Lab Sample ID:	PB167774TB	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142256.D	1	04/30/25 13:15	05/01/25 13:07	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	04/28/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/29/25
Client Sample ID:	WC-04282025	SDG No.:	Q1915
Lab Sample ID:	Q1915-01	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142261.D	1	04/30/25 13:15	05/01/25 15:34	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	50.0	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	50.0	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	50.0	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	100	U	11.0	100	ug/L
67-72-1	Hexachloroethane	50.0	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	50.0	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	50.0	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	50.0	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	50.0	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	50.0	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	50.0	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	100	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	101		10 - 139	67%	SPK: 150
13127-88-3	Phenol-d6	91.6		10 - 134	61%	SPK: 150
4165-60-0	Nitrobenzene-d5	86.8		49 - 133	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	75.3		52 - 132	75%	SPK: 100
118-79-6	2,4,6-Tribromophenol	116		44 - 137	78%	SPK: 150
1718-51-0	Terphenyl-d14	57.8		48 - 125	58%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	211000	6.91			
1146-65-2	Naphthalene-d8	797000	8.186			
15067-26-2	Acenaphthene-d10	393000	9.939			
1517-22-2	Phenanthrene-d10	545000	11.427			
1719-03-5	Chrysene-d12	445000	14.062			
1520-96-3	Perylene-d12	444000	15.556			

Report of Analysis

Client:	CDM Smith	Date Collected:	04/28/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/29/25
Client Sample ID:	WC-04282025	SDG No.:	Q1915
Lab Sample ID:	Q1915-01	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142261.D	1	04/30/25 13:15	05/01/25 15:34	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167774TB	PB167774TB	2-Fluorophenol	150	116	78		10	139
		Phenol-d6	150	116	77		10	134
		Nitrobenzene-d5	100	87.3	87		49	133
		2-Fluorobiphenyl	100	72.1	72		52	132
		2,4,6-Tribromophenol	150	131	87		44	137
		Terphenyl-d14	100	67.9	68		48	125
PB167810BL	PB167810BL	2-Fluorophenol	150	110	73		10	139
		Phenol-d6	150	109	73		10	134
		Nitrobenzene-d5	100	83.3	83		49	133
		2-Fluorobiphenyl	100	69.1	69		52	132
		2,4,6-Tribromophenol	150	124	83		44	137
		Terphenyl-d14	100	63.3	63		48	125
PB167810BS	PB167810BS	2-Fluorophenol	150	111	74		10	139
		Phenol-d6	150	113	75		10	134
		Nitrobenzene-d5	100	82.4	82		49	133
		2-Fluorobiphenyl	100	68.5	69		52	132
		2,4,6-Tribromophenol	150	132	88		44	137
		Terphenyl-d14	100	74.1	74		48	125
Q1901-08MS	B-167-SB01MS	2-Fluorophenol	150	96.6	64		10	139
		Phenol-d6	150	89.8	60		10	134
		Nitrobenzene-d5	100	84.0	84		49	133
		2-Fluorobiphenyl	100	70.7	71		52	132
		2,4,6-Tribromophenol	150	126	84		44	137
		Terphenyl-d14	100	70.4	70		48	125
Q1901-08MSD	B-167-SB01MSD	2-Fluorophenol	150	99.4	66		10	139
		Phenol-d6	150	92.8	62		10	134
		Nitrobenzene-d5	100	87.8	88		49	133
		2-Fluorobiphenyl	100	72.3	72		52	132
		2,4,6-Tribromophenol	150	133	88		44	137
		Terphenyl-d14	100	72.8	73		48	125
Q1915-01	WC-04282025	2-Fluorophenol	150	101	67		10	139
		Phenol-d6	150	91.6	61		10	134
		Nitrobenzene-d5	100	86.8	87		49	133
		2-Fluorobiphenyl	100	75.3	75		52	132
		2,4,6-Tribromophenol	150	116	78		44	137
		Terphenyl-d14	100	57.8	58		48	125

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q1901-08MS		Client Sample ID: B-167-SB01MS		DataFile: BF142277.D							
Pyridine	500	24.1	270	ug/L	49				10	109	
1,4-Dichlorobenzene	500	0	340	ug/L	68				55	125	
2-Methylphenol	500	0	360	ug/L	72				60	131	
3+4-Methylphenols	500	0	360	ug/L	72				54	136	
Hexachloroethane	500	0	350	ug/L	70				19	146	
Nitrobenzene	500	0	440	ug/L	88				62	112	
Hexachlorobutadiene	500	0	380	ug/L	76				52	125	
2,4,6-Trichlorophenol	500	0	470	ug/L	94				78	112	
2,4,5-Trichlorophenol	500	0	460	ug/L	92				71	111	
2,4-Dinitrotoluene	500	0	500	ug/L	100				74	137	
Hexachlorobenzene	500	0	440	ug/L	88				72	115	
Pentachlorophenol	1000	0	940	ug/L	94				52	162	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q1901-08MSD		Client Sample ID: B-167-SB01MSD		DataFile: BF142278.D							
Pyridine	500	24.1	280	ug/L	51	4			10	109	20
1,4-Dichlorobenzene	500	0	340	ug/L	68	0			55	125	20
2-Methylphenol	500	0	370	ug/L	74	3			60	131	20
3+4-Methylphenols	500	0	370	ug/L	74	3			54	136	20
Hexachloroethane	500	0	370	ug/L	74	6			19	146	20
Nitrobenzene	500	0	460	ug/L	92	4			62	112	20
Hexachlorobutadiene	500	0	390	ug/L	78	3			52	125	20
2,4,6-Trichlorophenol	500	0	480	ug/L	96	2			78	112	20
2,4,5-Trichlorophenol	500	0	470	ug/L	94	2			71	111	20
2,4-Dinitrotoluene	500	0	530	ug/L	106	6			74	137	20
Hexachlorobenzene	500	0	450	ug/L	90	2			72	115	20
Pentachlorophenol	1000	0	970	ug/L	97	3			52	162	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: 8270E DataFile: BF142255.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167810BS	Pyridine	50	37.2	ug/L	74				29	97	
	1,4-Dichlorobenzene	50	39.6	ug/L	79				76	103	
	2-Methylphenol	50	41.0	ug/L	82				69	109	
	3+4-Methylphenols	50	40.8	ug/L	82				67	106	
	Hexachloroethane	50	41.8	ug/L	84				76	118	
	Nitrobenzene	50	45.8	ug/L	92				58	106	
	Hexachlorobutadiene	50	41.5	ug/L	83				69	101	
	2,4,6-Trichlorophenol	50	43.4	ug/L	87				61	110	
	2,4,5-Trichlorophenol	50	44.9	ug/L	90				70	106	
	2,4-Dinitrotoluene	50	49.8	ug/L	100				60	115	
	Hexachlorobenzene	50	41.6	ug/L	83				73	106	
	Pentachlorophenol	100	93.4	ug/L	93				47	114	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167810BL

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q1915

SAS No.: Q1915 SDG NO.: Q1915

Lab File ID: BF142254.D

Lab Sample ID: PB167810BL

Instrument ID: BNA_F

Date Extracted: 04/30/2025

Matrix: (soil/water) water

Date Analyzed: 05/01/2025

Level: (low/med) LOW

Time Analyzed: 12:10

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167810BS	PB167810BS	BF142255.D	05/01/2025
WC-04282025	Q1915-01	BF142261.D	05/01/2025
B-167-SB01MS	Q1901-08MS	BF142277.D	05/02/2025
B-167-SB01MSD	Q1901-08MSD	BF142278.D	05/02/2025
PB167774TB	PB167774TB	BF142256.D	05/01/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q1915 SDG NO.: Q1915

Lab File ID: BF142238.D

DFTPP Injection Date: 04/30/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.6
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	27.3
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	37.1
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.3
275	10.0 - 60.0% of mass 198	22.8
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.9 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142239.D	04/30/2025	11:24
SSTDICC005	SSTDICC005	BF142240.D	04/30/2025	11:52
SSTDICC010	SSTDICC010	BF142241.D	04/30/2025	12:20
SSTDICC020	SSTDICC020	BF142242.D	04/30/2025	12:49
SSTDICCC040	SSTDICCC040	BF142243.D	04/30/2025	13:17
SSTDICC050	SSTDICC050	BF142244.D	04/30/2025	13:46
SSTDICC060	SSTDICC060	BF142245.D	04/30/2025	14:15
SSTDICC080	SSTDICC080	BF142246.D	04/30/2025	14:43

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q1915 SDG NO.: Q1915

Lab File ID: BF142249.D

DFTPP Injection Date: 05/01/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.4
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	31.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	42.4
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142250.D	05/01/2025	10:17
PB167810BL	PB167810BL	BF142254.D	05/01/2025	12:10
PB167810BS	PB167810BS	BF142255.D	05/01/2025	12:39
PB167774TB	PB167774TB	BF142256.D	05/01/2025	13:07
WC-04282025	Q1915-01	BF142261.D	05/01/2025	15:34

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q1915

SDG NO.: Q1915

Lab File ID: BF142273.D

DFTPP Injection Date: 05/02/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:11

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	34.7
68	Less than 2.0% of mass 69	0.4 (1.6) 1
69	Mass 69 relative abundance	27.6
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	37.6
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.6
275	10.0 - 60.0% of mass 198	23.7
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142274.D	05/02/2025	10:39
B-167-SB01MS	Q1901-08MS	BF142277.D	05/02/2025	12:27
B-167-SB01MSD	Q1901-08MSD	BF142278.D	05/02/2025	12:55

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	186578	6.91	719586	8.19	381198	9.95
UPPER LIMIT	373156	7.41	1439170	8.692	762396	10.445
LOWER LIMIT	93289	6.41	359793	7.692	190599	9.445
EPA SAMPLE NO.						
01 PB167774TB	217516	6.90	839699	8.19	443070	9.95
02 PB167810BL	223997	6.90	857825	8.19	449926	9.95
03 PB167810BS	228833	6.91	892782	8.19	472284	9.95
04 WC-04282025	210959	6.91	797035	8.19	392717	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	647816	11.433	384442	14.074	349272	15.562
UPPER LIMIT	1295630	11.933	768884	14.574	698544	16.062
LOWER LIMIT	323908	10.933	192221	13.574	174636	15.062
EPA SAMPLE NO.						
01 PB167774TB	773002	11.43	539235	14.07	394051	15.56
02 PB167810BL	794711	11.43	568207	14.07	404816	15.56
03 PB167810BS	804874	11.43	464382	14.07	446826	15.57
04 WC-04282025	545118	11.43	445392	14.06	444396	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/02/2025

Lab File ID: BF142274.D Time Analyzed: 10:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	189457	6.91	747235	8.19	389025	9.95
UPPER LIMIT	378914	7.41	1494470	8.692	778050	10.445
LOWER LIMIT	94728.5	6.41	373618	7.692	194513	9.445
EPA SAMPLE NO.						
01 B-167-SB01MS	212656	6.91	808361	8.19	402059	9.95
02 B-167-SB01MSD	217028	6.91	832936	8.19	419407	9.95

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/02/2025

Lab File ID: BF142274.D Time Analyzed: 10:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	652685	11.427	354752	14.068	327562	15.557
UPPER LIMIT	1305370	11.927	709504	14.568	655124	16.057
LOWER LIMIT	326343	10.927	177376	13.568	163781	15.057
EPA SAMPLE NO.						
01 B-167-SB01MS	629997	11.43	370460	14.07	379220	15.56
02 B-167-SB01MSD	684254	11.43	400079	14.07	401381	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BL	SDG No.:	Q1915
Lab Sample ID:	PB167810BL	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142254.D	1	04/30/25 13:15	05/01/25 12:10	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	5.00	U	1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	5.00	U	0.53	5.00	ug/L
95-48-7	2-Methylphenol	5.00	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	10.0	U	1.10	10.0	ug/L
67-72-1	Hexachloroethane	5.00	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	5.00	U	0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	5.00	U	0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	5.00	U	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	5.00	U	0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	5.00	U	1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	5.00	U	0.52	5.00	ug/L
87-86-5	Pentachlorophenol	10.0	U	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	110		10 - 139	73%	SPK: 150
13127-88-3	Phenol-d6	109		10 - 134	73%	SPK: 150
4165-60-0	Nitrobenzene-d5	83.3		49 - 133	83%	SPK: 100
321-60-8	2-Fluorobiphenyl	69.1		52 - 132	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	124		44 - 137	83%	SPK: 150
1718-51-0	Terphenyl-d14	63.3		48 - 125	63%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	224000	6.904			
1146-65-2	Naphthalene-d8	858000	8.187			
15067-26-2	Acenaphthene-d10	450000	9.945			
1517-22-2	Phenanthrene-d10	795000	11.428			
1719-03-5	Chrysene-d12	568000	14.069			
1520-96-3	Perylene-d12	405000	15.563			

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BL	SDG No.:	Q1915
Lab Sample ID:	PB167810BL	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142254.D	1	04/30/25 13:15	05/01/25 12:10	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BS	SDG No.:	Q1915
Lab Sample ID:	PB167810BS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142255.D	1	04/30/25 13:15	05/01/25 12:39	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	37.2		1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	39.6		0.53	5.00	ug/L
95-48-7	2-Methylphenol	41.0		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	40.8		1.10	10.0	ug/L
67-72-1	Hexachloroethane	41.8		0.65	5.00	ug/L
98-95-3	Nitrobenzene	45.8		0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	41.5		0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	43.4		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	44.9		0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	49.8		1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	41.6		0.52	5.00	ug/L
87-86-5	Pentachlorophenol	93.4	E	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	111		10 - 139	74%	SPK: 150
13127-88-3	Phenol-d6	113		10 - 134	75%	SPK: 150
4165-60-0	Nitrobenzene-d5	82.4		49 - 133	82%	SPK: 100
321-60-8	2-Fluorobiphenyl	68.5		52 - 132	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	132		44 - 137	88%	SPK: 150
1718-51-0	Terphenyl-d14	74.1		48 - 125	74%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	229000		6.91		
1146-65-2	Naphthalene-d8	893000		8.192		
15067-26-2	Acenaphthene-d10	472000		9.945		
1517-22-2	Phenanthrene-d10	805000		11.433		
1719-03-5	Chrysene-d12	464000		14.074		
1520-96-3	Perylene-d12	447000		15.568		

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BS	SDG No.:	Q1915
Lab Sample ID:	PB167810BS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142255.D	1	04/30/25 13:15	05/01/25 12:39	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142277.D	1	04/30/25 13:15	05/02/25 12:27	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	270		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	340		5.30	50.0	ug/L
95-48-7	2-Methylphenol	360		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	360		11.0	100	ug/L
67-72-1	Hexachloroethane	350		6.50	50.0	ug/L
98-95-3	Nitrobenzene	440		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	380		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	470		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	460		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	500		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	440		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	940	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	96.6		10 - 139	64%	SPK: 150
13127-88-3	Phenol-d6	89.8		10 - 134	60%	SPK: 150
4165-60-0	Nitrobenzene-d5	84.0		49 - 133	84%	SPK: 100
321-60-8	2-Fluorobiphenyl	70.7		52 - 132	71%	SPK: 100
118-79-6	2,4,6-Tribromophenol	126		44 - 137	84%	SPK: 150
1718-51-0	Terphenyl-d14	70.4		48 - 125	70%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	213000	6.91			
1146-65-2	Naphthalene-d8	808000	8.193			
15067-26-2	Acenaphthene-d10	402000	9.945			
1517-22-2	Phenanthrene-d10	630000	11.433			
1719-03-5	Chrysene-d12	370000	14.069			
1520-96-3	Perylene-d12	379000	15.557			

Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142277.D	1	04/30/25 13:15	05/02/25 12:27	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MSD	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142278.D	1	04/30/25 13:15	05/02/25 12:55	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	280		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	340		5.30	50.0	ug/L
95-48-7	2-Methylphenol	370		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	370		11.0	100	ug/L
67-72-1	Hexachloroethane	370		6.50	50.0	ug/L
98-95-3	Nitrobenzene	460		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	390		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	480		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	470		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	530		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	450		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	970	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	99.4		10 - 139	66%	SPK: 150
13127-88-3	Phenol-d6	92.8		10 - 134	62%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.8		49 - 133	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	72.3		52 - 132	72%	SPK: 100
118-79-6	2,4,6-Tribromophenol	133		44 - 137	88%	SPK: 150
1718-51-0	Terphenyl-d14	72.8		48 - 125	73%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	217000	6.91			
1146-65-2	Naphthalene-d8	833000	8.193			
15067-26-2	Acenaphthene-d10	419000	9.945			
1517-22-2	Phenanthrene-d10	684000	11.433			
1719-03-5	Chrysene-d12	400000	14.069			
1520-96-3	Perylene-d12	401000	15.563			

Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MSD	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142278.D	1	04/30/25 13:15	05/02/25 12:55	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF043025.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 30 16:00:01 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF142239.D 5 =BF142240.D 10 =BF142241.D 20 =BF142242.D 40 =BF142243.D 50 =BF142244.D 60 =BF142245.D 80 =BF142246.D

Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I 1,4-Dichlorobenzen...	-----ISTD-----									
2) 1,4-Dioxane	0.563	0.534	0.561	0.592	0.580	0.555	0.530	0.559	4.03	
3) Pyridine	1.465	1.391	1.454	1.522	1.503	1.460	1.380	1.454	3.62	
4) n-Nitrosodimet...	0.747	0.699	0.728	0.807	0.789	0.766	0.731	0.752	4.95	
5) S 2-Fluorophenol	1.324	1.238	1.265	1.314	1.285	1.210	1.131	1.252	5.35	
6) Aniline	2.212	2.085	2.157	2.245	2.203	2.101	1.981	2.140	4.29	
7) S Phenol-d6	1.559	1.505	1.534	1.593	1.550	1.475	1.370	1.512	4.85	
8) 2-Chlorophenol	1.252	1.232	1.304	1.379	1.354	1.313	1.225	1.294	4.64	
9) Benzaldehyde	1.154	1.088	1.095	1.117	1.078	0.993	0.875	1.057	8.91	
10) C Phenol	1.679	1.588	1.657	1.715	1.687	1.603	1.468	1.628	5.15	
11) bis(2-Chloroet...	1.345	1.250	1.270	1.318	1.285	1.231	1.132	1.262	5.46	
12) 1,3-Dichlorobe...	1.571	1.485	1.485	1.533	1.478	1.396	1.299	1.464	6.18	
13) C 1,4-Dichlorobe...	1.610	1.486	1.495	1.546	1.488	1.404	1.295	1.475	6.86	
14) 1,2-Dichlorobe...	1.489	1.418	1.431	1.475	1.432	1.356	1.255	1.408	5.67	
15) Benzyl Alcohol	1.047	1.017	1.064	1.149	1.126	1.082	1.023	1.073	4.69	
16) 2,2'-oxybis(1-...	2.368	2.287	2.303	2.386	2.344	2.216	2.043	2.278	5.19	
17) 2-Methylphenol	1.085	1.029	1.068	1.121	1.084	1.050	0.983	1.060	4.19	
18) Hexachloroethane	0.468	0.466	0.481	0.511	0.514	0.489	0.451	0.483	4.84	
19) P n-Nitroso-di-n...	0.953	0.987	0.937	0.949	0.967	0.939	0.898	0.844	0.934	4.77
20) 3+4-Methylphenols	1.357	1.323	1.372	1.400	1.366	1.296	1.166	1.326	5.90	
21) I Naphthalene-d8	-----ISTD-----									
22) Acetophenone	0.511	0.482	0.483	0.468	0.461	0.430	0.404	0.463	7.70	
23) S Nitrobenzene-d5	0.217	0.251	0.302	0.340	0.351	0.339	0.330	0.304	16.88	
24) Nitrobenzene	0.229	0.261	0.297	0.328	0.338	0.323	0.314	0.299	13.31	
25) Isophorone	0.660	0.630	0.642	0.660	0.654	0.626	0.601	0.639	3.36	
26) C 2-Nitrophenol	0.048	0.057	0.083	0.115	0.132	0.136	0.142	0.102	38.38#	
27) 2,4-Dimethylph...	0.315	0.319	0.333	0.337	0.333	0.315	0.302	0.322	3.96	
28) bis(2-Chloroet...	0.425	0.404	0.414	0.415	0.410	0.386	0.366	0.403	4.98	
29) C 2,4-Dichloroph...	0.246	0.253	0.274	0.291	0.286	0.272	0.263	0.269	6.12	
30) 1,2,4-Trichlor...	0.320	0.303	0.305	0.312	0.304	0.287	0.273	0.301	5.28	
31) Naphthalene	1.106	1.040	1.035	1.020	0.999	0.924	0.858	0.997	8.21	
32) Benzoic acid		0.048	0.087	0.130	0.146	0.159	0.179	0.125	39.24	
33) 4-Chloroaniline	0.446	0.414	0.429	0.429	0.423	0.396	0.371	0.415	5.98	
34) C Hexachlorobuta...	0.180	0.175	0.179	0.185	0.178	0.170	0.161	0.175	4.35	
35) Caprolactam	0.068	0.073	0.081	0.087	0.089	0.085	0.083	0.081	9.45	
36) C 4-Chloro-3-met...	0.262	0.265	0.279	0.295	0.293	0.276	0.266	0.277	4.84	
37) 2-Methylnaphth...	0.679	0.637	0.633	0.623	0.612	0.574	0.534	0.613	7.69	
38) 1-Methylnaphth...	0.700	0.655	0.667	0.652	0.641	0.590	0.542	0.635	8.29	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF043025.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.621	0.575	0.572	0.576	0.566	0.534	0.496	0.563	6.88	
41) P	Hexachlorocycl...	0.244	0.265	0.306	0.352	0.368	0.358	0.343	0.319	15.23	
42) S	2,4,6-Tribromo...	0.139	0.158	0.184	0.204	0.205	0.197	0.190	0.182	13.73	
43) C	2,4,6-Trichlor...	0.275	0.329	0.353	0.370	0.375	0.381	0.357	0.349	10.50	
44)	2,4,5-Trichlor...	0.322	0.330	0.362	0.404	0.411	0.371	0.365	0.367	9.17	
45) S	2-Fluorobiphenyl	1.738	1.580	1.520	1.409	1.347	1.221	1.114	1.418	15.11	
46)	1,1'-Biphenyl	1.773	1.621	1.612	1.581	1.547	1.435	1.316	1.555	9.38	
47)	2-Chloronaphth...	1.289	1.176	1.191	1.185	1.157	1.085	1.014	1.157	7.51	
48)	2-Nitroaniline	0.119	0.156	0.228	0.295	0.316	0.315	0.316	0.249	33.25	
49)	Acenaphthylene	2.149	2.025	2.025	1.997	1.965	1.828	1.675	1.952	7.93	
50)	Dimethylphthalate	1.362	1.284	1.301	1.321	1.309	1.227	1.152	1.280	5.42	
51)	2,6-Dinitrotol...	0.098	0.137	0.193	0.241	0.259	0.254	0.249	0.204	31.45	
52) C	Acenaphthene	1.271	1.175	1.156	1.172	1.138	1.072	0.995	1.140	7.63	
53)	3-Nitroaniline	0.143	0.187	0.252	0.298	0.311	0.310	0.299	0.257	26.09	
54) P	2,4-Dinitrophenol	0.025	0.036	0.057	0.069	0.076	0.085	0.058		40.36	
55)	Dibenzofuran	1.930	1.775	1.751	1.743	1.713	1.581	1.466	1.709	8.67	
56) P	4-Nitrophenol	0.131	0.177	0.216	0.233	0.230	0.220	0.201		19.82	
57)	2,4-Dinitrotol...	0.104	0.145	0.213	0.285	0.311	0.309	0.306	0.239	35.96	
58)	Fluorene	1.493	1.403	1.351	1.317	1.273	1.186	1.071	1.299	10.74	
59)	2,3,4,6-Tetrac...	0.252	0.267	0.312	0.334	0.334	0.323	0.305	0.304	10.65	
60)	Diethylphthalate	1.324	1.264	1.302	1.313	1.300	1.215	1.126	1.263	5.61	
61)	4-Chlorophenyl...	0.702	0.642	0.632	0.632	0.614	0.570	0.526	0.617	9.06	
62)	4-Nitroaniline	0.144	0.177	0.233	0.272	0.287	0.281	0.269	0.238	23.57	
63)	Azobenzene	1.346	1.260	1.275	1.279	1.278	1.198	1.098	1.248	6.33	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...	0.022	0.034	0.055	0.065	0.071	0.075	0.054		39.23	
66) c	n-Nitrosodiphe...	0.727	0.686	0.693	0.712	0.689	0.653	0.623	0.683	5.11	
67)	4-Bromophenyl-...	0.224	0.219	0.223	0.236	0.225	0.220	0.211	0.223	3.39	
68)	Hexachlorobenzene	0.256	0.245	0.248	0.261	0.255	0.244	0.237	0.249	3.28	
69)	Atrazine	0.166	0.172	0.186	0.190	0.190	0.182	0.174	0.180	5.29	
70) C	Pentachlorophenol	0.093	0.121	0.144	0.145	0.144	0.144	0.132		16.25	
71)	Phenanthrene	1.218	1.114	1.119	1.112	1.064	0.995	0.940	1.080	8.45	
72)	Anthracene	1.209	1.161	1.151	1.128	1.099	1.025	0.959	1.105	7.77	
73)	Carbazole	1.125	1.047	1.061	1.033	0.996	0.935	0.857	1.008	8.77	
74)	Di-n-butylphth...	0.986	1.006	1.088	1.069	1.045	0.988	0.912	1.013	5.88	
75) C	Fluoranthene	1.221	1.144	1.155	1.085	1.035	0.957	0.878	1.068	11.24	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.702	0.748	0.885	0.963	0.945	0.895	0.810	0.850	11.67	
78)	Pyrene	1.833	1.801	1.849	2.026	1.937	1.833	1.635	1.845	6.54	
79) S	Terphenyl-d14	1.472	1.413	1.422	1.481	1.392	1.290	1.155	1.375	8.41	
80)	Butylbenzylpht...	0.218	0.293	0.405	0.505	0.525	0.524	0.509	0.425	29.32	
81)	Benzo(a)anthra...	1.379	1.320	1.332	1.438	1.388	1.316	1.225	1.343	5.06	
82)	3,3'-Dichlorob...	0.286	0.317	0.380	0.442	0.450	0.451	0.435	0.394	17.37	
83)	Chrysene	1.302	1.210	1.249	1.232	1.227	1.197	1.138	1.222	4.12	
84)	Bis(2-ethylhex...	0.384	0.448	0.549	0.695	0.726	0.722	0.707	0.604	23.69	
85) c	Di-n-octyl pht...	0.668	0.875	1.262	1.320	1.347	1.329	1.133		25.55	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF043025.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.323	1.384	1.466	1.613	1.556	1.490	1.422	1.465	6.78
88)	Benzo(b)fluora...	1.317	1.220	1.208	1.302	1.271	1.177	1.202	1.242	4.36
89)	Benzo(k)fluora...	1.211	1.159	1.222	1.145	1.146	1.110	0.962	1.136	7.59
90) C	Benzo(a)pyrene	1.121	1.098	1.146	1.203	1.180	1.137	1.076	1.137	3.91
91)	Dibenzo(a,h)an...	1.075	1.131	1.197	1.312	1.271	1.192	1.144	1.189	6.90
92)	Benzo(g,h,i)pe...	1.076	1.138	1.204	1.308	1.267	1.213	1.162	1.195	6.55

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG No.: Q1915

Instrument ID: BNA_F Calibration Date/Time: 05/01/2025 10:17

Lab File ID: BF142250.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.532		5.4	
2-Fluorophenol	1.252	1.250		-0.2	
Phenol-d6	1.512	1.575		4.2	
1,4-Dichlorobenzene	1.475	1.526		3.5	20.0
2-Methylphenol	1.060	1.103		4.1	
3+4-Methylphenols	1.326	1.404		5.9	
Nitrobenzene-d5	0.304	0.333		9.5	
Hexachloroethane	0.483	0.505		4.6	
Nitrobenzene	0.299	0.320		7.0	
Hexachlorobutadiene	0.175	0.186		6.3	20.0
2,4,6-Trichlorophenol	0.349	0.354		1.4	20.0
2-Fluorobiphenyl	1.418	1.409		-0.6	
2,4,5-Trichlorophenol	0.367	0.386		5.2	
2,4-Dinitrotoluene	0.239	0.289		20.9	
2,4,6-Tribromophenol	0.182	0.196		7.7	
Hexachlorobenzene	0.249	0.257		3.2	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.422		3.4	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG No.: Q1915

Instrument ID: BNA_F Calibration Date/Time: 05/02/2025 10:39

Lab File ID: BF142274.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.384		-4.8	
2-Fluorophenol	1.252	1.195		-4.6	
Phenol-d6	1.512	1.457		-3.6	
1,4-Dichlorobenzene	1.475	1.364		-7.5	20.0
2-Methylphenol	1.060	1.005		-5.2	
3+4-Methylphenols	1.326	1.295		-2.3	
Nitrobenzene-d5	0.304	0.332		9.2	
Hexachloroethane	0.483	0.473		-2.1	
Nitrobenzene	0.299	0.315		5.4	
Hexachlorobutadiene	0.175	0.164		-6.3	20.0
2,4,6-Trichlorophenol	0.349	0.360		3.2	20.0
2-Fluorobiphenyl	1.418	1.293		-8.8	
2,4,5-Trichlorophenol	0.367	0.363		-1.1	
2,4-Dinitrotoluene	0.239	0.314		31.4	
2,4,6-Tribromophenol	0.182	0.190		4.4	
Hexachlorobenzene	0.249	0.232		-6.8	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.344		-2.3	

All other compounds must meet a minimum RRF of 0.010.

LAB CHRONICLE

OrderID:	Q1915	OrderDate:	4/29/2025 2:29:00 PM
Client:	CDM Smith	Project:	Con Ed UTEN Mount Vernon, NY
Contact:	Marcie Ann Encinas	Location:	L41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1915-01	WC-04282025	TCLP			04/28/25			04/29/25
			TCLP ICP Metals	6010D		04/30/25	05/01/25	
			TCLP Mercury	7470A		04/30/25	05/01/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1915 **Order ID:** Q1915
Client: CDM Smith **Project ID:** Con Ed UTEN Mount Vernon, NY

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : WC-04282025								
Q1915-01	WC-04282025	TCLP	Barium	4600		72.8	500	ug/L
Q1915-01	WC-04282025	TCLP	Cadmium	7.22	J	2.50	30.0	ug/L
Q1915-01	WC-04282025	TCLP	Lead	896		11.5	60.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	04/28/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/29/25
Client Sample ID:	WC-04282025	SDG No.:	Q1915
Lab Sample ID:	Q1915-01	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-38-2	Arsenic	100	U	1	25.6	100	ug/L	04/30/25 13:35	05/01/25 20:07	SW6010	SW3050
7440-39-3	Barium	4600		1	72.8	500	ug/L	04/30/25 13:35	05/01/25 20:07	SW6010	SW3050
7440-43-9	Cadmium	7.22	J	1	2.50	30.0	ug/L	04/30/25 13:35	05/01/25 20:07	SW6010	SW3050
7440-47-3	Chromium	50.0	U	1	10.6	50.0	ug/L	04/30/25 13:35	05/01/25 20:07	SW6010	SW3050
7439-92-1	Lead	896		1	11.5	60.0	ug/L	04/30/25 13:35	05/01/25 20:07	SW6010	SW3050
7439-97-6	Mercury	2.00	U	1	0.76	2.00	ug/L	04/30/25 13:30	05/01/25 14:43	SW7470A	
7782-49-2	Selenium	100	U	1	48.2	100	ug/L	04/30/25 13:35	05/01/25 20:07	SW6010	SW3050
7440-22-4	Silver	50.0	U	1	8.10	50.0	ug/L	04/30/25 13:35	05/01/25 20:07	SW6010	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP METALS			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits



METAL CALIBRATION DATA



—

- 2a -

1

1

1

1

1

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV62	Mercury	4.80	5.0	96	90 - 110	CV	05/01/2025	11:58	LB135623
CCV63	Mercury	5.01	5.0	100	90 - 110	CV	05/01/2025	12:49	LB135623
CCV64	Mercury	5.09	5.0	102	90 - 110	CV	05/01/2025	13:19	LB135623
CCV65	Mercury	4.77	5.0	95	90 - 110	CV	05/01/2025	13:58	LB135623
CCV66	Mercury	4.60	5.0	92	90 - 110	CV	05/01/2025	14:29	LB135623
CCV67	Mercury	4.61	5.0	92	90 - 110	CV	05/01/2025	15:23	LB135623

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q1915
 Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1010	1000	100	90 - 110	P	05/01/2025	13:44	LB135636
	Barium	495	520	95	90 - 110	P	05/01/2025	13:44	LB135636
	Cadmium	508	510	100	90 - 110	P	05/01/2025	13:44	LB135636
	Chromium	506	520	97	90 - 110	P	05/01/2025	13:44	LB135636
	Lead	989	1000	99	90 - 110	P	05/01/2025	13:44	LB135636
	Selenium	1040	1000	104	90 - 110	P	05/01/2025	13:44	LB135636
	Silver	240	250	96	90 - 110	P	05/01/2025	13:44	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	23.7	20.0	119	80 - 120	P	05/01/2025	14:01	LB135636
	Barium	95.8	100	96	80 - 120	P	05/01/2025	14:01	LB135636
	Cadmium	6.39	6.0	107	80 - 120	P	05/01/2025	14:01	LB135636
	Chromium	10.5	10.0	105	80 - 120	P	05/01/2025	14:01	LB135636
	Lead	13.1	12.0	109	80 - 120	P	05/01/2025	14:01	LB135636
	Selenium	22.9	20.0	114	80 - 120	P	05/01/2025	14:01	LB135636
	Silver	10.2	10.0	102	80 - 120	P	05/01/2025	14:01	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	15:05	LB135636
	Barium	9650	10000	96	90 - 110	P	05/01/2025	15:05	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	15:05	LB135636
	Chromium	1000	1000	100	90 - 110	P	05/01/2025	15:05	LB135636
	Lead	5010	5000	100	90 - 110	P	05/01/2025	15:05	LB135636
	Selenium	5140	5000	103	90 - 110	P	05/01/2025	15:05	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	15:05	LB135636
CCV02	Arsenic	4930	5000	99	90 - 110	P	05/01/2025	16:00	LB135636
	Barium	10200	10000	102	90 - 110	P	05/01/2025	16:00	LB135636
	Cadmium	2470	2500	99	90 - 110	P	05/01/2025	16:00	LB135636
	Chromium	995	1000	100	90 - 110	P	05/01/2025	16:00	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	16:00	LB135636
	Selenium	4960	5000	99	90 - 110	P	05/01/2025	16:00	LB135636
	Silver	1220	1250	98	90 - 110	P	05/01/2025	16:00	LB135636
CCV03	Arsenic	5040	5000	101	90 - 110	P	05/01/2025	17:03	LB135636
	Barium	9950	10000	100	90 - 110	P	05/01/2025	17:03	LB135636
	Cadmium	2510	2500	100	90 - 110	P	05/01/2025	17:03	LB135636
	Chromium	988	1000	99	90 - 110	P	05/01/2025	17:03	LB135636
	Lead	5010	5000	100	90 - 110	P	05/01/2025	17:03	LB135636
	Selenium	5060	5000	101	90 - 110	P	05/01/2025	17:03	LB135636
	Silver	1220	1250	97	90 - 110	P	05/01/2025	17:03	LB135636
CCV04	Arsenic	5160	5000	103	90 - 110	P	05/01/2025	17:49	LB135636
	Barium	10000	10000	100	90 - 110	P	05/01/2025	17:49	LB135636
	Cadmium	2470	2500	99	90 - 110	P	05/01/2025	17:49	LB135636
	Chromium	987	1000	99	90 - 110	P	05/01/2025	17:49	LB135636
	Lead	4950	5000	99	90 - 110	P	05/01/2025	17:49	LB135636
	Selenium	5240	5000	105	90 - 110	P	05/01/2025	17:49	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	17:49	LB135636
CCV05	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	18:36	LB135636
	Barium	9950	10000	100	90 - 110	P	05/01/2025	18:36	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	18:36	LB135636
	Chromium	984	1000	98	90 - 110	P	05/01/2025	18:36	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	18:36	LB135636
	Selenium	5150	5000	103	90 - 110	P	05/01/2025	18:36	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q1915
Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1230	1250	98	90 - 110	P	05/01/2025	18:36	LB135636
CCV06	Arsenic	4990	5000	100	90 - 110	P	05/01/2025	19:23	LB135636
	Barium	9830	10000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Cadmium	2460	2500	98	90 - 110	P	05/01/2025	19:23	LB135636
	Chromium	980	1000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Lead	4920	5000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Selenium	5030	5000	101	90 - 110	P	05/01/2025	19:23	LB135636
	Silver	1240	1250	99	90 - 110	P	05/01/2025	19:23	LB135636
CCV07	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	20:11	LB135636
	Barium	9790	10000	98	90 - 110	P	05/01/2025	20:11	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	20:11	LB135636
	Chromium	995	1000	100	90 - 110	P	05/01/2025	20:11	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	20:11	LB135636
	Selenium	5120	5000	102	90 - 110	P	05/01/2025	20:11	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	20:11	LB135636
CCV08	Arsenic	5180	5000	104	90 - 110	P	05/01/2025	20:24	LB135636
	Barium	9880	10000	99	90 - 110	P	05/01/2025	20:24	LB135636
	Cadmium	2490	2500	100	90 - 110	P	05/01/2025	20:24	LB135636
	Chromium	980	1000	98	90 - 110	P	05/01/2025	20:24	LB135636
	Lead	4990	5000	100	90 - 110	P	05/01/2025	20:24	LB135636
	Selenium	5270	5000	105	90 - 110	P	05/01/2025	20:24	LB135636
	Silver	1240	1250	99	90 - 110	P	05/01/2025	20:24	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q1915
Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1060	1000	106	90 - 110	P	05/02/2025	13:47	LB135654
	Barium	518	520	100	90 - 110	P	05/02/2025	13:47	LB135654
	Cadmium	493	510	97	90 - 110	P	05/02/2025	13:47	LB135654
	Chromium	503	520	97	90 - 110	P	05/02/2025	13:47	LB135654
	Lead	966	1000	97	90 - 110	P	05/02/2025	13:47	LB135654
	Selenium	1040	1000	104	90 - 110	P	05/02/2025	13:47	LB135654
	Silver	230	250	92	90 - 110	P	05/02/2025	13:47	LB135654

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client:	<u>CDM Smith</u>	SDG No.:	<u>Q1915</u>	
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q1915</u> SAS No.: <u>Q1915</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	Inorganic Ventures			

Sample ID	Analyte	Result	True Value	%	Acceptance	M	Analysis Date	Analysis Time	Run Number
		ug/L		Recovery	Window (%R)				
LLICV01	Arsenic	19.0	20.0	95	80 - 120	P	05/02/2025	15:03	LB135654
	Barium	81.5	100	82	80 - 120	P	05/02/2025	15:03	LB135654
	Cadmium	5.51	6.0	92	80 - 120	P	05/02/2025	15:03	LB135654
	Chromium	10.1	10.0	101	80 - 120	P	05/02/2025	15:03	LB135654
	Lead	11.1	12.0	93	80 - 120	P	05/02/2025	15:03	LB135654
	Selenium	16.8	20.0	84	80 - 120	P	05/02/2025	15:03	LB135654
	Silver	10.8	10.0	108	80 - 120	P	05/02/2025	15:03	LB135654

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q1915
Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5450	5000	109	90 - 110	P	05/02/2025	16:01	LB135654
	Barium	9540	10000	95	90 - 110	P	05/02/2025	16:01	LB135654
	Cadmium	2440	2500	98	90 - 110	P	05/02/2025	16:01	LB135654
	Chromium	1010	1000	101	90 - 110	P	05/02/2025	16:01	LB135654
	Lead	4820	5000	96	90 - 110	P	05/02/2025	16:01	LB135654
	Selenium	5460	5000	109	90 - 110	P	05/02/2025	16:01	LB135654
	Silver	1270	1250	102	90 - 110	P	05/02/2025	16:01	LB135654
CCV02	Arsenic	5240	5000	105	90 - 110	P	05/02/2025	16:34	LB135654
	Barium	9720	10000	97	90 - 110	P	05/02/2025	16:34	LB135654
	Cadmium	2390	2500	96	90 - 110	P	05/02/2025	16:34	LB135654
	Chromium	1000	1000	100	90 - 110	P	05/02/2025	16:34	LB135654
	Lead	4740	5000	95	90 - 110	P	05/02/2025	16:34	LB135654
	Selenium	5240	5000	105	90 - 110	P	05/02/2025	16:34	LB135654
	Silver	1260	1250	101	90 - 110	P	05/02/2025	16:34	LB135654

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	86	40 - 160	CV	05/01/2025	12:05	LB135623
CRI01	Arsenic	23.3	20.0	116	40 - 160	P	05/01/2025	14:16	LB135636
	Barium	93.6	100	94	40 - 160	P	05/01/2025	14:16	LB135636
	Cadmium	6.36	6.0	106	40 - 160	P	05/01/2025	14:16	LB135636
	Chromium	10.6	10.0	106	40 - 160	P	05/01/2025	14:16	LB135636
	Lead	13.3	12.0	111	40 - 160	P	05/01/2025	14:16	LB135636
	Selenium	23.1	20.0	115	40 - 160	P	05/01/2025	14:16	LB135636
	Silver	10.6	10.0	106	40 - 160	P	05/01/2025	14:16	LB135636
CRI01	Arsenic	21.5	20.0	108	40 - 160	P	05/02/2025	15:12	LB135654
	Barium	80.2	100	80	40 - 160	P	05/02/2025	15:12	LB135654
	Cadmium	5.66	6.0	94	40 - 160	P	05/02/2025	15:12	LB135654
	Chromium	10.6	10.0	106	40 - 160	P	05/02/2025	15:12	LB135654
	Lead	12.1	12.0	101	40 - 160	P	05/02/2025	15:12	LB135654
	Selenium	23.5	20.0	117	40 - 160	P	05/02/2025	15:12	LB135654
	Silver	10.7	10.0	107	40 - 160	P	05/02/2025	15:12	LB135654



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	CDM Smith				SDG No.:	Q1915				
Contract:	CAMP02	Lab Code:	CHEM		Case No.:	Q1915		SAS No.:	Q1915	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB112	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	11:49	LB135623	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB62	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	12:01	LB135623
CCB63	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	12:51	LB135623
CCB64	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	13:23	LB135623
CCB65	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	14:02	LB135623
CCB66	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	14:40	LB135623
CCB67	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	15:25	LB135623

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	14:11	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	14:11	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	14:11	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	14:11	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	14:11	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	14:11	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	14:11	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1915							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1915	SAS No.: Q1915						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	15:12	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	15:12	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	15:12	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	15:12	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	15:12	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	15:12	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	15:12	LB135636
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	16:04	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	16:04	LB135636
	Cadmium	0.60	+/-6.00	J	6.00	P	05/01/2025	16:04	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	16:04	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	16:04	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	16:04	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	16:04	LB135636
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	17:07	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	17:07	LB135636
	Cadmium	0.88	+/-6.00	J	6.00	P	05/01/2025	17:07	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	17:07	LB135636
	Lead	3.14	+/-12.0	J	12.0	P	05/01/2025	17:07	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	17:07	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	17:07	LB135636
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	17:53	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	17:53	LB135636
	Cadmium	0.92	+/-6.00	J	6.00	P	05/01/2025	17:53	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	17:53	LB135636
	Lead	2.63	+/-12.0	J	12.0	P	05/01/2025	17:53	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	17:53	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	17:53	LB135636
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	18:40	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	18:40	LB135636
	Cadmium	0.89	+/-6.00	J	6.00	P	05/01/2025	18:40	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	18:40	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	18:40	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	18:40	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	18:40	LB135636
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	19:27	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	19:27	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	19:27	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	19:27	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1915							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1915	SAS No.: Q1915						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	19:27	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	19:27	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	19:27	LB135636
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	20:15	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	20:15	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	20:15	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	20:15	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	20:15	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	20:15	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	20:15	LB135636
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	20:28	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	20:28	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	20:28	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	20:28	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	20:28	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	20:28	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	20:28	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/02/2025	15:08	LB135654
	Barium	100	+/-100	U	100	P	05/02/2025	15:08	LB135654
	Cadmium	6.00	+/-6.00	U	6.00	P	05/02/2025	15:08	LB135654
	Chromium	10.0	+/-10.0	U	10.0	P	05/02/2025	15:08	LB135654
	Lead	12.0	+/-12.0	U	12.0	P	05/02/2025	15:08	LB135654
	Selenium	20.0	+/-20.0	U	20.0	P	05/02/2025	15:08	LB135654
	Silver	10.0	+/-10.0	U	10.0	P	05/02/2025	15:08	LB135654

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1915							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1915	SAS No.: Q1915						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/02/2025	16:05	LB135654
	Barium	100	+/-100	U	100	P	05/02/2025	16:05	LB135654
	Cadmium	6.00	+/-6.00	U	6.00	P	05/02/2025	16:05	LB135654
	Chromium	10.0	+/-10.0	U	10.0	P	05/02/2025	16:05	LB135654
	Lead	12.0	+/-12.0	U	12.0	P	05/02/2025	16:05	LB135654
	Selenium	20.0	+/-20.0	U	20.0	P	05/02/2025	16:05	LB135654
	Silver	10.0	+/-10.0	U	10.0	P	05/02/2025	16:05	LB135654
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/02/2025	16:38	LB135654
	Barium	100	+/-100	U	100	P	05/02/2025	16:38	LB135654
	Cadmium	6.00	+/-6.00	U	6.00	P	05/02/2025	16:38	LB135654
	Chromium	10.0	+/-10.0	U	10.0	P	05/02/2025	16:38	LB135654
	Lead	12.0	+/-12.0	U	12.0	P	05/02/2025	16:38	LB135654
	Selenium	20.0	+/-20.0	U	20.0	P	05/02/2025	16:38	LB135654
	Silver	10.0	+/-10.0	U	10.0	P	05/02/2025	16:38	LB135654

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q1915

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167774TB	WATER			Batch Number:	PB167806		Prep Date:	04/30/2025	
	Mercury	2.00	<2.00	U	2.00	CV	05/01/2025	14:45	LB135623
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167806BL	WATER			Batch Number:	PB167806		Prep Date:	04/30/2025	
	Mercury	0.20	<0.20	U	0.20	CV	05/01/2025	12:28	LB135623

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q1915

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167774TB		WATER		Batch Number:	PB167808		Prep Date:	04/30/2025	
	Arsenic	100	<100	U	100	P	05/01/2025	16:37	LB135636
	Barium	500	<500	U	500	P	05/01/2025	16:37	LB135636
	Cadmium	30.0	<30.0	U	30.0	P	05/01/2025	16:37	LB135636
	Chromium	50.0	<50.0	U	50.0	P	05/01/2025	16:37	LB135636
	Lead	60.0	<60.0	U	60.0	P	05/01/2025	16:37	LB135636
	Selenium	100	<100	U	100	P	05/01/2025	16:37	LB135636
	Silver	50.0	<50.0	U	50.0	P	05/01/2025	16:37	LB135636
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167805TB		WATER		Batch Number:	PB167808		Prep Date:	04/30/2025	
	Arsenic	100	<100	U	100	P	05/01/2025	20:20	LB135636
	Barium	500	<500	U	500	P	05/01/2025	20:20	LB135636
	Cadmium	30.0	<30.0	U	30.0	P	05/01/2025	20:20	LB135636
	Chromium	15.9	<50.0	J	50.0	P	05/01/2025	20:20	LB135636
	Lead	60.0	<60.0	U	60.0	P	05/01/2025	20:20	LB135636
	Selenium	100	<100	U	100	P	05/01/2025	20:20	LB135636
	Silver	50.0	<50.0	U	50.0	P	05/01/2025	20:20	LB135636
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167808BL		WATER		Batch Number:	PB167808		Prep Date:	04/30/2025	
	Arsenic	100	<100	U	100	P	05/01/2025	16:23	LB135636
	Barium	500	<500	U	500	P	05/01/2025	16:23	LB135636
	Cadmium	30.0	<30.0	U	30.0	P	05/01/2025	16:23	LB135636
	Chromium	50.0	<50.0	U	50.0	P	05/01/2025	16:23	LB135636
	Lead	60.0	<60.0	U	60.0	P	05/01/2025	16:23	LB135636
	Selenium	100	<100	U	100	P	05/01/2025	16:23	LB135636
	Silver	50.0	<50.0	U	50.0	P	05/01/2025	16:23	LB135636

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: CDM Smith	SDG No.: Q1915
Contract: CAMP02	Lab Code: CHEM
ICS Source: EPA	Case No.: Q1915
	SAS No.: Q1915
	Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	2.11			-20	20	05/01/2025	14:35	LB135636
	Barium	0.52	6.0	9	-94	106	05/01/2025	14:35	LB135636
	Cadmium	-4.57	1.0	457	-5	7	05/01/2025	14:35	LB135636
	Chromium	53.0	52.0	102	42	62	05/01/2025	14:35	LB135636
	Lead	-4.15			-12	12	05/01/2025	14:35	LB135636
	Selenium	-9.38			-20	20	05/01/2025	14:35	LB135636
	Silver	1.45			-10	10	05/01/2025	14:35	LB135636
ICSAB01	Arsenic	120	104	115	88.4	120	05/01/2025	14:39	LB135636
	Barium	487	537	91	437	637	05/01/2025	14:39	LB135636
	Cadmium	1080	972	111	826	1120	05/01/2025	14:39	LB135636
	Chromium	583	542	108	460	624	05/01/2025	14:39	LB135636
	Lead	52.5	49.0	107	37	61	05/01/2025	14:39	LB135636
	Selenium	45.9	46.0	100	26	66	05/01/2025	14:39	LB135636
	Silver	203	201	101	170	232	05/01/2025	14:39	LB135636
ICSA01	Arsenic	0.098			-20	20	05/02/2025	15:30	LB135654
	Barium	-5.67	6.0	94	-94	106	05/02/2025	15:30	LB135654
	Cadmium	-2.89	1.0	289	-5	7	05/02/2025	15:30	LB135654
	Chromium	53.8	52.0	104	42	62	05/02/2025	15:30	LB135654
	Lead	4.14			-12	12	05/02/2025	15:30	LB135654
	Selenium	-1.90			-20	20	05/02/2025	15:30	LB135654
	Silver	-0.66			-10	10	05/02/2025	15:30	LB135654
ICSAB01	Arsenic	115	104	111	88.4	120	05/02/2025	15:46	LB135654
	Barium	464	537	86	437	637	05/02/2025	15:46	LB135654
	Cadmium	996	972	102	826	1120	05/02/2025	15:46	LB135654
	Chromium	560	542	103	460	624	05/02/2025	15:46	LB135654
	Lead	51.7	49.0	106	37	61	05/02/2025	15:46	LB135654
	Selenium	57.3	46.0	125	26	66	05/02/2025	15:46	LB135654
	Silver	193	201	96	170	232	05/02/2025	15:46	LB135654



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: CDM Smith	level: low	sdg no.: Q1915
contract: CAMP02	lab code: CHEM	case no.: Q1915 sas no.: Q1915
matrix: Water	sample id: Q1901-08	client id: B-167-SB01MS
Percent Solids for Sample: NA	Spiked ID: Q1901-08MS	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	39.1		2.00	U	40.0	98		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: CDM Smith	level: low	sdg no.: Q1915
contract: CAMP02	lab code: CHEM	case no.: Q1915 sas no.: Q1915
matrix: Water	sample id: Q1901-08	client id: B-167-SB01MSD
Percent Solids for Sample: NA	Spiked ID: Q1901-08MSD	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	40.9		2.00	U	40.0	102		CV

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>CDM Smith</u>	level: <u>low</u>	sdg no.: <u>Q1915</u>
contract: <u>CAMP02</u>	lab code: <u>CHEM</u>	case no.: <u>Q1915</u> sas no.: <u>Q1915</u>
matrix: <u>Water</u>	sample id: <u>Q1913-02</u>	client id: <u>WC-12-A-202504MS</u>
Percent Solids for Sample: NA	Spiked ID: Q1913-02MS	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	4220		100	U	4000	106		P
Barium	ug/L	75 - 125	1020		500	U	1000	102		P
Cadmium	ug/L	75 - 125	943		30.0	U	1000	94		P
Chromium	ug/L	75 - 125	2180		122		2000	103		P
Lead	ug/L	75 - 125	4570		60.0	U	5000	91		P
Selenium	ug/L	75 - 125	10100		100	U	10000	101		P
Silver	ug/L	75 - 125	385		50.0	U	380	101		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>CDM Smith</u>	level:	<u>low</u>	sdg no.:	<u>Q1915</u>
contract:	<u>CAMP02</u>	lab code:	<u>CHEM</u>	case no.:	<u>Q1915</u>
matrix:	<u>Water</u>	sample id:	<u>Q1913-02</u>	client id:	<u>WC-12-A-202504MSD</u>
Percent Solids for Sample:	<u>NA</u>	Spiked ID:	<u>Q1913-02MSD</u>	Percent Solids for Spike Sample:	<u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	4060		100	U	4000	102		P
Barium	ug/L	75 - 125	965		500	U	1000	96		P
Cadmium	ug/L	75 - 125	908		30.0	U	1000	91		P
Chromium	ug/L	75 - 125	2100		122		2000	99		P
Lead	ug/L	75 - 125	4400		60.0	U	5000	88		P
Selenium	ug/L	75 - 125	9630		100	U	10000	96		P
Silver	ug/L	75 - 125	366		50.0	U	380	96		P

Metals
- 5b -

Client: CDM Smith

SDG No.: Q1915

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1915

SAS No.: Q1915

Matrix:

Level: LOW

Client ID:

Sample ID:

Spiked ID:

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: CDM Smith **Level:** LOW **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Matrix: Water **Sample ID:** Q1901-08 **Client ID:** B-167-SB01DUP
Percent Solids for Sample: NA **Duplicate ID** Q1901-08DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	2.00	U	2.00	U			CV

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>CDM Smith</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q1915</u>
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1915</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q1901-08MS</u>	Client ID:	<u>B-167-SB01MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q1901-08MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	39.1		40.9		4		CV

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: CDM Smith **Level:** LOW **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Matrix: Water **Sample ID:** Q1913-02 **Client ID:** WC-12-A-202504DUP
Percent Solids for Sample: NA **Duplicate ID** Q1913-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	500	U	500	U			P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	122		113		8		P
Lead	ug/L	20	60.0	U	60.0	U			P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: CDM Smith **Level:** LOW **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Matrix: Water **Sample ID:** Q1913-02MS **Client ID:** WC-12-A-202504MSD
Percent Solids for Sample: NA **Duplicate ID** Q1913-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4220		4060		4		P
Barium	ug/L	20	1020		965		6		P
Cadmium	ug/L	20	943		908		4		P
Chromium	ug/L	20	2180		2100		4		P
Lead	ug/L	20	4570		4400		4		P
Selenium	ug/L	20	10100		9630		5		P
Silver	ug/L	20	385		366		5		P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167806BS Mercury	ug/L	4.0	3.34		84	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167808BS							
Arsenic	ug/L	4000	3770		94	80 - 120	P
Barium	ug/L	1000	826		83	80 - 120	P
Cadmium	ug/L	1000	943		94	80 - 120	P
Chromium	ug/L	2000	1870		94	80 - 120	P
Lead	ug/L	5000	4630		93	80 - 120	P
Selenium	ug/L	10000	9600		96	80 - 120	P
Silver	ug/L	380	346		91	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

B-167-SB01L

Lab Name: Chemtech Consulting Group **Contract:** CAMP02
Lab Code: CHEM **Lb No.:** lb135623 **Lab Sample ID :** Q1901-08L **SDG No.:** Q1915
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)	Serial Dilution Result (S)	% Differ- ence	Q	M
Mercury	2.00 U	10.0 U			CV

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

WC-12-A-202504L

Lab Name: Chemtech Consulting Group **Contract:** CAMP02
Lab Code: CHEM **Lb No.:** lb135654 **Lab Sample ID :** Q1913-02L **SDG No.:** Q1915
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Arsenic	100	U	500	U			P
Barium	500	U	2500	U			P
Cadmium	30.0	U	150	U			P
Chromium	122		122	J	0		P
Lead	60.0	U	300	U			P
Selenium	100	U	500	U			P
Silver	50.0	U	250	U			P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q1915

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1915

SAS No.: Q1915

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Arsenic	193.759	0.0000000	0.0000000	-0.0000440	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000930	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	-0.0000920	0.0000000	0.0000380	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	-0.0001440	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	-0.0001490	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith SDG No.: Q1915
 Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915
 Instrument ID: _____ Date: _____
 Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0002870
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0003170	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	-0.0003570
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q1915

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1915

SAS No.: Q1915

Instrument ID:

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

		ICP Interelement Correction Factors For:				
Analyte	Wave-Length (nm)	Cr	Cu	K	Mn	Mo
Arsenic	193.759	-0.0029000	0.0000000	0.0000000	0.0000000	0.0004900
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0001400	-0.0008600
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0007460	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	-0.0000120

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q1915

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1915

SAS No.: Q1915

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0006580	0.0000000	0.0000000	0.0001290
Selenium	196.090	0.0000000	0.0000000	0.0003330	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q1915

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1915

SAS No.: Q1915

Instrument ID:

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave-Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000630	0.0001280	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Lead	220.353	0.0000000	-0.0003610	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	-0.0007420	0.0000000	0.0000000	0.0000000



METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Method:**
Case No.: Q1915 **SAS No.:** Q1915

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB167806							
PB167774TB	PB167774TB	MB	WATER	04/30/2025	3.0	30.0	
PB167806BL	PB167806BL	MB	WATER	04/30/2025	30.0	30.0	
PB167806BS	PB167806BS	LCS	WATER	04/30/2025	30.0	30.0	
Q1901-08DUP	B-167-SB01DUP	DUP	WATER	04/30/2025	3.0	30.0	
Q1901-08MS	B-167-SB01MS	MS	WATER	04/30/2025	3.0	30.0	
Q1901-08MSD	B-167-SB01MSD	MSD	WATER	04/30/2025	3.0	30.0	
Q1915-01	WC-04282025	SAM	WATER	04/30/2025	3.0	30.0	

Metals
- 13 -

SAMPLE PREPARATION SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Method:**
Case No.: Q1915 **SAS No.:** Q1915

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB167808							
PB167774TB	PB167774TB	MB	WATER	04/30/2025	5.0	25.0	
PB167805TB	PB167805TB	MB	WATER	04/30/2025	5.0	25.0	
PB167808BL	PB167808BL	MB	WATER	04/30/2025	5.0	25.0	
PB167808BS	PB167808BS	LCS	WATER	04/30/2025	5.0	25.0	
Q1913-02DUP	WC-12-A-202504DUP	DUP	WATER	04/30/2025	5.0	25.0	
Q1913-02MS	WC-12-A-202504MS	MS	WATER	04/30/2025	5.0	25.0	
Q1913-02MSD	WC-12-A-202504MSD	MSD	WATER	04/30/2025	5.0	25.0	
Q1915-01	WC-04282025	SAM	WATER	04/30/2025	5.0	25.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: CDM Smith **Contract:** CAMP02

Lab code: CHEM **Case no.:** Q1915 **Sas no.:** Q1915 **Sdg no.:** Q1915

Instrument id number: **Method:** **Run number:** LB135623

Start date: 05/01/2025 **End date:** 05/01/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1117	HG
S0.2	S0.2	1	1119	HG
S2.5	S2.5	1	1122	HG
S5	S5	1	1126	HG
S7.5	S7.5	1	1131	HG
S10	S10	1	1136	HG
ICV112	ICV112	1	1147	HG
ICB112	ICB112	1	1149	HG
CCV62	CCV62	1	1158	HG
CCB62	CCB62	1	1201	HG
CRA	CRA	1	1205	HG
PB167806BL	PB167806BL	1	1228	HG
PB167806BS	PB167806BS	1	1230	HG
Q1901-08DUP	B-167-SB01DUP	1	1235	HG
Q1901-08MS	B-167-SB01MS	1	1239	HG
Q1901-08MSD	B-167-SB01MSD	1	1241	HG
CCV63	CCV63	1	1249	HG
CCB63	CCB63	1	1251	HG
CCV64	CCV64	1	1319	HG
CCB64	CCB64	1	1323	HG
CCV65	CCV65	1	1358	HG
CCB65	CCB65	1	1402	HG
CCV66	CCV66	1	1429	HG
CCB66	CCB66	1	1440	HG
Q1915-01	WC-04282025	1	1443	HG
PB167774TB	PB167774TB	1	1445	HG
Q1901-08L	B-167-SB01L	5	1449	HG
CCV67	CCV67	1	1523	HG
CCB67	CCB67	1	1525	HG

metals
- 14 -
ANALYSIS RUN LOG

Client: CDM Smith **Contract:** CAMP02

Lab code: CHEM **Case no.:** Q1915 **Sas no.:** Q1915 **Sdg no.:** Q1915

Instrument id number: **Method:** **Run number:** LB135636

Start date: 05/01/2025 **End date:** 05/01/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1227	Ag,As,Ba,Cd,Cr,Pb,Se
S1	S1	1	1231	Ag,As,Ba,Cd,Cr,Pb,Se
S2	S2	1	1236	Ag,As,Ba,Cd,Cr,Pb,Se
S3	S3	1	1240	Ag,As,Ba,Cd,Cr,Pb,Se
S4	S4	1	1244	Ag,As,Ba,Cd,Cr,Pb,Se
S5	S5	1	1248	Ag,As,Ba,Cd,Cr,Pb,Se
ICV01	ICV01	1	1344	Ag,As,Ba,Cd,Cr,Pb,Se
LLICV01	LLICV01	1	1401	Ag,As,Ba,Cd,Cr,Pb,Se
ICB01	ICB01	1	1411	Ag,As,Ba,Cd,Cr,Pb,Se
CRI01	CRI01	1	1416	Ag,As,Ba,Cd,Cr,Pb,Se
ICSA01	ICSA01	1	1435	Ag,As,Ba,Cd,Cr,Pb,Se
ICSAB01	ICSAB01	1	1439	Ag,As,Ba,Cd,Cr,Pb,Se
CCV01	CCV01	1	1505	Ag,As,Ba,Cd,Cr,Pb,Se
CCB01	CCB01	1	1512	Ag,As,Ba,Cd,Cr,Pb,Se
CCV02	CCV02	1	1600	Ag,As,Ba,Cd,Cr,Pb,Se
CCB02	CCB02	1	1604	Ag,As,Ba,Cd,Cr,Pb,Se
PB167808BL	PB167808BL	1	1623	Ag,As,Ba,Cd,Cr,Pb,Se
PB167808BS	PB167808BS	1	1633	Ag,As,Ba,Cd,Cr,Pb,Se
PB167774TB	PB167774TB	1	1637	Ag,As,Ba,Cd,Cr,Pb,Se
CCV03	CCV03	1	1703	Ag,As,Ba,Cd,Cr,Pb,Se
CCB03	CCB03	1	1707	Ag,As,Ba,Cd,Cr,Pb,Se
CCV04	CCV04	1	1749	Ag,As,Ba,Cd,Cr,Pb,Se
CCB04	CCB04	1	1753	Ag,As,Ba,Cd,Cr,Pb,Se
CCV05	CCV05	1	1836	Ag,As,Ba,Cd,Cr,Pb,Se
CCB05	CCB05	1	1840	Ag,As,Ba,Cd,Cr,Pb,Se
CCV06	CCV06	1	1923	Ag,As,Ba,Cd,Cr,Pb,Se
CCB06	CCB06	1	1927	Ag,As,Ba,Cd,Cr,Pb,Se
Q1915-01	WC-04282025	1	2007	Ag,As,Ba,Cd,Cr,Pb,Se
CCV07	CCV07	1	2011	Ag,As,Ba,Cd,Cr,Pb,Se
CCB07	CCB07	1	2015	Ag,As,Ba,Cd,Cr,Pb,Se
PB167805TB	PB167805TB	1	2020	Ag,As,Ba,Cd,Cr,Pb,Se
CCV08	CCV08	1	2024	Ag,As,Ba,Cd,Cr,Pb,Se
CCB08	CCB08	1	2028	Ag,As,Ba,Cd,Cr,Pb,Se

metals
- 14 -
ANALYSIS RUN LOG

Client: CDM Smith **Contract:** CAMP02

Lab code: CHEM **Case no.:** Q1915 **Sas no.:** Q1915 **Sdg no.:** Q1915

Instrument id number: **Method:** **Run number:** LB135654

Start date: 05/02/2025 **End date:** 05/02/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1234	Ag,As,Ba,Cd,Cr,Pb,Se
S1	S1	1	1238	Ag,As,Ba,Cd,Cr,Pb,Se
S2	S2	1	1242	Ag,As,Ba,Cd,Cr,Pb,Se
S3	S3	1	1246	Ag,As,Ba,Cd,Cr,Pb,Se
S4	S4	1	1251	Ag,As,Ba,Cd,Cr,Pb,Se
S5	S5	1	1255	Ag,As,Ba,Cd,Cr,Pb,Se
ICV01	ICV01	1	1347	Ag,As,Ba,Cd,Cr,Pb,Se
LLICV01	LLICV01	1	1503	Ag,As,Ba,Cd,Cr,Pb,Se
ICB01	ICB01	1	1508	Ag,As,Ba,Cd,Cr,Pb,Se
CRI01	CRI01	1	1512	Ag,As,Ba,Cd,Cr,Pb,Se
ICSA01	ICSA01	1	1530	Ag,As,Ba,Cd,Cr,Pb,Se
ICSAB01	ICSAB01	1	1546	Ag,As,Ba,Cd,Cr,Pb,Se
CCV01	CCV01	1	1601	Ag,As,Ba,Cd,Cr,Pb,Se
CCB01	CCB01	1	1605	Ag,As,Ba,Cd,Cr,Pb,Se
Q1913-02DUP	WC-12-A-202504DUP	1	1613	Ag,As,Ba,Cd,Cr,Pb,Se
Q1913-02L	WC-12-A-202504L	5	1618	Ag,As,Ba,Cd,Cr,Pb,Se
Q1913-02MS	WC-12-A-202504MS	1	1622	Ag,As,Ba,Cd,Cr,Pb,Se
Q1913-02MSD	WC-12-A-202504MSD	1	1626	Ag,As,Ba,Cd,Cr,Pb,Se
CCV02	CCV02	1	1634	Ag,As,Ba,Cd,Cr,Pb,Se
CCB02	CCB02	1	1638	Ag,As,Ba,Cd,Cr,Pb,Se

LAB CHRONICLE

OrderID:	Q1915	OrderDate:	4/29/2025 2:29:00 PM
Client:	CDM Smith	Project:	Con Ed UTEN Mount Vernon, NY
Contact:	Marcie Ann Encinas	Location:	L41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1915-01	WC-04282025	SOIL			04/28/25 14:15			04/29/25
			Corrosivity	9045D			04/29/25 18:00	
			Ignitability	1030			05/01/25 12:38	
			Reactive Cyanide	9012B		04/30/25	04/30/25 11:54	
			Reactive Sulfide	9034		05/01/25	05/01/25 11:20	



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	04/28/25 14:15
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/29/25
Client Sample ID:	WC-04282025	SDG No.:	Q1915
Lab Sample ID:	Q1915-01	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	8.41	H	1	0	0	pH		04/29/25 18:00	9045D
Ignitability	NO		1	0	0	oC		05/01/25 12:38	1030
Reactive Cyanide	0.050	U	1	0.0084	0.050	mg/Kg	04/30/25 08:50	04/30/25 11:54	9012B
Reactive Sulfide	1.58	J	1	0.20	10.0	mg/Kg	05/01/25 08:50	05/01/25 11:20	9034

Comments: pH result reported at temperature 22.5 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: CDM Smith

SDG No.: Q1915

Project: Con Ed UTEN Mount Vernon, NY

RunNo.: LB135607

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	6.99	7	100	90-110	04/29/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	04/29/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	04/29/2025

Initial and Continuing Calibration Verification

Client: CDM Smith

SDG No.: Q1915

Project: Con Ed UTEN Mount Vernon, NY

RunNo.: LB135608

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.092	0.099	93	85-115	04/30/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/30/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	04/30/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/30/2025

Initial and Continuing Calibration Blank Summary

Client: CDM Smith

SDG No.: Q1915

Project: Con Ed UTEN Mount Vernon, NY

RunNo.: LB135608

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/30/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/30/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/30/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/30/2025

Preparation Blank Summary

Client: CDM Smith

SDG No.: Q1915

Project: Con Ed UTEN Mount Vernon, NY

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: PB167792BL							
Reactive Cyanide	mg/Kg	< 0.0250	0.0250	U	0.0084	0.05	04/30/2025
Sample ID: PB167811BL							
Reactive Sulfide	mg/Kg	< 5.0000	5.0000	U	0.201	10	05/01/2025

Duplicate Sample Summary

Client:	CDM Smith	SDG No.:	Q1915
Project:	Con Ed UTEN Mount Vernon, NY	Sample ID:	Q1905-04
Client ID:	MH-GDUP	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Reactive Cyanide	mg/Kg	+/-20	0.0083	U	0.0083	U	1	0		04/30/2025
Reactive Sulfide	mg/Kg	+/-20	3.16	J	3.16	J	1	0		05/01/2025

Duplicate Sample Summary

Client:	CDM Smith	SDG No.:	Q1915
Project:	Con Ed UTEN Mount Vernon, NY	Sample ID:	Q1912-01
Client ID:	MH-EDUP	Percent Solids for Spike Sample:	92.5

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Ignitability	oC	+/-20	NO		NO		1	0		05/01/2025

Duplicate Sample Summary

Client: CDM Smith Project: Con Ed UTEN Mount Vernon, NY Client ID: WC-04282025DUP	SDG No.: Q1915 Sample ID: Q1915-01 Percent Solids for Spike Sample: 100
--	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Corrosivity	pH	+/-20	8.41		8.42		1	0.12		04/29/2025



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: CSM Smith
ADDRESS: 110 Fieldcrest Ave #8 6th Flr
CITY Edison STATE: NJ ZIP: 08837
ATTENTION: M Encinas
PHONE: 732 590 4697 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: UTEN VST
PROJECT NO.: LOCATION: MT Vernon NJ
PROJECT MANAGER: M. Encinas
e-mail: Encinas M@cdm.com
PHONE: 732 590 4697 FAX:

CLIENT BILLING INFORMATION

BILL TO: CSM Smith PO#:
ADDRESS: 110 Fieldcrest Ave #8 6th Flr
CITY Edison STATE: NJ ZIP: 08837
ATTENTION: M. Encinas PHONE: 732 590 4697

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC + Raw Data) ☐ NYS ASP A ☐ NYS ASP B
☐ EDD FORMAT ☐ Other _____

1. TCLP VOC
2. TCLP SVOC
3. TCLP metals
4. liquids
5. compositing
6. leachability
7. leachability
8. leachability
9. leachability

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		E	E	E	E	E	E					
1.	<u>WC 04282025</u>	<u>Soil</u>	☒	<u>only</u>	<u>4/28/25</u>	<u>1415</u>	<u>4</u>	☒	☒	☒	☒	☒	☒					
2.																		
3.																		
4.																		
5.																		
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP
1. <u>[Signature]</u>	<u>4/24/25 1345</u>	<u>[Signature]</u>	<u>2.5</u> °C
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Comments: <u>Temp 2.5°C Adjustment factor + 1 IR Gun #1</u>
2. <u>[Signature]</u>		<u>[Signature]</u>	
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3. <u>[Signature]</u>	<u>4-29-25</u>	<u>[Signature]</u>	

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete

☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488