



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: CDM Smith

SDG No.: Q1915

Project: Con Ed UTEN Mount Vernon, NY

RunNo.: LB135607

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	6.99	7	100	90-110	04/29/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	04/29/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	04/29/2025

Initial and Continuing Calibration Verification

Client: CDM Smith

SDG No.: Q1915

Project: Con Ed UTEN Mount Vernon, NY

RunNo.: LB135608

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.092	0.099	93	85-115	04/30/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/30/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	04/30/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/30/2025