


LB135623

7470A

INSTRUMENT 1A CV1

Linear



A= 0.0000e+000
 B= 1.8925e-004
 C= 1.5429e-002
 Rho= 0.9997379
 Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.014	0.014	-7	0.000	-7	0				
0.2	0.200	0.198	-0.002	966	0.0 %	966					
2.5	2.500	2.415	-0.085	12681	0.0 %	12681					
5.0	5.000	5.160	0.160	27185	0.0 %	27185					
7.5	7.500	7.402	-0.098	39029	0.0 %	39029					
10.0	10.000	10.011	0.011	52816	0.0 %	52816					

0/00
 1
 3
 3
 1
 0

01 May 2025 15:45