

LAB CHRONICLE

OrderID:	Q1915	OrderDate:	4/29/2025 2:29:00 PM
Client:	CDM Smith	Project:	Con Ed UTEN Mount Vernon, NY
Contact:	Marcie Ann Encinas	Location:	L41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1915-01	WC-04282025	TCLP			04/28/25			04/29/25
			TCLP ICP Metals	6010D		04/30/25	05/01/25	
			TCLP Mercury	7470A		04/30/25	05/01/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1915

Order ID: Q1915

Client: CDM Smith

Project ID: Con Ed UTEN Mount Vernon, NY

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : WC-04282025								
Q1915-01	WC-04282025	TCLP	Barium	4600		72.8	500	ug/L
Q1915-01	WC-04282025	TCLP	Cadmium	7.22	J	2.50	30.0	ug/L
Q1915-01	WC-04282025	TCLP	Lead	896		11.5	60.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	04/28/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/29/25
Client Sample ID:	WC-04282025	SDG No.:	Q1915
Lab Sample ID:	Q1915-01	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-38-2	Arsenic	100	U	1	25.6	100	ug/L	04/30/25 13:35	05/01/25 20:07	6010D	SW3050
7440-39-3	Barium	4600		1	72.8	500	ug/L	04/30/25 13:35	05/01/25 20:07	6010D	SW3050
7440-43-9	Cadmium	7.22	J	1	2.50	30.0	ug/L	04/30/25 13:35	05/01/25 20:07	6010D	SW3050
7440-47-3	Chromium	50.0	U	1	10.6	50.0	ug/L	04/30/25 13:35	05/01/25 20:07	6010D	SW3050
7439-92-1	Lead	896		1	11.5	60.0	ug/L	04/30/25 13:35	05/01/25 20:07	6010D	SW3050
7439-97-6	Mercury	2.00	U	1	0.76	2.00	ug/L	04/30/25 13:30	05/01/25 14:43	7470A	
7782-49-2	Selenium	100	U	1	48.2	100	ug/L	04/30/25 13:35	05/01/25 20:07	6010D	SW3050
7440-22-4	Silver	50.0	U	1	8.10	50.0	ug/L	04/30/25 13:35	05/01/25 20:07	6010D	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP METALS			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV112	Mercury	3.81	4.0	95	90 - 110	CV	05/01/2025	11:47	LB135623

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV62	Mercury	4.80	5.0	96	90 - 110	CV	05/01/2025	11:58	LB135623
CCV63	Mercury	5.01	5.0	100	90 - 110	CV	05/01/2025	12:49	LB135623
CCV64	Mercury	5.09	5.0	102	90 - 110	CV	05/01/2025	13:19	LB135623
CCV65	Mercury	4.77	5.0	95	90 - 110	CV	05/01/2025	13:58	LB135623
CCV66	Mercury	4.60	5.0	92	90 - 110	CV	05/01/2025	14:29	LB135623
CCV67	Mercury	4.61	5.0	92	90 - 110	CV	05/01/2025	15:23	LB135623



- 2a -

Client:	<u>CDM Smith</u>	SDG No.:	<u>Q1915</u>				
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1915</u>	SAS No.:	<u>Q1915</u>
Initial Calibration Source:	<u>EPA</u>						
Continuing Calibration Source:	Inorganic Ventures						

Sample ID	Analyte	Result	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
		ug/L							
ICV01	Arsenic	1010	1000	100	90 - 110	P	05/01/2025	13:44	LB135636
	Barium	495	520	95	90 - 110	P	05/01/2025	13:44	LB135636
	Cadmium	508	510	100	90 - 110	P	05/01/2025	13:44	LB135636
	Chromium	506	520	97	90 - 110	P	05/01/2025	13:44	LB135636
	Lead	989	1000	99	90 - 110	P	05/01/2025	13:44	LB135636
	Selenium	1040	1000	104	90 - 110	P	05/01/2025	13:44	LB135636
	Silver	240	250	96	90 - 110	P	05/01/2025	13:44	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	23.7	20.0	119	80 - 120	P	05/01/2025	14:01	LB135636
	Barium	95.8	100	96	80 - 120	P	05/01/2025	14:01	LB135636
	Cadmium	6.39	6.0	107	80 - 120	P	05/01/2025	14:01	LB135636
	Chromium	10.5	10.0	105	80 - 120	P	05/01/2025	14:01	LB135636
	Lead	13.1	12.0	109	80 - 120	P	05/01/2025	14:01	LB135636
	Selenium	22.9	20.0	114	80 - 120	P	05/01/2025	14:01	LB135636
	Silver	10.2	10.0	102	80 - 120	P	05/01/2025	14:01	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q1915
 Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	15:05	LB135636
	Barium	9650	10000	96	90 - 110	P	05/01/2025	15:05	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	15:05	LB135636
	Chromium	1000	1000	100	90 - 110	P	05/01/2025	15:05	LB135636
	Lead	5010	5000	100	90 - 110	P	05/01/2025	15:05	LB135636
	Selenium	5140	5000	103	90 - 110	P	05/01/2025	15:05	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	15:05	LB135636
CCV02	Arsenic	4930	5000	99	90 - 110	P	05/01/2025	16:00	LB135636
	Barium	10200	10000	102	90 - 110	P	05/01/2025	16:00	LB135636
	Cadmium	2470	2500	99	90 - 110	P	05/01/2025	16:00	LB135636
	Chromium	995	1000	100	90 - 110	P	05/01/2025	16:00	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	16:00	LB135636
	Selenium	4960	5000	99	90 - 110	P	05/01/2025	16:00	LB135636
	Silver	1220	1250	98	90 - 110	P	05/01/2025	16:00	LB135636
CCV03	Arsenic	5040	5000	101	90 - 110	P	05/01/2025	17:03	LB135636
	Barium	9950	10000	100	90 - 110	P	05/01/2025	17:03	LB135636
	Cadmium	2510	2500	100	90 - 110	P	05/01/2025	17:03	LB135636
	Chromium	988	1000	99	90 - 110	P	05/01/2025	17:03	LB135636
	Lead	5010	5000	100	90 - 110	P	05/01/2025	17:03	LB135636
	Selenium	5060	5000	101	90 - 110	P	05/01/2025	17:03	LB135636
	Silver	1220	1250	97	90 - 110	P	05/01/2025	17:03	LB135636
CCV04	Arsenic	5160	5000	103	90 - 110	P	05/01/2025	17:49	LB135636
	Barium	10000	10000	100	90 - 110	P	05/01/2025	17:49	LB135636
	Cadmium	2470	2500	99	90 - 110	P	05/01/2025	17:49	LB135636
	Chromium	987	1000	99	90 - 110	P	05/01/2025	17:49	LB135636
	Lead	4950	5000	99	90 - 110	P	05/01/2025	17:49	LB135636
	Selenium	5240	5000	105	90 - 110	P	05/01/2025	17:49	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	17:49	LB135636
CCV05	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	18:36	LB135636
	Barium	9950	10000	100	90 - 110	P	05/01/2025	18:36	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	18:36	LB135636
	Chromium	984	1000	98	90 - 110	P	05/01/2025	18:36	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	18:36	LB135636
	Selenium	5150	5000	103	90 - 110	P	05/01/2025	18:36	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q1915
 Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1230	1250	98	90 - 110	P	05/01/2025	18:36	LB135636
CCV06	Arsenic	4990	5000	100	90 - 110	P	05/01/2025	19:23	LB135636
	Barium	9830	10000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Cadmium	2460	2500	98	90 - 110	P	05/01/2025	19:23	LB135636
	Chromium	980	1000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Lead	4920	5000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Selenium	5030	5000	101	90 - 110	P	05/01/2025	19:23	LB135636
	Silver	1240	1250	99	90 - 110	P	05/01/2025	19:23	LB135636
CCV07	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	20:11	LB135636
	Barium	9790	10000	98	90 - 110	P	05/01/2025	20:11	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	20:11	LB135636
	Chromium	995	1000	100	90 - 110	P	05/01/2025	20:11	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	20:11	LB135636
	Selenium	5120	5000	102	90 - 110	P	05/01/2025	20:11	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	20:11	LB135636
CCV08	Arsenic	5180	5000	104	90 - 110	P	05/01/2025	20:24	LB135636
	Barium	9880	10000	99	90 - 110	P	05/01/2025	20:24	LB135636
	Cadmium	2490	2500	100	90 - 110	P	05/01/2025	20:24	LB135636
	Chromium	980	1000	98	90 - 110	P	05/01/2025	20:24	LB135636
	Lead	4990	5000	100	90 - 110	P	05/01/2025	20:24	LB135636
	Selenium	5270	5000	105	90 - 110	P	05/01/2025	20:24	LB135636
	Silver	1240	1250	99	90 - 110	P	05/01/2025	20:24	LB135636



- 2a -

Client:	<u>CDM Smith</u>	SDG No.:	<u>Q1915</u>				
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1915</u>	SAS No.:	<u>Q1915</u>
Initial Calibration Source:	<u>EPA</u>						
Continuing Calibration Source:	Inorganic Ventures						

Sample ID	Analyte	Result	True Value	%	Acceptance	M	Analysis Date	Analysis Time	Run Number
		ug/L		Recovery	Window (%R)				
ICV01	Arsenic	1060	1000	106	90 - 110	P	05/02/2025	13:47	LB135654
	Barium	518	520	100	90 - 110	P	05/02/2025	13:47	LB135654
	Cadmium	493	510	97	90 - 110	P	05/02/2025	13:47	LB135654
	Chromium	503	520	97	90 - 110	P	05/02/2025	13:47	LB135654
	Lead	966	1000	97	90 - 110	P	05/02/2025	13:47	LB135654
	Selenium	1040	1000	104	90 - 110	P	05/02/2025	13:47	LB135654
	Silver	230	250	92	90 - 110	P	05/02/2025	13:47	LB135654



- 2a -

Client:	<u>CDM Smith</u>	SDG No.:	<u>Q1915</u>	
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q1915</u> SAS No.: <u>Q1915</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	Inorganic Ventures			

Sample ID	Analyte	Result	True Value	%	Acceptance	M	Analysis Date	Analysis Time	Run Number
		ug/L		Recovery	Window (%R)				
LLICV01	Arsenic	19.0	20.0	95	80 - 120	P	05/02/2025	15:03	LB135654
	Barium	81.5	100	82	80 - 120	P	05/02/2025	15:03	LB135654
	Cadmium	5.51	6.0	92	80 - 120	P	05/02/2025	15:03	LB135654
	Chromium	10.1	10.0	101	80 - 120	P	05/02/2025	15:03	LB135654
	Lead	11.1	12.0	93	80 - 120	P	05/02/2025	15:03	LB135654
	Selenium	16.8	20.0	84	80 - 120	P	05/02/2025	15:03	LB135654
	Silver	10.8	10.0	108	80 - 120	P	05/02/2025	15:03	LB135654

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q1915
Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5450	5000	109	90 - 110	P	05/02/2025	16:01	LB135654
	Barium	9540	10000	95	90 - 110	P	05/02/2025	16:01	LB135654
	Cadmium	2440	2500	98	90 - 110	P	05/02/2025	16:01	LB135654
	Chromium	1010	1000	101	90 - 110	P	05/02/2025	16:01	LB135654
	Lead	4820	5000	96	90 - 110	P	05/02/2025	16:01	LB135654
	Selenium	5460	5000	109	90 - 110	P	05/02/2025	16:01	LB135654
	Silver	1270	1250	102	90 - 110	P	05/02/2025	16:01	LB135654
CCV02	Arsenic	5240	5000	105	90 - 110	P	05/02/2025	16:34	LB135654
	Barium	9720	10000	97	90 - 110	P	05/02/2025	16:34	LB135654
	Cadmium	2390	2500	96	90 - 110	P	05/02/2025	16:34	LB135654
	Chromium	1000	1000	100	90 - 110	P	05/02/2025	16:34	LB135654
	Lead	4740	5000	95	90 - 110	P	05/02/2025	16:34	LB135654
	Selenium	5240	5000	105	90 - 110	P	05/02/2025	16:34	LB135654
	Silver	1260	1250	101	90 - 110	P	05/02/2025	16:34	LB135654



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	86	40 - 160	CV	05/01/2025	12:05	LB135623
CRI01	Arsenic	23.3	20.0	116	40 - 160	P	05/01/2025	14:16	LB135636
	Barium	93.6	100	94	40 - 160	P	05/01/2025	14:16	LB135636
	Cadmium	6.36	6.0	106	40 - 160	P	05/01/2025	14:16	LB135636
	Chromium	10.6	10.0	106	40 - 160	P	05/01/2025	14:16	LB135636
	Lead	13.3	12.0	111	40 - 160	P	05/01/2025	14:16	LB135636
	Selenium	23.1	20.0	115	40 - 160	P	05/01/2025	14:16	LB135636
	Silver	10.6	10.0	106	40 - 160	P	05/01/2025	14:16	LB135636
CRI01	Arsenic	21.5	20.0	108	40 - 160	P	05/02/2025	15:12	LB135654
	Barium	80.2	100	80	40 - 160	P	05/02/2025	15:12	LB135654
	Cadmium	5.66	6.0	94	40 - 160	P	05/02/2025	15:12	LB135654
	Chromium	10.6	10.0	106	40 - 160	P	05/02/2025	15:12	LB135654
	Lead	12.1	12.0	101	40 - 160	P	05/02/2025	15:12	LB135654
	Selenium	23.5	20.0	117	40 - 160	P	05/02/2025	15:12	LB135654
	Silver	10.7	10.0	107	40 - 160	P	05/02/2025	15:12	LB135654



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB112	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	11:49	LB135623

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB62	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	12:01	LB135623
CCB63	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	12:51	LB135623
CCB64	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	13:23	LB135623
CCB65	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	14:02	LB135623
CCB66	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	14:40	LB135623
CCB67	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	15:25	LB135623

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	14:11	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	14:11	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	14:11	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	14:11	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	14:11	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	14:11	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	14:11	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q1915							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q1915	SAS No.: Q1915						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	15:12	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	15:12	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	15:12	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	15:12	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	15:12	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	15:12	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	15:12	LB135636
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	16:04	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	16:04	LB135636
	Cadmium	0.60	+/-6.00	J	6.00	P	05/01/2025	16:04	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	16:04	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	16:04	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	16:04	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	16:04	LB135636
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	17:07	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	17:07	LB135636
	Cadmium	0.88	+/-6.00	J	6.00	P	05/01/2025	17:07	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	17:07	LB135636
	Lead	3.14	+/-12.0	J	12.0	P	05/01/2025	17:07	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	17:07	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	17:07	LB135636
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	17:53	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	17:53	LB135636
	Cadmium	0.92	+/-6.00	J	6.00	P	05/01/2025	17:53	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	17:53	LB135636
	Lead	2.63	+/-12.0	J	12.0	P	05/01/2025	17:53	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	17:53	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	17:53	LB135636
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	18:40	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	18:40	LB135636
	Cadmium	0.89	+/-6.00	J	6.00	P	05/01/2025	18:40	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	18:40	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	18:40	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	18:40	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	18:40	LB135636
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	19:27	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	19:27	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	19:27	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	19:27	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	19:27	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	19:27	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	19:27	LB135636
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	20:15	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	20:15	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	20:15	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	20:15	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	20:15	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	20:15	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	20:15	LB135636
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	05/01/2025	20:28	LB135636
	Barium	100	+/-100	U	100	P	05/01/2025	20:28	LB135636
	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	20:28	LB135636
	Chromium	10.0	+/-10.0	U	10.0	P	05/01/2025	20:28	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	20:28	LB135636
	Selenium	20.0	+/-20.0	U	20.0	P	05/01/2025	20:28	LB135636
	Silver	10.0	+/-10.0	U	10.0	P	05/01/2025	20:28	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/02/2025	15:08	LB135654
	Barium	100	+/-100	U	100	P	05/02/2025	15:08	LB135654
	Cadmium	6.00	+/-6.00	U	6.00	P	05/02/2025	15:08	LB135654
	Chromium	10.0	+/-10.0	U	10.0	P	05/02/2025	15:08	LB135654
	Lead	12.0	+/-12.0	U	12.0	P	05/02/2025	15:08	LB135654
	Selenium	20.0	+/-20.0	U	20.0	P	05/02/2025	15:08	LB135654
	Silver	10.0	+/-10.0	U	10.0	P	05/02/2025	15:08	LB135654

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	05/02/2025	16:05	LB135654
	Barium	100	+/-100	U	100	P	05/02/2025	16:05	LB135654
	Cadmium	6.00	+/-6.00	U	6.00	P	05/02/2025	16:05	LB135654
	Chromium	10.0	+/-10.0	U	10.0	P	05/02/2025	16:05	LB135654
	Lead	12.0	+/-12.0	U	12.0	P	05/02/2025	16:05	LB135654
	Selenium	20.0	+/-20.0	U	20.0	P	05/02/2025	16:05	LB135654
	Silver	10.0	+/-10.0	U	10.0	P	05/02/2025	16:05	LB135654
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	05/02/2025	16:38	LB135654
	Barium	100	+/-100	U	100	P	05/02/2025	16:38	LB135654
	Cadmium	6.00	+/-6.00	U	6.00	P	05/02/2025	16:38	LB135654
	Chromium	10.0	+/-10.0	U	10.0	P	05/02/2025	16:38	LB135654
	Lead	12.0	+/-12.0	U	12.0	P	05/02/2025	16:38	LB135654
	Selenium	20.0	+/-20.0	U	20.0	P	05/02/2025	16:38	LB135654
	Silver	10.0	+/-10.0	U	10.0	P	05/02/2025	16:38	LB135654

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q1915

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167774TB		WATER		Batch Number:	PB167806		Prep Date:	04/30/2025	
	Mercury	2.00	<2.00	U	2.00	CV	05/01/2025	14:45	LB135623
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167806BL		WATER		Batch Number:	PB167806		Prep Date:	04/30/2025	
	Mercury	0.20	<0.20	U	0.20	CV	05/01/2025	12:28	LB135623

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q1915

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167774TB		WATER		Batch Number:	PB167808		Prep Date:	04/30/2025	
	Arsenic	100	<100	U	100	P	05/01/2025	16:37	LB135636
	Barium	500	<500	U	500	P	05/01/2025	16:37	LB135636
	Cadmium	30.0	<30.0	U	30.0	P	05/01/2025	16:37	LB135636
	Chromium	50.0	<50.0	U	50.0	P	05/01/2025	16:37	LB135636
	Lead	60.0	<60.0	U	60.0	P	05/01/2025	16:37	LB135636
	Selenium	100	<100	U	100	P	05/01/2025	16:37	LB135636
	Silver	50.0	<50.0	U	50.0	P	05/01/2025	16:37	LB135636
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167805TB		WATER		Batch Number:	PB167808		Prep Date:	04/30/2025	
	Arsenic	100	<100	U	100	P	05/01/2025	20:20	LB135636
	Barium	500	<500	U	500	P	05/01/2025	20:20	LB135636
	Cadmium	30.0	<30.0	U	30.0	P	05/01/2025	20:20	LB135636
	Chromium	15.9	<50.0	J	50.0	P	05/01/2025	20:20	LB135636
	Lead	60.0	<60.0	U	60.0	P	05/01/2025	20:20	LB135636
	Selenium	100	<100	U	100	P	05/01/2025	20:20	LB135636
	Silver	50.0	<50.0	U	50.0	P	05/01/2025	20:20	LB135636
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167808BL		WATER		Batch Number:	PB167808		Prep Date:	04/30/2025	
	Arsenic	100	<100	U	100	P	05/01/2025	16:23	LB135636
	Barium	500	<500	U	500	P	05/01/2025	16:23	LB135636
	Cadmium	30.0	<30.0	U	30.0	P	05/01/2025	16:23	LB135636
	Chromium	50.0	<50.0	U	50.0	P	05/01/2025	16:23	LB135636
	Lead	60.0	<60.0	U	60.0	P	05/01/2025	16:23	LB135636
	Selenium	100	<100	U	100	P	05/01/2025	16:23	LB135636
	Silver	50.0	<50.0	U	50.0	P	05/01/2025	16:23	LB135636

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>CDM Smith</u>	SDG No.: <u>Q1915</u>
Contract: <u>CAMP02</u>	Lab Code: <u>CHEM</u>
ICS Source: <u>EPA</u>	Case No.: <u>Q1915</u>
	SAS No.: <u>Q1915</u>
	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	2.11			-20	20	05/01/2025	14:35	LB135636
	Barium	0.52	6.0	9	-94	106	05/01/2025	14:35	LB135636
	Cadmium	-4.57	1.0	457	-5	7	05/01/2025	14:35	LB135636
	Chromium	53.0	52.0	102	42	62	05/01/2025	14:35	LB135636
	Lead	-4.15			-12	12	05/01/2025	14:35	LB135636
	Selenium	-9.38			-20	20	05/01/2025	14:35	LB135636
	Silver	1.45			-10	10	05/01/2025	14:35	LB135636
ICSAB01	Arsenic	120	104	115	88.4	120	05/01/2025	14:39	LB135636
	Barium	487	537	91	437	637	05/01/2025	14:39	LB135636
	Cadmium	1080	972	111	826	1120	05/01/2025	14:39	LB135636
	Chromium	583	542	108	460	624	05/01/2025	14:39	LB135636
	Lead	52.5	49.0	107	37	61	05/01/2025	14:39	LB135636
	Selenium	45.9	46.0	100	26	66	05/01/2025	14:39	LB135636
	Silver	203	201	101	170	232	05/01/2025	14:39	LB135636
ICSA01	Arsenic	0.098			-20	20	05/02/2025	15:30	LB135654
	Barium	-5.67	6.0	94	-94	106	05/02/2025	15:30	LB135654
	Cadmium	-2.89	1.0	289	-5	7	05/02/2025	15:30	LB135654
	Chromium	53.8	52.0	104	42	62	05/02/2025	15:30	LB135654
	Lead	4.14			-12	12	05/02/2025	15:30	LB135654
	Selenium	-1.90			-20	20	05/02/2025	15:30	LB135654
	Silver	-0.66			-10	10	05/02/2025	15:30	LB135654
ICSAB01	Arsenic	115	104	111	88.4	120	05/02/2025	15:46	LB135654
	Barium	464	537	86	437	637	05/02/2025	15:46	LB135654
	Cadmium	996	972	102	826	1120	05/02/2025	15:46	LB135654
	Chromium	560	542	103	460	624	05/02/2025	15:46	LB135654
	Lead	51.7	49.0	106	37	61	05/02/2025	15:46	LB135654
	Selenium	57.3	46.0	125	26	66	05/02/2025	15:46	LB135654
	Silver	193	201	96	170	232	05/02/2025	15:46	LB135654



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client:	CDM Smith	level:	low	sdg no.:	Q1915		
contract:	CAMP02	lab code:	CHEM	case no.:	Q1915	sas no.:	Q1915
matrix:	Water	sample id:	Q1901-08	client id:	B-167-SB01MS		
Percent Solids for Sample:	NA	Spiked ID:	Q1901-08MS	Percent Solids for Spike Sample:	NA		

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	39.1		2.00	U	40.0	98		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>CDM Smith</u>	level:	<u>low</u>	sdg no.:	<u>Q1915</u>				
contract:	<u>CAMP02</u>	lab code:	<u>CHEM</u>	case no.:	<u>Q1915</u>	sas no.:	<u>Q1915</u>		
matrix:	<u>Water</u>	sample id:	<u>Q1901-08</u>	client id:	<u>B-167-SB01MSD</u>				
Percent Solids for Sample:	<u>NA</u>	Spiked ID:	<u>Q1901-08MSD</u>	Percent Solids for Spike Sample:	<u>NA</u>				

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	40.9		2.00	U	40.0	102		CV

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>CDM Smith</u>	level: <u>low</u>	sdg no.: <u>Q1915</u>
contract: <u>CAMP02</u>	lab code: <u>CHEM</u>	case no.: <u>Q1915</u> sas no.: <u>Q1915</u>
matrix: <u>Water</u>	sample id: <u>Q1913-02</u>	client id: <u>WC-12-A-202504MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q1913-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	4220		100	U	4000	106		P
Barium	ug/L	75 - 125	1020		500	U	1000	102		P
Cadmium	ug/L	75 - 125	943		30.0	U	1000	94		P
Chromium	ug/L	75 - 125	2180		122		2000	103		P
Lead	ug/L	75 - 125	4570		60.0	U	5000	91		P
Selenium	ug/L	75 - 125	10100		100	U	10000	101		P
Silver	ug/L	75 - 125	385		50.0	U	380	101		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>CDM Smith</u>	level: <u>low</u>	sdg no.: <u>Q1915</u>
contract: <u>CAMP02</u>	lab code: <u>CHEM</u>	case no.: <u>Q1915</u> sas no.: <u>Q1915</u>
matrix: <u>Water</u>	sample id: <u>Q1913-02</u>	client id: <u>WC-12-A-202504MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q1913-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	4060		100	U	4000	102		P
Barium	ug/L	75 - 125	965		500	U	1000	96		P
Cadmium	ug/L	75 - 125	908		30.0	U	1000	91		P
Chromium	ug/L	75 - 125	2100		122		2000	99		P
Lead	ug/L	75 - 125	4400		60.0	U	5000	88		P
Selenium	ug/L	75 - 125	9630		100	U	10000	96		P
Silver	ug/L	75 - 125	366		50.0	U	380	96		P

Metals
- 5b -

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Matrix: **Level:** LOW **Client ID:**
Sample ID: **Spiked ID:**

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>CDM Smith</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q1915</u>
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1915</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q1901-08</u>	Client ID:	<u>B-167-SB01DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q1901-08DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	2.00	U	2.00	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>CDM Smith</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q1915</u>
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1915</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q1901-08MS</u>	Client ID:	<u>B-167-SB01MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q1901-08MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	39.1		40.9		4		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: CDM Smith **Level:** LOW **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Matrix: Water **Sample ID:** Q1913-02 **Client ID:** WC-12-A-202504DUP
Percent Solids for Sample: NA **Duplicate ID** Q1913-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	500	U	500	U			P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	122		113		8		P
Lead	ug/L	20	60.0	U	60.0	U			P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: CDM Smith **Level:** LOW **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Matrix: Water **Sample ID:** Q1913-02MS **Client ID:** WC-12-A-202504MSD
Percent Solids for Sample: NA **Duplicate ID** Q1913-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4220		4060		4		P
Barium	ug/L	20	1020		965		6		P
Cadmium	ug/L	20	943		908		4		P
Chromium	ug/L	20	2180		2100		4		P
Lead	ug/L	20	4570		4400		4		P
Selenium	ug/L	20	10100		9630		5		P
Silver	ug/L	20	385		366		5		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167806BS Mercury	ug/L	4.0	3.34		84	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith SDG No.: Q1915
Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167808BS							
Arsenic	ug/L	4000	3770		94	80 - 120	P
Barium	ug/L	1000	826		83	80 - 120	P
Cadmium	ug/L	1000	943		94	80 - 120	P
Chromium	ug/L	2000	1870		94	80 - 120	P
Lead	ug/L	5000	4630		93	80 - 120	P
Selenium	ug/L	10000	9600		96	80 - 120	P
Silver	ug/L	380	346		91	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

B-167-SB01L

Lab Name: Chemtech Consulting Group

Contract: CAMP02

Lab Code: CHEM

Lb No.: lb135623

Lab Sample ID : Q1901-08L

SDG No.: Q1915

Matrix (soil/water): Water

Level (low/med): LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Mercury	2.00 U	10.0 U			CV

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

WC-12-A-202504L

Lab Name: Chemtech Consulting Group **Contract:** CAMP02
Lab Code: CHEM **Lb No.:** lb135654 **Lab Sample ID :** Q1913-02L **SDG No.:** Q1915
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Arsenic	100	U	500	U			P
Barium	500	U	2500	U			P
Cadmium	30.0	U	150	U			P
Chromium	122		122	J	0		P
Lead	60.0	U	300	U			P
Selenium	100	U	500	U			P
Silver	50.0	U	250	U			P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith **SDG No.:** Q1915
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915
Instrument ID: _____ **Date:** _____
Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Arsenic	193.759	0.0000000	0.0000000	-0.0000440	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000930	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	-0.0000920	0.0000000	0.0000380	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	-0.0001440	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	-0.0001490	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith SDG No.: Q1915
 Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915
 Instrument ID: _____ Date: _____
 Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0002870
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0003170	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	-0.0003570
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q1915

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1915

SAS No.: Q1915

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Arsenic	193.759	-0.0029000	0.0000000	0.0000000	0.0000000	0.0004900
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0001400	-0.0008600
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0007460	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	-0.0000120

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q1915

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1915

SAS No.: Q1915

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0006580	0.0000000	0.0000000	0.0001290
Selenium	196.090	0.0000000	0.0000000	0.0003330	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: CDM Smith

SDG No.: Q1915

Contract: CAMP02

Lab Code: CHEM

Case No.: Q1915

SAS No.: Q1915

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000630	0.0001280	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Lead	220.353	0.0000000	-0.0003610	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	-0.0007420	0.0000000	0.0000000	0.0000000

METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client:	<u>CDM Smith</u>	SDG No.:	<u>Q1915</u>
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>
		Method:	<u></u>
		Case No.:	<u>Q1915</u>
		SAS No.:	<u>Q1915</u>

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB167806							
PB167774TB	PB167774TB	MB	WATER	04/30/2025	3.0	30.0	
PB167806BL	PB167806BL	MB	WATER	04/30/2025	30.0	30.0	
PB167806BS	PB167806BS	LCS	WATER	04/30/2025	30.0	30.0	
Q1901-08DUP	B-167-SB01DUP	DUP	WATER	04/30/2025	3.0	30.0	
Q1901-08MS	B-167-SB01MS	MS	WATER	04/30/2025	3.0	30.0	
Q1901-08MSD	B-167-SB01MSD	MSD	WATER	04/30/2025	3.0	30.0	
Q1915-01	WC-04282025	SAM	WATER	04/30/2025	3.0	30.0	

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client:	<u>CDM Smith</u>	SDG No.:	<u>Q1915</u>
Contract:	<u>CAMP02</u>	Lab Code:	<u>CHEM</u>
		Method:	<u></u>
		Case No.:	<u>Q1915</u>
		SAS No.:	<u>Q1915</u>

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB167808							
PB167774TB	PB167774TB	MB	WATER	04/30/2025	5.0	25.0	
PB167805TB	PB167805TB	MB	WATER	04/30/2025	5.0	25.0	
PB167808BL	PB167808BL	MB	WATER	04/30/2025	5.0	25.0	
PB167808BS	PB167808BS	LCS	WATER	04/30/2025	5.0	25.0	
Q1913-02DUP	WC-12-A-202504DUP	DUP	WATER	04/30/2025	5.0	25.0	
Q1913-02MS	WC-12-A-202504MS	MS	WATER	04/30/2025	5.0	25.0	
Q1913-02MSD	WC-12-A-202504MSD	MSD	WATER	04/30/2025	5.0	25.0	
Q1915-01	WC-04282025	SAM	WATER	04/30/2025	5.0	25.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: CDM Smith **Contract:** CAMP02

Lab code: CHEM **Case no.:** Q1915 **Sas no.:** Q1915 **Sdg no.:** Q1915

Instrument id number: **Method:** **Run number:** LB135623

Start date: 05/01/2025 **End date:** 05/01/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1117	HG
S0.2	S0.2	1	1119	HG
S2.5	S2.5	1	1122	HG
S5	S5	1	1126	HG
S7.5	S7.5	1	1131	HG
S10	S10	1	1136	HG
ICV112	ICV112	1	1147	HG
ICB112	ICB112	1	1149	HG
CCV62	CCV62	1	1158	HG
CCB62	CCB62	1	1201	HG
CRA	CRA	1	1205	HG
PB167806BL	PB167806BL	1	1228	HG
PB167806BS	PB167806BS	1	1230	HG
Q1901-08DUP	B-167-SB01DUP	1	1235	HG
Q1901-08MS	B-167-SB01MS	1	1239	HG
Q1901-08MSD	B-167-SB01MSD	1	1241	HG
CCV63	CCV63	1	1249	HG
CCB63	CCB63	1	1251	HG
CCV64	CCV64	1	1319	HG
CCB64	CCB64	1	1323	HG
CCV65	CCV65	1	1358	HG
CCB65	CCB65	1	1402	HG
CCV66	CCV66	1	1429	HG
CCB66	CCB66	1	1440	HG
Q1915-01	WC-04282025	1	1443	HG
PB167774TB	PB167774TB	1	1445	HG
Q1901-08L	B-167-SB01L	5	1449	HG
CCV67	CCV67	1	1523	HG
CCB67	CCB67	1	1525	HG

metals
- 14 -
ANALYSIS RUN LOG

Client: CDM Smith **Contract:** CAMP02

Lab code: CHEM **Case no.:** Q1915 **Sas no.:** Q1915 **Sdg no.:** Q1915

Instrument id number: **Method:** **Run number:** LB135636

Start date: 05/01/2025 **End date:** 05/01/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1227	Ag,As,Ba,Cd,Cr,Pb,Se
S1	S1	1	1231	Ag,As,Ba,Cd,Cr,Pb,Se
S2	S2	1	1236	Ag,As,Ba,Cd,Cr,Pb,Se
S3	S3	1	1240	Ag,As,Ba,Cd,Cr,Pb,Se
S4	S4	1	1244	Ag,As,Ba,Cd,Cr,Pb,Se
S5	S5	1	1248	Ag,As,Ba,Cd,Cr,Pb,Se
ICV01	ICV01	1	1344	Ag,As,Ba,Cd,Cr,Pb,Se
LLICV01	LLICV01	1	1401	Ag,As,Ba,Cd,Cr,Pb,Se
ICB01	ICB01	1	1411	Ag,As,Ba,Cd,Cr,Pb,Se
CRI01	CRI01	1	1416	Ag,As,Ba,Cd,Cr,Pb,Se
ICSA01	ICSA01	1	1435	Ag,As,Ba,Cd,Cr,Pb,Se
ICSAB01	ICSAB01	1	1439	Ag,As,Ba,Cd,Cr,Pb,Se
CCV01	CCV01	1	1505	Ag,As,Ba,Cd,Cr,Pb,Se
CCB01	CCB01	1	1512	Ag,As,Ba,Cd,Cr,Pb,Se
CCV02	CCV02	1	1600	Ag,As,Ba,Cd,Cr,Pb,Se
CCB02	CCB02	1	1604	Ag,As,Ba,Cd,Cr,Pb,Se
PB167808BL	PB167808BL	1	1623	Ag,As,Ba,Cd,Cr,Pb,Se
PB167808BS	PB167808BS	1	1633	Ag,As,Ba,Cd,Cr,Pb,Se
PB167774TB	PB167774TB	1	1637	Ag,As,Ba,Cd,Cr,Pb,Se
CCV03	CCV03	1	1703	Ag,As,Ba,Cd,Cr,Pb,Se
CCB03	CCB03	1	1707	Ag,As,Ba,Cd,Cr,Pb,Se
CCV04	CCV04	1	1749	Ag,As,Ba,Cd,Cr,Pb,Se
CCB04	CCB04	1	1753	Ag,As,Ba,Cd,Cr,Pb,Se
CCV05	CCV05	1	1836	Ag,As,Ba,Cd,Cr,Pb,Se
CCB05	CCB05	1	1840	Ag,As,Ba,Cd,Cr,Pb,Se
CCV06	CCV06	1	1923	Ag,As,Ba,Cd,Cr,Pb,Se
CCB06	CCB06	1	1927	Ag,As,Ba,Cd,Cr,Pb,Se
Q1915-01	WC-04282025	1	2007	Ag,As,Ba,Cd,Cr,Pb,Se
CCV07	CCV07	1	2011	Ag,As,Ba,Cd,Cr,Pb,Se
CCB07	CCB07	1	2015	Ag,As,Ba,Cd,Cr,Pb,Se
PB167805TB	PB167805TB	1	2020	Ag,As,Ba,Cd,Cr,Pb,Se
CCV08	CCV08	1	2024	Ag,As,Ba,Cd,Cr,Pb,Se
CCB08	CCB08	1	2028	Ag,As,Ba,Cd,Cr,Pb,Se

metals
- 14 -
ANALYSIS RUN LOG

Client: CDM Smith **Contract:** CAMP02

Lab code: CHEM **Case no.:** Q1915 **Sas no.:** Q1915 **Sdg no.:** Q1915

Instrument id number: **Method:** **Run number:** LB135654

Start date: 05/02/2025 **End date:** 05/02/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1234	Ag,As,Ba,Cd,Cr,Pb,Se
S1	S1	1	1238	Ag,As,Ba,Cd,Cr,Pb,Se
S2	S2	1	1242	Ag,As,Ba,Cd,Cr,Pb,Se
S3	S3	1	1246	Ag,As,Ba,Cd,Cr,Pb,Se
S4	S4	1	1251	Ag,As,Ba,Cd,Cr,Pb,Se
S5	S5	1	1255	Ag,As,Ba,Cd,Cr,Pb,Se
ICV01	ICV01	1	1347	Ag,As,Ba,Cd,Cr,Pb,Se
LLICV01	LLICV01	1	1503	Ag,As,Ba,Cd,Cr,Pb,Se
ICB01	ICB01	1	1508	Ag,As,Ba,Cd,Cr,Pb,Se
CRI01	CRI01	1	1512	Ag,As,Ba,Cd,Cr,Pb,Se
ICSA01	ICSA01	1	1530	Ag,As,Ba,Cd,Cr,Pb,Se
ICSAB01	ICSAB01	1	1546	Ag,As,Ba,Cd,Cr,Pb,Se
CCV01	CCV01	1	1601	Ag,As,Ba,Cd,Cr,Pb,Se
CCB01	CCB01	1	1605	Ag,As,Ba,Cd,Cr,Pb,Se
Q1913-02DUP	WC-12-A-202504DUP	1	1613	Ag,As,Ba,Cd,Cr,Pb,Se
Q1913-02L	WC-12-A-202504L	5	1618	Ag,As,Ba,Cd,Cr,Pb,Se
Q1913-02MS	WC-12-A-202504MS	1	1622	Ag,As,Ba,Cd,Cr,Pb,Se
Q1913-02MSD	WC-12-A-202504MSD	1	1626	Ag,As,Ba,Cd,Cr,Pb,Se
CCV02	CCV02	1	1634	Ag,As,Ba,Cd,Cr,Pb,Se
CCB02	CCB02	1	1638	Ag,As,Ba,Cd,Cr,Pb,Se