

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** CDM Smith **SDG No.:** Q1915  
**Contract:** CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** PLASMA-PURE

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|----|------------------|------------------|---------------|
| ICV112    | Mercury | 3.81           | 4.0        | 95            | 90 - 110                  | CV | 05/01/2025       | 11:47            | LB135623      |

---

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** CDM Smith **SDG No.:** Q1915  
**Contract:** CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** PLASMA-PURE

---

| Sample ID | Analyte | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M  | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|---------|----------------|------------|---------------|---------------------------|----|------------------|------------------|---------------|
| CCV62     | Mercury | 4.80           | 5.0        | 96            | 90 - 110                  | CV | 05/01/2025       | 11:58            | LB135623      |
| CCV63     | Mercury | 5.01           | 5.0        | 100           | 90 - 110                  | CV | 05/01/2025       | 12:49            | LB135623      |
| CCV64     | Mercury | 5.09           | 5.0        | 102           | 90 - 110                  | CV | 05/01/2025       | 13:19            | LB135623      |
| CCV65     | Mercury | 4.77           | 5.0        | 95            | 90 - 110                  | CV | 05/01/2025       | 13:58            | LB135623      |
| CCV66     | Mercury | 4.60           | 5.0        | 92            | 90 - 110                  | CV | 05/01/2025       | 14:29            | LB135623      |
| CCV67     | Mercury | 4.61           | 5.0        | 92            | 90 - 110                  | CV | 05/01/2025       | 15:23            | LB135623      |



- 2a -

|                                       |                    |                  |              |                  |              |                 |              |
|---------------------------------------|--------------------|------------------|--------------|------------------|--------------|-----------------|--------------|
| <b>Client:</b>                        | <u>CDM Smith</u>   | <b>SDG No.:</b>  | <u>Q1915</u> |                  |              |                 |              |
| <b>Contract:</b>                      | <u>CAMP02</u>      | <b>Lab Code:</b> | <u>CHEM</u>  | <b>Case No.:</b> | <u>Q1915</u> | <b>SAS No.:</b> | <u>Q1915</u> |
| <b>Initial Calibration Source:</b>    | <u>EPA</u>         |                  |              |                  |              |                 |              |
| <b>Continuing Calibration Source:</b> | Inorganic Ventures |                  |              |                  |              |                 |              |

| Sample ID | Analyte  | Result | True Value | % Recovery | Acceptance Window (%R) | M | Analysis Date | Analysis Time | Run Number |
|-----------|----------|--------|------------|------------|------------------------|---|---------------|---------------|------------|
|           |          | ug/L   |            |            |                        |   |               |               |            |
| ICV01     | Arsenic  | 1010   | 1000       | 100        | 90 - 110               | P | 05/01/2025    | 13:44         | LB135636   |
|           | Barium   | 495    | 520        | 95         | 90 - 110               | P | 05/01/2025    | 13:44         | LB135636   |
|           | Cadmium  | 508    | 510        | 100        | 90 - 110               | P | 05/01/2025    | 13:44         | LB135636   |
|           | Chromium | 506    | 520        | 97         | 90 - 110               | P | 05/01/2025    | 13:44         | LB135636   |
|           | Lead     | 989    | 1000       | 99         | 90 - 110               | P | 05/01/2025    | 13:44         | LB135636   |
|           | Selenium | 1040   | 1000       | 104        | 90 - 110               | P | 05/01/2025    | 13:44         | LB135636   |
|           | Silver   | 240    | 250        | 96         | 90 - 110               | P | 05/01/2025    | 13:44         | LB135636   |



- 2a -

|                                       |                    |                  |              |                  |              |                 |              |
|---------------------------------------|--------------------|------------------|--------------|------------------|--------------|-----------------|--------------|
| <b>Client:</b>                        | <u>CDM Smith</u>   | <b>SDG No.:</b>  | <u>Q1915</u> |                  |              |                 |              |
| <b>Contract:</b>                      | <u>CAMP02</u>      | <b>Lab Code:</b> | <u>CHEM</u>  | <b>Case No.:</b> | <u>Q1915</u> | <b>SAS No.:</b> | <u>Q1915</u> |
| <b>Initial Calibration Source:</b>    | <u>EPA</u>         |                  |              |                  |              |                 |              |
| <b>Continuing Calibration Source:</b> | Inorganic Ventures |                  |              |                  |              |                 |              |

| Sample ID | Analyte  | Result | True Value | % Recovery | Acceptance Window (%R) | M | Analysis Date | Analysis Time | Run Number |
|-----------|----------|--------|------------|------------|------------------------|---|---------------|---------------|------------|
|           |          | ug/L   |            |            |                        |   |               |               |            |
| LLICV01   | Arsenic  | 23.7   | 20.0       | 119        | 80 - 120               | P | 05/01/2025    | 14:01         | LB135636   |
|           | Barium   | 95.8   | 100        | 96         | 80 - 120               | P | 05/01/2025    | 14:01         | LB135636   |
|           | Cadmium  | 6.39   | 6.0        | 107        | 80 - 120               | P | 05/01/2025    | 14:01         | LB135636   |
|           | Chromium | 10.5   | 10.0       | 105        | 80 - 120               | P | 05/01/2025    | 14:01         | LB135636   |
|           | Lead     | 13.1   | 12.0       | 109        | 80 - 120               | P | 05/01/2025    | 14:01         | LB135636   |
|           | Selenium | 22.9   | 20.0       | 114        | 80 - 120               | P | 05/01/2025    | 14:01         | LB135636   |
|           | Silver   | 10.2   | 10.0       | 102        | 80 - 120               | P | 05/01/2025    | 14:01         | LB135636   |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

**Client:** CDM Smith **SDG No.:** Q1915  
**Contract:** CAMP02 **Lab Code:** CHEM **Case No.:** Q1915 **SAS No.:** Q1915  
**Initial Calibration Source:** EPA  
**Continuing Calibration Source:** Inorganic Ventures

| Sample ID | Analyte  | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Arsenic  | 5090           | 5000       | 102           | 90 - 110                  | P | 05/01/2025       | 15:05            | LB135636      |
|           | Barium   | 9650           | 10000      | 96            | 90 - 110                  | P | 05/01/2025       | 15:05            | LB135636      |
|           | Cadmium  | 2500           | 2500       | 100           | 90 - 110                  | P | 05/01/2025       | 15:05            | LB135636      |
|           | Chromium | 1000           | 1000       | 100           | 90 - 110                  | P | 05/01/2025       | 15:05            | LB135636      |
|           | Lead     | 5010           | 5000       | 100           | 90 - 110                  | P | 05/01/2025       | 15:05            | LB135636      |
|           | Selenium | 5140           | 5000       | 103           | 90 - 110                  | P | 05/01/2025       | 15:05            | LB135636      |
|           | Silver   | 1250           | 1250       | 100           | 90 - 110                  | P | 05/01/2025       | 15:05            | LB135636      |
| CCV02     | Arsenic  | 4930           | 5000       | 99            | 90 - 110                  | P | 05/01/2025       | 16:00            | LB135636      |
|           | Barium   | 10200          | 10000      | 102           | 90 - 110                  | P | 05/01/2025       | 16:00            | LB135636      |
|           | Cadmium  | 2470           | 2500       | 99            | 90 - 110                  | P | 05/01/2025       | 16:00            | LB135636      |
|           | Chromium | 995            | 1000       | 100           | 90 - 110                  | P | 05/01/2025       | 16:00            | LB135636      |
|           | Lead     | 4980           | 5000       | 100           | 90 - 110                  | P | 05/01/2025       | 16:00            | LB135636      |
|           | Selenium | 4960           | 5000       | 99            | 90 - 110                  | P | 05/01/2025       | 16:00            | LB135636      |
|           | Silver   | 1220           | 1250       | 98            | 90 - 110                  | P | 05/01/2025       | 16:00            | LB135636      |
| CCV03     | Arsenic  | 5040           | 5000       | 101           | 90 - 110                  | P | 05/01/2025       | 17:03            | LB135636      |
|           | Barium   | 9950           | 10000      | 100           | 90 - 110                  | P | 05/01/2025       | 17:03            | LB135636      |
|           | Cadmium  | 2510           | 2500       | 100           | 90 - 110                  | P | 05/01/2025       | 17:03            | LB135636      |
|           | Chromium | 988            | 1000       | 99            | 90 - 110                  | P | 05/01/2025       | 17:03            | LB135636      |
|           | Lead     | 5010           | 5000       | 100           | 90 - 110                  | P | 05/01/2025       | 17:03            | LB135636      |
|           | Selenium | 5060           | 5000       | 101           | 90 - 110                  | P | 05/01/2025       | 17:03            | LB135636      |
|           | Silver   | 1220           | 1250       | 97            | 90 - 110                  | P | 05/01/2025       | 17:03            | LB135636      |
| CCV04     | Arsenic  | 5160           | 5000       | 103           | 90 - 110                  | P | 05/01/2025       | 17:49            | LB135636      |
|           | Barium   | 10000          | 10000      | 100           | 90 - 110                  | P | 05/01/2025       | 17:49            | LB135636      |
|           | Cadmium  | 2470           | 2500       | 99            | 90 - 110                  | P | 05/01/2025       | 17:49            | LB135636      |
|           | Chromium | 987            | 1000       | 99            | 90 - 110                  | P | 05/01/2025       | 17:49            | LB135636      |
|           | Lead     | 4950           | 5000       | 99            | 90 - 110                  | P | 05/01/2025       | 17:49            | LB135636      |
|           | Selenium | 5240           | 5000       | 105           | 90 - 110                  | P | 05/01/2025       | 17:49            | LB135636      |
|           | Silver   | 1250           | 1250       | 100           | 90 - 110                  | P | 05/01/2025       | 17:49            | LB135636      |
| CCV05     | Arsenic  | 5090           | 5000       | 102           | 90 - 110                  | P | 05/01/2025       | 18:36            | LB135636      |
|           | Barium   | 9950           | 10000      | 100           | 90 - 110                  | P | 05/01/2025       | 18:36            | LB135636      |
|           | Cadmium  | 2500           | 2500       | 100           | 90 - 110                  | P | 05/01/2025       | 18:36            | LB135636      |
|           | Chromium | 984            | 1000       | 98            | 90 - 110                  | P | 05/01/2025       | 18:36            | LB135636      |
|           | Lead     | 4980           | 5000       | 100           | 90 - 110                  | P | 05/01/2025       | 18:36            | LB135636      |
|           | Selenium | 5150           | 5000       | 103           | 90 - 110                  | P | 05/01/2025       | 18:36            | LB135636      |

**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

Client: CDM Smith SDG No.: Q1915  
Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915  
Initial Calibration Source: EPA  
Continuing Calibration Source: Inorganic Ventures

| Sample ID | Analyte  | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV05     | Silver   | 1230           | 1250       | 98            | 90 - 110                  | P | 05/01/2025       | 18:36            | LB135636      |
| CCV06     | Arsenic  | 4990           | 5000       | 100           | 90 - 110                  | P | 05/01/2025       | 19:23            | LB135636      |
|           | Barium   | 9830           | 10000      | 98            | 90 - 110                  | P | 05/01/2025       | 19:23            | LB135636      |
|           | Cadmium  | 2460           | 2500       | 98            | 90 - 110                  | P | 05/01/2025       | 19:23            | LB135636      |
|           | Chromium | 980            | 1000       | 98            | 90 - 110                  | P | 05/01/2025       | 19:23            | LB135636      |
|           | Lead     | 4920           | 5000       | 98            | 90 - 110                  | P | 05/01/2025       | 19:23            | LB135636      |
|           | Selenium | 5030           | 5000       | 101           | 90 - 110                  | P | 05/01/2025       | 19:23            | LB135636      |
|           | Silver   | 1240           | 1250       | 99            | 90 - 110                  | P | 05/01/2025       | 19:23            | LB135636      |
| CCV07     | Arsenic  | 5090           | 5000       | 102           | 90 - 110                  | P | 05/01/2025       | 20:11            | LB135636      |
|           | Barium   | 9790           | 10000      | 98            | 90 - 110                  | P | 05/01/2025       | 20:11            | LB135636      |
|           | Cadmium  | 2500           | 2500       | 100           | 90 - 110                  | P | 05/01/2025       | 20:11            | LB135636      |
|           | Chromium | 995            | 1000       | 100           | 90 - 110                  | P | 05/01/2025       | 20:11            | LB135636      |
|           | Lead     | 4980           | 5000       | 100           | 90 - 110                  | P | 05/01/2025       | 20:11            | LB135636      |
|           | Selenium | 5120           | 5000       | 102           | 90 - 110                  | P | 05/01/2025       | 20:11            | LB135636      |
|           | Silver   | 1250           | 1250       | 100           | 90 - 110                  | P | 05/01/2025       | 20:11            | LB135636      |
| CCV08     | Arsenic  | 5180           | 5000       | 104           | 90 - 110                  | P | 05/01/2025       | 20:24            | LB135636      |
|           | Barium   | 9880           | 10000      | 99            | 90 - 110                  | P | 05/01/2025       | 20:24            | LB135636      |
|           | Cadmium  | 2490           | 2500       | 100           | 90 - 110                  | P | 05/01/2025       | 20:24            | LB135636      |
|           | Chromium | 980            | 1000       | 98            | 90 - 110                  | P | 05/01/2025       | 20:24            | LB135636      |
|           | Lead     | 4990           | 5000       | 100           | 90 - 110                  | P | 05/01/2025       | 20:24            | LB135636      |
|           | Selenium | 5270           | 5000       | 105           | 90 - 110                  | P | 05/01/2025       | 20:24            | LB135636      |
|           | Silver   | 1240           | 1250       | 99            | 90 - 110                  | P | 05/01/2025       | 20:24            | LB135636      |



- 2a -

|                                       |                    |                  |              |                                                            |
|---------------------------------------|--------------------|------------------|--------------|------------------------------------------------------------|
| <b>Client:</b>                        | <u>CDM Smith</u>   | <b>SDG No.:</b>  | <u>Q1915</u> |                                                            |
| <b>Contract:</b>                      | <u>CAMP02</u>      | <b>Lab Code:</b> | <u>CHEM</u>  | <b>Case No.:</b> <u>Q1915</u> <b>SAS No.:</b> <u>Q1915</u> |
| <b>Initial Calibration Source:</b>    | <u>EPA</u>         |                  |              |                                                            |
| <b>Continuing Calibration Source:</b> | Inorganic Ventures |                  |              |                                                            |

| Sample ID | Analyte  | Result | True Value | %        | Acceptance  | M | Analysis Date | Analysis Time | Run Number |
|-----------|----------|--------|------------|----------|-------------|---|---------------|---------------|------------|
|           |          | ug/L   |            | Recovery | Window (%R) |   |               |               |            |
| ICV01     | Arsenic  | 1060   | 1000       | 106      | 90 - 110    | P | 05/02/2025    | 13:47         | LB135654   |
|           | Barium   | 518    | 520        | 100      | 90 - 110    | P | 05/02/2025    | 13:47         | LB135654   |
|           | Cadmium  | 493    | 510        | 97       | 90 - 110    | P | 05/02/2025    | 13:47         | LB135654   |
|           | Chromium | 503    | 520        | 97       | 90 - 110    | P | 05/02/2025    | 13:47         | LB135654   |
|           | Lead     | 966    | 1000       | 97       | 90 - 110    | P | 05/02/2025    | 13:47         | LB135654   |
|           | Selenium | 1040   | 1000       | 104      | 90 - 110    | P | 05/02/2025    | 13:47         | LB135654   |
|           | Silver   | 230    | 250        | 92       | 90 - 110    | P | 05/02/2025    | 13:47         | LB135654   |



- 2a -

|                                       |                    |                  |              |                                                            |
|---------------------------------------|--------------------|------------------|--------------|------------------------------------------------------------|
| <b>Client:</b>                        | <u>CDM Smith</u>   | <b>SDG No.:</b>  | <u>Q1915</u> |                                                            |
| <b>Contract:</b>                      | <u>CAMP02</u>      | <b>Lab Code:</b> | <u>CHEM</u>  | <b>Case No.:</b> <u>Q1915</u> <b>SAS No.:</b> <u>Q1915</u> |
| <b>Initial Calibration Source:</b>    | <u>EPA</u>         |                  |              |                                                            |
| <b>Continuing Calibration Source:</b> | Inorganic Ventures |                  |              |                                                            |

| Sample ID | Analyte  | Result | True Value | %        | Acceptance  | M | Analysis Date | Analysis Time | Run Number |
|-----------|----------|--------|------------|----------|-------------|---|---------------|---------------|------------|
|           |          | ug/L   |            | Recovery | Window (%R) |   |               |               |            |
| LLICV01   | Arsenic  | 19.0   | 20.0       | 95       | 80 - 120    | P | 05/02/2025    | 15:03         | LB135654   |
|           | Barium   | 81.5   | 100        | 82       | 80 - 120    | P | 05/02/2025    | 15:03         | LB135654   |
|           | Cadmium  | 5.51   | 6.0        | 92       | 80 - 120    | P | 05/02/2025    | 15:03         | LB135654   |
|           | Chromium | 10.1   | 10.0       | 101      | 80 - 120    | P | 05/02/2025    | 15:03         | LB135654   |
|           | Lead     | 11.1   | 12.0       | 93       | 80 - 120    | P | 05/02/2025    | 15:03         | LB135654   |
|           | Selenium | 16.8   | 20.0       | 84       | 80 - 120    | P | 05/02/2025    | 15:03         | LB135654   |
|           | Silver   | 10.8   | 10.0       | 108      | 80 - 120    | P | 05/02/2025    | 15:03         | LB135654   |



**Metals**

- 2a -

**INITIAL AND CONTINUING CALIBRATION VERIFICATION**

Client: CDM Smith SDG No.: Q1915  
Contract: CAMP02 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915  
Initial Calibration Source: EPA  
Continuing Calibration Source: Inorganic Ventures

| Sample ID | Analyte  | Result<br>ug/L | True Value | %<br>Recovery | Acceptance<br>Window (%R) | M | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|-----------|----------|----------------|------------|---------------|---------------------------|---|------------------|------------------|---------------|
| CCV01     | Arsenic  | 5450           | 5000       | 109           | 90 - 110                  | P | 05/02/2025       | 16:01            | LB135654      |
|           | Barium   | 9540           | 10000      | 95            | 90 - 110                  | P | 05/02/2025       | 16:01            | LB135654      |
|           | Cadmium  | 2440           | 2500       | 98            | 90 - 110                  | P | 05/02/2025       | 16:01            | LB135654      |
|           | Chromium | 1010           | 1000       | 101           | 90 - 110                  | P | 05/02/2025       | 16:01            | LB135654      |
|           | Lead     | 4820           | 5000       | 96            | 90 - 110                  | P | 05/02/2025       | 16:01            | LB135654      |
|           | Selenium | 5460           | 5000       | 109           | 90 - 110                  | P | 05/02/2025       | 16:01            | LB135654      |
|           | Silver   | 1270           | 1250       | 102           | 90 - 110                  | P | 05/02/2025       | 16:01            | LB135654      |
| CCV02     | Arsenic  | 5240           | 5000       | 105           | 90 - 110                  | P | 05/02/2025       | 16:34            | LB135654      |
|           | Barium   | 9720           | 10000      | 97            | 90 - 110                  | P | 05/02/2025       | 16:34            | LB135654      |
|           | Cadmium  | 2390           | 2500       | 96            | 90 - 110                  | P | 05/02/2025       | 16:34            | LB135654      |
|           | Chromium | 1000           | 1000       | 100           | 90 - 110                  | P | 05/02/2025       | 16:34            | LB135654      |
|           | Lead     | 4740           | 5000       | 95            | 90 - 110                  | P | 05/02/2025       | 16:34            | LB135654      |
|           | Selenium | 5240           | 5000       | 105           | 90 - 110                  | P | 05/02/2025       | 16:34            | LB135654      |
|           | Silver   | 1260           | 1250       | 101           | 90 - 110                  | P | 05/02/2025       | 16:34            | LB135654      |