

LAB CHRONICLE

OrderID:	Q1915	OrderDate:	4/29/2025 2:29:00 PM
Client:	CDM Smith	Project:	Con Ed UTEN Mount Vernon, NY
Contact:	Marcie Ann Encinas	Location:	L41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1915-01	WC-04282025	TCLP	TCLP BNA	8270E	04/28/25	04/30/25	05/01/25	04/29/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1915
Client: CDM Smith

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
Total Svoc :					0.00			
Total Concentration:					0.00			



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	04/30/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/30/25
Client Sample ID:	PB167774TB	SDG No.:	Q1915
Lab Sample ID:	PB167774TB	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142256.D	1	04/30/25 13:15	05/01/25 13:07	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	50.0	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	50.0	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	50.0	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	100	U	11.0	100	ug/L
67-72-1	Hexachloroethane	50.0	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	50.0	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	50.0	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	50.0	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	50.0	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	50.0	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	50.0	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	100	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	116		10 - 139	78%	SPK: 150
13127-88-3	Phenol-d6	116		10 - 134	77%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.3		49 - 133	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	72.1		52 - 132	72%	SPK: 100
118-79-6	2,4,6-Tribromophenol	131		44 - 137	87%	SPK: 150
1718-51-0	Terphenyl-d14	67.9		48 - 125	68%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	218000		6.904		
1146-65-2	Naphthalene-d8	840000		8.186		
15067-26-2	Acenaphthene-d10	443000		9.945		
1517-22-2	Phenanthrene-d10	773000		11.427		
1719-03-5	Chrysene-d12	539000		14.068		
1520-96-3	Perylene-d12	394000		15.562		

Report of Analysis

Client:	CDM Smith	Date Collected:	04/30/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/30/25
Client Sample ID:	PB167774TB	SDG No.:	Q1915
Lab Sample ID:	PB167774TB	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142256.D	1	04/30/25 13:15	05/01/25 13:07	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	04/28/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/29/25
Client Sample ID:	WC-04282025	SDG No.:	Q1915
Lab Sample ID:	Q1915-01	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142261.D	1	04/30/25 13:15	05/01/25 15:34	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	50.0	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	50.0	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	50.0	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	100	U	11.0	100	ug/L
67-72-1	Hexachloroethane	50.0	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	50.0	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	50.0	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	50.0	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	50.0	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	50.0	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	50.0	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	100	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	101		10 - 139	67%	SPK: 150
13127-88-3	Phenol-d6	91.6		10 - 134	61%	SPK: 150
4165-60-0	Nitrobenzene-d5	86.8		49 - 133	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	75.3		52 - 132	75%	SPK: 100
118-79-6	2,4,6-Tribromophenol	116		44 - 137	78%	SPK: 150
1718-51-0	Terphenyl-d14	57.8		48 - 125	58%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	211000	6.91			
1146-65-2	Naphthalene-d8	797000	8.186			
15067-26-2	Acenaphthene-d10	393000	9.939			
1517-22-2	Phenanthrene-d10	545000	11.427			
1719-03-5	Chrysene-d12	445000	14.062			
1520-96-3	Perylene-d12	444000	15.556			

Report of Analysis

Client:	CDM Smith	Date Collected:	04/28/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/29/25
Client Sample ID:	WC-04282025	SDG No.:	Q1915
Lab Sample ID:	Q1915-01	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142261.D	1	04/30/25 13:15	05/01/25 15:34	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167774TB	PB167774TB	2-Fluorophenol	150	116	78		10	139
		Phenol-d6	150	116	77		10	134
		Nitrobenzene-d5	100	87.3	87		49	133
		2-Fluorobiphenyl	100	72.1	72		52	132
		2,4,6-Tribromophenol	150	131	87		44	137
		Terphenyl-d14	100	67.9	68		48	125
PB167810BL	PB167810BL	2-Fluorophenol	150	110	73		10	139
		Phenol-d6	150	109	73		10	134
		Nitrobenzene-d5	100	83.3	83		49	133
		2-Fluorobiphenyl	100	69.1	69		52	132
		2,4,6-Tribromophenol	150	124	83		44	137
		Terphenyl-d14	100	63.3	63		48	125
PB167810BS	PB167810BS	2-Fluorophenol	150	111	74		10	139
		Phenol-d6	150	113	75		10	134
		Nitrobenzene-d5	100	82.4	82		49	133
		2-Fluorobiphenyl	100	68.5	69		52	132
		2,4,6-Tribromophenol	150	132	88		44	137
		Terphenyl-d14	100	74.1	74		48	125
Q1901-08MS	B-167-SB01MS	2-Fluorophenol	150	96.6	64		10	139
		Phenol-d6	150	89.8	60		10	134
		Nitrobenzene-d5	100	84.0	84		49	133
		2-Fluorobiphenyl	100	70.7	71		52	132
		2,4,6-Tribromophenol	150	126	84		44	137
		Terphenyl-d14	100	70.4	70		48	125
Q1901-08MSD	B-167-SB01MSD	2-Fluorophenol	150	99.4	66		10	139
		Phenol-d6	150	92.8	62		10	134
		Nitrobenzene-d5	100	87.8	88		49	133
		2-Fluorobiphenyl	100	72.3	72		52	132
		2,4,6-Tribromophenol	150	133	88		44	137
		Terphenyl-d14	100	72.8	73		48	125
Q1915-01	WC-04282025	2-Fluorophenol	150	101	67		10	139
		Phenol-d6	150	91.6	61		10	134
		Nitrobenzene-d5	100	86.8	87		49	133
		2-Fluorobiphenyl	100	75.3	75		52	132
		2,4,6-Tribromophenol	150	116	78		44	137
		Terphenyl-d14	100	57.8	58		48	125

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1901-08MS	Client Sample ID:	B-167-SB01MS					DataFile:	BF142277.D		
Pyridine	500	24.1	270	ug/L	49				10	109	
1,4-Dichlorobenzene	500	0	340	ug/L	68				55	125	
2-Methylphenol	500	0	360	ug/L	72				60	131	
3+4-Methylphenols	500	0	360	ug/L	72				54	136	
Hexachloroethane	500	0	350	ug/L	70				19	146	
Nitrobenzene	500	0	440	ug/L	88				62	112	
Hexachlorobutadiene	500	0	380	ug/L	76				52	125	
2,4,6-Trichlorophenol	500	0	470	ug/L	94				78	112	
2,4,5-Trichlorophenol	500	0	460	ug/L	92				71	111	
2,4-Dinitrotoluene	500	0	500	ug/L	100				74	137	
Hexachlorobenzene	500	0	440	ug/L	88				72	115	
Pentachlorophenol	1000	0	940	ug/L	94				52	162	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1901-08MSD	Client Sample ID:	B-167-SB01MSD					DataFile:	BF142278.D		
Pyridine	500	24.1	280	ug/L	51		4		10	109	20
1,4-Dichlorobenzene	500	0	340	ug/L	68		0		55	125	20
2-Methylphenol	500	0	370	ug/L	74		3		60	131	20
3+4-Methylphenols	500	0	370	ug/L	74		3		54	136	20
Hexachloroethane	500	0	370	ug/L	74		6		19	146	20
Nitrobenzene	500	0	460	ug/L	92		4		62	112	20
Hexachlorobutadiene	500	0	390	ug/L	78		3		52	125	20
2,4,6-Trichlorophenol	500	0	480	ug/L	96		2		78	112	20
2,4,5-Trichlorophenol	500	0	470	ug/L	94		2		71	111	20
2,4-Dinitrotoluene	500	0	530	ug/L	106		6		74	137	20
Hexachlorobenzene	500	0	450	ug/L	90		2		72	115	20
Pentachlorophenol	1000	0	970	ug/L	97		3		52	162	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: 8270E DataFile: BF142255.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167810BS	Pyridine	50	37.2	ug/L	74				29	97	
	1,4-Dichlorobenzene	50	39.6	ug/L	79				76	103	
	2-Methylphenol	50	41.0	ug/L	82				69	109	
	3+4-Methylphenols	50	40.8	ug/L	82				67	106	
	Hexachloroethane	50	41.8	ug/L	84				76	118	
	Nitrobenzene	50	45.8	ug/L	92				58	106	
	Hexachlorobutadiene	50	41.5	ug/L	83				69	101	
	2,4,6-Trichlorophenol	50	43.4	ug/L	87				61	110	
	2,4,5-Trichlorophenol	50	44.9	ug/L	90				70	106	
	2,4-Dinitrotoluene	50	49.8	ug/L	100				60	115	
	Hexachlorobenzene	50	41.6	ug/L	83				73	106	
	Pentachlorophenol	100	93.4	ug/L	93				47	114	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167810BL

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q1915

SAS No.: Q1915 SDG NO.: Q1915

Lab File ID: BF142254.D

Lab Sample ID: PB167810BL

Instrument ID: BNA_F

Date Extracted: 04/30/2025

Matrix: (soil/water) water

Date Analyzed: 05/01/2025

Level: (low/med) LOW

Time Analyzed: 12:10

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167810BS	PB167810BS	BF142255.D	05/01/2025
WC-04282025	Q1915-01	BF142261.D	05/01/2025
B-167-SB01MS	Q1901-08MS	BF142277.D	05/02/2025
B-167-SB01MSD	Q1901-08MSD	BF142278.D	05/02/2025
PB167774TB	PB167774TB	BF142256.D	05/01/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q1915 SDG NO.: Q1915

Lab File ID: BF142238.D

DFTPP Injection Date: 04/30/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.6
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	27.3
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	37.1
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.3
275	10.0 - 60.0% of mass 198	22.8
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.9 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142239.D	04/30/2025	11:24
SSTDICC005	SSTDICC005	BF142240.D	04/30/2025	11:52
SSTDICC010	SSTDICC010	BF142241.D	04/30/2025	12:20
SSTDICC020	SSTDICC020	BF142242.D	04/30/2025	12:49
SSTDICCC040	SSTDICCC040	BF142243.D	04/30/2025	13:17
SSTDICC050	SSTDICC050	BF142244.D	04/30/2025	13:46
SSTDICC060	SSTDICC060	BF142245.D	04/30/2025	14:15
SSTDICC080	SSTDICC080	BF142246.D	04/30/2025	14:43

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q1915 SDG NO.: Q1915

Lab File ID: BF142249.D

DFTPP Injection Date: 05/01/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.4
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	31.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	42.4
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142250.D	05/01/2025	10:17
PB167810BL	PB167810BL	BF142254.D	05/01/2025	12:10
PB167810BS	PB167810BS	BF142255.D	05/01/2025	12:39
PB167774TB	PB167774TB	BF142256.D	05/01/2025	13:07
WC-04282025	Q1915-01	BF142261.D	05/01/2025	15:34

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q1915 SDG NO.: Q1915

Lab File ID: BF142273.D

DFTPP Injection Date: 05/02/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:11

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	34.7
68	Less than 2.0% of mass 69	0.4 (1.6) 1
69	Mass 69 relative abundance	27.6
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	37.6
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.6
275	10.0 - 60.0% of mass 198	23.7
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142274.D	05/02/2025	10:39
B-167-SB01MS	Q1901-08MS	BF142277.D	05/02/2025	12:27
B-167-SB01MSD	Q1901-08MSD	BF142278.D	05/02/2025	12:55

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	186578	6.91	719586	8.19	381198	9.95
UPPER LIMIT	373156	7.41	1439170	8.692	762396	10.445
LOWER LIMIT	93289	6.41	359793	7.692	190599	9.445
EPA SAMPLE NO.						
01 PB167774TB	217516	6.90	839699	8.19	443070	9.95
02 PB167810BL	223997	6.90	857825	8.19	449926	9.95
03 PB167810BS	228833	6.91	892782	8.19	472284	9.95
04 WC-04282025	210959	6.91	797035	8.19	392717	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	647816	11.433	384442	14.074	349272	15.562
UPPER LIMIT	1295630	11.933	768884	14.574	698544	16.062
LOWER LIMIT	323908	10.933	192221	13.574	174636	15.062
EPA SAMPLE NO.						
01 PB167774TB	773002	11.43	539235	14.07	394051	15.56
02 PB167810BL	794711	11.43	568207	14.07	404816	15.56
03 PB167810BS	804874	11.43	464382	14.07	446826	15.57
04 WC-04282025	545118	11.43	445392	14.06	444396	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/02/2025

Lab File ID: BF142274.D Time Analyzed: 10:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	189457	6.91	747235	8.19	389025	9.95
UPPER LIMIT	378914	7.41	1494470	8.692	778050	10.445
LOWER LIMIT	94728.5	6.41	373618	7.692	194513	9.445
EPA SAMPLE NO.						
01 B-167-SB01MS	212656	6.91	808361	8.19	402059	9.95
02 B-167-SB01MSD	217028	6.91	832936	8.19	419407	9.95

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/02/2025

Lab File ID: BF142274.D Time Analyzed: 10:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	652685	11.427	354752	14.068	327562	15.557
UPPER LIMIT	1305370	11.927	709504	14.568	655124	16.057
LOWER LIMIT	326343	10.927	177376	13.568	163781	15.057
EPA SAMPLE NO.						
01 B-167-SB01MS	629997	11.43	370460	14.07	379220	15.56
02 B-167-SB01MSD	684254	11.43	400079	14.07	401381	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BL	SDG No.:	Q1915
Lab Sample ID:	PB167810BL	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142254.D	1	04/30/25 13:15	05/01/25 12:10	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	5.00	U	1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	5.00	U	0.53	5.00	ug/L
95-48-7	2-Methylphenol	5.00	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	10.0	U	1.10	10.0	ug/L
67-72-1	Hexachloroethane	5.00	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	5.00	U	0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	5.00	U	0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	5.00	U	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	5.00	U	0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	5.00	U	1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	5.00	U	0.52	5.00	ug/L
87-86-5	Pentachlorophenol	10.0	U	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	110		10 - 139	73%	SPK: 150
13127-88-3	Phenol-d6	109		10 - 134	73%	SPK: 150
4165-60-0	Nitrobenzene-d5	83.3		49 - 133	83%	SPK: 100
321-60-8	2-Fluorobiphenyl	69.1		52 - 132	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	124		44 - 137	83%	SPK: 150
1718-51-0	Terphenyl-d14	63.3		48 - 125	63%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	224000		6.904		
1146-65-2	Naphthalene-d8	858000		8.187		
15067-26-2	Acenaphthene-d10	450000		9.945		
1517-22-2	Phenanthrene-d10	795000		11.428		
1719-03-5	Chrysene-d12	568000		14.069		
1520-96-3	Perylene-d12	405000		15.563		

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BL	SDG No.:	Q1915
Lab Sample ID:	PB167810BL	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142254.D	1	04/30/25 13:15	05/01/25 12:10	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BS	SDG No.:	Q1915
Lab Sample ID:	PB167810BS	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142255.D	1	04/30/25 13:15	05/01/25 12:39	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	37.2		1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	39.6		0.53	5.00	ug/L
95-48-7	2-Methylphenol	41.0		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	40.8		1.10	10.0	ug/L
67-72-1	Hexachloroethane	41.8		0.65	5.00	ug/L
98-95-3	Nitrobenzene	45.8		0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	41.5		0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	43.4		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	44.9		0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	49.8		1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	41.6		0.52	5.00	ug/L
87-86-5	Pentachlorophenol	93.4	E	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	111		10 - 139	74%	SPK: 150
13127-88-3	Phenol-d6	113		10 - 134	75%	SPK: 150
4165-60-0	Nitrobenzene-d5	82.4		49 - 133	82%	SPK: 100
321-60-8	2-Fluorobiphenyl	68.5		52 - 132	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	132		44 - 137	88%	SPK: 150
1718-51-0	Terphenyl-d14	74.1		48 - 125	74%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	229000		6.91		
1146-65-2	Naphthalene-d8	893000		8.192		
15067-26-2	Acenaphthene-d10	472000		9.945		
1517-22-2	Phenanthrene-d10	805000		11.433		
1719-03-5	Chrysene-d12	464000		14.074		
1520-96-3	Perylene-d12	447000		15.568		

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BS	SDG No.:	Q1915
Lab Sample ID:	PB167810BS	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142255.D	1	04/30/25 13:15	05/01/25 12:39	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MS	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142277.D	1	04/30/25 13:15	05/02/25 12:27	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	270		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	340		5.30	50.0	ug/L
95-48-7	2-Methylphenol	360		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	360		11.0	100	ug/L
67-72-1	Hexachloroethane	350		6.50	50.0	ug/L
98-95-3	Nitrobenzene	440		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	380		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	470		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	460		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	500		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	440		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	940	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	96.6		10 - 139	64%	SPK: 150
13127-88-3	Phenol-d6	89.8		10 - 134	60%	SPK: 150
4165-60-0	Nitrobenzene-d5	84.0		49 - 133	84%	SPK: 100
321-60-8	2-Fluorobiphenyl	70.7		52 - 132	71%	SPK: 100
118-79-6	2,4,6-Tribromophenol	126		44 - 137	84%	SPK: 150
1718-51-0	Terphenyl-d14	70.4		48 - 125	70%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	213000		6.91		
1146-65-2	Naphthalene-d8	808000		8.193		
15067-26-2	Acenaphthene-d10	402000		9.945		
1517-22-2	Phenanthrene-d10	630000		11.433		
1719-03-5	Chrysene-d12	370000		14.069		
1520-96-3	Perylene-d12	379000		15.557		

Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MS	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142277.D	1	04/30/25 13:15	05/02/25 12:27	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MSD	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142278.D	1	04/30/25 13:15	05/02/25 12:55	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	280		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	340		5.30	50.0	ug/L
95-48-7	2-Methylphenol	370		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	370		11.0	100	ug/L
67-72-1	Hexachloroethane	370		6.50	50.0	ug/L
98-95-3	Nitrobenzene	460		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	390		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	480		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	470		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	530		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	450		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	970	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	99.4		10 - 139	66%	SPK: 150
13127-88-3	Phenol-d6	92.8		10 - 134	62%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.8		49 - 133	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	72.3		52 - 132	72%	SPK: 100
118-79-6	2,4,6-Tribromophenol	133		44 - 137	88%	SPK: 150
1718-51-0	Terphenyl-d14	72.8		48 - 125	73%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	217000		6.91		
1146-65-2	Naphthalene-d8	833000		8.193		
15067-26-2	Acenaphthene-d10	419000		9.945		
1517-22-2	Phenanthrene-d10	684000		11.433		
1719-03-5	Chrysene-d12	400000		14.069		
1520-96-3	Perylene-d12	401000		15.563		

Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MSD	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142278.D	1	04/30/25 13:15	05/02/25 12:55	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF043025.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 30 16:00:01 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF142239.D 5 =BF142240.D 10 =BF142241.D 20 =BF142242.D 40 =BF142243.D 50 =BF142244.D 60 =BF142245.D 80 =BF142246.D

Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I 1,4-Dichlorobenzen...	-----ISTD-----									
2) 1,4-Dioxane	0.563	0.534	0.561	0.592	0.580	0.555	0.530	0.559	4.03	
3) Pyridine	1.465	1.391	1.454	1.522	1.503	1.460	1.380	1.454	3.62	
4) n-Nitrosodimet...	0.747	0.699	0.728	0.807	0.789	0.766	0.731	0.752	4.95	
5) S 2-Fluorophenol	1.324	1.238	1.265	1.314	1.285	1.210	1.131	1.252	5.35	
6) Aniline	2.212	2.085	2.157	2.245	2.203	2.101	1.981	2.140	4.29	
7) S Phenol-d6	1.559	1.505	1.534	1.593	1.550	1.475	1.370	1.512	4.85	
8) 2-Chlorophenol	1.252	1.232	1.304	1.379	1.354	1.313	1.225	1.294	4.64	
9) Benzaldehyde	1.154	1.088	1.095	1.117	1.078	0.993	0.875	1.057	8.91	
10) C Phenol	1.679	1.588	1.657	1.715	1.687	1.603	1.468	1.628	5.15	
11) bis(2-Chloroet...	1.345	1.250	1.270	1.318	1.285	1.231	1.132	1.262	5.46	
12) 1,3-Dichlorobe...	1.571	1.485	1.485	1.533	1.478	1.396	1.299	1.464	6.18	
13) C 1,4-Dichlorobe...	1.610	1.486	1.495	1.546	1.488	1.404	1.295	1.475	6.86	
14) 1,2-Dichlorobe...	1.489	1.418	1.431	1.475	1.432	1.356	1.255	1.408	5.67	
15) Benzyl Alcohol	1.047	1.017	1.064	1.149	1.126	1.082	1.023	1.073	4.69	
16) 2,2'-oxybis(1-...	2.368	2.287	2.303	2.386	2.344	2.216	2.043	2.278	5.19	
17) 2-Methylphenol	1.085	1.029	1.068	1.121	1.084	1.050	0.983	1.060	4.19	
18) Hexachloroethane	0.468	0.466	0.481	0.511	0.514	0.489	0.451	0.483	4.84	
19) P n-Nitroso-di-n...	0.953	0.987	0.937	0.949	0.967	0.939	0.898	0.844	0.934	4.77
20) 3+4-Methylphenols	1.357	1.323	1.372	1.400	1.366	1.296	1.166	1.326	5.90	

21) I Naphthalene-d8	-----ISTD-----									
22) Acetophenone	0.511	0.482	0.483	0.468	0.461	0.430	0.404	0.463	7.70	
23) S Nitrobenzene-d5	0.217	0.251	0.302	0.340	0.351	0.339	0.330	0.304	16.88	
24) Nitrobenzene	0.229	0.261	0.297	0.328	0.338	0.323	0.314	0.299	13.31	
25) Isophorone	0.660	0.630	0.642	0.660	0.654	0.626	0.601	0.639	3.36	
26) C 2-Nitrophenol	0.048	0.057	0.083	0.115	0.132	0.136	0.142	0.102	38.38#	
27) 2,4-Dimethylph...	0.315	0.319	0.333	0.337	0.333	0.315	0.302	0.322	3.96	
28) bis(2-Chloroet...	0.425	0.404	0.414	0.415	0.410	0.386	0.366	0.403	4.98	
29) C 2,4-Dichloroph...	0.246	0.253	0.274	0.291	0.286	0.272	0.263	0.269	6.12	
30) 1,2,4-Trichlor...	0.320	0.303	0.305	0.312	0.304	0.287	0.273	0.301	5.28	
31) Naphthalene	1.106	1.040	1.035	1.020	0.999	0.924	0.858	0.997	8.21	
32) Benzoic acid	0.048	0.087	0.130	0.146	0.159	0.179	0.125	0.125	39.24	
33) 4-Chloroaniline	0.446	0.414	0.429	0.429	0.423	0.396	0.371	0.415	5.98	
34) C Hexachlorobuta...	0.180	0.175	0.179	0.185	0.178	0.170	0.161	0.175	4.35	
35) Caprolactam	0.068	0.073	0.081	0.087	0.089	0.085	0.083	0.081	9.45	
36) C 4-Chloro-3-met...	0.262	0.265	0.279	0.295	0.293	0.276	0.266	0.277	4.84	
37) 2-Methylnaphth...	0.679	0.637	0.633	0.623	0.612	0.574	0.534	0.613	7.69	
38) 1-Methylnaphth...	0.700	0.655	0.667	0.652	0.641	0.590	0.542	0.635	8.29	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF043025.M

39) I	Acenaphthene-d10	-----ISTD-----								
40)	1,2,4,5-Tetrac...	0.621	0.575	0.572	0.576	0.566	0.534	0.496	0.563	6.88
41) P	Hexachlorocycl...	0.244	0.265	0.306	0.352	0.368	0.358	0.343	0.319	15.23
42) S	2,4,6-Tribromo...	0.139	0.158	0.184	0.204	0.205	0.197	0.190	0.182	13.73
43) C	2,4,6-Trichlor...	0.275	0.329	0.353	0.370	0.375	0.381	0.357	0.349	10.50
44)	2,4,5-Trichlor...	0.322	0.330	0.362	0.404	0.411	0.371	0.365	0.367	9.17
45) S	2-Fluorobiphenyl	1.738	1.580	1.520	1.409	1.347	1.221	1.114	1.418	15.11
46)	1,1'-Biphenyl	1.773	1.621	1.612	1.581	1.547	1.435	1.316	1.555	9.38
47)	2-Chloronaphth...	1.289	1.176	1.191	1.185	1.157	1.085	1.014	1.157	7.51
48)	2-Nitroaniline	0.119	0.156	0.228	0.295	0.316	0.315	0.316	0.249	33.25
49)	Acenaphthylene	2.149	2.025	2.025	1.997	1.965	1.828	1.675	1.952	7.93
50)	Dimethylphthalate	1.362	1.284	1.301	1.321	1.309	1.227	1.152	1.280	5.42
51)	2,6-Dinitrotol...	0.098	0.137	0.193	0.241	0.259	0.254	0.249	0.204	31.45
52) C	Acenaphthene	1.271	1.175	1.156	1.172	1.138	1.072	0.995	1.140	7.63
53)	3-Nitroaniline	0.143	0.187	0.252	0.298	0.311	0.310	0.299	0.257	26.09
54) P	2,4-Dinitrophenol		0.025	0.036	0.057	0.069	0.076	0.085	0.058	40.36
55)	Dibenzofuran	1.930	1.775	1.751	1.743	1.713	1.581	1.466	1.709	8.67
56) P	4-Nitrophenol		0.131	0.177	0.216	0.233	0.230	0.220	0.201	19.82
57)	2,4-Dinitrotol...	0.104	0.145	0.213	0.285	0.311	0.309	0.306	0.239	35.96
58)	Fluorene	1.493	1.403	1.351	1.317	1.273	1.186	1.071	1.299	10.74
59)	2,3,4,6-Tetrac...	0.252	0.267	0.312	0.334	0.334	0.323	0.305	0.304	10.65
60)	Diethylphthalate	1.324	1.264	1.302	1.313	1.300	1.215	1.126	1.263	5.61
61)	4-Chlorophenyl...	0.702	0.642	0.632	0.632	0.614	0.570	0.526	0.617	9.06
62)	4-Nitroaniline	0.144	0.177	0.233	0.272	0.287	0.281	0.269	0.238	23.57
63)	Azobenzene	1.346	1.260	1.275	1.279	1.278	1.198	1.098	1.248	6.33
64) I	Phenanthrene-d10	-----ISTD-----								
65)	4,6-Dinitro-2-...		0.022	0.034	0.055	0.065	0.071	0.075	0.054	39.23
66) c	n-Nitrosodiphe...	0.727	0.686	0.693	0.712	0.689	0.653	0.623	0.683	5.11
67)	4-Bromophenyl-...	0.224	0.219	0.223	0.236	0.225	0.220	0.211	0.223	3.39
68)	Hexachlorobenzene	0.256	0.245	0.248	0.261	0.255	0.244	0.237	0.249	3.28
69)	Atrazine	0.166	0.172	0.186	0.190	0.190	0.182	0.174	0.180	5.29
70) C	Pentachlorophenol		0.093	0.121	0.144	0.145	0.144	0.144	0.132	16.25
71)	Phenanthrene	1.218	1.114	1.119	1.112	1.064	0.995	0.940	1.080	8.45
72)	Anthracene	1.209	1.161	1.151	1.128	1.099	1.025	0.959	1.105	7.77
73)	Carbazole	1.125	1.047	1.061	1.033	0.996	0.935	0.857	1.008	8.77
74)	Di-n-butylphth...	0.986	1.006	1.088	1.069	1.045	0.988	0.912	1.013	5.88
75) C	Fluoranthene	1.221	1.144	1.155	1.085	1.035	0.957	0.878	1.068	11.24
76) I	Chrysene-d12	-----ISTD-----								
77)	Benzidine	0.702	0.748	0.885	0.963	0.945	0.895	0.810	0.850	11.67
78)	Pyrene	1.833	1.801	1.849	2.026	1.937	1.833	1.635	1.845	6.54
79) S	Terphenyl-d14	1.472	1.413	1.422	1.481	1.392	1.290	1.155	1.375	8.41
80)	Butylbenzylphth...	0.218	0.293	0.405	0.505	0.525	0.524	0.509	0.425	29.32
81)	Benzo(a)anthra...	1.379	1.320	1.332	1.438	1.388	1.316	1.225	1.343	5.06
82)	3,3'-Dichlorob...	0.286	0.317	0.380	0.442	0.450	0.451	0.435	0.394	17.37
83)	Chrysene	1.302	1.210	1.249	1.232	1.227	1.197	1.138	1.222	4.12
84)	Bis(2-ethylhex...	0.384	0.448	0.549	0.695	0.726	0.722	0.707	0.604	23.69
85) c	Di-n-octyl pht...		0.668	0.875	1.262	1.320	1.347	1.329	1.133	25.55

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF043025.M

86) I	Perylene-d12	-----ISTD-----									
87)	Indeno(1,2,3-c...	1.323	1.384	1.466	1.613	1.556	1.490	1.422	1.465		6.78
88)	Benzo(b)fluora...	1.317	1.220	1.208	1.302	1.271	1.177	1.202	1.242		4.36
89)	Benzo(k)fluora...	1.211	1.159	1.222	1.145	1.146	1.110	0.962	1.136		7.59
90) C	Benzo(a)pyrene	1.121	1.098	1.146	1.203	1.180	1.137	1.076	1.137		3.91
91)	Dibenzo(a,h)an...	1.075	1.131	1.197	1.312	1.271	1.192	1.144	1.189		6.90
92)	Benzo(g,h,i)pe...	1.076	1.138	1.204	1.308	1.267	1.213	1.162	1.195		6.55

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG No.: Q1915

Instrument ID: BNA_F Calibration Date/Time: 05/01/2025 10:17

Lab File ID: BF142250.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.532		5.4	
2-Fluorophenol	1.252	1.250		-0.2	
Phenol-d6	1.512	1.575		4.2	
1,4-Dichlorobenzene	1.475	1.526		3.5	20.0
2-Methylphenol	1.060	1.103		4.1	
3+4-Methylphenols	1.326	1.404		5.9	
Nitrobenzene-d5	0.304	0.333		9.5	
Hexachloroethane	0.483	0.505		4.6	
Nitrobenzene	0.299	0.320		7.0	
Hexachlorobutadiene	0.175	0.186		6.3	20.0
2,4,6-Trichlorophenol	0.349	0.354		1.4	20.0
2-Fluorobiphenyl	1.418	1.409		-0.6	
2,4,5-Trichlorophenol	0.367	0.386		5.2	
2,4-Dinitrotoluene	0.239	0.289		20.9	
2,4,6-Tribromophenol	0.182	0.196		7.7	
Hexachlorobenzene	0.249	0.257		3.2	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.422		3.4	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG No.: Q1915

Instrument ID: BNA_F Calibration Date/Time: 05/02/2025 10:39

Lab File ID: BF142274.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.384		-4.8	
2-Fluorophenol	1.252	1.195		-4.6	
Phenol-d6	1.512	1.457		-3.6	
1,4-Dichlorobenzene	1.475	1.364		-7.5	20.0
2-Methylphenol	1.060	1.005		-5.2	
3+4-Methylphenols	1.326	1.295		-2.3	
Nitrobenzene-d5	0.304	0.332		9.2	
Hexachloroethane	0.483	0.473		-2.1	
Nitrobenzene	0.299	0.315		5.4	
Hexachlorobutadiene	0.175	0.164		-6.3	20.0
2,4,6-Trichlorophenol	0.349	0.360		3.2	20.0
2-Fluorobiphenyl	1.418	1.293		-8.8	
2,4,5-Trichlorophenol	0.367	0.363		-1.1	
2,4-Dinitrotoluene	0.239	0.314		31.4	
2,4,6-Tribromophenol	0.182	0.190		4.4	
Hexachlorobenzene	0.249	0.232		-6.8	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.344		-2.3	

All other compounds must meet a minimum RRF of 0.010.