

Cover Page

Order ID : Q1915

Project ID : Con Ed UTEN Mount Vernon, NY

Client : CDM Smith

Lab Sample Number

Q1915-01

Client Sample Number

WC-04282025

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/6/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1915

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: PRATHA PARGHI

Date: 05/06/2025



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q1915	OrderDate:	4/29/2025 2:29:00 PM
Client:	CDM Smith	Project:	Con Ed UTEN Mount Vernon, NY
Contact:	Marcie Ann Encinas	Location:	L41

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1915-01	WC-04282025	TCLP	TCLP BNA	8270E	04/28/25	04/30/25	05/01/25	04/29/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1915
Client: CDM Smith

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :								
				0.000				
			Total Svoc :			0.00		
			Total Concentration:			0.00		



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167774TB	PB167774TB	2-Fluorophenol	150	116	78		10	139
		Phenol-d6	150	116	77		10	134
		Nitrobenzene-d5	100	87.3	87		49	133
		2-Fluorobiphenyl	100	72.1	72		52	132
		2,4,6-Tribromophenol	150	131	87		44	137
		Terphenyl-d14	100	67.9	68		48	125
PB167810BL	PB167810BL	2-Fluorophenol	150	110	73		10	139
		Phenol-d6	150	109	73		10	134
		Nitrobenzene-d5	100	83.3	83		49	133
		2-Fluorobiphenyl	100	69.1	69		52	132
		2,4,6-Tribromophenol	150	124	83		44	137
		Terphenyl-d14	100	63.3	63		48	125
PB167810BS	PB167810BS	2-Fluorophenol	150	111	74		10	139
		Phenol-d6	150	113	75		10	134
		Nitrobenzene-d5	100	82.4	82		49	133
		2-Fluorobiphenyl	100	68.5	69		52	132
		2,4,6-Tribromophenol	150	132	88		44	137
		Terphenyl-d14	100	74.1	74		48	125
Q1901-08MS	B-167-SB01MS	2-Fluorophenol	150	96.6	64		10	139
		Phenol-d6	150	89.8	60		10	134
		Nitrobenzene-d5	100	84.0	84		49	133
		2-Fluorobiphenyl	100	70.7	71		52	132
		2,4,6-Tribromophenol	150	126	84		44	137
		Terphenyl-d14	100	70.4	70		48	125
Q1901-08MSD	B-167-SB01MSD	2-Fluorophenol	150	99.4	66		10	139
		Phenol-d6	150	92.8	62		10	134
		Nitrobenzene-d5	100	87.8	88		49	133
		2-Fluorobiphenyl	100	72.3	72		52	132
		2,4,6-Tribromophenol	150	133	88		44	137
		Terphenyl-d14	100	72.8	73		48	125
Q1915-01	WC-04282025	2-Fluorophenol	150	101	67		10	139
		Phenol-d6	150	91.6	61		10	134
		Nitrobenzene-d5	100	86.8	87		49	133
		2-Fluorobiphenyl	100	75.3	75		52	132
		2,4,6-Tribromophenol	150	116	78		44	137
		Terphenyl-d14	100	57.8	58		48	125

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1901-08MS	Client Sample ID:	B-167-SB01MS					DataFile:	BF142277.D		
Pyridine	500	24.1	270	ug/L	49				10	109	
1,4-Dichlorobenzene	500	0	340	ug/L	68				55	125	
2-Methylphenol	500	0	360	ug/L	72				60	131	
3+4-Methylphenols	500	0	360	ug/L	72				54	136	
Hexachloroethane	500	0	350	ug/L	70				19	146	
Nitrobenzene	500	0	440	ug/L	88				62	112	
Hexachlorobutadiene	500	0	380	ug/L	76				52	125	
2,4,6-Trichlorophenol	500	0	470	ug/L	94				78	112	
2,4,5-Trichlorophenol	500	0	460	ug/L	92				71	111	
2,4-Dinitrotoluene	500	0	500	ug/L	100				74	137	
Hexachlorobenzene	500	0	440	ug/L	88				72	115	
Pentachlorophenol	1000	0	940	ug/L	94				52	162	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1901-08MSD	Client Sample ID:	B-167-SB01MSD					DataFile:	BF142278.D		
Pyridine	500	24.1	280	ug/L	51		4		10	109	20
1,4-Dichlorobenzene	500	0	340	ug/L	68		0		55	125	20
2-Methylphenol	500	0	370	ug/L	74		3		60	131	20
3+4-Methylphenols	500	0	370	ug/L	74		3		54	136	20
Hexachloroethane	500	0	370	ug/L	74		6		19	146	20
Nitrobenzene	500	0	460	ug/L	92		4		62	112	20
Hexachlorobutadiene	500	0	390	ug/L	78		3		52	125	20
2,4,6-Trichlorophenol	500	0	480	ug/L	96		2		78	112	20
2,4,5-Trichlorophenol	500	0	470	ug/L	94		2		71	111	20
2,4-Dinitrotoluene	500	0	530	ug/L	106		6		74	137	20
Hexachlorobenzene	500	0	450	ug/L	90		2		72	115	20
Pentachlorophenol	1000	0	970	ug/L	97		3		52	162	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1915

Client: CDM Smith

Analytical Method: 8270E DataFile: BF142255.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167810BS	Pyridine	50	37.2	ug/L	74				29	97	
	1,4-Dichlorobenzene	50	39.6	ug/L	79				76	103	
	2-Methylphenol	50	41.0	ug/L	82				69	109	
	3+4-Methylphenols	50	40.8	ug/L	82				67	106	
	Hexachloroethane	50	41.8	ug/L	84				76	118	
	Nitrobenzene	50	45.8	ug/L	92				58	106	
	Hexachlorobutadiene	50	41.5	ug/L	83				69	101	
	2,4,6-Trichlorophenol	50	43.4	ug/L	87				61	110	
	2,4,5-Trichlorophenol	50	44.9	ug/L	90				70	106	
	2,4-Dinitrotoluene	50	49.8	ug/L	100				60	115	
	Hexachlorobenzene	50	41.6	ug/L	83				73	106	
	Pentachlorophenol	100	93.4	ug/L	93				47	114	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167810BL

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q1915

SAS No.: Q1915 SDG NO.: Q1915

Lab File ID: BF142254.D

Lab Sample ID: PB167810BL

Instrument ID: BNA_F

Date Extracted: 04/30/2025

Matrix: (soil/water) water

Date Analyzed: 05/01/2025

Level: (low/med) LOW

Time Analyzed: 12:10

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167810BS	PB167810BS	BF142255.D	05/01/2025
WC-04282025	Q1915-01	BF142261.D	05/01/2025
B-167-SB01MS	Q1901-08MS	BF142277.D	05/02/2025
B-167-SB01MSD	Q1901-08MSD	BF142278.D	05/02/2025
PB167774TB	PB167774TB	BF142256.D	05/01/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: CAMP02Lab Code: CHEMSAS No.: Q1915SDG NO.: Q1915Lab File ID: BF142238.DDFTPP Injection Date: 04/30/2025Instrument ID: BNA_FDFTPP Injection Time: 10:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	32.6
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	27.3
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	37.1
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.3
275	10.0 - 60.0% of mass 198	22.8
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.9 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142239.D	04/30/2025	11:24
SSTDICC005	SSTDICC005	BF142240.D	04/30/2025	11:52
SSTDICC010	SSTDICC010	BF142241.D	04/30/2025	12:20
SSTDICC020	SSTDICC020	BF142242.D	04/30/2025	12:49
SSTDICCC040	SSTDICCC040	BF142243.D	04/30/2025	13:17
SSTDICC050	SSTDICC050	BF142244.D	04/30/2025	13:46
SSTDICC060	SSTDICC060	BF142245.D	04/30/2025	14:15
SSTDICC080	SSTDICC080	BF142246.D	04/30/2025	14:43



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: CAMP02Lab Code: CHEMSAS No.: Q1915SDG NO.: Q1915Lab File ID: BF142249.DDFTPP Injection Date: 05/01/2025Instrument ID: BNA_FDFTPP Injection Time: 09:48

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.4
68	Less than 2.0% of mass 69	0.5 (1.6) 1
69	Mass 69 relative abundance	31.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	42.4
197	Less than 2.0% of mass 198	0.6
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.9
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142250.D	05/01/2025	10:17
PB167810BL	PB167810BL	BF142254.D	05/01/2025	12:10
PB167810BS	PB167810BS	BF142255.D	05/01/2025	12:39
PB167774TB	PB167774TB	BF142256.D	05/01/2025	13:07
WC-04282025	Q1915-01	BF142261.D	05/01/2025	15:34



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: CAMP02Lab Code: CHEMSAS No.: Q1915 SDG NO.: Q1915Lab File ID: BF142273.DDFTPP Injection Date: 05/02/2025Instrument ID: BNA_FDFTPP Injection Time: 10:11

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	34.7
68	Less than 2.0% of mass 69	0.4 (1.6) 1
69	Mass 69 relative abundance	27.6
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	37.6
197	Less than 2.0% of mass 198	0.7
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.6
275	10.0 - 60.0% of mass 198	23.7
365	Greater than 1% of mass 198	3
441	Present, but less than mass 443	15.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142274.D	05/02/2025	10:39
B-167-SB01MS	Q1901-08MS	BF142277.D	05/02/2025	12:27
B-167-SB01MSD	Q1901-08MSD	BF142278.D	05/02/2025	12:55

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	186578	6.91	719586	8.19	381198	9.95
UPPER LIMIT	373156	7.41	1439170	8.692	762396	10.445
LOWER LIMIT	93289	6.41	359793	7.692	190599	9.445
EPA SAMPLE NO.						
01 PB167774TB	217516	6.90	839699	8.19	443070	9.95
02 PB167810BL	223997	6.90	857825	8.19	449926	9.95
03 PB167810BS	228833	6.91	892782	8.19	472284	9.95
04 WC-04282025	210959	6.91	797035	8.19	392717	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/01/2025

Lab File ID: BF142250.D Time Analyzed: 10:17

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	647816	11.433	384442	14.074	349272	15.562
UPPER LIMIT	1295630	11.933	768884	14.574	698544	16.062
LOWER LIMIT	323908	10.933	192221	13.574	174636	15.062
EPA SAMPLE NO.						
01 PB167774TB	773002	11.43	539235	14.07	394051	15.56
02 PB167810BL	794711	11.43	568207	14.07	404816	15.56
03 PB167810BS	804874	11.43	464382	14.07	446826	15.57
04 WC-04282025	545118	11.43	445392	14.06	444396	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/02/2025

Lab File ID: BF142274.D Time Analyzed: 10:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	189457	6.91	747235	8.19	389025	9.95
UPPER LIMIT	378914	7.41	1494470	8.692	778050	10.445
LOWER LIMIT	94728.5	6.41	373618	7.692	194513	9.445
EPA SAMPLE NO.						
01 B-167-SB01MS	212656	6.91	808361	8.19	402059	9.95
02 B-167-SB01MSD	217028	6.91	832936	8.19	419407	9.95

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/02/2025

Lab File ID: BF142274.D Time Analyzed: 10:39

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	652685	11.427	354752	14.068	327562	15.557
UPPER LIMIT	1305370	11.927	709504	14.568	655124	16.057
LOWER LIMIT	326343	10.927	177376	13.568	163781	15.057
EPA SAMPLE NO.						
01 B-167-SB01MS	629997	11.43	370460	14.07	379220	15.56
02 B-167-SB01MSD	684254	11.43	400079	14.07	401381	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	04/28/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/29/25
Client Sample ID:	WC-04282025	SDG No.:	Q1915
Lab Sample ID:	Q1915-01	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142261.D	1	04/30/25 13:15	05/01/25 15:34	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	50.0	U	12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	50.0	U	5.30	50.0	ug/L
95-48-7	2-Methylphenol	50.0	U	11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	100	U	11.0	100	ug/L
67-72-1	Hexachloroethane	50.0	U	6.50	50.0	ug/L
98-95-3	Nitrobenzene	50.0	U	7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	50.0	U	5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	50.0	U	5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	50.0	U	6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	50.0	U	12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	50.0	U	5.20	50.0	ug/L
87-86-5	Pentachlorophenol	100	U	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	101		10 - 139	67%	SPK: 150
13127-88-3	Phenol-d6	91.6		10 - 134	61%	SPK: 150
4165-60-0	Nitrobenzene-d5	86.8		49 - 133	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	75.3		52 - 132	75%	SPK: 100
118-79-6	2,4,6-Tribromophenol	116		44 - 137	78%	SPK: 150
1718-51-0	Terphenyl-d14	57.8		48 - 125	58%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	211000	6.91			
1146-65-2	Naphthalene-d8	797000	8.186			
15067-26-2	Acenaphthene-d10	393000	9.939			
1517-22-2	Phenanthrene-d10	545000	11.427			
1719-03-5	Chrysene-d12	445000	14.062			
1520-96-3	Perylene-d12	444000	15.556			

Report of Analysis

Client:	CDM Smith	Date Collected:	04/28/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/29/25
Client Sample ID:	WC-04282025	SDG No.:	Q1915
Lab Sample ID:	Q1915-01	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142261.D	1	04/30/25 13:15	05/01/25 15:34	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142261.D
Acq On : 01 May 2025 15:34
Operator : RC/JU
Sample : Q1915-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-04282025

Quant Time: May 01 16:06:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	210959	20.000	ng	0.00
21) Naphthalene-d8	8.186	136	797035	20.000	ng	0.00
39) Acenaphthene-d10	9.939	164	392717	20.000	ng	-0.01
64) Phenanthrene-d10	11.427	188	545118	20.000	ng	0.00
76) Chrysene-d12	14.062	240	445392	20.000	ng	-0.01
86) Perylene-d12	15.556	264	444396	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	1332052	100.843	ng	0.01
7) Phenol-d6	6.522	99	1461054	91.596	ng	0.00
23) Nitrobenzene-d5	7.469	82	1053543	86.841	ng	0.00
42) 2,4,6-Tribromophenol	10.727	330	416794	116.418	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2097570	75.309	ng	0.00
79) Terphenyl-d14	13.015	244	1769965	57.797	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142261.D
Acq On : 01 May 2025 15:34
Operator : RC/JU
Sample : Q1915-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-04282025

Quant Time: May 01 16:06:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.910 min Scan# 80
 Delta R.T. -0.000 min
 Lab File: BF142261.D
 Acq: 01 May 2025 15:34




Instrument :
 BNA_F
 ClientSampleId :
 WC-04282025


Tgt Ion:152 Resp: 210959
 Ion Ratio Lower Upper
 152 100
 150 156.1 130.2 195.2
 115 57.9 47.0 70.4



#5
 2-Fluorophenol
 Concen: 100.843 ng
 RT: 5.528 min Scan# 567
 Delta R.T. 0.011 min
 Lab File: BF142261.D
 Acq: 01 May 2025 15:34

Tgt Ion:112 Resp: 1332052
 Ion Ratio Lower Upper
 112 100
 64 60.6 48.4 72.6
 63 31.5 24.8 37.2









#7
Phenol-d6
Concen: 91.596 ng
RT: 6.522 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF142261.D
Acq: 01 May 2025 15:34




Instrument :
BNA_F
ClientSampleId :
WC-04282025


Tgt Ion: 99 Resp: 1461054
Ion Ratio Lower Upper
99 100
42 21.0 16.3 24.5
71 32.4 25.0 37.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.186 min Scan# 1019
Delta R.T. -0.007 min
Lab File: BF142261.D
Acq: 01 May 2025 15:34

Tgt Ion: 136 Resp: 797035
Ion Ratio Lower Upper
136 100
137 11.1 8.7 13.1
54 9.7 7.7 11.5
68 6.2 5.0 7.6





#23

Nitrobenzene-d5

Concen: 86.841 ng

RT: 7.469 min Scan# 89

Delta R.T. -0.006 min

Lab File: BF142261.D

Acq: 01 May 2025 15:34

Instrument :





BNA_F

ClientSampleId :

WC-04282025

Tgt Ion: 82 Resp: 1053543

Ion	Ratio	Lower	Upper
82	100		
128	47.3	37.0	55.6
54	59.9	48.0	72.0



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.939 min Scan# 1317




Delta R.T. -0.012 min


Lab File: BF142261.D

Acq: 01 May 2025 15:34

Tgt Ion:164 Resp: 392717

Ion	Ratio	Lower	Upper
164	100		
162	98.9	78.6	117.8
160	44.7	35.7	53.5





#42

2,4,6-Tribromophenol

Concen: 116.418 ng

RT: 10.727 min Scan# 1451

Delta R.T. -0.006 min

Lab File: BF142261.D


Acq: 01 May 2025 15:34

Instrument :

BNA_F

ClientSampleId :

WC-04282025






Tgt Ion:330 Resp: 416794

Ion Ratio Lower Upper

330 100

332 95.2 76.5 114.7

141 29.5 24.3 36.5



#45

2-Fluorobiphenyl

Concen: 75.309 ng

RT: 9.263 min Scan# 1202

Delta R.T. -0.006 min

Lab File: BF142261.D

Acq: 01 May 2025 15:34




Tgt Ion:172 Resp: 2097570


Ion Ratio Lower Upper

172 100

171 35.7 29.0 43.4

170 23.7 19.4 29.0









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.427 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142261.D
Acq: 01 May 2025 15:34




Instrument :
BNA_F
ClientSampleId :
WC-04282025


Tgt Ion:188 Resp: 545118
Ion Ratio Lower Upper
188 100
94 10.1 8.2 12.2
80 10.3 8.6 12.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.062 min Scan# 2018
Delta R.T. -0.012 min
Lab File: BF142261.D
Acq: 01 May 2025 15:34

Tgt Ion:240 Resp: 445392
Ion Ratio Lower Upper
240 100
120 12.8 10.1 15.1
236 25.5 20.5 30.7





#79

Terphenyl-d14

Concen: 57.797 ng

RT: 13.015 min Scan# 1840

Delta R.T. -0.006 min

Lab File: BF142261.D


Acq: 01 May 2025 15:34

Instrument :

BNA_F

ClientSampleId :

WC-04282025






Tgt Ion:244 Resp: 1769965

Ion Ratio Lower Upper

244 100

212 6.6 5.3 7.9

122 11.4 9.1 13.7



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.556 min Scan# 2272

Delta R.T. -0.012 min

Lab File: BF142261.D

Acq: 01 May 2025 15:34




Tgt Ion:264 Resp: 444396

Ion Ratio Lower Upper

264 100

260 23.1 18.4 27.6

265 21.7 17.3 25.9





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG No.: Q1915
 Instrument ID: BNA_F Calibration Date(s): 04/30/2025 04/30/2025
 Calibration Time(s): 11:24 14:43

LAB FILE ID:		RRF2.5 = BF142239.D		RRF005 = BF142240.D		RRF010 = BF142241.D		
		RRF020 = BF142242.D		RRF040 = BF142243.D		RRF050 = BF142244.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Pyridine		1.465	1.391	1.454	1.522	1.503	1.454	3.6
2-Fluorophenol		1.324	1.238	1.265	1.314	1.285	1.252	5.3
Phenol-d6		1.559	1.505	1.534	1.593	1.550	1.512	4.9
1,4-Dichlorobenzene		1.610	1.486	1.495	1.546	1.488	1.475	6.9
2-Methylphenol		1.085	1.029	1.068	1.121	1.084	1.060	4.2
3+4-Methylphenols		1.357	1.323	1.372	1.400	1.366	1.326	5.9
Nitrobenzene-d5		0.217	0.251	0.302	0.340	0.351	0.304	16.9
Hexachloroethane		0.468	0.466	0.481	0.511	0.514	0.483	4.8
Nitrobenzene		0.229	0.261	0.297	0.328	0.338	0.299	13.3
Hexachlorobutadiene		0.180	0.175	0.179	0.185	0.178	0.175	4.3
2,4,6-Trichlorophenol		0.275	0.329	0.353	0.370	0.375	0.349	10.5
2-Fluorobiphenyl		1.738	1.580	1.520	1.409	1.347	1.418	15.1
2,4,5-Trichlorophenol		0.322	0.330	0.362	0.404	0.411	0.367	9.2
2,4-Dinitrotoluene		0.104	0.145	0.213	0.285	0.311	0.239	36.0
2,4,6-Tribromophenol		0.139	0.158	0.184	0.204	0.205	0.182	13.7
Hexachlorobenzene		0.256	0.245	0.248	0.261	0.255	0.249	3.3
Pentachlorophenol			0.093	0.121	0.144	0.145	0.132	16.2
Terphenyl-d14		1.472	1.413	1.422	1.481	1.392	1.375	8.4

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF043025.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 30 16:00:01 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF142239.D 5 =BF142240.D 10 =BF142241.D 20 =BF142242.D 40 =BF142243.D 50 =BF142244.D 60 =BF142245.D 80 =BF142246.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.563	0.534	0.561	0.592	0.580	0.555	0.530	0.559		4.03
3)	Pyridine	1.465	1.391	1.454	1.522	1.503	1.460	1.380	1.454		3.62
4)	n-Nitrosodimet...	0.747	0.699	0.728	0.807	0.789	0.766	0.731	0.752		4.95
5) S	2-Fluorophenol	1.324	1.238	1.265	1.314	1.285	1.210	1.131	1.252		5.35
6)	Aniline	2.212	2.085	2.157	2.245	2.203	2.101	1.981	2.140		4.29
7) S	Phenol-d6	1.559	1.505	1.534	1.593	1.550	1.475	1.370	1.512		4.85
8)	2-Chlorophenol	1.252	1.232	1.304	1.379	1.354	1.313	1.225	1.294		4.64
9)	Benzaldehyde	1.154	1.088	1.095	1.117	1.078	0.993	0.875	1.057		8.91
10) C	Phenol	1.679	1.588	1.657	1.715	1.687	1.603	1.468	1.628		5.15
11)	bis(2-Chloroet...	1.345	1.250	1.270	1.318	1.285	1.231	1.132	1.262		5.46
12)	1,3-Dichlorobe...	1.571	1.485	1.485	1.533	1.478	1.396	1.299	1.464		6.18
13) C	1,4-Dichlorobe...	1.610	1.486	1.495	1.546	1.488	1.404	1.295	1.475		6.86
14)	1,2-Dichlorobe...	1.489	1.418	1.431	1.475	1.432	1.356	1.255	1.408		5.67
15)	Benzyl Alcohol	1.047	1.017	1.064	1.149	1.126	1.082	1.023	1.073		4.69
16)	2,2'-oxybis(1-...	2.368	2.287	2.303	2.386	2.344	2.216	2.043	2.278		5.19
17)	2-Methylphenol	1.085	1.029	1.068	1.121	1.084	1.050	0.983	1.060		4.19
18)	Hexachloroethane	0.468	0.466	0.481	0.511	0.514	0.489	0.451	0.483		4.84
19) P	n-Nitroso-di-n...	0.953	0.987	0.937	0.949	0.967	0.939	0.898	0.844	0.934	4.77
20)	3+4-Methylphenols	1.357	1.323	1.372	1.400	1.366	1.296	1.166	1.326		5.90
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.511	0.482	0.483	0.468	0.461	0.430	0.404	0.463		7.70
23) S	Nitrobenzene-d5	0.217	0.251	0.302	0.340	0.351	0.339	0.330	0.304		16.88
24)	Nitrobenzene	0.229	0.261	0.297	0.328	0.338	0.323	0.314	0.299		13.31
25)	Isophorone	0.660	0.630	0.642	0.660	0.654	0.626	0.601	0.639		3.36
26) C	2-Nitrophenol	0.048	0.057	0.083	0.115	0.132	0.136	0.142	0.102		38.38#
27)	2,4-Dimethylph...	0.315	0.319	0.333	0.337	0.333	0.315	0.302	0.322		3.96
28)	bis(2-Chloroet...	0.425	0.404	0.414	0.415	0.410	0.386	0.366	0.403		4.98
29) C	2,4-Dichloroph...	0.246	0.253	0.274	0.291	0.286	0.272	0.263	0.269		6.12
30)	1,2,4-Trichlor...	0.320	0.303	0.305	0.312	0.304	0.287	0.273	0.301		5.28
31)	Naphthalene	1.106	1.040	1.035	1.020	0.999	0.924	0.858	0.997		8.21
32)	Benzoic acid		0.048	0.087	0.130	0.146	0.159	0.179	0.125		39.24
33)	4-Chloroaniline	0.446	0.414	0.429	0.429	0.423	0.396	0.371	0.415		5.98
34) C	Hexachlorobuta...	0.180	0.175	0.179	0.185	0.178	0.170	0.161	0.175		4.35
35)	Caprolactam	0.068	0.073	0.081	0.087	0.089	0.085	0.083	0.081		9.45
36) C	4-Chloro-3-met...	0.262	0.265	0.279	0.295	0.293	0.276	0.266	0.277		4.84
37)	2-Methylnaphth...	0.679	0.637	0.633	0.623	0.612	0.574	0.534	0.613		7.69
38)	1-Methylnaphth...	0.700	0.655	0.667	0.652	0.641	0.590	0.542	0.635		8.29

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF043025.M

39) I	Acenaphthene-d10	-----ISTD-----								
40)	1,2,4,5-Tetrac...	0.621	0.575	0.572	0.576	0.566	0.534	0.496	0.563	6.88
41) P	Hexachlorocycl...	0.244	0.265	0.306	0.352	0.368	0.358	0.343	0.319	15.23
42) S	2,4,6-Tribromo...	0.139	0.158	0.184	0.204	0.205	0.197	0.190	0.182	13.73
43) C	2,4,6-Trichlor...	0.275	0.329	0.353	0.370	0.375	0.381	0.357	0.349	10.50
44)	2,4,5-Trichlor...	0.322	0.330	0.362	0.404	0.411	0.371	0.365	0.367	9.17
45) S	2-Fluorobiphenyl	1.738	1.580	1.520	1.409	1.347	1.221	1.114	1.418	15.11
46)	1,1'-Biphenyl	1.773	1.621	1.612	1.581	1.547	1.435	1.316	1.555	9.38
47)	2-Chloronaphth...	1.289	1.176	1.191	1.185	1.157	1.085	1.014	1.157	7.51
48)	2-Nitroaniline	0.119	0.156	0.228	0.295	0.316	0.315	0.316	0.249	33.25
49)	Acenaphthylene	2.149	2.025	2.025	1.997	1.965	1.828	1.675	1.952	7.93
50)	Dimethylphthalate	1.362	1.284	1.301	1.321	1.309	1.227	1.152	1.280	5.42
51)	2,6-Dinitrotol...	0.098	0.137	0.193	0.241	0.259	0.254	0.249	0.204	31.45
52) C	Acenaphthene	1.271	1.175	1.156	1.172	1.138	1.072	0.995	1.140	7.63
53)	3-Nitroaniline	0.143	0.187	0.252	0.298	0.311	0.310	0.299	0.257	26.09
54) P	2,4-Dinitrophenol		0.025	0.036	0.057	0.069	0.076	0.085	0.058	40.36
55)	Dibenzofuran	1.930	1.775	1.751	1.743	1.713	1.581	1.466	1.709	8.67
56) P	4-Nitrophenol		0.131	0.177	0.216	0.233	0.230	0.220	0.201	19.82
57)	2,4-Dinitrotol...	0.104	0.145	0.213	0.285	0.311	0.309	0.306	0.239	35.96
58)	Fluorene	1.493	1.403	1.351	1.317	1.273	1.186	1.071	1.299	10.74
59)	2,3,4,6-Tetrac...	0.252	0.267	0.312	0.334	0.334	0.323	0.305	0.304	10.65
60)	Diethylphthalate	1.324	1.264	1.302	1.313	1.300	1.215	1.126	1.263	5.61
61)	4-Chlorophenyl...	0.702	0.642	0.632	0.632	0.614	0.570	0.526	0.617	9.06
62)	4-Nitroaniline	0.144	0.177	0.233	0.272	0.287	0.281	0.269	0.238	23.57
63)	Azobenzene	1.346	1.260	1.275	1.279	1.278	1.198	1.098	1.248	6.33
64) I	Phenanthrene-d10	-----ISTD-----								
65)	4,6-Dinitro-2-...		0.022	0.034	0.055	0.065	0.071	0.075	0.054	39.23
66) c	n-Nitrosodiphe...	0.727	0.686	0.693	0.712	0.689	0.653	0.623	0.683	5.11
67)	4-Bromophenyl-...	0.224	0.219	0.223	0.236	0.225	0.220	0.211	0.223	3.39
68)	Hexachlorobenzene	0.256	0.245	0.248	0.261	0.255	0.244	0.237	0.249	3.28
69)	Atrazine	0.166	0.172	0.186	0.190	0.190	0.182	0.174	0.180	5.29
70) C	Pentachlorophenol		0.093	0.121	0.144	0.145	0.144	0.144	0.132	16.25
71)	Phenanthrene	1.218	1.114	1.119	1.112	1.064	0.995	0.940	1.080	8.45
72)	Anthracene	1.209	1.161	1.151	1.128	1.099	1.025	0.959	1.105	7.77
73)	Carbazole	1.125	1.047	1.061	1.033	0.996	0.935	0.857	1.008	8.77
74)	Di-n-butylphth...	0.986	1.006	1.088	1.069	1.045	0.988	0.912	1.013	5.88
75) C	Fluoranthene	1.221	1.144	1.155	1.085	1.035	0.957	0.878	1.068	11.24
76) I	Chrysene-d12	-----ISTD-----								
77)	Benzidine	0.702	0.748	0.885	0.963	0.945	0.895	0.810	0.850	11.67
78)	Pyrene	1.833	1.801	1.849	2.026	1.937	1.833	1.635	1.845	6.54
79) S	Terphenyl-d14	1.472	1.413	1.422	1.481	1.392	1.290	1.155	1.375	8.41
80)	Butylbenzylphth...	0.218	0.293	0.405	0.505	0.525	0.524	0.509	0.425	29.32
81)	Benzo(a)anthra...	1.379	1.320	1.332	1.438	1.388	1.316	1.225	1.343	5.06
82)	3,3'-Dichlorob...	0.286	0.317	0.380	0.442	0.450	0.451	0.435	0.394	17.37
83)	Chrysene	1.302	1.210	1.249	1.232	1.227	1.197	1.138	1.222	4.12
84)	Bis(2-ethylhex...	0.384	0.448	0.549	0.695	0.726	0.722	0.707	0.604	23.69
85) c	Di-n-octyl pht...		0.668	0.875	1.262	1.320	1.347	1.329	1.133	25.55


Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF043025.M

86) I	Perylene-d12	-----ISTD-----	
87)	Indeno(1,2,3-c...	1.323 1.384 1.466 1.613 1.556 1.490 1.422 1.465	6.78
88)	Benzo(b)fluora...	1.317 1.220 1.208 1.302 1.271 1.177 1.202 1.242	4.36
89)	Benzo(k)fluora...	1.211 1.159 1.222 1.145 1.146 1.110 0.962 1.136	7.59
90) C	Benzo(a)pyrene	1.121 1.098 1.146 1.203 1.180 1.137 1.076 1.137	3.91
91)	Dibenzo(a,h)an...	1.075 1.131 1.197 1.312 1.271 1.192 1.144 1.189	6.90
92)	Benzo(g,h,i)pe...	1.076 1.138 1.204 1.308 1.267 1.213 1.162 1.195	6.55

(#) = Out of Range

Di-n-octyl phthalate

Response Ratio



$$\text{Response} = 1.466\text{e}+000 * \text{Amt} - 4.448\text{e}-001$$


Coef of Det (r^2) = 0.997236 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M

Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 7.529\text{e-}001 * \text{Amt} - 1.300\text{e-}001$$


Coef of Det (r^2) = 0.997585 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M

Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

Butylbenzylphthalate

Response Ratio




$$\text{Response} = 5.468\text{e-}001 * \text{Amt} - 1.016\text{e-}001$$

Coef of Det (r^2) = 0.997533 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M

Calibration Table Last Updated: Wed Apr 30 16:00:01 2025


4,6-Dinitro-2-methylphenol



Response = $8.458 \times 10^{-2} \times \text{Amt} - 4.487 \times 10^{-2}$
Coef of Det (r^2) = 0.991555 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

4-Nitroaniline

Response Ratio



$$\text{Response} = 2.879\text{e-}001 * \text{Amt} - 3.958\text{e-}002$$


Coef of Det (r^2) = 0.996450 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M

Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

2,4-Dinitrotoluene

Response Ratio




$$\text{Response} = 3.318\text{e-}001 * \text{Amt} - 8.340\text{e-}002$$

Coef of Det (r^2) = 0.997011 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M

Calibration Table Last Updated: Wed Apr 30 16:00:01 2025


2,4-Dinitrophenol



$R = 1.037e-002 A^2 + 4.909e-002 A - 1.884e-002$
 Coef of Det (r^2) = 0.998097 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
 Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

3-Nitroaniline

Response Ratio



$$\text{Response} = 3.198\text{e-}001 * \text{Amt} - 5.099\text{e-}002$$


Coef of Det (r^2) = 0.997645 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M

Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

2,6-Dinitrotoluene

Response Ratio



$$\text{Response} = 2.679\text{e-}001 * \text{Amt} - 5.390\text{e-}002$$


Coef of Det (r^2) = 0.997116 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M

Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

2-Nitroaniline

Response Ratio




$$\text{Response} = 3.387\text{e-}001 * \text{Amt} - 8.076\text{e-}002$$

Coef of Det (r^2) = 0.998123 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M


Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

Benzoic acid



Response = $1.978 \times 10^{-1} \times \text{Amt} - 1.069 \times 10^{-1}$
Coef of Det (r^2) = 0.989640 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

2-Nitrophenol



$R = 8.463e-003 A^2 + 1.177e-001 A - 2.955e-002$
 Coef of Det (r^2) = 0.997290 Curve Fit: Quadratic
 Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF043025.M
 Calibration Table Last Updated: Wed Apr 30 16:00:01 2025

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142239.D
Acq On : 30 Apr 2025 11:24
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 30 15:56:10 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	204993	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	780462	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	414360	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	737147	20.000	ng	0.00
76) Chrysene-d12	14.074	240	525717	20.000	ng	0.00
86) Perylene-d12	15.568	264	366865	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						
19) n-Nitroso-di-n-propyla...	7.304	70	24421	2.550	ng	Qvalue 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142239.D
Acq On : 30 Apr 2025 11:24
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Apr 30 15:56:10 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.000 ng

RT: 6.910 min Scan# 80

Delta R.T. 0.000 min

Lab File: BF142239.D


Acq: 30 Apr 2025 11:24

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5






Tgt Ion:152 Resp: 204993

Ion Ratio Lower Upper

152 100

150 154.4 130.2 195.2

115 57.4 47.0 70.4



#5

2-Fluorophenol

Concen: 0.000 ng

Expected RT: 5.52 min

Lab File: BF142239.D

Acq: 30 Apr 2025 11:24


Tgt Ion: 112


Sig Exp Ratio

112 100

64 60.5

63 31.0







#7
Phenol-d6
Concen: 0.000 ng
Expected RT: 6.53 min

Lab File: BF142239.D
Acq: 30 Apr 2025 11:24




Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5


Tgt Ion: 99
Sig Exp Ratio
99 100
42 20.4
71 31.2



#19
n-Nitroso-di-n-propylamine
Concen: 2.550 ng
RT: 7.304 min Scan# 869
Delta R.T. -0.018 min
Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Tgt Ion: 70 Resp: 24421
Ion Ratio Lower Upper
70 100
42 67.3 53.4 80.2
101 9.1 7.7 11.5
130 20.5 17.2 25.8





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.192 min Scan# 1

Delta R.T. -0.001 min

Lab File: BF142239.D


Acq: 30 Apr 2025 11:24

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5



Tgt Ion:136 Resp: 780462



Ion Ratio Lower Upper

136 100

137 10.7 8.7 13.1

54 9.2 7.7 11.5

68 6.2 5.0 7.6



#23

Nitrobenzene-d5

Concen: 0.000 ng

Expected RT: 7.47 min

Lab File: BF142239.D

Acq: 30 Apr 2025 11:24


Tgt Ion: 82


Sig Exp Ratio

82 100

128 46.3

54 60.0





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.945 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF142239.D


Acq: 30 Apr 2025 11:24

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5






Tgt Ion:164 Resp: 414360

Ion Ratio Lower Upper

164 100

162 98.5 78.6 117.8

160 44.5 35.7 53.5



#42

2,4,6-Tribromophenol

Concen: 0.000 ng

Expected RT: 10.73 min

Lab File: BF142239.D

Acq: 30 Apr 2025 11:24


Tgt Ion: 330


Sig Exp Ratio

330 100

332 95.6

141 30.4





#45

2-Fluorobiphenyl

Concen: 0.000 ng

Expected RT: 9.27 min

Lab File: BF142239.D

Acq: 30 Apr 2025 11:24

Instrument :

BNA_F

ClientSampleId :

SSTDICC2.5



Tgt Ion: 172

Sig Exp Ratio

172 100

171 36.2

170 24.2



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.433 min Scan# 1571

Delta R.T. -0.000 min

Lab File: BF142239.D

Acq: 30 Apr 2025 11:24




Tgt Ion:188 Resp: 737147


Ion Ratio Lower Upper

188 100

94 9.9 8.2 12.2


80 10.3 8.6 12.8








#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 20
Delta R.T. -0.000 min
Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5




Tgt Ion:240 Resp: 525717
Ion Ratio Lower Upper
240 100
120 11.8 10.1 15.1
236 24.9 20.5 30.7



#79
Terphenyl-d14
Concen: 0.000 ng
Expected RT: 13.02 min


Lab File: BF142239.D
Acq: 30 Apr 2025 11:24

Tgt Ion: 244
Sig Exp Ratio
244 100
212 6.6
122 11.4



Ion 212.00 (211.70 to 212.70): BF142239.D\data

Ion 122.00 (121.70 to 122.70): BF142239.D\data



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.568 min Scan# 21

Delta R.T. -0.000 min

Lab File: BF142239.D


Acq: 30 Apr 2025 11:24

Instrument :

BNA_F



ClientSampleId :

SSTDICC2.5



Tgt Ion:264 Resp: 366865

Ion	Ratio	Lower	Upper
264	100		
260	23.7	18.4	27.6
265	21.3	17.3	25.9



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142240.D
 Acq On : 30 Apr 2025 11:52
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:40:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	187794	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	726978	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	380022	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	664931	20.000	ng	0.00
76) Chrysene-d12	14.074	240	454911	20.000	ng	0.00
86) Perylene-d12	15.568	264	322126	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	124307	10.571	ng	0.00
7) Phenol-d6	6.510	99	146414	10.311	ng	-0.02
23) Nitrobenzene-d5	7.463	82	78765	7.118	ng	-0.01
42) 2,4,6-Tribromophenol	10.727	330	26375	7.613	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	330323	12.256	ng	0.00
79) Terphenyl-d14	13.016	244	334852	10.706	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.699	88	26443	5.035	ng	97
3) Pyridine	3.469	79	68779	5.039	ng	99
4) n-Nitrosodimethylamine	3.393	42	35057	4.963	ng	# 97
6) Aniline	6.563	93	103866	5.168	ng	99
8) 2-Chlorophenol	6.687	128	58780	4.838	ng	99
9) Benzaldehyde	6.457	77	54160	5.457	ng	98
10) Phenol	6.528	94	78823	5.157	ng	96
11) bis(2-Chloroethyl)ether	6.634	93	63148	5.331	ng	99
12) 1,3-Dichlorobenzene	6.845	146	73777	5.367	ng	100
13) 1,4-Dichlorobenzene	6.928	146	75604	5.459	ng	97
14) 1,2-Dichlorobenzene	7.075	146	69920	5.289	ng	99
15) Benzyl Alcohol	7.039	79	49136	4.879	ng	97
16) 2,2'-oxybis(1-Chloropr...	7.181	45	111179	5.198	ng	100
17) 2-Methylphenol	7.145	107	50945	5.118	ng	98
18) Hexachloroethane	7.422	117	21958	4.844	ng	96
19) n-Nitroso-di-n-propyla...	7.310	70	46341	5.281	ng	99
20) 3+4-Methylphenols	7.292	107	63725	5.119	ng	93
22) Acetophenone	7.310	105	92806	5.518	ng	96
24) Nitrobenzene	7.481	77	41660	3.837	ng	94
25) Isophorone	7.722	82	119954	5.165	ng	100
26) 2-Nitrophenol	7.804	139	8775	6.901	ng	99
27) 2,4-Dimethylphenol	7.828	122	57217	4.887	ng	97
28) bis(2-Chloroethoxy)met...	7.934	93	77155	5.268	ng	100
29) 2,4-Dichlorophenol	8.034	162	44643	4.560	ng	98
30) 1,2,4-Trichlorobenzene	8.128	180	58177	5.326	ng	99
31) Naphthalene	8.210	128	200969	5.543	ng	99
33) 4-Chloroaniline	8.251	127	81024	5.367	ng	98
34) Hexachlorobutadiene	8.334	225	32676	5.127	ng	97
35) Caprolactam	8.575	113	12427	4.223	ng	100
36) 4-Chloro-3-methylphenol	8.722	107	47673	4.743	ng	99
37) 2-Methylnaphthalene	8.904	142	123458	5.539	ng	97
38) 1-Methylnaphthalene	8.998	142	127274	5.512	ng	99
40) 1,2,4,5-Tetrachloroben...	9.063	216	58967	5.513	ng	97
41) Hexachlorocyclopentadiene	9.057	237	23228	3.827	ng	98
43) 2,4,6-Trichlorophenol	9.175	196	26171	3.949	ng	98
44) 2,4,5-Trichlorophenol	9.204	196	30587	4.390	ng	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142240.D
 Acq On : 30 Apr 2025 11:52
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:40:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 1,1'-Biphenyl	9.363	154	168456	5.702	ng	98
47) 2-Chloronaphthalene	9.386	162	122469	5.572	ng	98
48) 2-Nitroaniline	9.475	65	11346	6.532	ng	89
49) Acenaphthylene	9.804	152	204157	5.504	ng	99
50) Dimethylphthalate	9.657	163	129376	5.322	ng	100
51) 2,6-Dinitrotoluene	9.716	165	9271	5.846	ng	# 81
52) Acenaphthene	9.975	154	120775	5.577	ng	99
53) 3-Nitroaniline	9.886	138	13576	5.423	ng	# 88
55) Dibenzofuran	10.151	168	183405	5.649	ng	98
57) 2,4-Dinitrotoluene	10.116	165	9928	6.601	ng	# 85
58) Fluorene	10.492	166	141835	5.746	ng	98
59) 2,3,4,6-Tetrachlorophenol	10.263	232	23974m	4.154	ng	
60) Diethylphthalate	10.363	149	125748	5.238	ng	99
61) 4-Chlorophenyl-phenyle...	10.486	204	66676	5.688	ng	98
62) 4-Nitroaniline	10.492	138	13695	5.254	ng	# 76
63) Azobenzene	10.645	77	127873	5.394	ng	99
66) n-Nitrosodiphenylamine	10.598	169	120800	5.317	ng	100
67) 4-Bromophenyl-phenylether	10.975	248	37248	5.030	ng	97
68) Hexachlorobenzene	11.039	284	42558	5.131	ng	98
69) Atrazine	11.122	200	27629	4.614	ng	98
71) Phenanthrene	11.457	178	202460	5.637	ng	98
72) Anthracene	11.504	178	200946	5.472	ng	98
73) Carbazole	11.657	167	186963	5.581	ng	98
74) Di-n-butylphthalate	11.998	149	163860	4.864	ng	99
75) Fluoranthene	12.645	202	202911	5.715	ng	98
77) Benzidine	12.769	184	79877	4.132	ng	98
78) Pyrene	12.874	202	208462	4.968	ng	99
80) Butylbenzylphthalate	13.498	149	24840	5.714	ng	96
81) Benzo(a)anthracene	14.063	228	156840	5.136	ng	99
82) 3,3'-Dichlorobenzidine	14.027	252	32503	3.623	ng	94
83) Chrysene	14.098	228	148119	5.328	ng	98
84) Bis(2-ethylhexyl)phtha...	14.063	149	43692	6.006	ng	99
87) Indeno(1,2,3-cd)pyrene	17.068	276	106554	4.516	ng	98
88) Benzo(b)fluoranthene	15.121	252	106072	5.301	ng	100
89) Benzo(k)fluoranthene	15.151	252	97496	5.327	ng	99
90) Benzo(a)pyrene	15.498	252	90284	4.928	ng	99
91) Dibenzo(a,h)anthracene	17.086	278	86557	4.520	ng	100
92) Benzo(g,h,i)perylene	17.521	276	86631	4.500	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142240.D
Acq On : 30 Apr 2025 11:52
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025


Quant Time: Apr 30 15:40:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion:152 Resp: 187794
Ion Ratio Lower Upper
152 100
150 157.6 130.2 195.2
115 58.9 47.0 70.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#2
1,4-Dioxane
Concen: 5.035 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 88 Resp: 26443
Ion Ratio Lower Upper
88 100
58 91.8 72.2 108.2
43 37.4 27.0 40.6





#3

Pyridine

Concen: 5.039 ng

RT: 3.469 min Scan# 216

Delta R.T. 0.006 min

Lab File: BF142240.D


Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005



Tgt Ion: 79 Resp: 68779

Ion Ratio Lower Upper

79 100

52 88.7 71.5 107.3




51 42.1 34.5 51.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#4

n-Nitrosodimethylamine


Concen: 4.963 ng

RT: 3.393 min Scan# 204

Delta R.T. -0.018 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52





Tgt Ion: 42 Resp: 35057


Ion Ratio Lower Upper

42 100

74 110.4 89.5 134.3


44 18.5 6.6 9.8





#5
2-Fluorophenol
Concen: 10.571 ng
RT: 5.510 min Scan# 51
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion: 112 Resp: 12430
Ion Ratio Lower Upper
112 100
64 58.2 48.4 72.6
63 30.8 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#6
Aniline
Concen: 5.168 ng
RT: 6.563 min Scan# 743
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 93 Resp: 103866
Ion Ratio Lower Upper
93 100
66 39.6 32.1 48.1
65 20.7 16.6 24.8





#7

Phenol-d6

Concen: 10.311 ng

RT: 6.510 min Scan# 71

Delta R.T. -0.018 min

Lab File: BF142240.D


Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005



Tgt Ion: 99 Resp: 146414

Ion Ratio Lower Upper

99 100

42 21.1 16.3 24.5




71 30.6 25.0 37.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#8

2-Chlorophenol

Concen: 4.838 ng

RT: 6.687 min Scan# 764

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52




Tgt Ion:128 Resp: 58780


Ion Ratio Lower Upper

128 100

130 33.7 13.1 53.1

64 49.0 28.6 68.6





#9

Benzaldehyde

Concen: 5.457 ng

RT: 6.457 min Scan# 71

Delta R.T. -0.000 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005

Tgt Ion: 77 Resp: 54160

Ion Ratio Lower Upper

77 100

105 99.9 79.1 119.1





106 98.3 76.0 116.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#10

Phenol

Concen: 5.157 ng

RT: 6.528 min Scan# 737

Delta R.T. -0.012 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52




Tgt Ion: 94 Resp: 78823


Ion Ratio Lower Upper

94 100

65 28.3 8.9 48.9

66 34.6 18.7 58.7









#11
bis(2-Chloroethyl)ether
Concen: 5.331 ng
RT: 6.634 min Scan# 755
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion: 93 Resp: 63148
Ion Ratio Lower Upper
93 100
63 81.0 62.2 102.2
95 32.0 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#12
1,3-Dichlorobenzene
Concen: 5.367 ng
RT: 6.845 min Scan# 791
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:146 Resp: 73777
Ion Ratio Lower Upper
146 100
148 62.7 50.5 75.7
75 29.2 23.2 34.8





#13
1,4-Dichlorobenzene
Concen: 5.459 ng
RT: 6.928 min Scan# 805
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005




Tgt Ion:146 Resp: 75604
Ion Ratio Lower Upper
146 100
148 60.5 50.6 76.0
111 39.6 33.4 50.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#14
1,2-Dichlorobenzene
Concen: 5.289 ng
RT: 7.075 min Scan# 830
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Tgt Ion:146 Resp: 69920
Ion Ratio Lower Upper
146 100
148 64.6 50.5 75.7
111 44.0 34.9 52.3





#15
Benzyl Alcohol
Concen: 4.879 ng
RT: 7.039 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion: 79 Resp: 49130
Ion Ratio Lower Upper
79 100
108 82.4 64.5 96.7
77 65.4 49.2 73.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#16
2,2'-oxybis(1-chloropropane)
Concen: 5.198 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion: 45 Resp: 111179
Ion Ratio Lower Upper
45 100
77 11.5 0.0 31.3
79 8.5 0.0 28.5





#17
2-Methylphenol
Concen: 5.118 ng
RT: 7.145 min Scan# 843
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion:107 Resp: 5094
Ion Ratio Lower Upper
107 100
108 111.4 91.0 136.4
77 41.1 34.4 51.6
79 42.1 34.4 51.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#18
Hexachloroethane
Concen: 4.844 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:117 Resp: 21958
Ion Ratio Lower Upper
117 100
119 97.2 77.1 115.7
201 91.6 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 5.281 ng

RT: 7.310 min Scan# 81

Delta R.T. -0.012 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005

Tgt Ion: 70 Resp: 4634

Ion Ratio Lower Upper

70 100

42 68.0 53.4 80.2

101 8.9 7.7 11.5





130 21.2 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#20

3+4-Methylphenols

Concen: 5.119 ng

RT: 7.292 min Scan# 867

Delta R.T. -0.012 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Tgt Ion:107 Resp: 63725




Ion Ratio Lower Upper


107 100

108 91.8 71.2 111.2

77 48.0 41.1 81.1

79 26.0 8.1 48.1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.192 min Scan# 1020

Delta R.T. -0.001 min

Lab File: BF142240.D


Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005



Tgt Ion:136 Resp: 726978

Ion Ratio Lower Upper

136 100

137 11.1 8.7 13.1

54 8.9 7.7 11.5




68 6.0 5.0 7.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#22

Acetophenone


Concen: 5.518 ng

RT: 7.310 min Scan# 870

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52



Tgt Ion:105 Resp: 92806



Ion Ratio Lower Upper


105 100

71 7.3 5.7 8.5

51 33.0 30.1 45.1

120 23.8 19.0 28.4





#23

Nitrobenzene-d5

Concen: 7.118 ng

RT: 7.463 min Scan# 896

Delta R.T. -0.012 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005

Tgt Ion: 82 Resp: 7876

Ion Ratio Lower Upper

82 100

128 40.7 37.0 55.6





54 62.4 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#24

Nitrobenzene

Concen: 3.837 ng

RT: 7.481 min Scan# 899

Delta R.T. -0.012 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52




Tgt Ion: 77 Resp: 41660


Ion Ratio Lower Upper

77 100

123 42.7 38.6 58.0


65 14.1 11.5 17.3





#25
Isophorone
Concen: 5.165 ng
RT: 7.722 min Scan# 941
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion: 82 Resp: 119954
Ion Ratio Lower Upper
82 100
95 7.2 5.6 8.4
138 19.0 15.3 22.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#26
2-Nitrophenol
Concen: 6.901 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:139 Resp: 8775
Ion Ratio Lower Upper
139 100
109 27.5 21.8 32.8
65 52.6 41.1 61.7





#27

2,4-Dimethylphenol

Concen: 4.887 ng

RT: 7.828 min Scan# 91

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:	122	Resp:	5721
Ion	Ratio	Lower	Upper
122	100		
107	111.3	86.0	129.0
121	59.1	46.3	69.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#28

bis(2-Chloroethoxy)methane

Concen: 5.268 ng




RT: 7.934 min Scan# 976


Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Tgt Ion:	93	Resp:	77155
Ion	Ratio	Lower	Upper
93	100		
95	32.8	26.2	39.2
123	11.8	9.7	14.5





#29

2,4-Dichlorophenol

Concen: 4.560 ng

RT: 8.034 min Scan# 994

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F





ClientSampleId :

SSTDICC005

Tgt Ion:162	Resp:	4464
Ion Ratio	Lower	Upper
162	100	
164	60.8	43.0 83.0
98	36.0	16.7 56.7

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#30

1,2,4-Trichlorobenzene

Concen: 5.326 ng




RT: 8.128 min Scan# 1009


Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52


Tgt Ion:180	Resp:	58177
Ion Ratio	Lower	Upper
180	100	
182	94.8	75.7 113.5
145	30.3	25.4 38.0





#31
Naphthalene
Concen: 5.543 ng
RT: 8.210 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion:128 Resp: 200969
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.4
127 12.4 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#33
4-Chloroaniline
Concen: 5.367 ng
RT: 8.251 min Scan# 1030
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:127 Resp: 81024
Ion Ratio Lower Upper
127 100
129 31.2 26.1 39.1
65 32.3 25.5 38.3
92 18.7 15.4 23.2





#34

Hexachlorobutadiene

Concen: 5.127 ng

RT: 8.334 min Scan# 1044

Delta R.T. -0.000 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005

Tgt Ion:225 Resp: 32670

Ion Ratio Lower Upper

225 100

223 64.0 49.7 74.5





227 67.0 51.0 76.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#35

Caprolactam

Concen: 4.223 ng

RT: 8.575 min Scan# 1085

Delta R.T. -0.047 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52




Tgt Ion:113 Resp: 12427


Ion Ratio Lower Upper

113 100

55 197.9 177.7 217.7

56 146.0 127.1 167.1









#36
4-Chloro-3-methylphenol
Concen: 4.743 ng
RT: 8.722 min Scan# 1111
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:107 Resp: 4767
Ion Ratio Lower Upper
107 100
144 26.6 21.0 31.4
142 83.1 66.1 99.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#37
2-Methylnaphthalene
Concen: 5.539 ng
RT: 8.904 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:142 Resp: 123458
Ion Ratio Lower Upper
142 100
141 85.8 70.5 105.7





#38
1-Methylnaphthalene
Concen: 5.512 ng
RT: 8.998 min Scan# 1158
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion:142 Resp: 127274
Ion Ratio Lower Upper
142 100
141 90.5 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:164 Resp: 380022
Ion Ratio Lower Upper
164 100
162 100.5 78.6 117.8
160 44.8 35.7 53.5





#40

1,2,4,5-Tetrachlorobenzene

Concen: 5.513 ng

RT: 9.063 min Scan# 1169

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :





SSTDICC005

Tgt Ion:	216	Resp:	5896
Ion Ratio	Lower	Upper	
216	100		
214	76.4	63.8	95.6
179	20.8	16.7	25.1
108	19.3	15.6	23.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#41

Hexachlorocyclopentadiene

Concen: 3.827 ng




RT: 9.057 min Scan# 1167


Delta R.T. -0.000 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52


Tgt Ion:	237	Resp:	23228
Ion Ratio	Lower	Upper	
237	100		
235	61.6	43.5	83.5
272	12.0	0.0	32.9





#42
2,4,6-Tribromophenol
Concen: 7.613 ng
RT: 10.727 min Scan# 1451
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion:330 Resp: 2637
Ion Ratio Lower Upper
330 100
332 94.2 76.5 114.7
141 34.3 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#43
2,4,6-Trichlorophenol
Concen: 3.949 ng
RT: 9.175 min Scan# 1187
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:196 Resp: 26171
Ion Ratio Lower Upper
196 100
198 98.3 77.4 116.2
200 33.2 24.6 37.0





#44
2,4,5-Trichlorophenol
Concen: 4.390 ng
RT: 9.204 min Scan# 1193
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005




Tgt Ion:196 Resp: 3058
Ion Ratio Lower Upper
196 100
198 99.9 75.2 112.8
97 52.4 39.8 59.8
132 32.5 23.8 35.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#45
2-Fluorobiphenyl
Concen: 12.256 ng
RT: 9.263 min Scan# 1202
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52



Tgt Ion:172 Resp: 330323
Ion Ratio Lower Upper
172 100
171 35.0 29.0 43.4
170 22.9 19.4 29.0





#46

1,1'-Biphenyl

Concen: 5.702 ng

RT: 9.363 min Scan# 1219

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005

Tgt Ion:154 Resp: 168450

Ion Ratio Lower Upper

154 100

153 38.7 20.5 60.5





76 13.0 0.0 32.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#47

2-Chloronaphthalene

Concen: 5.572 ng

RT: 9.386 min Scan# 1223

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52




Tgt Ion:162 Resp: 122469


Ion Ratio Lower Upper

162 100

127 39.2 32.6 49.0


164 32.9 26.1 39.1





#48
2-Nitroaniline
Concen: 6.532 ng
RT: 9.475 min Scan# 1134
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion: 65 Resp: 1134
Ion Ratio Lower Upper
65 100
92 64.6 54.3 81.5
138 94.3 88.6 133.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#49
Acenaphthylene
Concen: 5.504 ng
RT: 9.804 min Scan# 1294
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:152 Resp: 204157
Ion Ratio Lower Upper
152 100
151 19.6 16.2 24.2
153 12.9 10.7 16.1





#50
Dimethylphthalate
Concen: 5.322 ng
RT: 9.657 min Scan# 1270
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion:163 Resp: 129370
Ion Ratio Lower Upper
163 100
194 3.9 3.2 4.8
164 10.1 8.2 12.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#51
2,6-Dinitrotoluene
Concen: 5.846 ng
RT: 9.716 min Scan# 1279
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:165 Resp: 9271
Ion Ratio Lower Upper
165 100
63 77.9 48.5 72.7#
89 65.9 43.9 65.9#









#52
Acenaphthene
Concen: 5.577 ng
RT: 9.975 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Instrument :
BNA_F
ClientSampleId :
SSTDICC005

Tgt Ion:154 Resp: 120775
Ion Ratio Lower Upper
154 100
153 110.1 88.6 133.0
152 52.8 42.1 63.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#53
3-Nitroaniline
Concen: 5.423 ng
RT: 9.886 min Scan# 1308
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:138 Resp: 13576
Ion Ratio Lower Upper
138 100
108 37.8 11.7 17.5#
92 120.9 91.2 136.8





#55

Dibenzenofuran

Concen: 5.649 ng

RT: 10.151 min Scan# 1353

Delta R.T. -0.000 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005

Tgt Ion:168 Resp: 18340

Ion Ratio Lower Upper

168 100

139 35.1 29.4 44.0





169 13.2 11.0 16.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#57

2,4-Dinitrotoluene

Concen: 6.601 ng

RT: 10.116 min Scan# 1347

Delta R.T. -0.012 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Tgt Ion:165 Resp: 9928




Ion Ratio Lower Upper


165 100

63 55.7 38.8 58.2

89 86.5 57.1 85.7#


182 2.6 2.2 3.2





#58
Fluorene
Concen: 5.746 ng
RT: 10.492 min Scan# 1412
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005




Tgt Ion:166 Resp: 141835
Ion Ratio Lower Upper
166 100
165 94.0 73.9 110.9
167 13.3 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#59
2,3,4,6-Tetrachlorophenol
Concen: 4.154 ng m
RT: 10.263 min Scan# 1372
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52



Tgt Ion:232 Resp: 23974
Ion Ratio Lower Upper
232 100
131 33.9 35.9 53.9#
130 2.0 1.8 2.6
166 20.6 24.0 36.0#





#60

Diethylphthalate

Concen: 5.238 ng

RT: 10.363 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005

Tgt Ion:149 Resp: 125748

Ion Ratio Lower Upper

149 100

177 21.0 17.6 26.4





150 12.8 10.2 15.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#61

4-Chlorophenyl-phenylether

Concen: 5.688 ng

RT: 10.486 min Scan# 1410

Delta R.T. -0.000 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52




Tgt Ion:204 Resp: 66676


Ion Ratio Lower Upper

204 100

206 31.7 25.6 38.4

141 58.7 48.9 73.3






#62
4-Nitroaniline
Concen: 5.254 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005






Tgt Ion:138 Resp: 1369
Ion Ratio Lower Upper
138 100
92 54.7 29.2 69.2
108 97.8 49.1 89.1


Manual Integrations

APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#63
Azobenzene
Concen: 5.394 ng
RT: 10.645 min Scan# 1437
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Tgt Ion: 77 Resp: 127873
Ion Ratio Lower Upper
77 100
182 26.1 7.3 47.3
105 22.7 2.5 42.5
51 34.7 14.4 54.4





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 11429
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion:188 Resp: 66493
Ion Ratio Lower Upper
188 100
94 10.0 8.2 12.2
80 10.5 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#66
n-Nitrosodiphenylamine
Concen: 5.317 ng
RT: 10.598 min Scan# 1429
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:169 Resp: 120800
Ion Ratio Lower Upper
169 100
168 66.4 53.0 79.6
167 36.3 28.8 43.2





#67
4-Bromophenyl-phenylether
Concen: 5.030 ng
RT: 10.975 min Scan# 1493
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005




Tgt Ion:248 Resp: 37248
Ion Ratio Lower Upper
248 100
250 97.7 77.2 115.8
141 77.2 58.4 87.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#68
Hexachlorobenzene
Concen: 5.131 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52



Tgt Ion:284 Resp: 42558
Ion Ratio Lower Upper
284 100
142 32.4 24.6 37.0
249 28.6 23.2 34.8





#69

Atrazine

Concen: 4.614 ng

RT: 11.122 min Scan# 1518

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDIC005

Tgt Ion:200 Resp: 27629

Ion Ratio Lower Upper

200 100

173 28.4 8.6 48.6





215 46.6 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#71

Phenanthrene

Concen: 5.637 ng

RT: 11.457 min Scan# 1575

Delta R.T. -0.000 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52




Tgt Ion:178 Resp: 202460


Ion Ratio Lower Upper

178 100

176 18.6 15.5 23.3


179 14.7 12.6 19.0





#72
Anthracene
Concen: 5.472 ng
RT: 11.504 min Scan# 11504
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005




Tgt Ion:178 Resp: 200940
Ion Ratio Lower Upper
178 100
176 18.2 15.3 22.9
179 15.3 12.7 19.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#73
Carbazole
Concen: 5.581 ng
RT: 11.657 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Tgt Ion:167 Resp: 186963
Ion Ratio Lower Upper
167 100
166 20.5 17.3 25.9
139 12.3 10.3 15.5





#74
Di-n-butylphthalate
Concen: 4.864 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005




Tgt Ion:149 Resp: 163860
Ion Ratio Lower Upper
149 100
150 8.9 7.5 11.3
104 4.6 3.6 5.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#75
Fluoranthene
Concen: 5.715 ng
RT: 12.645 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Tgt Ion:202 Resp: 202911
Ion Ratio Lower Upper
202 100
101 11.3 0.0 32.4
203 17.2 0.0 37.5





#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion:240 Resp: 45491
Ion Ratio Lower Upper
240 100
120 12.1 10.1 15.1
236 25.2 20.5 30.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#77
Benidine
Concen: 4.132 ng
RT: 12.769 min Scan# 1798
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52


Tgt Ion:184 Resp: 79877
Ion Ratio Lower Upper
184 100
185 13.8 11.8 17.6
183 10.7 8.9 13.3





#78
Pyrene
Concen: 4.968 ng
RT: 12.874 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005




Tgt Ion:202 Resp: 208462
Ion Ratio Lower Upper
202 100
200 20.0 16.1 24.1
203 17.2 14.1 21.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#79
Terphenyl-d14
Concen: 10.706 ng
RT: 13.016 min Scan# 1840
Delta R.T. -0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52



Tgt Ion:244 Resp: 334852
Ion Ratio Lower Upper
244 100
212 6.3 5.3 7.9
122 12.4 9.1 13.7





#80

Butylbenzylphthalate

Concen: 5.714 ng

RT: 13.498 min Scan# 1922

Delta R.T. -0.000 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005





Tgt Ion:149	Resp:	24840
Ion Ratio	Lower	Upper
149	100	
91	71.9	54.4 81.6
206	20.4	15.9 23.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#81

Benzo(a)anthracene

Concen: 5.136 ng




RT: 14.063 min Scan# 2018


Delta R.T. -0.000 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Tgt Ion:228	Resp:	156840
Ion Ratio	Lower	Upper
228	100	
226	27.1	21.9 32.9
229	18.8	16.0 24.0





#82

3,3'-Dichlorobenzidine

Concen: 3.623 ng

RT: 14.027 min Scan# 2012

Delta R.T. -0.000 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :





SSTDICC005

Tgt Ion:	252	Resp:	3250
Ion Ratio	Lower	Upper	
252	100		
254	68.6	51.0	76.6
126	14.1	9.4	14.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#83

Chrysene

Concen: 5.328 ng




RT: 14.098 min Scan# 2024


Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52


Tgt Ion:	228	Resp:	148119
Ion Ratio	Lower	Upper	
228	100		
226	29.2	24.0	36.0
229	18.8	15.7	23.5





#84
Bis(2-ethylhexyl)phthalate
Concen: 6.006 ng
RT: 14.063 min Scan# 2017
Delta R.T. 0.006 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion:149 Resp: 43692
Ion Ratio Lower Upper
149 100
167 26.5 21.7 32.5
279 3.3 3.0 4.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion:264 Resp: 322126
Ion Ratio Lower Upper
264 100
260 22.9 18.4 27.6
265 22.0 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 4.516 ng

RT: 17.068 min Scan# 2198

Delta R.T. -0.012 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :

SSTDICC005

Tgt Ion:276 Resp: 106554

Ion Ratio Lower Upper

276 100

138 29.4 22.6 33.8





277 24.7 20.1 30.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#88

Benzo(b)fluoranthene

Concen: 5.301 ng

RT: 15.121 min Scan# 2198

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52




Tgt Ion:252 Resp: 106072


Ion Ratio Lower Upper

252 100

253 22.2 17.7 26.5

125 11.6 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 5.327 ng

RT: 15.151 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52

Instrument :

BNA_F

ClientSampleId :





SSTDICC005

Tgt Ion:	252	Resp:	97490
Ion Ratio	Lower	Upper	
252	100		
253	21.9	17.8	26.8
125	10.8	9.0	13.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#90

Benzo(a)pyrene

Concen: 4.928 ng




RT: 15.498 min Scan# 2262


Delta R.T. -0.006 min

Lab File: BF142240.D

Acq: 30 Apr 2025 11:52


Tgt Ion:	252	Resp:	90284
Ion Ratio	Lower	Upper	
252	100		
253	21.1	17.6	26.4
125	11.9	9.5	14.3





#91
Dibenzo(a,h)anthracene
Concen: 4.520 ng
RT: 17.086 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52




Instrument :
BNA_F
ClientSampleId :
SSTDICC005



Tgt Ion: 278 Resp: 8655
Ion Ratio Lower Upper
278 100
139 19.0 15.2 22.8
279 23.7 19.0 28.6




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 4.500 ng
RT: 17.521 min Scan# 2606
Delta R.T. -0.018 min
Lab File: BF142240.D
Acq: 30 Apr 2025 11:52

Tgt Ion: 276 Resp: 86631
Ion Ratio Lower Upper
276 100
277 23.3 18.7 28.1
138 25.0 19.4 29.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142241.D
 Acq On : 30 Apr 2025 12:20
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Quant Time: Apr 30 15:41:16 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	198828	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	764114	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	402628	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	692116	20.000	ng	0.00
76) Chrysene-d12	14.074	240	449534	20.000	ng	0.00
86) Perylene-d12	15.568	264	332067	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	246136	19.771	ng	0.00
7) Phenol-d6	6.516	99	299244	19.905	ng	-0.01
23) Nitrobenzene-d5	7.463	82	192172	16.523	ng	-0.01
42) 2,4,6-Tribromophenol	10.727	330	63470	17.292	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	636078	22.275	ng	0.00
79) Terphenyl-d14	13.016	244	635233	20.552	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.699	88	53052	9.541	ng	97
3) Pyridine	3.469	79	138279	9.569	ng	99
4) n-Nitrosodimethylamine	3.399	42	69527	9.297	ng	# 98
6) Aniline	6.563	93	207232	9.739	ng	100
8) 2-Chlorophenol	6.687	128	122451	9.519	ng	98
9) Benzaldehyde	6.457	77	108198	10.297	ng	99
10) Phenol	6.528	94	157877	9.755	ng	99
11) bis(2-Chloroethyl)ether	6.634	93	124281	9.909	ng	99
12) 1,3-Dichlorobenzene	6.851	146	147587	10.141	ng	99
13) 1,4-Dichlorobenzene	6.928	146	147728	10.075	ng	99
14) 1,2-Dichlorobenzene	7.081	146	140999	10.073	ng	99
15) Benzyl Alcohol	7.039	79	101094	9.481	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.181	45	227340	10.038	ng	99
17) 2-Methylphenol	7.145	107	102299	9.707	ng	99
18) Hexachloroethane	7.422	117	46369	9.662	ng	94
19) n-Nitroso-di-n-propyla...	7.310	70	93168	10.029	ng	99
20) 3+4-Methylphenols	7.298	107	131546	9.981	ng	97
22) Acetophenone	7.310	105	184253	10.423	ng	96
24) Nitrobenzene	7.481	77	99900	8.754	ng	95
25) Isophorone	7.722	82	240659	9.859	ng	100
26) 2-Nitrophenol	7.804	139	21744	9.530	ng	98
27) 2,4-Dimethylphenol	7.834	122	121707	9.889	ng	97
28) bis(2-Chloroethoxy)met...	7.934	93	154349	10.026	ng	99
29) 2,4-Dichlorophenol	8.034	162	96761	9.404	ng	98
30) 1,2,4-Trichlorobenzene	8.128	180	115815	10.087	ng	99
31) Naphthalene	8.210	128	397159	10.422	ng	99
32) Benzoic acid	7.881	122	18272	13.045	ng	96
33) 4-Chloroaniline	8.257	127	157999	9.957	ng	99
34) Hexachlorobutadiene	8.334	225	66829	9.976	ng	98
35) Caprolactam	8.586	113	27800	8.989	ng	98
36) 4-Chloro-3-methylphenol	8.722	107	101179	9.576	ng	99
37) 2-Methylnaphthalene	8.904	142	243429	10.390	ng	99
38) 1-Methylnaphthalene	9.004	142	250299	10.312	ng	99
40) 1,2,4,5-Tetrachloroben...	9.069	216	115685	10.209	ng	98
41) Hexachlorocyclopentadiene	9.057	237	53365	8.298	ng	98
43) 2,4,6-Trichlorophenol	9.175	196	66271	9.438	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142241.D
 Acq On : 30 Apr 2025 12:20
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Quant Time: Apr 30 15:41:16 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	66509	9.009	ng	98
46) 1,1'-Biphenyl	9.363	154	326241	10.423	ng	99
47) 2-Chloronaphthalene	9.392	162	236725	10.165	ng	99
48) 2-Nitroaniline	9.480	65	31455	9.382	ng	98
49) Acenaphthylene	9.804	152	407630	10.373	ng	98
50) Dimethylphthalate	9.657	163	258512	10.036	ng	99
51) 2,6-Dinitrotoluene	9.722	165	27624	9.147	ng	91
52) Acenaphthene	9.980	154	236444	10.305	ng	99
53) 3-Nitroaniline	9.886	138	37727	9.049	ng	# 97
54) 2,4-Dinitrophenol	9.986	184	4993	11.362	ng	# 1
55) Dibenzofuran	10.151	168	357431	10.392	ng	99
56) 4-Nitrophenol	10.028	139	26399	6.520	ng	91
57) 2,4-Dinitrotoluene	10.122	165	29245	9.405	ng	94
58) Fluorene	10.492	166	282449	10.800	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	53660	8.775	ng	99
60) Diethylphthalate	10.363	149	254431	10.003	ng	98
61) 4-Chlorophenyl-phenyle...	10.486	204	129149	10.398	ng	99
62) 4-Nitroaniline	10.492	138	35703	8.911	ng	88
63) Azobenzene	10.645	77	253731	10.102	ng	99
65) 4,6-Dinitro-2-methylph...	10.527	198	7762	13.263	ng	95
66) n-Nitrosodiphenylamine	10.598	169	237471	10.042	ng	100
67) 4-Bromophenyl-phenylether	10.975	248	75692	9.820	ng	97
68) Hexachlorobenzene	11.039	284	84881	9.832	ng	96
69) Atrazine	11.122	200	59553	9.554	ng	98
70) Pentachlorophenol	11.227	266	32058	7.029	ng	99
71) Phenanthrene	11.457	178	385677	10.316	ng	99
72) Anthracene	11.510	178	401700	10.509	ng	98
73) Carbazole	11.657	167	362249	10.389	ng	98
74) Di-n-butylphthalate	11.998	149	348111	9.927	ng	99
75) Fluoranthene	12.645	202	395757	10.709	ng	99
77) Benzidine	12.769	184	168038	8.797	ng	99
78) Pyrene	12.874	202	404860	9.763	ng	100
80) Butylbenzylphthalate	13.498	149	65964	9.084	ng	98
81) Benzo(a)anthracene	14.063	228	296710	9.832	ng	99
82) 3,3'-Dichlorobenzidine	14.027	252	71321	8.046	ng	98
83) Chrysene	14.098	228	271875	9.897	ng	100
84) Bis(2-ethylhexyl)phtha...	14.063	149	100777	9.409	ng	99
85) Di-n-octyl phthalate	14.680	149	150072	10.622	ng	99
87) Indeno(1,2,3-cd)pyrene	17.068	276	229818	9.449	ng	100
88) Benzo(b)fluoranthene	15.121	252	202615	9.823	ng	99
89) Benzo(k)fluoranthene	15.157	252	192500	10.203	ng	98
90) Benzo(a)pyrene	15.498	252	182258	9.651	ng	99
91) Dibenzo(a,h)anthracene	17.092	278	187788	9.513	ng	99
92) Benzo(g,h,i)perylene	17.521	276	188891	9.518	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_F
ClientSampleId :
SSTDICC010





[illegible]



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:152 Resp: 198828
Ion Ratio Lower Upper
152 100
150 155.6 130.2 195.2
115 58.2 47.0 70.4



#2
1,4-Dioxane
Concen: 9.541 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 88 Resp: 53052
Ion Ratio Lower Upper
88 100
58 92.8 72.2 108.2
43 35.3 27.0 40.6





#3

Pyridine

Concen: 9.569 ng

RT: 3.469 min Scan# 21

Delta R.T. 0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion: 79 Resp: 138279

Ion Ratio Lower Upper

79 100

52 89.5 71.5 107.3

51 41.3 34.5 51.7



#4

n-Nitrosodimethylamine

Concen: 9.297 ng

RT: 3.399 min Scan# 205

Delta R.T. -0.012 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion: 42 Resp: 69527


Ion Ratio Lower Upper

42 100

74 113.7 89.5 134.3

44 12.7 6.6 9.8#









#5
2-Fluorophenol
Concen: 19.771 ng
RT: 5.510 min Scan# 50
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:112 Resp: 246136
Ion Ratio Lower Upper
112 100
64 60.2 48.4 72.6
63 31.3 24.8 37.2



#6
Aniline
Concen: 9.739 ng
RT: 6.563 min Scan# 743
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 93 Resp: 207232
Ion Ratio Lower Upper
93 100
66 40.2 32.1 48.1
65 20.5 16.6 24.8





#7

Phenol-d6

Concen: 19.905 ng

RT: 6.516 min Scan# 71

Delta R.T. -0.012 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion: 99 Resp: 299244

Ion Ratio Lower Upper

99 100

42 20.6 16.3 24.5

71 30.9 25.0 37.4



#8

2-Chlorophenol

Concen: 9.519 ng

RT: 6.687 min Scan# 764

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion: 128 Resp: 122451


Ion Ratio Lower Upper

128 100

130 31.1 13.1 53.1

64 49.2 28.6 68.6





#9

Benzaldehyde

Concen: 10.297 ng

RT: 6.457 min Scan# 71 Instrument :

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA F

ClientSampleId :

SSTDICC010



Tgt Ion: 77 Resp: 108198

Ion	Ratio	Lower	Upper

Lower	Upper
77	100

105	100.7	79.1	119.1

106	95.9	76.0	116.0



#10

Phenol

Concen: 9.755 ng

RT: 6.528 min Scan# 737

Delta R.T. -0.012 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion: 94 Resp: 157877


Ion	Ratio	Lower	Upper

Ion	Ratio	Lower	Upper
94	100		

65	28.9	8.9	48.9

66	37.6	18.7	58.7









#11
bis(2-Chloroethyl)ether
Concen: 9.909 ng
RT: 6.634 min Scan# 755
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion: 93 Resp: 124281
Ion Ratio Lower Upper
93 100
63 83.6 62.2 102.2
95 32.8 12.7 52.7



#12
1,3-Dichlorobenzene
Concen: 10.141 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 146 Resp: 147587
Ion Ratio Lower Upper
146 100
148 63.5 50.5 75.7
75 28.0 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 10.075 ng

RT: 6.928 min Scan# 805

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDIC010





Tgt Ion:146 Resp: 147728

Ion Ratio Lower Upper

146 100

148 62.2 50.6 76.0

111 41.4 33.4 50.0



#14

1,2-Dichlorobenzene

Concen: 10.073 ng

RT: 7.081 min Scan# 831

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion:146 Resp: 140999


Ion Ratio Lower Upper

146 100

148 63.7 50.5 75.7

111 42.8 34.9 52.3









#15
Benzyl Alcohol
Concen: 9.481 ng
RT: 7.039 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion: 79 Resp: 101094
Ion Ratio Lower Upper
79 100
108 80.3 64.5 96.7
77 63.9 49.2 73.8



#16
2,2'-oxybis(1-chloropropane)
Concen: 10.038 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20


Tgt Ion: 45 Resp: 227340
Ion Ratio Lower Upper
45 100
77 10.8 0.0 31.3
79 8.2 0.0 28.5








#17
2-Methylphenol
Concen: 9.707 ng
RT: 7.145 min Scan# 843
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Instrument :
BNA_F
ClientSampleId :
SSTDICC010







Tgt Ion	Ratio	Lower	Upper
107	100		
108	113.6	91.0	136.4
77	42.5	34.4	51.6
79	42.0	34.4	51.6



#18
Hexachloroethane
Concen: 9.662 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion	Ratio	Lower	Upper
117	100		
119	96.1	77.1	115.7
201	88.1	79.4	119.2





#19

n-Nitroso-di-n-propylamine

Concen: 10.029 ng

RT: 7.310 min Scan# 81

Delta R.T. -0.012 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010

Tgt Ion: 70 Resp: 93168





Ion Ratio Lower Upper

70 100

42 68.0 53.4 80.2

101 9.5 7.7 11.5

130 21.4 17.2 25.8



#20

3+4-Methylphenols

Concen: 9.981 ng

RT: 7.298 min Scan# 868

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Tgt Ion: 107 Resp: 131546




Ion Ratio Lower Upper


107 100

108 91.7 71.2 111.2

77 68.1 41.1 81.1

79 27.8 8.1 48.1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.192 min Scan# 1

Delta R.T. -0.001 min

Lab File: BF142241.D


Acq: 30 Apr 2025 12:20

Instrument :

BNA_F




ClientSampleId :

SSTDICC010



Tgt Ion:136 Resp: 764114

Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.7	13.1
54	9.1	7.7	11.5
68	6.1	5.0	7.6



#22

Acetophenone

Concen: 10.423 ng

RT: 7.310 min Scan# 870




Delta R.T. -0.006 min


Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Tgt Ion:105 Resp: 184253

Ion	Ratio	Lower	Upper
105	100		
71	8.1	5.7	8.5
51	34.2	30.1	45.1
120	22.7	19.0	28.4





#23

Nitrobenzene-d5

Concen: 16.523 ng

RT: 7.463 min Scan# 896

Delta R.T. -0.012 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :





BNA_F

ClientSampleId :

SSTDICC010

Tgt Ion: 82 Resp: 192172

Ion	Ratio	Lower	Upper
82	100		
128	41.8	37.0	55.6
54	60.8	48.0	72.0



#24

Nitrobenzene

Concen: 8.754 ng

RT: 7.481 min Scan# 899




Delta R.T. -0.012 min


Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Tgt Ion: 77 Resp: 99900

Ion	Ratio	Lower	Upper
77	100		
123	43.7	38.6	58.0
65	14.3	11.5	17.3









#25
Isophorone
Concen: 9.859 ng
RT: 7.722 min Scan# 941
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion: 82 Resp: 240659
Ion Ratio Lower Upper
82 100
95 6.9 5.6 8.4
138 18.8 15.3 22.9



#26
2-Nitrophenol
Concen: 9.530 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 139 Resp: 21744
Ion Ratio Lower Upper
139 100
109 26.6 21.8 32.8
65 53.0 41.1 61.7





#27

2,4-Dimethylphenol

Concen: 9.889 ng

RT: 7.834 min Scan# 91

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:122 Resp: 121707

Ion Ratio Lower Upper

122 100

107 102.8 86.0 129.0

121 57.1 46.3 69.5



#28

bis(2-Chloroethoxy)methane

Concen: 10.026 ng

RT: 7.934 min Scan# 976

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion: 93 Resp: 154349


Ion Ratio Lower Upper

93 100

95 32.9 26.2 39.2

123 11.2 9.7 14.5





#29

2,4-Dichlorophenol

Concen: 9.404 ng

RT: 8.034 min Scan# 994

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:162 Resp: 96761

Ion Ratio Lower Upper

162 100

164 63.7 43.0 83.0

98 38.4 16.7 56.7



#30

1,2,4-Trichlorobenzene

Concen: 10.087 ng

RT: 8.128 min Scan# 1009

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion:180 Resp: 115815


Ion Ratio Lower Upper

180 100

182 93.3 75.7 113.5

145 31.6 25.4 38.0









#31
Naphthalene
Concen: 10.422 ng
RT: 8.210 min Scan# 1024
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:128 Resp: 397159
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.4
127 12.8 10.6 15.8



#32
Benzoic acid
Concen: 13.045 ng
RT: 7.881 min Scan# 967
Delta R.T. -0.053 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:122 Resp: 18272
Ion Ratio Lower Upper
122 100
105 120.0 98.2 138.2
77 92.6 66.8 106.8









#33
4-Chloroaniline
Concen: 9.957 ng
RT: 8.257 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:	127	Resp:	157999
Ion	Ratio	Lower	Upper
127	100		
129	31.9	26.1	39.1
65	30.9	25.5	38.3
92	18.9	15.4	23.2



#34
Hexachlorobutadiene
Concen: 9.976 ng
RT: 8.334 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:	225	Resp:	66829
Ion	Ratio	Lower	Upper
225	100		
223	63.9	49.7	74.5
227	62.6	51.0	76.4









#35
Caprolactam
Concen: 8.989 ng
RT: 8.586 min Scan# 1107
Delta R.T. -0.035 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:113 Resp: 27800
Ion Ratio Lower Upper
113 100
55 194.0 177.7 217.7
56 144.7 127.1 167.1



#36
4-Chloro-3-methylphenol
Concen: 9.576 ng
RT: 8.722 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:107 Resp: 101179
Ion Ratio Lower Upper
107 100
144 27.3 21.0 31.4
142 82.8 66.1 99.1





#37

2-Methylnaphthalene

Concen: 10.390 ng

RT: 8.904 min Scan# 1141

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :





SSTDICC010

Tgt Ion:142 Resp: 243429

Ion Ratio Lower Upper

142 100

141 86.8 70.5 105.7



#38

1-Methylnaphthalene

Concen: 10.312 ng

RT: 9.004 min Scan# 1158

Delta R.T. -0.000 min

Lab File: BF142241.D




Acq: 30 Apr 2025 12:20


Tgt Ion:142 Resp: 250299

Ion Ratio Lower Upper

142 100

141 92.1 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.945 min Scan# 1118

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:164 Resp: 402628

Ion Ratio Lower Upper

164 100

162 99.1 78.6 117.8

160 43.8 35.7 53.5



#40

1,2,4,5-Tetrachlorobenzene

Concen: 10.209 ng

RT: 9.069 min Scan# 1169

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Tgt Ion:216 Resp: 115685




Ion Ratio Lower Upper


216 100

214 77.5 63.8 95.6

179 21.0 16.7 25.1

108 18.9 15.6 23.4









#41
Hexachlorocyclopentadiene
Concen: 8.298 ng
RT: 9.057 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion: 237 Resp: 53365
Ion Ratio Lower Upper
237 100
235 62.1 43.5 83.5
272 12.8 0.0 32.9



#42
2,4,6-Tribromophenol
Concen: 17.292 ng
RT: 10.727 min Scan# 1451
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 330 Resp: 63470
Ion Ratio Lower Upper
330 100
332 95.0 76.5 114.7
141 32.9 24.3 36.5





#43

2,4,6-Trichlorophenol

Concen: 9.438 ng

RT: 9.175 min Scan# 1187

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:196 Resp: 66271

Ion Ratio Lower Upper

196 100

198 92.9 77.4 116.2

200 30.2 24.6 37.0



#44

2,4,5-Trichlorophenol

Concen: 9.009 ng

RT: 9.210 min Scan# 1193

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Tgt Ion:196 Resp: 66509




Ion Ratio Lower Upper


196 100

198 95.6 75.2 112.8

97 47.5 39.8 59.8

132 29.2 23.8 35.8









#45
2-Fluorobiphenyl
Concen: 22.275 ng
RT: 9.263 min Scan# 1119
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:172 Resp: 636078
Ion Ratio Lower Upper
172 100
171 35.4 29.0 43.4
170 22.6 19.4 29.0



#46
1,1'-Biphenyl
Concen: 10.423 ng
RT: 9.363 min Scan# 1219
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:154 Resp: 326241
Ion Ratio Lower Upper
154 100
153 39.5 20.5 60.5
76 12.7 0.0 32.8





#47

2-Chloronaphthalene

Concen: 10.165 ng

RT: 9.392 min Scan# 11

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:162 Resp: 236725

Ion Ratio Lower Upper

162 100

127 40.0 32.6 49.0

164 33.3 26.1 39.1



#48

2-Nitroaniline

Concen: 9.382 ng

RT: 9.480 min Scan# 1239

Delta R.T. -0.007 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion: 65 Resp: 31455


Ion Ratio Lower Upper

65 100

92 68.4 54.3 81.5

138 108.0 88.6 133.0





#49

Acenaphthylene

Concen: 10.373 ng

RT: 9.804 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:152 Resp: 407630

Ion Ratio Lower Upper

152 100

151 19.3 16.2 24.2

153 13.1 10.7 16.1



#50

Dimethylphthalate

Concen: 10.036 ng

RT: 9.657 min Scan# 1269

Delta R.T. -0.012 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion:163 Resp: 258512


Ion Ratio Lower Upper

163 100

194 3.8 3.2 4.8

164 10.0 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 9.147 ng
RT: 9.722 min Scan# 1281
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:165 Resp: 27624
Ion Ratio Lower Upper
165 100
63 66.0 48.5 72.7
89 63.2 43.9 65.9



#52
Acenaphthene
Concen: 10.305 ng
RT: 9.980 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:154 Resp: 236444
Ion Ratio Lower Upper
154 100
153 110.2 88.6 133.0
152 52.1 42.1 63.1









#53
3-Nitroaniline
Concen: 9.049 ng
RT: 9.886 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:138 Resp: 37727
Ion Ratio Lower Upper
138 100
108 25.3 11.7 17.5#
92 113.9 91.2 136.8



#54
2,4-Dinitrophenol
Concen: 11.362 ng
RT: 9.986 min Scan# 1325
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:184 Resp: 4993
Ion Ratio Lower Upper
184 100
63 302.8 69.2 103.8#
154 3030.9 271.4 407.2#





#55

Dibenzenofuran

Concen: 10.392 ng

RT: 10.151 min Scan# 11

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:168 Resp: 357431

Ion Ratio Lower Upper

168 100

139 35.7 29.4 44.0

169 13.3 11.0 16.4



#56

4-Nitrophenol

Concen: 6.520 ng

RT: 10.028 min Scan# 1332

Delta R.T. -0.011 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion:139 Resp: 26399


Ion Ratio Lower Upper

139 100

109 77.2 44.9 84.9

65 99.0 74.7 114.7









#57
2,4-Dinitrotoluene
Concen: 9.405 ng
RT: 10.122 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion	Ratio	Lower	Upper
165	100		
63	51.5	38.8	58.2
89	76.8	57.1	85.7
182	2.7	2.2	3.2



#58
Fluorene
Concen: 10.800 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion	Ratio	Lower	Upper
166	100		
165	93.3	73.9	110.9
167	13.4	10.8	16.2









#59
2,3,4,6-Tetrachlorophenol
Concen: 8.775 ng
RT: 10.263 min Scan# 1372
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:232 Resp: 53660
Ion Ratio Lower Upper
232 100
131 44.3 35.9 53.9
130 2.4 1.8 2.6
166 30.8 24.0 36.0



#60
Diethylphthalate
Concen: 10.003 ng
RT: 10.363 min Scan# 1389
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:149 Resp: 254431
Ion Ratio Lower Upper
149 100
177 21.2 17.6 26.4
150 12.3 10.2 15.2









#61
4-Chlorophenyl-phenylether
Concen: 10.398 ng
RT: 10.486 min Scan# 1410
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:204 Resp: 129149
Ion Ratio Lower Upper
204 100
206 31.9 25.6 38.4
141 60.4 48.9 73.3



#62
4-Nitroaniline
Concen: 8.911 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:138 Resp: 35703
Ion Ratio Lower Upper
138 100
92 52.3 29.2 69.2
108 82.7 49.1 89.1









#63
Azobenzene
Concen: 10.102 ng
RT: 10.645 min Scan# 1437
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion: 77 Resp: 253731
Ion Ratio Lower Upper
77 100
182 25.8 7.3 47.3
105 22.6 2.5 42.5
51 34.7 14.4 54.4



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion: 188 Resp: 692116
Ion Ratio Lower Upper
188 100
94 10.1 8.2 12.2
80 10.4 8.6 12.8









#65
4,6-Dinitro-2-methylphenol
Concen: 13.263 ng
RT: 10.527 min Scan# 1417
Delta R.T. -0.007 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:198 Resp: 7762
Ion Ratio Lower Upper
198 100
51 65.4 40.6 80.6
105 52.1 29.2 69.2



#66
n-Nitrosodiphenylamine
Concen: 10.042 ng
RT: 10.598 min Scan# 1429
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:169 Resp: 237471
Ion Ratio Lower Upper
169 100
168 66.5 53.0 79.6
167 35.9 28.8 43.2









#67
4-Bromophenyl-phenylether
Concen: 9.820 ng
RT: 10.975 min Scan# 1493
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:248 Resp: 75692
Ion Ratio Lower Upper
248 100
250 96.8 77.2 115.8
141 79.1 58.4 87.6



#68
Hexachlorobenzene
Concen: 9.832 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:284 Resp: 84881
Ion Ratio Lower Upper
284 100
142 33.6 24.6 37.0
249 30.2 23.2 34.8





#69

Atrazine

Concen: 9.554 ng

RT: 11.122 min Scan# 1518

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDIC010





Tgt Ion:200 Resp: 59553

Ion Ratio Lower Upper

200 100

173 26.8 8.6 48.6

215 47.6 28.5 68.5



#70

Pentachlorophenol

Concen: 7.029 ng

RT: 11.227 min Scan# 1536

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion:266 Resp: 32058


Ion Ratio Lower Upper

266 100

268 63.5 50.0 75.0

264 61.4 49.3 73.9









#71
Phenanthrene
Concen: 10.316 ng
RT: 11.457 min Scan# 1575
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:178 Resp: 385677
Ion Ratio Lower Upper
178 100
176 19.1 15.5 23.3
179 15.2 12.6 19.0



#72
Anthracene
Concen: 10.509 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:178 Resp: 401700
Ion Ratio Lower Upper
178 100
176 18.2 15.3 22.9
179 15.1 12.7 19.1









#73
Carbazole
Concen: 10.389 ng
RT: 11.657 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:167 Resp: 362249
Ion Ratio Lower Upper
167 100
166 20.7 17.3 25.9
139 12.5 10.3 15.5



#74
Di-n-butylphthalate
Concen: 9.927 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:149 Resp: 348111
Ion Ratio Lower Upper
149 100
150 9.0 7.5 11.3
104 4.4 3.6 5.4









#75
Fluoranthene
Concen: 10.709 ng
RT: 12.645 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:202 Resp: 395757
Ion Ratio Lower Upper
202 100
101 11.6 0.0 32.4
203 17.3 0.0 37.5



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:240 Resp: 449534
Ion Ratio Lower Upper
240 100
120 12.0 10.1 15.1
236 25.6 20.5 30.7









#77
Benzidine
Concen: 8.797 ng
RT: 12.769 min Scan# 1798
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:184 Resp: 168038
Ion Ratio Lower Upper
184 100
185 14.3 11.8 17.6
183 10.4 8.9 13.3



#78
Pyrene
Concen: 9.763 ng
RT: 12.874 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:202 Resp: 404860
Ion Ratio Lower Upper
202 100
200 19.8 16.1 24.1
203 17.6 14.1 21.1





#79

Terphenyl-d14

Concen: 20.552 ng

RT: 13.016 min Scan# 1841

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDIC010





Tgt Ion:244 Resp: 635233

Ion Ratio Lower Upper

244 100

212 6.5 5.3 7.9

122 12.3 9.1 13.7



#80

Butylbenzylphthalate

Concen: 9.084 ng

RT: 13.498 min Scan# 1922

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion:149 Resp: 65964


Ion Ratio Lower Upper

149 100

91 66.6 54.4 81.6

206 20.5 15.9 23.9









#81
Benzo(a)anthracene
Concen: 9.832 ng
RT: 14.063 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:228 Resp: 296710
Ion Ratio Lower Upper
228 100
226 27.0 21.9 32.9
229 19.7 16.0 24.0



#82
3,3'-Dichlorobenzidine
Concen: 8.046 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:252 Resp: 71321
Ion Ratio Lower Upper
252 100
254 65.6 51.0 76.6
126 12.5 9.4 14.2





#83

Chrysene

Concen: 9.897 ng

RT: 14.098 min Scan# 2018

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:228 Resp: 271875

Ion Ratio Lower Upper

228 100

226 30.3 24.0 36.0

229 19.8 15.7 23.5



#84

Bis(2-ethylhexyl)phthalate

Concen: 9.409 ng

RT: 14.063 min Scan# 2018

Delta R.T. 0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion:149 Resp: 100777


Ion Ratio Lower Upper

149 100

167 26.3 21.7 32.5

279 4.0 3.0 4.4









#85
Di-n-octyl phthalate
Concen: 10.622 ng
RT: 14.680 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:149 Resp: 150072
Ion Ratio Lower Upper
149 100
167 1.5 1.1 1.7
43 12.1 9.3 13.9



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:264 Resp: 332067
Ion Ratio Lower Upper
264 100
260 23.4 18.4 27.6
265 21.0 17.3 25.9









#87
Indeno(1,2,3-cd)pyrene
Concen: 9.449 ng
RT: 17.068 min Scan# 2198
Delta R.T. -0.012 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:276 Resp: 229818
Ion Ratio Lower Upper
276 100
138 28.5 22.6 33.8
277 25.1 20.1 30.1



#88
Benzo(b)fluoranthene
Concen: 9.823 ng
RT: 15.121 min Scan# 2198
Delta R.T. -0.006 min
Lab File: BF142241.D
Acq: 30 Apr 2025 12:20

Tgt Ion:252 Resp: 202615
Ion Ratio Lower Upper
252 100
253 21.6 17.7 26.5
125 11.1 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 10.203 ng

RT: 15.157 min Scan# 21

Delta R.T. -0.000 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:252 Resp: 192500

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.8

125 10.4 9.0 13.6



#90

Benzo(a)pyrene

Concen: 9.651 ng

RT: 15.498 min Scan# 2262

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion:252 Resp: 182258


Ion Ratio Lower Upper

252 100

253 21.3 17.6 26.4

125 12.5 9.5 14.3





#91

Dibenzo(a,h)anthracene

Concen: 9.513 ng

RT: 17.092 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:278 Resp: 187788

Ion Ratio Lower Upper

278 100

139 19.3 15.2 22.8

279 24.0 19.0 28.6



#92

Benzo(g,h,i)perylene

Concen: 9.518 ng

RT: 17.521 min Scan# 2606

Delta R.T. -0.018 min

Lab File: BF142241.D

Acq: 30 Apr 2025 12:20




Tgt Ion:276 Resp: 188891

Ion Ratio Lower Upper

276 100

277 23.9 18.7 28.1

138 24.4 19.4 29.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142242.D
 Acq On : 30 Apr 2025 12:49
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:42:06 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Apr 30 15:28:40 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	194641	20.000	ng	0.00
21) Naphthalene-d8	8.193	136	746567	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	398805	20.000	ng	0.00
64) Phenanthrene-d10	11.434	188	679379	20.000	ng	0.00
76) Chrysene-d12	14.075	240	430351	20.000	ng	0.00
86) Perylene-d12	15.569	264	350840	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	492356	40.399	ng	0.00
7) Phenol-d6	6.516	99	597221	40.580	ng	-0.01
23) Nitrobenzene-d5	7.469	82	450837	39.673	ng	0.00
42) 2,4,6-Tribromophenol	10.734	330	146779	40.372	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	1212743	42.876	ng	0.00
79) Terphenyl-d14	13.022	244	1223735	41.357	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.693	88	109201	20.062	ng	100
3) Pyridine	3.458	79	282987	20.004	ng	100
4) n-Nitrosodimethylamine	3.399	42	141737	19.361	ng	97
6) Aniline	6.569	93	419797	20.152	ng	99
8) 2-Chlorophenol	6.687	128	253840	20.156	ng	99
9) Benzaldehyde	6.457	77	213052	20.712	ng	98
10) Phenol	6.534	94	322435	20.352	ng	98
11) bis(2-Chloroethyl)ether	6.640	93	247179	20.132	ng	99
12) 1,3-Dichlorobenzene	6.851	146	289129	20.294	ng	100
13) 1,4-Dichlorobenzene	6.928	146	290927	20.268	ng	99
14) 1,2-Dichlorobenzene	7.081	146	278592	20.331	ng	99
15) Benzyl Alcohol	7.040	79	207042	19.835	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.181	45	448250	20.218	ng	99
17) 2-Methylphenol	7.146	107	207957	20.157	ng	99
18) Hexachloroethane	7.422	117	93572	19.917	ng	96
19) n-Nitroso-di-n-propyla...	7.316	70	184776	20.318	ng	99
20) 3+4-Methylphenols	7.298	107	267057	20.700	ng	96
22) Acetophenone	7.310	105	360484	20.872	ng	98
24) Nitrobenzene	7.487	77	221499	19.865	ng	98
25) Isophorone	7.722	82	479096	20.088	ng	99
26) 2-Nitrophenol	7.804	139	61615	17.893	ng	98
27) 2,4-Dimethylphenol	7.834	122	248864	20.696	ng	97
28) bis(2-Chloroethoxy)met...	7.934	93	309073	20.549	ng	99
29) 2,4-Dichlorophenol	8.040	162	204821	20.374	ng	99
30) 1,2,4-Trichlorobenzene	8.128	180	227898	20.315	ng	100
31) Naphthalene	8.210	128	772839	20.758	ng	100
32) Benzoic acid	7.904	122	64656m	19.412	ng	
33) 4-Chloroaniline	8.257	127	320218	20.654	ng	99
34) Hexachlorobutadiene	8.334	225	133474	20.393	ng	99
35) Caprolactam	8.604	113	60791	20.118	ng	99
36) 4-Chloro-3-methylphenol	8.722	107	208419	20.190	ng	99
37) 2-Methylnaphthalene	8.904	142	472896	20.658	ng	100
38) 1-Methylnaphthalene	9.004	142	497635	20.985	ng	99
40) 1,2,4,5-Tetrachloroben...	9.069	216	228311	20.342	ng	99
41) Hexachlorocyclopentadiene	9.057	237	121953	19.145	ng	99
43) 2,4,6-Trichlorophenol	9.175	196	140807	20.246	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142242.D
 Acq On : 30 Apr 2025 12:49
 Operator : RC/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC020

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:42:06 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	144462	19.756	ng	99
46) 1,1'-Biphenyl	9.369	154	643017	20.740	ng	99
47) 2-Chloronaphthalene	9.392	162	474899	20.588	ng	98
48) 2-Nitroaniline	9.481	65	90798	18.213	ng	97
49) Acenaphthylene	9.810	152	807654	20.749	ng	99
50) Dimethylphthalate	9.663	163	518802	20.334	ng	100
51) 2,6-Dinitrotoluene	9.722	165	76955	18.432	ng	93
52) Acenaphthene	9.981	154	460927	20.281	ng	100
53) 3-Nitroaniline	9.892	138	100560	18.959	ng	# 95
54) 2,4-Dinitrophenol	9.992	184	14553	18.804	ng	# 1
55) Dibenzofuran	10.151	168	698155	20.493	ng	100
56) 4-Nitrophenol	10.034	139	70423	17.561	ng	97
57) 2,4-Dinitrotoluene	10.122	165	84812	17.844	ng	95
58) Fluorene	10.492	166	538647	20.793	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	124353	20.530	ng	99
60) Diethylphthalate	10.363	149	519294	20.611	ng	98
61) 4-Chlorophenyl-phenyle...	10.486	204	252213	20.502	ng	99
62) 4-Nitroaniline	10.498	138	92884	18.933	ng	97
63) Azobenzene	10.645	77	508482	20.438	ng	98
65) 4,6-Dinitro-2-methylph...	10.528	198	23400	18.756	ng	98
66) n-Nitrosodiphenylamine	10.604	169	470767	20.280	ng	100
67) 4-Bromophenyl-phenylether	10.981	248	151720	20.053	ng	99
68) Hexachlorobenzene	11.039	284	168767	19.916	ng	97
69) Atrazine	11.128	200	126444	20.665	ng	99
70) Pentachlorophenol	11.228	266	81929	18.301	ng	99
71) Phenanthrene	11.457	178	760476	20.723	ng	99
72) Anthracene	11.510	178	782215	20.848	ng	99
73) Carbazole	11.657	167	720775	21.059	ng	99
74) Di-n-butylphthalate	11.998	149	739011	21.469	ng	99
75) Fluoranthene	12.645	202	784880	21.636	ng	100
77) Benzidine	12.769	184	381034	20.837	ng	99
78) Pyrene	12.875	202	795806	20.046	ng	100
80) Butylbenzylphthalate	13.498	149	174268	18.529	ng	99
81) Benzo(a)anthracene	14.063	228	573438	19.849	ng	99
82) 3,3'-Dichlorobenzidine	14.027	252	163359	19.251	ng	100
83) Chrysene	14.098	228	537388	20.435	ng	100
84) Bis(2-ethylhexyl)phtha...	14.057	149	236325	18.042	ng	100
85) Di-n-octyl phthalate	14.674	149	376362	17.998	ng	98
87) Indeno(1,2,3-cd)pyrene	17.074	276	514496	20.021	ng	99
88) Benzo(b)fluoranthene	15.127	252	423694	19.442	ng	98
89) Benzo(k)fluoranthene	15.157	252	428646	21.504	ng	98
90) Benzo(a)pyrene	15.504	252	402101	20.154	ng	99
91) Dibenzo(a,h)anthracene	17.092	278	420029	20.139	ng	98
92) Benzo(g,h,i)perylene	17.527	276	422497	20.150	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142242.D
Acq On : 30 Apr 2025 12:49
Operator : RC/JU
Sample : SSTDICC020
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025


Quant Time: Apr 30 15:42:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 15:28:40 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:152 Resp: 19464
Ion Ratio Lower Upper
152 100
150 157.2 130.2 195.2
115 57.3 47.0 70.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#2
1,4-Dioxane
Concen: 20.062 ng
RT: 2.693 min Scan# 85
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 88 Resp: 109201
Ion Ratio Lower Upper
88 100
58 90.0 72.2 108.2
43 34.2 27.0 40.6





#3

Pyridine

Concen: 20.004 ng

RT: 3.458 min Scan# 216

Delta R.T. -0.006 min

Lab File: BF142242.D


Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020






Tgt Ion	Ratio	Lower	Upper
79	100		
52	89.0	71.5	107.3
51	43.1	34.5	51.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#4

n-Nitrosodimethylamine


Concen: 19.361 ng

RT: 3.399 min Scan# 205



Delta R.T. -0.012 min


Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion	Ratio	Lower	Upper
42	100		
74	114.9	89.5	134.3
44	9.6	6.6	9.8





#5
2-Fluorophenol
Concen: 40.399 ng
RT: 5.510 min Scan# 565
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:112 Resp: 492350
Ion Ratio Lower Upper
112 100
64 60.7 48.4 72.6
63 31.7 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#6
Aniline
Concen: 20.152 ng
RT: 6.569 min Scan# 744
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion: 93 Resp: 419797
Ion Ratio Lower Upper
93 100
66 39.3 32.1 48.1
65 20.1 16.6 24.8





#7
Phenol-d6
Concen: 40.580 ng
RT: 6.516 min Scan# 71
Delta R.T. -0.012 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020




Tgt Ion: 99 Resp: 59722
Ion Ratio Lower Upper
99 100
42 21.3 16.3 24.5
71 31.6 25.0 37.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#8
2-Chlorophenol
Concen: 20.156 ng
RT: 6.687 min Scan# 764
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49



Tgt Ion:128 Resp: 253840
Ion Ratio Lower Upper
128 100
130 32.8 13.1 53.1
64 50.0 28.6 68.6





#9

Benzaldehyde

Concen: 20.712 ng

RT: 6.457 min Scan# 71

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion: 77 Resp: 213052

Ion Ratio Lower Upper

77 100

105 101.1 79.1 119.1





106 97.4 76.0 116.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#10

Phenol

Concen: 20.352 ng

RT: 6.534 min Scan# 738

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion: 94 Resp: 322435


Ion Ratio Lower Upper

94 100

65 28.4 8.9 48.9

66 37.1 18.7 58.7





#11

bis(2-Chloroethyl)ether

Concen: 20.132 ng

RT: 6.640 min Scan# 757

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion: 93 Resp: 247179

Ion Ratio Lower Upper

93 100

63 83.3 62.2 102.2





95 32.3 12.7 52.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#12

1,3-Dichlorobenzene

Concen: 20.294 ng

RT: 6.851 min Scan# 792

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:146 Resp: 289129


Ion Ratio Lower Upper

146 100

148 63.2 50.5 75.7

75 28.6 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 20.268 ng

RT: 6.928 min Scan# 805

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:146 Resp: 29092

Ion Ratio Lower Upper

146 100

148 64.1 50.6 76.0





111 41.6 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#14

1,2-Dichlorobenzene

Concen: 20.331 ng

RT: 7.081 min Scan# 831

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:146 Resp: 278592


Ion Ratio Lower Upper

146 100

148 63.6 50.5 75.7

111 42.7 34.9 52.3





#15

Benzyl Alcohol

Concen: 19.835 ng

RT: 7.040 min Scan# 81

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion: 79 Resp: 20704

Ion Ratio Lower Upper

79 100

108 81.7 64.5 96.7





77 63.5 49.2 73.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 20.218 ng

RT: 7.181 min Scan# 848

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion: 45 Resp: 448250


Ion Ratio Lower Upper

45 100

77 10.9 0.0 31.3


79 8.4 0.0 28.5





#17
2-Methylphenol
Concen: 20.157 ng
RT: 7.146 min Scan# 843
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:107 Resp: 20795
Ion Ratio Lower Upper
107 100
108 114.6 91.0 136.4
77 44.2 34.4 51.6
79 42.9 34.4 51.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#18
Hexachloroethane
Concen: 19.917 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:117 Resp: 93572
Ion Ratio Lower Upper
117 100
119 97.2 77.1 115.7
201 92.2 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 20.318 ng

RT: 7.316 min Scan# 81

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion: 70 Resp: 184770

Ion Ratio Lower Upper

70 100

42 67.2 53.4 80.2

101 9.2 7.7 11.5





130 21.3 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#20

3+4-Methylphenols

Concen: 20.700 ng

RT: 7.298 min Scan# 868

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Tgt Ion:107 Resp: 267057




Ion Ratio Lower Upper


107 100

108 89.7 71.2 111.2

77 54.6 41.1 81.1

79 26.8 8.1 48.1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.193 min Scan# 1020

Delta R.T. -0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:136 Resp: 74656

Ion Ratio Lower Upper

136 100

137 10.9 8.7 13.1

54 9.1 7.7 11.5





68 6.1 5.0 7.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#22

Acetophenone

Concen: 20.872 ng

RT: 7.310 min Scan# 870

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Tgt Ion:105 Resp: 360484




Ion Ratio Lower Upper


105 100

71 8.3 5.7 8.5

51 36.4 30.1 45.1

120 23.4 19.0 28.4





#23

Nitrobenzene-d5

Concen: 39.673 ng

RT: 7.469 min Scan# 89

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion: 82 Resp: 45083

Ion Ratio Lower Upper

82 100

128 44.3 37.0 55.6





54 61.1 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#24

Nitrobenzene

Concen: 19.865 ng

RT: 7.487 min Scan# 900

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion: 77 Resp: 221499


Ion Ratio Lower Upper

77 100

123 46.9 38.6 58.0

65 14.3 11.5 17.3









#25
Isophorone
Concen: 20.088 ng
RT: 7.722 min Scan# 941
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion: 82 Resp: 479090
Ion Ratio Lower Upper
82 100
95 6.9 5.6 8.4
138 19.4 15.3 22.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#26
2-Nitrophenol
Concen: 17.893 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:139 Resp: 61615
Ion Ratio Lower Upper
139 100
109 26.4 21.8 32.8
65 49.6 41.1 61.7





#27
2,4-Dimethylphenol
Concen: 20.696 ng
RT: 7.834 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:122 Resp: 248864
Ion Ratio Lower Upper
122 100
107 104.1 86.0 129.0
121 56.5 46.3 69.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#28
bis(2-Chloroethoxy)methane
Concen: 20.549 ng
RT: 7.934 min Scan# 976
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion: 93 Resp: 309073
Ion Ratio Lower Upper
93 100
95 32.6 26.2 39.2
123 11.5 9.7 14.5





#29

2,4-Dichlorophenol

Concen: 20.374 ng

RT: 8.040 min Scan# 994

Delta R.T. 0.000 min

Lab File: BF142242.D


Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020



Tgt Ion:162 Resp: 20482

Ion Ratio Lower Upper

162 100

164 64.3 43.0 83.0




98 36.1 16.7 56.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#30

1,2,4-Trichlorobenzene

Concen: 20.315 ng

RT: 8.128 min Scan# 1009

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:180 Resp: 227898


Ion Ratio Lower Upper

180 100

182 94.7 75.7 113.5

145 31.7 25.4 38.0





#31

Naphthalene

Concen: 20.758 ng

RT: 8.210 min Scan# 1023

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:128 Resp: 772839

Ion Ratio Lower Upper

128 100

129 11.2 9.0 13.4





127 13.1 10.6 15.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#32

Benzoic acid

Concen: 19.412 ng m

RT: 7.904 min Scan# 971

Delta R.T. -0.030 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:122 Resp: 64656


Ion Ratio Lower Upper

122 100

105 117.1 98.2 138.2


77 84.0 66.8 106.8





#33
4-Chloroaniline
Concen: 20.654 ng
RT: 8.257 min Scan# 1031
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020




Tgt Ion:127 Resp: 320218
Ion Ratio Lower Upper
127 100
129 32.2 26.1 39.1
65 30.7 25.5 38.3
92 18.9 15.4 23.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#34
Hexachlorobutadiene
Concen: 20.393 ng
RT: 8.334 min Scan# 1044
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Tgt Ion:225 Resp: 133474
Ion Ratio Lower Upper
225 100
223 62.8 49.7 74.5
227 63.4 51.0 76.4





#35
Caprolactam
Concen: 20.118 ng
RT: 8.604 min Scan# 1090
Delta R.T. -0.018 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:113 Resp: 6079
Ion Ratio Lower Upper
113 100
55 198.2 177.7 217.7
56 144.8 127.1 167.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#36
4-Chloro-3-methylphenol
Concen: 20.190 ng
RT: 8.722 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:107 Resp: 208419
Ion Ratio Lower Upper
107 100
144 26.4 21.0 31.4
142 83.4 66.1 99.1





#37

2-Methylnaphthalene

Concen: 20.658 ng

RT: 8.904 min Scan# 1141

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:142 Resp: 472890

Ion Ratio Lower Upper

142 100





141 87.8 70.5 105.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#38

1-Methylnaphthalene

Concen: 20.985 ng

RT: 9.004 min Scan# 1158

Delta R.T. 0.000 min

Lab File: BF142242.D




Acq: 30 Apr 2025 12:49


Tgt Ion:142 Resp: 497635

Ion Ratio Lower Upper

142 100

141 90.1 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.945 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:164 Resp: 39880

Ion Ratio Lower Upper

164 100

162 98.3 78.6 117.8





160 44.1 35.7 53.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#40

1,2,4,5-Tetrachlorobenzene

Concen: 20.342 ng

RT: 9.069 min Scan# 1169

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Tgt Ion:216 Resp: 228311




Ion Ratio Lower Upper


216 100

214 78.9 63.8 95.6

179 20.5 16.7 25.1


108 19.4 15.6 23.4





#41
Hexachlorocyclopentadiene
Concen: 19.145 ng
RT: 9.057 min Scan# 1167
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020




Tgt Ion: 237 Resp: 12195
Ion Ratio Lower Upper
237 100
235 62.4 43.5 83.5
272 12.8 0.0 32.9

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#42
2,4,6-Tribromophenol
Concen: 40.372 ng
RT: 10.734 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Tgt Ion: 330 Resp: 146779
Ion Ratio Lower Upper
330 100
332 96.7 76.5 114.7
141 32.8 24.3 36.5





#43
2,4,6-Trichlorophenol
Concen: 20.246 ng
RT: 9.175 min Scan# 1187
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020




Tgt Ion:196 Resp: 14080
Ion Ratio Lower Upper
196 100
198 96.4 77.4 116.2
200 30.7 24.6 37.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#44
2,4,5-Trichlorophenol
Concen: 19.756 ng
RT: 9.210 min Scan# 1193
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Tgt Ion:196 Resp: 144462
Ion Ratio Lower Upper
196 100
198 96.1 75.2 112.8
97 49.7 39.8 59.8
132 29.1 23.8 35.8





#45
2-Fluorobiphenyl
Concen: 42.876 ng
RT: 9.263 min Scan# 1202
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:172 Resp: 121274
Ion Ratio Lower Upper
172 100
171 34.7 29.0 43.4
170 23.3 19.4 29.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#46
1,1'-Biphenyl
Concen: 20.740 ng
RT: 9.369 min Scan# 1220
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:154 Resp: 643017
Ion Ratio Lower Upper
154 100
153 39.9 20.5 60.5
76 12.0 0.0 32.8





#47

2-Chloronaphthalene

Concen: 20.588 ng

RT: 9.392 min Scan# 1224

Delta R.T. 0.000 min

Lab File: BF142242.D


Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020






Tgt Ion:	162	Resp:	47489
Ion	Ratio	Lower	Upper
162	100		
127	39.1	32.6	49.0
164	32.2	26.1	39.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#48

2-Nitroaniline


Concen: 18.213 ng

RT: 9.481 min Scan# 1239



Delta R.T. -0.006 min


Lab File: BF142242.D

Acq: 30 Apr 2025 12:49



Tgt Ion:	65	Resp:	90798
Ion	Ratio	Lower	Upper
65	100		
92	66.4	54.3	81.5
138	107.2	88.6	133.0





#49

Acenaphthylene

Concen: 20.749 ng

RT: 9.810 min Scan# 1295

Delta R.T. 0.000 min

Lab File: BF142242.D


Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020



Tgt Ion:152 Resp: 807654

Ion Ratio Lower Upper

152 100

151 19.6 16.2 24.2




153 13.0 10.7 16.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#50

Dimethylphthalate

Concen: 20.334 ng

RT: 9.663 min Scan# 1270

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:163 Resp: 518802


Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8

164 10.0 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 18.432 ng
RT: 9.722 min Scan# 1281
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 7695
Ion Ratio Lower Upper
165 100
63 66.2 48.5 72.7
89 59.6 43.9 65.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#52
Acenaphthene
Concen: 20.281 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:154 Resp: 460927
Ion Ratio Lower Upper
154 100
153 110.6 88.6 133.0
152 52.4 42.1 63.1





#53
3-Nitroaniline
Concen: 18.959 ng
RT: 9.892 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:138 Resp: 100560
Ion Ratio Lower Upper
138 100
108 19.4 11.7 17.5
92 110.1 91.2 136.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#54
2,4-Dinitrophenol
Concen: 18.804 ng
RT: 9.992 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:184 Resp: 14553
Ion Ratio Lower Upper
184 100
63 131.7 69.2 103.8#
154 869.6 271.4 407.2#





#55

Dibenzenofuran

Concen: 20.493 ng

RT: 10.151 min Scan# 1353

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:168 Resp: 69815

Ion Ratio Lower Upper

168 100

139 36.8 29.4 44.0





169 13.3 11.0 16.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#56

4-Nitrophenol

Concen: 17.561 ng

RT: 10.034 min Scan# 1333

Delta R.T. -0.005 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:139 Resp: 70423


Ion Ratio Lower Upper

139 100

109 71.2 44.9 84.9

65 94.8 74.7 114.7









#57
2,4-Dinitrotoluene
Concen: 17.844 ng
RT: 10.122 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:165 Resp: 8481
Ion Ratio Lower Upper
165 100
63 50.7 38.8 58.2
89 76.2 57.1 85.7
182 2.8 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#58
Fluorene
Concen: 20.793 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:166 Resp: 538647
Ion Ratio Lower Upper
166 100
165 93.4 73.9 110.9
167 13.6 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 20.530 ng
RT: 10.263 min Scan# 1372
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020




Tgt Ion:232 Resp: 12435
Ion Ratio Lower Upper
232 100
131 45.0 35.9 53.9
130 2.1 1.8 2.6
166 28.8 24.0 36.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#60
Diethylphthalate
Concen: 20.611 ng
RT: 10.363 min Scan# 1389
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Tgt Ion:149 Resp: 519294
Ion Ratio Lower Upper
149 100
177 21.3 17.6 26.4
150 12.0 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 20.502 ng
RT: 10.486 min Scan# 1410
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020




Tgt Ion:204 Resp: 25221
Ion Ratio Lower Upper
204 100
206 33.0 25.6 38.4
141 60.1 48.9 73.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#62
4-Nitroaniline
Concen: 18.933 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Tgt Ion:138 Resp: 92884
Ion Ratio Lower Upper
138 100
92 49.5 29.2 69.2
108 73.2 49.1 89.1





#63
Azobenzene
Concen: 20.438 ng
RT: 10.645 min Scan# 1437
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion: 77 Resp: 50848
Ion Ratio Lower Upper
77 100
182 25.1 7.3 47.3
105 22.2 2.5 42.5
51 34.2 14.4 54.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.434 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:188 Resp: 679379
Ion Ratio Lower Upper
188 100
94 9.8 8.2 12.2
80 10.5 8.6 12.8









#65
4,6-Dinitro-2-methylphenol
Concen: 18.756 ng
RT: 10.528 min Scan# 1418
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:198 Resp: 23400
Ion Ratio Lower Upper
198 100
51 62.5 40.6 80.6
105 50.4 29.2 69.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#66
n-Nitrosodiphenylamine
Concen: 20.280 ng
RT: 10.604 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:169 Resp: 470767
Ion Ratio Lower Upper
169 100
168 66.2 53.0 79.6
167 36.4 28.8 43.2





#67
4-Bromophenyl-phenylether
Concen: 20.053 ng
RT: 10.981 min Scan# 1494
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020




Tgt Ion:248 Resp: 151720
Ion Ratio Lower Upper
248 100
250 98.1 77.2 115.8
141 72.5 58.4 87.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#68
Hexachlorobenzene
Concen: 19.916 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49



Tgt Ion:284 Resp: 168767
Ion Ratio Lower Upper
284 100
142 33.8 24.6 37.0
249 29.3 23.2 34.8





#69

Atrazine

Concen: 20.665 ng

RT: 11.128 min Scan# 1519

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:200 Resp: 12644

Ion Ratio Lower Upper

200 100

173 27.9 8.6 48.6





215 49.1 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#70

Pentachlorophenol

Concen: 18.301 ng

RT: 11.228 min Scan# 1536

Delta R.T. -0.005 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:266 Resp: 81929


Ion Ratio Lower Upper

266 100

268 63.6 50.0 75.0

264 62.8 49.3 73.9









#71
Phenanthrene
Concen: 20.723 ng
RT: 11.457 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Instrument :
BNA_F
ClientSampleId :
SSTDICC020

Tgt Ion:178 Resp: 760470
Ion Ratio Lower Upper
178 100
176 19.0 15.5 23.3
179 15.3 12.6 19.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#72
Anthracene
Concen: 20.848 ng
RT: 11.510 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:178 Resp: 782215
Ion Ratio Lower Upper
178 100
176 18.4 15.3 22.9
179 15.5 12.7 19.1





#73
Carbazole
Concen: 21.059 ng
RT: 11.657 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020




Tgt Ion:167 Resp: 720775
Ion Ratio Lower Upper
167 100
166 21.2 17.3 25.9
139 12.7 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 21.469 ng
RT: 11.998 min Scan# 1667
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Tgt Ion:149 Resp: 739011
Ion Ratio Lower Upper
149 100
150 8.9 7.5 11.3
104 4.3 3.6 5.4





#75
Fluoranthene
Concen: 21.636 ng
RT: 12.645 min Scan# 1777
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:202 Resp: 784880
Ion Ratio Lower Upper
202 100
101 12.1 0.0 32.4
203 17.6 0.0 37.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.075 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49


Tgt Ion:240 Resp: 430351
Ion Ratio Lower Upper
240 100
120 12.4 10.1 15.1
236 25.6 20.5 30.7





#77
Benzidine
Concen: 20.837 ng
RT: 12.769 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:184 Resp: 381034
Ion Ratio Lower Upper
184 100
185 14.2 11.8 17.6
183 11.1 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#78
Pyrene
Concen: 20.046 ng
RT: 12.875 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:202 Resp: 795806
Ion Ratio Lower Upper
202 100
200 20.1 16.1 24.1
203 17.9 14.1 21.1





#79

Terphenyl-d14

Concen: 41.357 ng

RT: 13.022 min Scan# 1841

Delta R.T. 0.000 min

Lab File: BF142242.D


Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020



Tgt Ion:244 Resp: 122373

Ion Ratio Lower Upper

244 100

212 6.3 5.3 7.9




122 11.1 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#80

Butylbenzylphthalate

Concen: 18.529 ng

RT: 13.498 min Scan# 1922

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:149 Resp: 174268


Ion Ratio Lower Upper

149 100

91 67.9 54.4 81.6


206 20.8 15.9 23.9





#81
Benzo(a)anthracene
Concen: 19.849 ng
RT: 14.063 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020




Tgt Ion:228 Resp: 573438
Ion Ratio Lower Upper
228 100
226 26.7 21.9 32.9
229 19.7 16.0 24.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#82
3,3'-Dichlorobenzidine
Concen: 19.251 ng
RT: 14.027 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49



Tgt Ion:252 Resp: 163359
Ion Ratio Lower Upper
252 100
254 63.8 51.0 76.6
126 11.8 9.4 14.2





#83

Chrysene

Concen: 20.435 ng

RT: 14.098 min Scan# 2024

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:228 Resp: 537388

Ion Ratio Lower Upper

228 100

226 30.1 24.0 36.0





229 19.6 15.7 23.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#84

Bis(2-ethylhexyl)phthalate

Concen: 18.042 ng

RT: 14.057 min Scan# 2017

Delta R.T. -0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:149 Resp: 236325


Ion Ratio Lower Upper

149 100

167 27.1 21.7 32.5


279 3.8 3.0 4.4





#85
Di-n-octyl phthalate
Concen: 17.998 ng
RT: 14.674 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:149 Resp: 376362
Ion Ratio Lower Upper
149 100
167 1.5 1.1 1.7
43 12.4 9.3 13.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.569 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:264 Resp: 350840
Ion Ratio Lower Upper
264 100
260 23.9 18.4 27.6
265 21.8 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 20.021 ng

RT: 17.074 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:276 Resp: 514490

Ion Ratio Lower Upper

276 100

138 28.9 22.6 33.8





277 25.1 20.1 30.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#88

Benzo(b)fluoranthene

Concen: 19.442 ng

RT: 15.127 min Scan# 2199

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:252 Resp: 423694


Ion Ratio Lower Upper

252 100

253 21.3 17.7 26.5

125 10.8 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 21.504 ng

RT: 15.157 min Scan# 2204

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49

Instrument :

BNA_F

ClientSampleId :

SSTDICC020

Tgt Ion:252 Resp: 428640

Ion Ratio Lower Upper

252 100

253 21.5 17.8 26.8





125 10.5 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#90

Benzo(a)pyrene

Concen: 20.154 ng

RT: 15.504 min Scan# 2263

Delta R.T. 0.000 min

Lab File: BF142242.D

Acq: 30 Apr 2025 12:49




Tgt Ion:252 Resp: 402101


Ion Ratio Lower Upper

252 100

253 21.5 17.6 26.4


125 11.7 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 20.139 ng
RT: 17.092 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49




Instrument :
BNA_F
ClientSampleId :
SSTDICC020



Tgt Ion:278 Resp: 420029
Ion Ratio Lower Upper
278 100
139 18.2 15.2 22.8
279 23.0 19.0 28.6




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 20.150 ng
RT: 17.527 min Scan# 2607
Delta R.T. -0.012 min
Lab File: BF142242.D
Acq: 30 Apr 2025 12:49

Tgt Ion:276 Resp: 422497
Ion Ratio Lower Upper
276 100
277 23.3 18.7 28.1
138 24.4 19.4 29.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142243.D
 Acq On : 30 Apr 2025 13:17
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICCC040

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:43:02 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	180007	20.000	ng	0.00
21) Naphthalene-d8	8.193	136	706244	20.000	ng	0.00
39) Acenaphthene-d10	9.951	164	370869	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	616408	20.000	ng	0.00
76) Chrysene-d12	14.074	240	331192	20.000	ng	0.00
86) Perylene-d12	15.568	264	328405	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	946150	83.944	ng	0.00
7) Phenol-d6	6.528	99	1146926	84.267	ng	0.00
23) Nitrobenzene-d5	7.475	82	961009	89.396	ng	0.00
42) 2,4,6-Tribromophenol	10.734	330	302223	89.389	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2089707	79.446	ng	0.00
79) Terphenyl-d14	13.022	244	1962470	86.181	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.699	88	213029	42.319	ng	100
3) Pyridine	3.463	79	547989	41.886	ng	100
4) n-Nitrosodimethylamine	3.410	42	290357	42.887	ng	100
6) Aniline	6.569	93	808338	41.959	ng	100
8) 2-Chlorophenol	6.693	128	496290	42.612	ng	100
9) Benzaldehyde	6.457	77	402047	42.263	ng	100
10) Phenol	6.540	94	617257	42.128	ng	100
11) bis(2-Chloroethyl)ether	6.646	93	474404	41.780	ng	100
12) 1,3-Dichlorobenzene	6.851	146	551813	41.881	ng	100
13) 1,4-Dichlorobenzene	6.928	146	556691	41.937	ng	100
14) 1,2-Dichlorobenzene	7.081	146	530895	41.892	ng	100
15) Benzyl Alcohol	7.046	79	413830	42.870	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.187	45	858875	41.888	ng	100
17) 2-Methylphenol	7.151	107	403467	42.287	ng	100
18) Hexachloroethane	7.428	117	183881	42.322	ng	100
19) n-Nitroso-di-n-propyla...	7.322	70	348288	41.410	ng	100
20) 3+4-Methylphenols	7.304	107	503856	42.229	ng	100
22) Acetophenone	7.316	105	660959	40.454	ng	100
24) Nitrobenzene	7.493	77	463357	43.929	ng	100
25) Isophorone	7.728	82	931763	41.299	ng	100
26) 2-Nitrophenol	7.804	139	162563	38.737	ng	100
27) 2,4-Dimethylphenol	7.834	122	476186	41.862	ng	100
28) bis(2-Chloroethoxy)met...	7.940	93	586887	41.247	ng	100
29) 2,4-Dichlorophenol	8.040	162	410487	43.163	ng	100
30) 1,2,4-Trichlorobenzene	8.134	180	440825	41.539	ng	100
31) Naphthalene	8.216	128	1441240	40.920	ng	100
32) Benzoic acid	7.934	122	184052m	37.083	ng	
33) 4-Chloroaniline	8.257	127	606136	41.328	ng	100
34) Hexachlorobutadiene	8.334	225	260821	42.125	ng	100
35) Caprolactam	8.622	113	122598	42.889	ng	100
36) 4-Chloro-3-methylphenol	8.728	107	417126	42.715	ng	100
37) 2-Methylnaphthalene	8.904	142	880315	40.652	ng	100
38) 1-Methylnaphthalene	9.004	142	920983	41.054	ng	100
40) 1,2,4,5-Tetrachloroben...	9.069	216	427286	40.938	ng	100
41) Hexachlorocyclopentadiene	9.057	237	260979	44.056	ng	100
43) 2,4,6-Trichlorophenol	9.175	196	274555	42.450	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142243.D
 Acq On : 30 Apr 2025 13:17
 Operator : RC/JU
 Sample : SSTDICCC040
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICCC040

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:43:02 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	299967	44.113	ng	100
46) 1,1'-Biphenyl	9.369	154	1172400	40.664	ng	100
47) 2-Chloronaphthalene	9.392	162	879152	40.984	ng	100
48) 2-Nitroaniline	9.487	65	218766	39.600	ng	100
49) Acenaphthylene	9.810	152	1481543	40.929	ng	100
50) Dimethylphthalate	9.669	163	980172	41.312	ng	100
51) 2,6-Dinitrotoluene	9.728	165	178867	40.034	ng	100
52) Acenaphthene	9.981	154	869604	41.145	ng	100
53) 3-Nitroaniline	9.892	138	221168	40.485	ng	100
54) 2,4-Dinitrophenol	9.998	184	42260	38.466	ng	100
55) Dibenzofuran	10.151	168	1293035	40.813	ng	100
56) 4-Nitrophenol	10.039	139	159989	42.900	ng	100
57) 2,4-Dinitrotoluene	10.128	165	211218	39.352	ng	100
58) Fluorene	10.498	166	976948	40.554	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.269	232	247655	43.967	ng	100
60) Diethylphthalate	10.369	149	974036	41.573	ng	100
61) 4-Chlorophenyl-phenyle...	10.486	204	468898	40.986	ng	100
62) 4-Nitroaniline	10.504	138	201491	40.498	ng	100
63) Azobenzene	10.651	77	948440	40.994	ng	100
65) 4,6-Dinitro-2-methylph...	10.534	198	67678	36.574	ng	100
66) n-Nitrosodiphenylamine	10.604	169	877338	41.656	ng	100
67) 4-Bromophenyl-phenylether	10.981	248	291063	42.400	ng	100
68) Hexachlorobenzene	11.045	284	321448	41.809	ng	100
69) Atrazine	11.128	200	234825	42.299	ng	100
70) Pentachlorophenol	11.233	266	177755	43.764	ng	100
71) Phenanthrene	11.457	178	1371009	41.176	ng	100
72) Anthracene	11.510	178	1390308	40.840	ng	100
73) Carbazole	11.663	167	1273184	40.999	ng	100
74) Di-n-butylphthalate	11.998	149	1317363	42.180	ng	100
75) Fluoranthene	12.645	202	1338006	40.652	ng	100
77) Benzidine	12.769	184	638142	45.346	ng	100
78) Pyrene	12.880	202	1341800	43.920	ng	100
80) Butylbenzylphthalate	13.498	149	334175	40.624	ng	100
81) Benzo(a)anthracene	14.063	228	952290	42.831	ng	100
82) 3,3'-Dichlorobenzidine	14.027	252	292958	44.859	ng	100
83) Chrysene	14.104	228	816041	40.321	ng	100
84) Bis(2-ethylhexyl)phtha...	14.057	149	460074	40.356	ng	100
85) Di-n-octyl phthalate	14.680	149	836078	40.504	ng	100
87) Indeno(1,2,3-cd)pyrene	17.080	276	1059591	44.049	ng	100
88) Benzo(b)fluoranthene	15.127	252	854971	41.912	ng	100
89) Benzo(k)fluoranthene	15.157	252	752135	40.310	ng	100
90) Benzo(a)pyrene	15.504	252	790330	42.318	ng	100
91) Dibenzo(a,h)anthracene	17.098	278	861783	44.142	ng	100
92) Benzo(g,h,i)perylene	17.539	276	858804	43.758	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Manual Integrations


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:152 Resp: 180000
Ion Ratio Lower Upper
152 100
150 162.7 130.2 195.2
115 58.7 47.0 70.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#2
1,4-Dioxane
Concen: 42.319 ng
RT: 2.699 min Scan# 86
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 88 Resp: 213029
Ion Ratio Lower Upper
88 100
58 90.2 72.2 108.2
43 33.8 27.0 40.6





#3

Pyridine

Concen: 41.886 ng

RT: 3.463 min Scan# 216

Delta R.T. 0.000 min

Lab File: BF142243.D


Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040






Tgt Ion	Ratio	Lower	Upper
79	100		
52	89.4	71.5	107.3
51	43.1	34.5	51.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#4

n-Nitrosodimethylamine


Concen: 42.887 ng

RT: 3.410 min Scan# 207



Delta R.T. 0.000 min


Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion	Ratio	Lower	Upper
42	100		
74	111.9	89.5	134.3
44	8.2	6.6	9.8





#5
2-Fluorophenol
Concen: 83.944 ng
RT: 5.516 min Scan# 565
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion: 112 Resp: 946150
Ion Ratio Lower Upper
112 100
64 60.5 48.4 72.6
63 31.0 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#6
Aniline
Concen: 41.959 ng
RT: 6.569 min Scan# 744
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion: 93 Resp: 808338
Ion Ratio Lower Upper
93 100
66 40.1 32.1 48.1
65 20.7 16.6 24.8





#7

Phenol-d6

Concen: 84.267 ng

RT: 6.528 min Scan# 737

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion: 99 Resp: 1146920

Ion Ratio Lower Upper

99 100

42 20.4 16.3 24.5





71 31.2 25.0 37.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#8

2-Chlorophenol

Concen: 42.612 ng

RT: 6.693 min Scan# 765

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion:128 Resp: 496290


Ion Ratio Lower Upper

128 100

130 33.1 13.1 53.1

64 48.6 28.6 68.6





#9

Benzaldehyde

Concen: 42.263 ng

RT: 6.457 min Scan# 725

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion: 77 Resp: 40204

Ion Ratio Lower Upper

77 100

105 99.1 79.1 119.1





106 96.0 76.0 116.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#10

Phenol

Concen: 42.128 ng

RT: 6.540 min Scan# 739

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion: 94 Resp: 617257


Ion Ratio Lower Upper

94 100

65 28.9 8.9 48.9

66 38.7 18.7 58.7





#11

bis(2-Chloroethyl)ether

Concen: 41.780 ng

RT: 6.646 min Scan# 757

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion: 93 Resp: 474404

Ion Ratio Lower Upper

93 100

63 82.2 62.2 102.2





95 32.7 12.7 52.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#12

1,3-Dichlorobenzene

Concen: 41.881 ng

RT: 6.851 min Scan# 792

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion:146 Resp: 551813


Ion Ratio Lower Upper

146 100

148 63.1 50.5 75.7

75 29.0 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 41.937 ng

RT: 6.928 min Scan# 805

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion:146 Resp: 55669

Ion Ratio Lower Upper

146 100

148 63.3 50.6 76.0





111 41.7 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#14

1,2-Dichlorobenzene

Concen: 41.892 ng

RT: 7.081 min Scan# 831

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion:146 Resp: 530895


Ion Ratio Lower Upper

146 100

148 63.1 50.5 75.7

111 43.6 34.9 52.3





#15

Benzyl Alcohol

Concen: 42.870 ng

RT: 7.046 min Scan# 81

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion: 79 Resp: 413830

Ion Ratio Lower Upper

79 100

108 80.6 64.5 96.7





77 61.5 49.2 73.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 41.888 ng

RT: 7.187 min Scan# 849

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion: 45 Resp: 858875


Ion Ratio Lower Upper

45 100

77 11.3 0.0 31.3


79 8.5 0.0 28.5





#17
2-Methylphenol
Concen: 42.287 ng
RT: 7.151 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:107 Resp: 40346
Ion Ratio Lower Upper
107 100
108 113.7 91.0 136.4
77 43.0 34.4 51.6
79 43.0 34.4 51.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#18
Hexachloroethane
Concen: 42.322 ng
RT: 7.428 min Scan# 890
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:117 Resp: 183881
Ion Ratio Lower Upper
117 100
119 96.4 77.1 115.7
201 99.3 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 41.410 ng

RT: 7.322 min Scan# 872

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion: 70 Resp: 34828

Ion Ratio Lower Upper

70 100

42 66.8 53.4 80.2

101 9.6 7.7 11.5





130 21.5 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#20

3+4-Methylphenols

Concen: 42.229 ng

RT: 7.304 min Scan# 869

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Tgt Ion:107 Resp: 503856




Ion Ratio Lower Upper


107 100

108 91.2 71.2 111.2

77 61.1 41.1 81.1

79 28.1 8.1 48.1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.193 min Scan# 1020

Delta R.T. -0.000 min

Lab File: BF142243.D


Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040



Tgt Ion:136 Resp: 706244

Ion Ratio Lower Upper

136 100

137 10.9 8.7 13.1

54 9.6 7.7 11.5




68 6.3 5.0 7.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#22

Acetophenone


Concen: 40.454 ng

RT: 7.316 min Scan# 871

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17



Tgt Ion:105 Resp: 660959



Ion Ratio Lower Upper


105 100

71 7.1 5.7 8.5

51 37.6 30.1 45.1

120 23.7 19.0 28.4





#23

Nitrobenzene-d5

Concen: 89.396 ng

RT: 7.475 min Scan# 89

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion: 82 Resp: 961009

Ion Ratio Lower Upper

82 100

128 46.3 37.0 55.6





54 60.0 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#24

Nitrobenzene

Concen: 43.929 ng

RT: 7.493 min Scan# 901

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion: 77 Resp: 463357


Ion Ratio Lower Upper

77 100

123 48.3 38.6 58.0

65 14.4 11.5 17.3









#25
Isophorone
Concen: 41.299 ng
RT: 7.728 min Scan# 941
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion: 82 Resp: 931763
Ion Ratio Lower Upper
82 100
95 7.0 5.6 8.4
138 19.1 15.3 22.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#26
2-Nitrophenol
Concen: 38.737 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:139 Resp: 162563
Ion Ratio Lower Upper
139 100
109 27.3 21.8 32.8
65 51.4 41.1 61.7





#27

2,4-Dimethylphenol

Concen: 41.862 ng

RT: 7.834 min Scan# 91

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion:122 Resp: 476180

Ion Ratio Lower Upper

122 100

107 107.5 86.0 129.0





121 57.9 46.3 69.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#28

bis(2-Chloroethoxy)methane

Concen: 41.247 ng

RT: 7.940 min Scan# 977

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion: 93 Resp: 586887


Ion Ratio Lower Upper

93 100

95 32.7 26.2 39.2

123 12.1 9.7 14.5





#29

2,4-Dichlorophenol

Concen: 43.163 ng

RT: 8.040 min Scan# 994

Delta R.T. 0.000 min

Lab File: BF142243.D


Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040






Tgt Ion:162	Resp: 41048
Ion Ratio	Lower Upper
162	100
164	63.0 43.0 83.0
98	36.7 16.7 56.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#30

1,2,4-Trichlorobenzene


Concen: 41.539 ng

RT: 8.134 min Scan# 1010



Delta R.T. 0.000 min


Lab File: BF142243.D

Acq: 30 Apr 2025 13:17



Tgt Ion:180	Resp: 440825
Ion Ratio	Lower Upper
180	100
182	94.6 75.7 113.5
145	31.7 25.4 38.0





#31

Naphthalene

Concen: 40.920 ng

RT: 8.216 min Scan# 1024

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion:128 Resp: 1441240

Ion Ratio Lower Upper

128 100

129 11.2 9.0 13.4





127 13.2 10.6 15.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#32

Benzoic acid

Concen: 37.083 ng m

RT: 7.934 min Scan# 976

Delta R.T. -0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion:122 Resp: 184052


Ion Ratio Lower Upper

122 100

105 118.2 98.2 138.2

77 86.8 66.8 106.8









#33
4-Chloroaniline
Concen: 41.328 ng
RT: 8.257 min Scan# 1031
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:127 Resp: 606130
Ion Ratio Lower Upper
127 100
129 32.6 26.1 39.1
65 31.9 25.5 38.3
92 19.3 15.4 23.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#34
Hexachlorobutadiene
Concen: 42.125 ng
RT: 8.334 min Scan# 1044
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:225 Resp: 260821
Ion Ratio Lower Upper
225 100
223 62.1 49.7 74.5
227 63.7 51.0 76.4





#35
Caprolactam
Concen: 42.889 ng
RT: 8.622 min Scan# 1093
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:113 Resp: 122598
Ion Ratio Lower Upper
113 100
55 197.7 177.7 217.7
56 147.1 127.1 167.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#36
4-Chloro-3-methylphenol
Concen: 42.715 ng
RT: 8.728 min Scan# 1111
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:107 Resp: 417126
Ion Ratio Lower Upper
107 100
144 26.2 21.0 31.4
142 82.6 66.1 99.1





#37
2-Methylnaphthalene
Concen: 40.652 ng
RT: 8.904 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:142 Resp: 880315
Ion Ratio Lower Upper
142 100
141 88.1 70.5 105.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#38
1-Methylnaphthalene
Concen: 41.054 ng
RT: 9.004 min Scan# 1158
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:142 Resp: 920983
Ion Ratio Lower Upper
142 100
141 91.0 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.951 min Scan# 1169

Delta R.T. 0.000 min

Lab File: BF142243.D


Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040



Tgt Ion:164 Resp: 370869

Ion Ratio Lower Upper

164 100

162 98.2 78.6 117.8




160 44.6 35.7 53.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#40

1,2,4,5-Tetrachlorobenzene

Concen: 40.938 ng

RT: 9.069 min Scan# 1169

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Tgt Ion:216 Resp: 427286




Ion Ratio Lower Upper


216 100

214 79.7 63.8 95.6

179 20.9 16.7 25.1

108 19.5 15.6 23.4





#41

Hexachlorocyclopentadiene

Concen: 44.056 ng

RT: 9.057 min Scan# 1167

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :





SSTDICCC040

Tgt Ion:	237	Resp:	260979
Ion Ratio	Lower	Upper	
237	100		
235	63.5	43.5	83.5
272	12.9	0.0	32.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#42

2,4,6-Tribromophenol

Concen: 89.389 ng




RT: 10.734 min Scan# 1452


Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Tgt Ion:	330	Resp:	302223
Ion Ratio	Lower	Upper	
330	100		
332	95.6	76.5	114.7
141	30.4	24.3	36.5





#43

2,4,6-Trichlorophenol

Concen: 42.450 ng

RT: 9.175 min Scan# 1187

Delta R.T. 0.000 min

Lab File: BF142243.D


Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040



Tgt Ion:196 Resp: 27455

Ion Ratio Lower Upper

196 100

198 96.8 77.4 116.2



200 30.8 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#44

2,4,5-Trichlorophenol


Concen: 44.113 ng

RT: 9.210 min Scan# 1193

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17



Tgt Ion:196 Resp: 299967



Ion Ratio Lower Upper


196 100

198 94.0 75.2 112.8

97 49.8 39.8 59.8


132 29.8 23.8 35.8





#45
2-Fluorobiphenyl
Concen: 79.446 ng
RT: 9.269 min Scan# 1203
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:172 Resp: 208970
Ion Ratio Lower Upper
172 100
171 36.2 29.0 43.4
170 24.2 19.4 29.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#46
1,1'-Biphenyl
Concen: 40.664 ng
RT: 9.369 min Scan# 1220
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:154 Resp: 1172400
Ion Ratio Lower Upper
154 100
153 40.5 20.5 60.5
76 12.8 0.0 32.8





#47

2-Chloronaphthalene

Concen: 40.984 ng

RT: 9.392 min Scan# 1224

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion:162 Resp: 879152

Ion Ratio Lower Upper

162 100

127 40.8 32.6 49.0





164 32.6 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#48

2-Nitroaniline

Concen: 39.600 ng

RT: 9.487 min Scan# 1240

Delta R.T. -0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion: 65 Resp: 218766


Ion Ratio Lower Upper

65 100

92 67.9 54.3 81.5

138 110.8 88.6 133.0





#49

Acenaphthylene

Concen: 40.929 ng

RT: 9.810 min Scan# 1295

Delta R.T. 0.000 min

Lab File: BF142243.D


Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040



Tgt Ion:152 Resp: 148154

Ion Ratio Lower Upper

152 100

151 20.2 16.2 24.2




153 13.4 10.7 16.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#50

Dimethylphthalate


Concen: 41.312 ng

RT: 9.669 min Scan# 1271

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17





Tgt Ion:163 Resp: 980172


Ion Ratio Lower Upper

163 100

194 4.0 3.2 4.8


164 10.2 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 40.034 ng
RT: 9.728 min Scan# 1281
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:165 Resp: 17886
Ion Ratio Lower Upper
165 100
63 60.6 48.5 72.7
89 54.9 43.9 65.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#52
Acenaphthene
Concen: 41.145 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:154 Resp: 869604
Ion Ratio Lower Upper
154 100
153 110.8 88.6 133.0
152 52.6 42.1 63.1





#53
3-Nitroaniline
Concen: 40.485 ng
RT: 9.892 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:138 Resp: 221168
Ion Ratio Lower Upper
138 100
108 14.6 11.7 17.5
92 114.0 91.2 136.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#54
2,4-Dinitrophenol
Concen: 38.466 ng
RT: 9.998 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:184 Resp: 42260
Ion Ratio Lower Upper
184 100
63 86.5 69.2 103.8
154 339.3 271.4 407.2





#55

Dibenzenofuran

Concen: 40.813 ng

RT: 10.151 min Scan# 1353

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion:168 Resp: 1293035

Ion Ratio Lower Upper

168 100

139 36.7 29.4 44.0





169 13.7 11.0 16.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#56

4-Nitrophenol

Concen: 42.900 ng

RT: 10.039 min Scan# 1334

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion:139 Resp: 159989


Ion Ratio Lower Upper

139 100

109 64.9 44.9 84.9

65 94.7 74.7 114.7









#57
2,4-Dinitrotoluene
Concen: 39.352 ng
RT: 10.128 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:165 Resp: 211218
Ion Ratio Lower Upper
165 100
63 48.5 38.8 58.2
89 71.4 57.1 85.7
182 2.7 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#58
Fluorene
Concen: 40.554 ng
RT: 10.498 min Scan# 1412
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:166 Resp: 976948
Ion Ratio Lower Upper
166 100
165 92.4 73.9 110.9
167 13.5 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 43.967 ng
RT: 10.269 min Scan# 1373
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:232 Resp: 247651
Ion Ratio Lower Upper
232 100
131 44.9 35.9 53.9
130 2.2 1.8 2.6
166 30.0 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#60
Diethylphthalate
Concen: 41.573 ng
RT: 10.369 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:149 Resp: 974036
Ion Ratio Lower Upper
149 100
177 22.0 17.6 26.4
150 12.7 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 40.986 ng
RT: 10.486 min Scan# 1410
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040




Tgt Ion:204 Resp: 468898
Ion Ratio Lower Upper
204 100
206 32.0 25.6 38.4
141 61.1 48.9 73.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#62
4-Nitroaniline
Concen: 40.498 ng
RT: 10.504 min Scan# 1413
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Tgt Ion:138 Resp: 201491
Ion Ratio Lower Upper
138 100
92 49.2 29.2 69.2
108 69.1 49.1 89.1





#63
Azobenzene
Concen: 40.994 ng
RT: 10.651 min Scan# 1438
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion: 77 Resp: 948440
Ion Ratio Lower Upper
77 100
182 27.3 7.3 47.3
105 22.5 2.5 42.5
51 34.4 14.4 54.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:188 Resp: 616408
Ion Ratio Lower Upper
188 100
94 10.2 8.2 12.2
80 10.7 8.6 12.8









#65
4,6-Dinitro-2-methylphenol
Concen: 36.574 ng
RT: 10.534 min Scan# 1418
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Instrument :
BNA_F
ClientSampleId :
SSTDICCC040

Tgt Ion:198 Resp: 67678
Ion Ratio Lower Upper
198 100
51 60.6 40.6 80.6
105 49.2 29.2 69.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#66
n-Nitrosodiphenylamine
Concen: 41.656 ng
RT: 10.604 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:169 Resp: 877338
Ion Ratio Lower Upper
169 100
168 66.3 53.0 79.6
167 36.0 28.8 43.2





#67
4-Bromophenyl-phenylether
Concen: 42.400 ng
RT: 10.981 min Scan# 1494
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040




Tgt Ion:248 Resp: 291063
Ion Ratio Lower Upper
248 100
250 96.5 77.2 115.8
141 73.0 58.4 87.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#68
Hexachlorobenzene
Concen: 41.809 ng
RT: 11.045 min Scan# 1505
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17



Tgt Ion:284 Resp: 321448
Ion Ratio Lower Upper
284 100
142 30.8 24.6 37.0
249 29.0 23.2 34.8





#69

Atrazine

Concen: 42.299 ng

RT: 11.128 min Scan# 1519

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion:200 Resp: 234825

Ion Ratio Lower Upper

200 100

173 28.6 8.6 48.6





215 48.5 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#70

Pentachlorophenol

Concen: 43.764 ng

RT: 11.233 min Scan# 1537

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion:266 Resp: 177755


Ion Ratio Lower Upper

266 100

268 62.5 50.0 75.0


264 61.6 49.3 73.9





#71
Phenanthrene
Concen: 41.176 ng
RT: 11.457 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:178 Resp: 1371009
Ion Ratio Lower Upper
178 100
176 19.4 15.5 23.3
179 15.8 12.6 19.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#72
Anthracene
Concen: 40.840 ng
RT: 11.510 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:178 Resp: 1390308
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.9 12.7 19.1





#73
Carbazole
Concen: 40.999 ng
RT: 11.663 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040




Tgt Ion:167 Resp: 1273184
Ion Ratio Lower Upper
167 100
166 21.6 17.3 25.9
139 12.9 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 42.180 ng
RT: 11.998 min Scan# 1667
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Tgt Ion:149 Resp: 1317363
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 4.5 3.6 5.4





#75
Fluoranthene
Concen: 40.652 ng
RT: 12.645 min Scan# 1777
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:202 Resp: 1338000
Ion Ratio Lower Upper
202 100
101 12.4 0.0 32.4
203 17.5 0.0 37.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:240 Resp: 331192
Ion Ratio Lower Upper
240 100
120 12.6 10.1 15.1
236 25.6 20.5 30.7





#77
Benzidine
Concen: 45.346 ng
RT: 12.769 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:184 Resp: 63814
Ion Ratio Lower Upper
184 100
185 14.7 11.8 17.6
183 11.1 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#78
Pyrene
Concen: 43.920 ng
RT: 12.880 min Scan# 1817
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:202 Resp: 1341800
Ion Ratio Lower Upper
202 100
200 20.1 16.1 24.1
203 17.6 14.1 21.1





#79

Terphenyl-d14

Concen: 86.181 ng

RT: 13.022 min Scan# 1841

Delta R.T. 0.000 min

Lab File: BF142243.D


Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040



Tgt Ion:244 Resp: 1962470

Ion Ratio Lower Upper

244 100

212 6.6 5.3 7.9




122 11.4 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#80

Butylbenzylphthalate


Concen: 40.624 ng

RT: 13.498 min Scan# 1922

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17





Tgt Ion:149 Resp: 334175


Ion Ratio Lower Upper

149 100

91 68.0 54.4 81.6


206 19.9 15.9 23.9





#81
Benzo(a)anthracene
Concen: 42.831 ng
RT: 14.063 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040




Tgt Ion:228 Resp: 952290
Ion Ratio Lower Upper
228 100
226 27.4 21.9 32.9
229 20.0 16.0 24.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#82
3,3'-Dichlorobenzidine
Concen: 44.859 ng
RT: 14.027 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Tgt Ion:252 Resp: 292958
Ion Ratio Lower Upper
252 100
254 63.8 51.0 76.6
126 11.8 9.4 14.2





#83
Chrysene
Concen: 40.321 ng
RT: 14.104 min Scan# 2025
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:228 Resp: 81604
Ion Ratio Lower Upper
228 100
226 30.0 24.0 36.0
229 19.6 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 40.356 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:149 Resp: 460074
Ion Ratio Lower Upper
149 100
167 27.1 21.7 32.5
279 3.7 3.0 4.4





#85
Di-n-octyl phthalate
Concen: 40.504 ng
RT: 14.680 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:149 Resp: 836078
Ion Ratio Lower Upper
149 100
167 1.4 1.1 1.7
43 11.6 9.3 13.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17


Tgt Ion:264 Resp: 328405
Ion Ratio Lower Upper
264 100
260 23.0 18.4 27.6
265 21.6 17.3 25.9





#87
 Indeno(1,2,3-cd)pyrene
 Concen: 44.049 ng
 RT: 17.080 min Scan# 2199
 Delta R.T. 0.000 min
 Lab File: BF142243.D
 Acq: 30 Apr 2025 13:17




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC040



Tgt Ion: 276 Resp: 105959
 Ion Ratio Lower Upper
 276 100
 138 28.2 22.6 33.8
 277 25.1 20.1 30.1




Manual Integrations
 APPROVED


Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025



#88
 Benzo(b)fluoranthene
 Concen: 41.912 ng
 RT: 15.127 min Scan# 2199
 Delta R.T. 0.000 min
 Lab File: BF142243.D
 Acq: 30 Apr 2025 13:17

Tgt Ion: 252 Resp: 854971
 Ion Ratio Lower Upper
 252 100
 253 22.1 17.7 26.5
 125 11.2 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 40.310 ng

RT: 15.157 min Scan# 2204

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17

Instrument :

BNA_F

ClientSampleId :

SSTDICCC040

Tgt Ion:252 Resp: 75213

Ion Ratio Lower Upper

252 100

253 22.3 17.8 26.8





125 11.3 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#90

Benzo(a)pyrene

Concen: 42.318 ng

RT: 15.504 min Scan# 2263

Delta R.T. 0.000 min

Lab File: BF142243.D

Acq: 30 Apr 2025 13:17




Tgt Ion:252 Resp: 790330


Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4


125 11.9 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 44.142 ng
RT: 17.098 min Scan# 2534
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17




Instrument :
BNA_F
ClientSampleId :
SSTDICCC040



Tgt Ion:278 Resp: 86178
Ion Ratio Lower Upper
278 100
139 19.0 15.2 22.8
279 23.8 19.0 28.6




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 43.758 ng
RT: 17.539 min Scan# 2609
Delta R.T. 0.000 min
Lab File: BF142243.D
Acq: 30 Apr 2025 13:17

Tgt Ion:276 Resp: 858804
Ion Ratio Lower Upper
276 100
277 23.4 18.7 28.1
138 24.2 19.4 29.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142244.D
 Acq On : 30 Apr 2025 13:46
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:43:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	179073	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	691236	20.000	ng	0.00
39) Acenaphthene-d10	9.951	164	360144	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	604171	20.000	ng	0.00
76) Chrysene-d12	14.074	240	322957	20.000	ng	0.00
86) Perylene-d12	15.568	264	337726	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1150166	102.577	ng	0.00
7) Phenol-d6	6.528	99	1387418	102.468	ng	0.00
23) Nitrobenzene-d5	7.475	82	1213856	115.369	ng	0.00
42) 2,4,6-Tribromophenol	10.739	330	369823	112.641	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2425708	94.966	ng	0.00
79) Terphenyl-d14	13.021	244	2247982	101.236	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.693	88	259869	51.894	ng	99
3) Pyridine	3.463	79	672711	51.688	ng	100
4) n-Nitrosodimethylamine	3.416	42	353028	52.416	ng	96
6) Aniline	6.575	93	986118	51.454	ng	99
8) 2-Chlorophenol	6.692	128	606252	52.325	ng	99
9) Benzaldehyde	6.463	77	482619	50.998	ng	98
10) Phenol	6.539	94	755326m	51.820	ng	
11) bis(2-Chloroethyl)ether	6.645	93	575347	50.934	ng	100
12) 1,3-Dichlorobenzene	6.851	146	661612	50.476	ng	99
13) 1,4-Dichlorobenzene	6.928	146	666083	50.439	ng	99
14) 1,2-Dichlorobenzene	7.081	146	640983	50.843	ng	100
15) Benzyl Alcohol	7.045	79	504297	52.514	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.186	45	1049319	51.444	ng	100
17) 2-Methylphenol	7.151	107	485145	51.113	ng	100
18) Hexachloroethane	7.428	117	229911	53.193	ng	97
19) n-Nitroso-di-n-propyla...	7.328	70	420535	50.261	ng	99
20) 3+4-Methylphenols	7.310	107	611404	51.510	ng	# 89
22) Acetophenone	7.322	105	796460	49.805	ng	96
24) Nitrobenzene	7.492	77	584670	56.634	ng	99
25) Isophorone	7.734	82	1129933	51.170	ng	99
26) 2-Nitrophenol	7.810	139	228021	51.526	ng	97
27) 2,4-Dimethylphenol	7.839	122	576054	51.741	ng	100
28) bis(2-Chloroethoxy)met...	7.939	93	708439	50.870	ng	99
29) 2,4-Dichlorophenol	8.045	162	494488	53.125	ng	99
30) 1,2,4-Trichlorobenzene	8.133	180	524508	50.497	ng	100
31) Naphthalene	8.216	128	1725818	50.064	ng	100
32) Benzoic acid	7.939	122	252184	47.667	ng	98
33) 4-Chloroaniline	8.257	127	730860	50.914	ng	99
34) Hexachlorobutadiene	8.333	225	307118	50.679	ng	99
35) Caprolactam	8.628	113	154579	55.251	ng	97
36) 4-Chloro-3-methylphenol	8.733	107	505527	52.891	ng	100
37) 2-Methylnaphthalene	8.904	142	1057924	49.915	ng	100
38) 1-Methylnaphthalene	9.004	142	1107217	50.427	ng	98
40) 1,2,4,5-Tetrachloroben...	9.069	216	509426	50.261	ng	99
41) Hexachlorocyclopentadiene	9.063	237	331287	57.591	ng	99
43) 2,4,6-Trichlorophenol	9.180	196	337910	53.801	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142244.D
 Acq On : 30 Apr 2025 13:46
 Operator : RC/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC050

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:43:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.216	196	370489	56.106	ng	98
46) 1,1'-Biphenyl	9.369	154	1392513	49.736	ng	100
47) 2-Chloronaphthalene	9.398	162	1041842	50.015	ng	100
48) 2-Nitroaniline	9.486	65	284370	51.394	ng	100
49) Acenaphthylene	9.810	152	1769441	50.339	ng	100
50) Dimethylphthalate	9.669	163	1178638	51.156	ng	99
51) 2,6-Dinitrotoluene	9.727	165	232906	52.309	ng	99
52) Acenaphthene	9.986	154	1024409	49.912	ng	100
53) 3-Nitroaniline	9.898	138	279608	51.745	ng	97
54) 2,4-Dinitrophenol	9.998	184	62086	50.725	ng	98
55) Dibenzofuran	10.157	168	1542097	50.123	ng	100
56) 4-Nitrophenol	10.039	139	210153	58.029	ng	99
57) 2,4-Dinitrotoluene	10.127	165	280193	51.918	ng	98
58) Fluorene	10.498	166	1146269	49.000	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.269	232	300782	54.989	ng	99
60) Diethylphthalate	10.369	149	1170793	51.459	ng	99
61) 4-Chlorophenyl-phenyle...	10.492	204	553271	49.802	ng	98
62) 4-Nitroaniline	10.510	138	258117	52.547	ng	100
63) Azobenzene	10.651	77	1150899	51.226	ng	99
65) 4,6-Dinitro-2-methylph...	10.539	198	98114	49.012	ng	95
66) n-Nitrosodiphenylamine	10.610	169	1041277	50.442	ng	100
67) 4-Bromophenyl-phenylether	10.980	248	340298	50.576	ng	97
68) Hexachlorobenzene	11.045	284	384807	51.064	ng	99
69) Atrazine	11.133	200	287405	52.819	ng	100
70) Pentachlorophenol	11.233	266	219414	55.114	ng	98
71) Phenanthrene	11.463	178	1607091	49.244	ng	99
72) Anthracene	11.510	178	1659788	49.744	ng	100
73) Carbazole	11.663	167	1504715	49.436	ng	100
74) Di-n-butylphthalate	11.998	149	1578769	51.573	ng	100
75) Fluoranthene	12.651	202	1563998	48.481	ng	99
77) Benzidine	12.768	184	762858	55.590	ng	99
78) Pyrene	12.880	202	1563568	52.484	ng	100
80) Butylbenzylphthalate	13.498	149	423529	51.685	ng	100
81) Benzo(a)anthracene	14.068	228	1120699	51.691	ng	100
82) 3,3'-Dichlorobenzidine	14.027	252	363124	57.021	ng	98
83) Chrysene	14.104	228	990991	50.214	ng	100
84) Bis(2-ethylhexyl)phtha...	14.057	149	585800	51.638	ng	100
85) Di-n-octyl phthalate	14.680	149	1065626	51.078	ng	99
87) Indeno(1,2,3-cd)pyrene	17.080	276	1313336	53.091	ng	99
88) Benzo(b)fluoranthene	15.127	252	1073096	51.153	ng	100
89) Benzo(k)fluoranthene	15.162	252	967248	50.408	ng	98
90) Benzo(a)pyrene	15.504	252	996599	51.890	ng	100
91) Dibenzo(a,h)anthracene	17.103	278	1073246	53.457	ng	99
92) Benzo(g,h,i)perylene	17.539	276	1069402	52.984	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Manual Integrations APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050




Tgt Ion:152 Resp: 17907
Ion Ratio Lower Upper
152 100
150 157.6 130.2 195.2
115 59.6 47.0 70.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#2
1,4-Dioxane
Concen: 51.894 ng
RT: 2.693 min Scan# 85
Delta R.T. -0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46



Tgt Ion: 88 Resp: 259869
Ion Ratio Lower Upper
88 100
58 90.5 72.2 108.2
43 35.0 27.0 40.6





#3

Pyridine

Concen: 51.688 ng

RT: 3.463 min Scan# 216

Delta R.T. -0.000 min

Lab File: BF142244.D


Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050






Tgt Ion:	79	Resp:	67271
Ion	Ratio	Lower	Upper
79	100		
52	89.2	71.5	107.3
51	42.6	34.5	51.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#4

n-Nitrosodimethylamine


Concen: 52.416 ng

RT: 3.416 min Scan# 208



Delta R.T. 0.006 min


Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion:	42	Resp:	353028
Ion	Ratio	Lower	Upper
42	100		
74	116.0	89.5	134.3
44	7.9	6.6	9.8





#5
2-Fluorophenol
Concen: 102.577 ng
RT: 5.516 min Scan# 50
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion: 112 Resp: 115016
Ion Ratio Lower Upper
112 100
64 60.8 48.4 72.6
63 31.3 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#6
Aniline
Concen: 51.454 ng
RT: 6.575 min Scan# 745
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion: 93 Resp: 986118
Ion Ratio Lower Upper
93 100
66 39.7 32.1 48.1
65 20.2 16.6 24.8





#7
Phenol-d6
Concen: 102.468 ng
RT: 6.528 min Scan# 737
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion: 99 Resp: 1387418
Ion Ratio Lower Upper
99 100
42 20.5 16.3 24.5
71 31.7 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#8
2-Chlorophenol
Concen: 52.325 ng
RT: 6.692 min Scan# 765
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:128 Resp: 606252
Ion Ratio Lower Upper
128 100
130 32.3 13.1 53.1
64 48.9 28.6 68.6





#9

Benzaldehyde

Concen: 50.998 ng

RT: 6.463 min Scan# 726

Delta R.T. 0.006 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion: 77 Resp: 482619

Ion Ratio Lower Upper

77 100

105 100.3 79.1 119.1





106 97.8 76.0 116.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#10

Phenol

Concen: 51.820 ng m

RT: 6.539 min Scan# 739

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion: 94 Resp: 755326


Ion Ratio Lower Upper

94 100

65 28.8 8.9 48.9

66 40.2 18.7 58.7






#11
bis(2-Chloroethyl)ether
Concen: 50.934 ng
RT: 6.645 min Scan# 71
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :




SSTDICC050



Tgt Ion: 93 Resp: 57534
Ion Ratio Lower Upper
93 100
63 82.3 62.2 102.2
95 32.5 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#12
1,3-Dichlorobenzene
Concen: 50.476 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:146 Resp: 661612
Ion Ratio Lower Upper
146 100
148 63.6 50.5 75.7
75 30.6 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 50.439 ng

RT: 6.928 min Scan# 805

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:146 Resp: 666083

Ion Ratio Lower Upper

146 100

148 63.4 50.6 76.0





111 42.4 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#14

1,2-Dichlorobenzene

Concen: 50.843 ng

RT: 7.081 min Scan# 831

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion:146 Resp: 640983


Ion Ratio Lower Upper

146 100

148 63.4 50.5 75.7

111 43.7 34.9 52.3





#15

Benzyl Alcohol

Concen: 52.514 ng

RT: 7.045 min Scan# 81

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion: 79 Resp: 50429

Ion Ratio Lower Upper

79 100

108 80.2 64.5 96.7





77 62.5 49.2 73.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 51.444 ng

RT: 7.186 min Scan# 849

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion: 45 Resp: 1049319


Ion Ratio Lower Upper

45 100

77 11.3 0.0 31.3


79 8.6 0.0 28.5





#17
2-Methylphenol
Concen: 51.113 ng
RT: 7.151 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:107 Resp: 485145
Ion Ratio Lower Upper
107 100
108 113.7 91.0 136.4
77 43.6 34.4 51.6
79 43.1 34.4 51.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#18
Hexachloroethane
Concen: 53.193 ng
RT: 7.428 min Scan# 890
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:117 Resp: 229911
Ion Ratio Lower Upper
117 100
119 95.9 77.1 115.7
201 94.8 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 50.261 ng

RT: 7.328 min Scan# 81

Delta R.T. 0.006 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion: 70 Resp: 42053

Ion Ratio Lower Upper

70 100

42 66.4 53.4 80.2

101 9.9 7.7 11.5





130 22.3 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#20

3+4-Methylphenols

Concen: 51.510 ng

RT: 7.310 min Scan# 870

Delta R.T. 0.006 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Tgt Ion:107 Resp: 611404




Ion Ratio Lower Upper


107 100

108 91.7 71.2 111.2

77 84.0 41.1 81.1


79 28.3 8.1 48.1





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1020
Delta R.T. -0.001 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050




Tgt Ion:136 Resp: 691230
Ion Ratio Lower Upper
136 100
137 10.8 8.7 13.1
54 9.7 7.7 11.5
68 6.3 5.0 7.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#22
Acetophenone
Concen: 49.805 ng
RT: 7.322 min Scan# 872
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46



Tgt Ion:105 Resp: 796460
Ion Ratio Lower Upper
105 100
71 7.9 5.7 8.5
51 33.2 30.1 45.1
120 23.5 19.0 28.4





#23

Nitrobenzene-d5

Concen: 115.369 ng

RT: 7.475 min Scan# 898

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion: 82 Resp: 121385

Ion Ratio Lower Upper

82 100

128 46.3 37.0 55.6





54 59.8 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#24

Nitrobenzene

Concen: 56.634 ng

RT: 7.492 min Scan# 901

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion: 77 Resp: 584670


Ion Ratio Lower Upper

77 100

123 47.9 38.6 58.0

65 14.1 11.5 17.3









#25
Isophorone
Concen: 51.170 ng
RT: 7.734 min Scan# 941
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Instrument :
BNA_F
ClientSampleId :
SSTDICC050

Tgt Ion: 82 Resp: 112993
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 19.9 15.3 22.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#26
2-Nitrophenol
Concen: 51.526 ng
RT: 7.810 min Scan# 955
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:139 Resp: 228021
Ion Ratio Lower Upper
139 100
109 26.7 21.8 32.8
65 48.1 41.1 61.7






#27
2,4-Dimethylphenol
Concen: 51.741 ng
RT: 7.839 min Scan# 90
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050






Tgt Ion:122 Resp: 576054
Ion Ratio Lower Upper
122 100
107 107.3 86.0 129.0
121 57.9 46.3 69.5


Manual Integrations

APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#28
bis(2-Chloroethoxy)methane
Concen: 50.870 ng
RT: 7.939 min Scan# 977
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46



Tgt Ion: 93 Resp: 708439
Ion Ratio Lower Upper
93 100
95 32.9 26.2 39.2
123 12.4 9.7 14.5





#29

2,4-Dichlorophenol

Concen: 53.125 ng

RT: 8.045 min Scan# 994

Delta R.T. 0.006 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:162 Resp: 49448

Ion Ratio Lower Upper

162 100

164 64.7 43.0 83.0





98 36.5 16.7 56.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#30

1,2,4-Trichlorobenzene

Concen: 50.497 ng

RT: 8.133 min Scan# 1010

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion:180 Resp: 524508


Ion Ratio Lower Upper

180 100

182 94.9 75.7 113.5

145 32.0 25.4 38.0





#31

Naphthalene

Concen: 50.064 ng

RT: 8.216 min Scan# 1024

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:128 Resp: 1725818

Ion Ratio Lower Upper

128 100

129 11.4 9.0 13.4





127 13.3 10.6 15.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#32

Benzoic acid

Concen: 47.667 ng

RT: 7.939 min Scan# 977

Delta R.T. 0.005 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion:122 Resp: 252184


Ion Ratio Lower Upper

122 100

105 120.1 98.2 138.2

77 89.3 66.8 106.8





#33

4-Chloroaniline

Concen: 50.914 ng

RT: 8.257 min Scan# 1031

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:127 Resp: 730860

Ion Ratio Lower Upper

127 100

129 32.2 26.1 39.1

65 32.0 25.5 38.3





92 18.9 15.4 23.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#34

Hexachlorobutadiene

Concen: 50.679 ng

RT: 8.333 min Scan# 1044

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion:225 Resp: 307118


Ion Ratio Lower Upper

225 100

223 63.2 49.7 74.5


227 63.8 51.0 76.4





#35
Caprolactam
Concen: 55.251 ng
RT: 8.628 min Scan# 1093
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050




Tgt Ion:113 Resp: 154579
Ion Ratio Lower Upper
113 100
55 194.4 177.7 217.7
56 142.7 127.1 167.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#36
4-Chloro-3-methylphenol
Concen: 52.891 ng
RT: 8.733 min Scan# 1112
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Tgt Ion:107 Resp: 505527
Ion Ratio Lower Upper
107 100
144 26.9 21.0 31.4
142 82.6 66.1 99.1





#37
2-Methylnaphthalene
Concen: 49.915 ng
RT: 8.904 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:142 Resp: 1057924
Ion Ratio Lower Upper
142 100
141 87.8 70.5 105.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#38
1-Methylnaphthalene
Concen: 50.427 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:142 Resp: 1107217
Ion Ratio Lower Upper
142 100
141 89.6 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.951 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF142244.D


Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050






Tgt Ion:164 Resp: 360144
Ion Ratio Lower Upper
164 100
162 99.8 78.6 117.8
160 44.5 35.7 53.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#40

1,2,4,5-Tetrachlorobenzene


Concen: 50.261 ng

RT: 9.069 min Scan# 1169



Delta R.T. -0.000 min


Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion:216 Resp: 509426
Ion Ratio Lower Upper
216 100
214 80.5 63.8 95.6
179 21.3 16.7 25.1
108 19.3 15.6 23.4





#41
Hexachlorocyclopentadiene
Concen: 57.591 ng
RT: 9.063 min Scan# 1168
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
Client Sample Id :
SSTDICC050




Tgt Ion: 237 Resp: 33128
Ion Ratio Lower Upper
237 100
235 62.8 43.5 83.5
272 12.9 0.0 32.9

Manual Integrations
APPROVED



Reviewed By : Rahul Chavli 05/01/2025
Supervised By : Jagrut Upadhyay 05/01/2025




#42
2,4,6-Tribromophenol
Concen: 112.641 ng
RT: 10.739 min Scan# 1453
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46



Tgt Ion: 330 Resp: 369823
Ion Ratio Lower Upper
330 100
332 97.4 76.5 114.7
141 30.5 24.3 36.5





#43

2,4,6-Trichlorophenol

Concen: 53.801 ng

RT: 9.180 min Scan# 1187

Delta R.T. 0.006 min

Lab File: BF142244.D


Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050



Tgt Ion:196 Resp: 337910

Ion Ratio Lower Upper

196 100

198 96.7 77.4 116.2




200 30.9 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#44

2,4,5-Trichlorophenol


Concen: 56.106 ng

RT: 9.216 min Scan# 1194

Delta R.T. 0.006 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46



Tgt Ion:196 Resp: 370489



Ion Ratio Lower Upper


196 100

198 94.7 75.2 112.8

97 47.7 39.8 59.8

132 28.5 23.8 35.8





#45

2-Fluorobiphenyl

Concen: 94.966 ng

RT: 9.269 min Scan# 1203

Delta R.T. -0.000 min

Lab File: BF142244.D


Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050



Tgt Ion:172 Resp: 2425708

Ion Ratio Lower Upper

172 100

171 36.1 29.0 43.4




170 24.1 19.4 29.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#46

1,1'-Biphenyl


Concen: 49.736 ng

RT: 9.369 min Scan# 1220

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46





Tgt Ion:154 Resp: 1392513


Ion Ratio Lower Upper

154 100

153 40.6 20.5 60.5

76 13.0 0.0 32.8





#47

2-Chloronaphthalene

Concen: 50.015 ng

RT: 9.398 min Scan# 1224

Delta R.T. 0.006 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:162 Resp: 104184

Ion Ratio Lower Upper

162 100

127 40.6 32.6 49.0





164 32.9 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#48

2-Nitroaniline

Concen: 51.394 ng

RT: 9.486 min Scan# 1240

Delta R.T. -0.001 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion: 65 Resp: 284370


Ion Ratio Lower Upper

65 100

92 67.9 54.3 81.5

138 110.9 88.6 133.0





#49

Acenaphthylene

Concen: 50.339 ng

RT: 9.810 min Scan# 1295

Delta R.T. -0.000 min

Lab File: BF142244.D


Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050



Tgt Ion:152 Resp: 176944

Ion Ratio Lower Upper

152 100

151 20.3 16.2 24.2




153 13.7 10.7 16.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#50

Dimethylphthalate


Concen: 51.156 ng

RT: 9.669 min Scan# 1271

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46





Tgt Ion:163 Resp: 1178638


Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8


164 10.6 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 52.309 ng
RT: 9.727 min Scan# 1281
Delta R.T. -0.001 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:165 Resp: 232900
Ion Ratio Lower Upper
165 100
63 61.1 48.5 72.7
89 55.8 43.9 65.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#52
Acenaphthene
Concen: 49.912 ng
RT: 9.986 min Scan# 1325
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:154 Resp: 1024409
Ion Ratio Lower Upper
154 100
153 110.9 88.6 133.0
152 52.7 42.1 63.1





#53
3-Nitroaniline
Concen: 51.745 ng
RT: 9.898 min Scan# 1309
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:138 Resp: 279608
Ion Ratio Lower Upper
138 100
108 13.4 11.7 17.5
92 111.0 91.2 136.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#54
2,4-Dinitrophenol
Concen: 50.725 ng
RT: 9.998 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:184 Resp: 62086
Ion Ratio Lower Upper
184 100
63 92.9 69.2 103.8
154 342.0 271.4 407.2





#55

Dibenzenofuran

Concen: 50.123 ng

RT: 10.157 min Scan# 11

Delta R.T. 0.006 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:168 Resp: 154209

Ion Ratio Lower Upper

168 100

139 36.7 29.4 44.0





169 13.6 11.0 16.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#56

4-Nitrophenol

Concen: 58.029 ng

RT: 10.039 min Scan# 1334

Delta R.T. 0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion:139 Resp: 210153


Ion Ratio Lower Upper

139 100

109 65.7 44.9 84.9


65 94.5 74.7 114.7





#57
2,4-Dinitrotoluene
Concen: 51.918 ng
RT: 10.127 min Scan# 11
Delta R.T. -0.001 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:165 Resp: 28019
Ion Ratio Lower Upper
165 100
63 50.2 38.8 58.2
89 72.8 57.1 85.7
182 2.8 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#58
Fluorene
Concen: 49.000 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:166 Resp: 1146269
Ion Ratio Lower Upper
166 100
165 93.2 73.9 110.9
167 13.6 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 54.989 ng
RT: 10.269 min Scan# 1373
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:232 Resp: 300781
Ion Ratio Lower Upper
232 100
131 44.4 35.9 53.9
130 2.3 1.8 2.6
166 29.3 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#60
Diethylphthalate
Concen: 51.459 ng
RT: 10.369 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:149 Resp: 1170793
Ion Ratio Lower Upper
149 100
177 21.7 17.6 26.4
150 12.8 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 49.802 ng
RT: 10.492 min Scan# 1410
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050




Tgt Ion:204 Resp: 55327
Ion Ratio Lower Upper
204 100
206 32.6 25.6 38.4
141 58.9 48.9 73.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#62
4-Nitroaniline
Concen: 52.547 ng
RT: 10.510 min Scan# 1414
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Tgt Ion:138 Resp: 258117
Ion Ratio Lower Upper
138 100
92 49.0 29.2 69.2
108 69.4 49.1 89.1





#63
Azobenzene
Concen: 51.226 ng
RT: 10.651 min Scan# 1438
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDIC050



Tgt Ion: 77 Resp: 1150899
Ion Ratio Lower Upper
77 100
182 26.8 7.3 47.3
105 22.6 2.5 42.5
51 33.9 14.4 54.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:188 Resp: 604171
Ion Ratio Lower Upper
188 100
94 10.2 8.2 12.2
80 10.8 8.6 12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 49.012 ng
RT: 10.539 min Scan# 1418
Delta R.T. 0.005 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050




Tgt Ion:198 Resp: 98114
Ion Ratio Lower Upper
198 100
51 55.5 40.6 80.6
105 47.3 29.2 69.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#66
n-Nitrosodiphenylamine
Concen: 50.442 ng
RT: 10.610 min Scan# 1431
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46



Tgt Ion:169 Resp: 1041277
Ion Ratio Lower Upper
169 100
168 65.9 53.0 79.6
167 36.1 28.8 43.2





#67

4-Bromophenyl-phenylether

Concen: 50.576 ng

RT: 10.980 min Scan# 1494

Delta R.T. -0.000 min

Lab File: BF142244.D


Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050



Tgt Ion:248 Resp: 340298

Ion Ratio Lower Upper

248 100

250 100.0 77.2 115.8




141 75.5 58.4 87.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#68

Hexachlorobenzene


Concen: 51.064 ng

RT: 11.045 min Scan# 1505

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46





Tgt Ion:284 Resp: 384807


Ion Ratio Lower Upper

284 100

142 31.4 24.6 37.0

249 30.0 23.2 34.8





#69

Atrazine

Concen: 52.819 ng

RT: 11.133 min Scan# 1520

Delta R.T. 0.006 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:200 Resp: 28740

Ion Ratio Lower Upper

200 100

173 28.4 8.6 48.6





215 48.4 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#70

Pentachlorophenol

Concen: 55.114 ng

RT: 11.233 min Scan# 1537

Delta R.T. 0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion:266 Resp: 219414


Ion Ratio Lower Upper

266 100

268 63.5 50.0 75.0


264 63.2 49.3 73.9





#71
Phenanthrene
Concen: 49.244 ng
RT: 11.463 min Scan# 1575
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:178 Resp: 1607091
Ion Ratio Lower Upper
178 100
176 19.9 15.5 23.3
179 15.7 12.6 19.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#72
Anthracene
Concen: 49.744 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:178 Resp: 1659788
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.9 12.7 19.1





#73
Carbazole
Concen: 49.436 ng
RT: 11.663 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050




Tgt Ion:167 Resp: 1504715
Ion Ratio Lower Upper
167 100
166 21.4 17.3 25.9
139 13.0 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 51.573 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Tgt Ion:149 Resp: 1578769
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 4.6 3.6 5.4





#75
Fluoranthene
Concen: 48.481 ng
RT: 12.651 min Scan# 1777
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:202 Resp: 1563998
Ion Ratio Lower Upper
202 100
101 11.7 0.0 32.4
203 17.9 0.0 37.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:240 Resp: 322957
Ion Ratio Lower Upper
240 100
120 12.6 10.1 15.1
236 25.3 20.5 30.7





#77
Benzidine
Concen: 55.590 ng
RT: 12.768 min Scan# 1798
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:184 Resp: 762858
Ion Ratio Lower Upper
184 100
185 15.0 11.8 17.6
183 10.8 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#78
Pyrene
Concen: 52.484 ng
RT: 12.880 min Scan# 1817
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:202 Resp: 1563568
Ion Ratio Lower Upper
202 100
200 20.3 16.1 24.1
203 17.7 14.1 21.1





#79

Terphenyl-d14

Concen: 101.236 ng

RT: 13.021 min Scan# 1841

Delta R.T. -0.000 min

Lab File: BF142244.D


Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050



Tgt Ion:244 Resp: 224798

Ion Ratio Lower Upper

244 100

212 6.8 5.3 7.9




122 11.6 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#80

Butylbenzylphthalate

Concen: 51.685 ng

RT: 13.498 min Scan# 1922

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion:149 Resp: 423529


Ion Ratio Lower Upper

149 100

91 67.9 54.4 81.6


206 20.3 15.9 23.9





#81
Benzo(a)anthracene
Concen: 51.691 ng
RT: 14.068 min Scan# 2018
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050




Tgt Ion:228 Resp: 1120699
Ion Ratio Lower Upper
228 100
226 27.3 21.9 32.9
229 19.8 16.0 24.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#82
3,3'-Dichlorobenzidine
Concen: 57.021 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46



Tgt Ion:252 Resp: 363124
Ion Ratio Lower Upper
252 100
254 65.7 51.0 76.6
126 12.3 9.4 14.2





#83

Chrysene

Concen: 50.214 ng

RT: 14.104 min Scan# 2025

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :





SSTDICC050

Tgt Ion:	228	Resp:	99099
Ion Ratio	Lower	Upper	
228	100		
226	29.9	24.0	36.0
229	19.7	15.7	23.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#84

Bis(2-ethylhexyl)phthalate

Concen: 51.638 ng




RT: 14.057 min Scan# 2017


Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46


Tgt Ion:	149	Resp:	585800
Ion Ratio	Lower	Upper	
149	100		
167	27.4	21.7	32.5
279	3.7	3.0	4.4





#85
Di-n-octyl phthalate
Concen: 51.078 ng
RT: 14.680 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:149 Resp: 1065620
Ion Ratio Lower Upper
149 100
167 1.5 1.1 1.7
43 11.3 9.3 13.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46


Tgt Ion:264 Resp: 337726
Ion Ratio Lower Upper
264 100
260 23.0 18.4 27.6
265 21.4 17.3 25.9





#87
Indeno(1,2,3-cd)pyrene
Concen: 53.091 ng
RT: 17.080 min Scan# 2199
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion: 276 Resp: 1313330
Ion Ratio Lower Upper
276 100
138 29.0 22.6 33.8
277 25.1 20.1 30.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#88
Benzo(b)fluoranthene
Concen: 51.153 ng
RT: 15.127 min Scan# 2199
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion: 252 Resp: 1073096
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 11.2 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 50.408 ng

RT: 15.162 min Scan# 21

Delta R.T. 0.006 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46

Instrument :

BNA_F

ClientSampleId :

SSTDICC050

Tgt Ion:252 Resp: 96724

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.8





125 10.0 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#90

Benzo(a)pyrene

Concen: 51.890 ng

RT: 15.504 min Scan# 2263

Delta R.T. -0.000 min

Lab File: BF142244.D

Acq: 30 Apr 2025 13:46




Tgt Ion:252 Resp: 996599


Ion Ratio Lower Upper

252 100

253 22.1 17.6 26.4


125 12.2 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 53.457 ng
RT: 17.103 min Scan# 2535
Delta R.T. 0.006 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46




Instrument :
BNA_F
ClientSampleId :
SSTDICC050



Tgt Ion:278 Resp: 1073240
Ion Ratio Lower Upper
278 100
139 18.2 15.2 22.8
279 23.7 19.0 28.6




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 52.984 ng
RT: 17.539 min Scan# 2609
Delta R.T. -0.000 min
Lab File: BF142244.D
Acq: 30 Apr 2025 13:46

Tgt Ion:276 Resp: 1069402
Ion Ratio Lower Upper
276 100
277 23.8 18.7 28.1
138 25.3 19.4 29.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142245.D
 Acq On : 30 Apr 2025 14:15
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:46:00 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	187963	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	730767	20.000	ng	0.00
39) Acenaphthene-d10	9.951	164	379270	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	624729	20.000	ng	0.00
76) Chrysene-d12	14.080	240	328990	20.000	ng	0.00
86) Perylene-d12	15.568	264	361305	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	1364962	115.976	ng	0.00
7) Phenol-d6	6.528	99	1663210	117.026	ng	0.00
23) Nitrobenzene-d5	7.475	82	1488103	133.783	ng	0.00
42) 2,4,6-Tribromophenol	10.739	330	447862	129.531	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2777485	103.255	ng	0.00
79) Terphenyl-d14	13.022	244	2546694	112.585	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.699	88	312897	59.528	ng	99
3) Pyridine	3.469	79	823442	60.277	ng	98
4) n-Nitrosodimethylamine	3.422	42	431685	61.063	ng	98
6) Aniline	6.575	93	1184649	58.890	ng	100
8) 2-Chlorophenol	6.693	128	740327	60.874	ng	99
9) Benzaldehyde	6.463	77	559766	56.352	ng	98
10) Phenol	6.545	94	903976m	59.085	ng	
11) bis(2-Chloroethyl)ether	6.646	93	693997	58.532	ng	99
12) 1,3-Dichlorobenzene	6.851	146	787463	57.236	ng	99
13) 1,4-Dichlorobenzene	6.928	146	791509	57.102	ng	99
14) 1,2-Dichlorobenzene	7.081	146	764397	57.765	ng	99
15) Benzyl Alcohol	7.045	79	609871	60.504	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.187	45	1249716	58.370	ng	100
17) 2-Methylphenol	7.151	107	592171	59.438	ng	99
18) Hexachloroethane	7.428	117	275661	60.761	ng	99
19) n-Nitroso-di-n-propyla...	7.328	70	506389	57.660	ng	100
20) 3+4-Methylphenols	7.310	107	730966	58.670	ng	94
22) Acetophenone	7.322	105	943225	55.792	ng	97
24) Nitrobenzene	7.498	77	708847	64.948	ng	99
25) Isophorone	7.734	82	1371824	58.764	ng	100
26) 2-Nitrophenol	7.810	139	299107	61.132	ng	97
27) 2,4-Dimethylphenol	7.840	122	691268	58.731	ng	100
28) bis(2-Chloroethoxy)met...	7.940	93	847188	57.543	ng	100
29) 2,4-Dichlorophenol	8.045	162	596890	60.658	ng	99
30) 1,2,4-Trichlorobenzene	8.134	180	628720	57.256	ng	100
31) Naphthalene	8.216	128	2026390	55.603	ng	99
32) Benzoic acid	7.951	122	349577	59.198	ng	99
33) 4-Chloroaniline	8.263	127	868158	57.207	ng	99
34) Hexachlorobutadiene	8.334	225	373284	58.266	ng	100
35) Caprolactam	8.639	113	186124	62.928	ng	94
36) 4-Chloro-3-methylphenol	8.734	107	605133	59.887	ng	99
37) 2-Methylnaphthalene	8.904	142	1257822	56.136	ng	99
38) 1-Methylnaphthalene	9.004	142	1294318	55.760	ng	100
40) 1,2,4,5-Tetrachloroben...	9.069	216	607614	56.925	ng	100
41) Hexachlorocyclopentadiene	9.063	237	407387	67.248	ng	100
43) 2,4,6-Trichlorophenol	9.181	196	433323	65.514	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142245.D
 Acq On : 30 Apr 2025 14:15
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:46:00 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.216	196	422145	60.705	ng	98
46) 1,1'-Biphenyl	9.369	154	1632628	55.372	ng	100
47) 2-Chloronaphthalene	9.398	162	1234670	56.283	ng	99
48) 2-Nitroaniline	9.486	65	357897	60.490	ng	99
49) Acenaphthylene	9.810	152	2079452	56.175	ng	99
50) Dimethylphthalate	9.675	163	1395869	57.529	ng	100
51) 2,6-Dinitrotoluene	9.728	165	289353	60.986	ng	98
52) Acenaphthene	9.986	154	1219341	56.414	ng	99
53) 3-Nitroaniline	9.898	138	352271	61.278	ng	97
54) 2,4-Dinitrophenol	9.998	184	86520	61.133	ng	83
55) Dibenzofuran	10.157	168	1798627	55.513	ng	99
56) 4-Nitrophenol	10.045	139	261987	68.694	ng	98
57) 2,4-Dinitrotoluene	10.134	165	352001	60.964	ng	98
58) Fluorene	10.498	166	1348902	54.754	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.269	232	367314	63.765	ng	99
60) Diethylphthalate	10.369	149	1382905	57.717	ng	100
61) 4-Chlorophenyl-phenyle...	10.492	204	648920	55.466	ng	97
62) 4-Nitroaniline	10.516	138	319292	61.243	ng	99
63) Azobenzene	10.651	77	1362906	57.603	ng	99
65) 4,6-Dinitro-2-methylph...	10.539	198	132189	60.646	ng	96
66) n-Nitrosodiphenylamine	10.610	169	1224189	57.351	ng	99
67) 4-Bromophenyl-phenylether	10.980	248	413069	59.371	ng	100
68) Hexachlorobenzene	11.045	284	457317	58.689	ng	99
69) Atrazine	11.133	200	340471	60.513	ng	99
70) Pentachlorophenol	11.233	266	270497	65.710	ng	98
71) Phenanthrene	11.463	178	1864413	55.249	ng	100
72) Anthracene	11.516	178	1920817	55.673	ng	100
73) Carbazole	11.663	167	1751538	55.651	ng	100
74) Di-n-butylphthalate	11.998	149	1852105	58.511	ng	100
75) Fluoranthene	12.651	202	1794425	53.793	ng	98
77) Benzidine	12.769	184	883568	63.206	ng	99
78) Pyrene	12.880	202	1809576	59.627	ng	99
80) Butylbenzylphthalate	13.498	149	516723	61.167	ng	99
81) Benzo(a)anthracene	14.069	228	1299270	58.829	ng	100
82) 3,3'-Dichlorobenzidine	14.027	252	444901	68.582	ng	99
83) Chrysene	14.104	228	1181729	58.781	ng	99
84) Bis(2-ethylhexyl)phtha...	14.057	149	712716	61.002	ng	99
85) Di-n-octyl phthalate	14.680	149	1329733	61.204	ng	99
87) Indeno(1,2,3-cd)pyrene	17.086	276	1614650	61.012	ng	100
88) Benzo(b)fluoranthene	15.127	252	1275552	56.836	ng	99
89) Benzo(k)fluoranthene	15.163	252	1202952	58.601	ng	99
90) Benzo(a)pyrene	15.510	252	1232454	59.982	ng	100
91) Dibenzo(a,h)anthracene	17.110	278	1292219	60.163	ng	99
92) Benzo(g,h,i)perylene	17.545	276	1315066	60.904	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142245.D
 Acq On : 30 Apr 2025 14:15
 Operator : RC/JU
 Sample : SSTDICC060
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC060

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025


Quant Time: Apr 30 15:46:00 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion:152 Resp: 18796
Ion Ratio Lower Upper
152 100
150 160.9 130.2 195.2
115 59.8 47.0 70.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#2
1,4-Dioxane
Concen: 59.528 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion: 88 Resp: 312897
Ion Ratio Lower Upper
88 100
58 90.8 72.2 108.2
43 35.0 27.0 40.6





#3

Pyridine

Concen: 60.277 ng

RT: 3.469 min Scan# 216

Delta R.T. 0.006 min

Lab File: BF142245.D


Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060



Tgt Ion: 79 Resp: 82344

Ion Ratio Lower Upper

79 100

52 87.7 71.5 107.3




51 41.9 34.5 51.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#4

n-Nitrosodimethylamine

Concen: 61.063 ng

RT: 3.422 min Scan# 209

Delta R.T. 0.012 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion: 42 Resp: 431685


Ion Ratio Lower Upper

42 100

74 113.5 89.5 134.3


44 7.5 6.6 9.8





#5
2-Fluorophenol
Concen: 115.976 ng
RT: 5.516 min Scan# 565
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion: 112 Resp: 1364962
Ion Ratio Lower Upper
112 100
64 61.6 48.4 72.6
63 31.5 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#6
Aniline
Concen: 58.890 ng
RT: 6.575 min Scan# 745
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion: 93 Resp: 1184649
Ion Ratio Lower Upper
93 100
66 40.1 32.1 48.1
65 20.2 16.6 24.8





#7
Phenol-d6
Concen: 117.026 ng
RT: 6.528 min Scan# 737
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion: 99 Resp: 1663210
Ion Ratio Lower Upper
99 100
42 20.7 16.3 24.5
71 32.3 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#8
2-Chlorophenol
Concen: 60.874 ng
RT: 6.693 min Scan# 765
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:128 Resp: 740327
Ion Ratio Lower Upper
128 100
130 32.8 13.1 53.1
64 49.4 28.6 68.6





#9

Benzaldehyde

Concen: 56.352 ng

RT: 6.463 min Scan# 71

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion: 77 Resp: 55976

Ion Ratio Lower Upper

77 100

105 100.6 79.1 119.1





106 98.1 76.0 116.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#10

Phenol

Concen: 59.085 ng m

RT: 6.545 min Scan# 740

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion: 94 Resp: 903976


Ion Ratio Lower Upper

94 100

65 29.0 8.9 48.9

66 39.2 18.7 58.7









#11
bis(2-Chloroethyl)ether
Concen: 58.532 ng
RT: 6.646 min Scan# 757
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Instrument :
BNA_F
Client Sample Id :
SSTDICC060

Tgt Ion: 93 Resp: 693997
Ion Ratio Lower Upper
93 100
63 83.0 62.2 102.2
95 32.3 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#12
1,3-Dichlorobenzene
Concen: 57.236 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:146 Resp: 787463
Ion Ratio Lower Upper
146 100
148 63.6 50.5 75.7
75 30.1 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 57.102 ng

RT: 6.928 min Scan# 805

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:146 Resp: 791509

Ion Ratio Lower Upper

146 100

148 64.1 50.6 76.0





111 42.7 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#14

1,2-Dichlorobenzene

Concen: 57.765 ng

RT: 7.081 min Scan# 831

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion:146 Resp: 764397


Ion Ratio Lower Upper

146 100

148 63.9 50.5 75.7


111 44.2 34.9 52.3





#15
Benzyl Alcohol
Concen: 60.504 ng
RT: 7.045 min Scan# 81
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion: 79 Resp: 60987
Ion Ratio Lower Upper
79 100
108 79.3 64.5 96.7
77 61.4 49.2 73.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#16
2,2'-oxybis(1-Chloropropane)
Concen: 58.370 ng
RT: 7.187 min Scan# 849
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion: 45 Resp: 1249716
Ion Ratio Lower Upper
45 100
77 11.2 0.0 31.3
79 8.5 0.0 28.5





#17
2-Methylphenol
Concen: 59.438 ng
RT: 7.151 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060




Tgt Ion:107 Resp: 59217
Ion Ratio Lower Upper
107 100
108 113.3 91.0 136.4
77 43.7 34.4 51.6
79 43.5 34.4 51.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#18
Hexachloroethane
Concen: 60.761 ng
RT: 7.428 min Scan# 890
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15



Tgt Ion:117 Resp: 275661
Ion Ratio Lower Upper
117 100
119 96.1 77.1 115.7
201 97.0 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 57.660 ng

RT: 7.328 min Scan# 81

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion: 70 Resp: 506389

Ion Ratio Lower Upper

70 100

42 66.9 53.4 80.2

101 10.0 7.7 11.5





130 21.9 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#20

3+4-Methylphenols

Concen: 58.670 ng

RT: 7.310 min Scan# 870

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Tgt Ion:107 Resp: 730966




Ion Ratio Lower Upper


107 100

108 90.6 71.2 111.2

77 73.0 41.1 81.1


79 28.2 8.1 48.1





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1020
Delta R.T. -0.001 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060




Tgt Ion:136 Resp: 73076
Ion Ratio Lower Upper
136 100
137 11.2 8.7 13.1
54 9.7 7.7 11.5
68 6.3 5.0 7.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#22
Acetophenone
Concen: 55.792 ng
RT: 7.322 min Scan# 872
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15



Tgt Ion:105 Resp: 943225
Ion Ratio Lower Upper
105 100
71 6.5 5.7 8.5
51 34.7 30.1 45.1
120 23.8 19.0 28.4





#23

Nitrobenzene-d5

Concen: 133.783 ng

RT: 7.475 min Scan# 89

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion: 82 Resp: 148810

Ion Ratio Lower Upper

82 100

128 46.0 37.0 55.6





54 60.6 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#24

Nitrobenzene

Concen: 64.948 ng

RT: 7.498 min Scan# 902

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion: 77 Resp: 708847


Ion Ratio Lower Upper

77 100

123 48.8 38.6 58.0

65 14.4 11.5 17.3









#25
Isophorone
Concen: 58.764 ng
RT: 7.734 min Scan# 941
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion: 82 Resp: 1371824
Ion Ratio Lower Upper
82 100
95 7.2 5.6 8.4
138 19.2 15.3 22.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#26
2-Nitrophenol
Concen: 61.132 ng
RT: 7.810 min Scan# 955
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:139 Resp: 299107
Ion Ratio Lower Upper
139 100
109 26.5 21.8 32.8
65 49.1 41.1 61.7





#27

2,4-Dimethylphenol

Concen: 58.731 ng

RT: 7.840 min Scan# 959

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:122 Resp: 691268

Ion Ratio Lower Upper

122 100

107 107.8 86.0 129.0





121 57.5 46.3 69.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#28

bis(2-Chloroethoxy)methane

Concen: 57.543 ng

RT: 7.940 min Scan# 977

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion: 93 Resp: 847188


Ion Ratio Lower Upper

93 100

95 32.9 26.2 39.2

123 12.3 9.7 14.5





#29

2,4-Dichlorophenol

Concen: 60.658 ng

RT: 8.045 min Scan# 994

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:162 Resp: 596890

Ion Ratio Lower Upper

162 100

164 64.4 43.0 83.0





98 36.7 16.7 56.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#30

1,2,4-Trichlorobenzene

Concen: 57.256 ng

RT: 8.134 min Scan# 1010

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion:180 Resp: 628720


Ion Ratio Lower Upper

180 100

182 94.6 75.7 113.5

145 32.1 25.4 38.0





#31

Naphthalene

Concen: 55.603 ng

RT: 8.216 min Scan# 1024

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:128 Resp: 2026390

Ion Ratio Lower Upper

128 100

129 11.6 9.0 13.4





127 13.2 10.6 15.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#32

Benzoic acid

Concen: 59.198 ng

RT: 7.951 min Scan# 979

Delta R.T. 0.017 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion:122 Resp: 349577


Ion Ratio Lower Upper

122 100

105 118.0 98.2 138.2


77 88.8 66.8 106.8





#33
4-Chloroaniline
Concen: 57.207 ng
RT: 8.263 min Scan# 1031
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion:127 Resp: 86815
Ion Ratio Lower Upper
127 100
129 32.8 26.1 39.1
65 30.5 25.5 38.3
92 18.8 15.4 23.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#34
Hexachlorobutadiene
Concen: 58.266 ng
RT: 8.334 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:225 Resp: 373284
Ion Ratio Lower Upper
225 100
223 62.5 49.7 74.5
227 63.6 51.0 76.4





#35
Caprolactam
Concen: 62.928 ng
RT: 8.639 min Scan# 1096
Delta R.T. 0.018 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion:113 Resp: 186124
Ion Ratio Lower Upper
113 100
55 189.5 177.7 217.7
56 139.8 127.1 167.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#36
4-Chloro-3-methylphenol
Concen: 59.887 ng
RT: 8.734 min Scan# 1112
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:107 Resp: 605133
Ion Ratio Lower Upper
107 100
144 26.6 21.0 31.4
142 83.2 66.1 99.1





#37

2-Methylnaphthalene

Concen: 56.136 ng

RT: 8.904 min Scan# 1141

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:142 Resp: 125782

Ion Ratio Lower Upper

142 100





141 87.6 70.5 105.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#38

1-Methylnaphthalene

Concen: 55.760 ng

RT: 9.004 min Scan# 1158

Delta R.T. -0.000 min

Lab File: BF142245.D




Acq: 30 Apr 2025 14:15


Tgt Ion:142 Resp: 1294318

Ion Ratio Lower Upper

142 100

141 90.6 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.951 min Scan# 1169

Delta R.T. 0.000 min

Lab File: BF142245.D


Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060



Tgt Ion:164 Resp: 379270

Ion Ratio Lower Upper

164 100

162 99.2 78.6 117.8




160 44.3 35.7 53.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#40

1,2,4,5-Tetrachlorobenzene


Concen: 56.925 ng

RT: 9.069 min Scan# 1169

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15



Tgt Ion:216 Resp: 607614



Ion Ratio Lower Upper


216 100

214 79.4 63.8 95.6

179 20.9 16.7 25.1

108 19.1 15.6 23.4









#41
Hexachlorocyclopentadiene
Concen: 67.248 ng
RT: 9.063 min Scan# 1167
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:237 Resp: 40738
Ion Ratio Lower Upper
237 100
235 63.6 43.5 83.5
272 13.1 0.0 32.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#42
2,4,6-Tribromophenol
Concen: 129.531 ng
RT: 10.739 min Scan# 1453
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:330 Resp: 447862
Ion Ratio Lower Upper
330 100
332 96.2 76.5 114.7
141 29.9 24.3 36.5





#43

2,4,6-Trichlorophenol

Concen: 65.514 ng

RT: 9.181 min Scan# 1187

Delta R.T. 0.006 min

Lab File: BF142245.D


Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060



Tgt Ion:196 Resp: 43332

Ion Ratio Lower Upper

196 100

198 95.9 77.4 116.2




200 30.8 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#44

2,4,5-Trichlorophenol


Concen: 60.705 ng

RT: 9.216 min Scan# 1194

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15



Tgt Ion:196 Resp: 422145



Ion Ratio Lower Upper


196 100

198 95.8 75.2 112.8

97 48.1 39.8 59.8

132 29.4 23.8 35.8





#45

2-Fluorobiphenyl

Concen: 103.255 ng

RT: 9.269 min Scan# 1203

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:172 Resp: 277748

Ion Ratio Lower Upper

172 100

171 36.5 29.0 43.4





170 24.4 19.4 29.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#46

1,1'-Biphenyl

Concen: 55.372 ng

RT: 9.369 min Scan# 1220

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion:154 Resp: 1632628


Ion Ratio Lower Upper

154 100

153 40.7 20.5 60.5

76 13.3 0.0 32.8





#47

2-Chloronaphthalene

Concen: 56.283 ng

RT: 9.398 min Scan# 1224

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:162 Resp: 1234670

Ion Ratio Lower Upper

162 100

127 41.6 32.6 49.0





164 33.2 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#48

2-Nitroaniline

Concen: 60.490 ng

RT: 9.486 min Scan# 1240

Delta R.T. -0.001 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion: 65 Resp: 357897


Ion Ratio Lower Upper

65 100

92 68.1 54.3 81.5

138 111.9 88.6 133.0





#49

Acenaphthylene

Concen: 56.175 ng

RT: 9.810 min Scan# 1295

Delta R.T. -0.000 min

Lab File: BF142245.D


Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060



Tgt Ion:152 Resp: 207945

Ion Ratio Lower Upper

152 100

151 20.5 16.2 24.2




153 13.6 10.7 16.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#50

Dimethylphthalate


Concen: 57.529 ng

RT: 9.675 min Scan# 1272

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15





Tgt Ion:163 Resp: 1395869


Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8


164 10.2 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 60.986 ng
RT: 9.728 min Scan# 1281
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060




Tgt Ion:165 Resp: 28935
Ion Ratio Lower Upper
165 100
63 62.4 48.5 72.7
89 55.4 43.9 65.9

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#52
Acenaphthene
Concen: 56.414 ng
RT: 9.986 min Scan# 1325
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Tgt Ion:154 Resp: 1219341
Ion Ratio Lower Upper
154 100
153 109.8 88.6 133.0
152 52.4 42.1 63.1





#53
3-Nitroaniline
Concen: 61.278 ng
RT: 9.898 min Scan# 1309
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion:138 Resp: 35227
Ion Ratio Lower Upper
138 100
108 12.4 11.7 17.5
92 110.7 91.2 136.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#54
2,4-Dinitrophenol
Concen: 61.133 ng
RT: 9.998 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:184 Resp: 86520
Ion Ratio Lower Upper
184 100
63 93.7 69.2 103.8
154 379.9 271.4 407.2





#55

Dibenzenofuran

Concen: 55.513 ng

RT: 10.157 min Scan# 11

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:168 Resp: 179862

Ion Ratio Lower Upper

168 100

139 37.1 29.4 44.0





169 13.9 11.0 16.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#56

4-Nitrophenol

Concen: 68.694 ng

RT: 10.045 min Scan# 1335

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion:139 Resp: 261987


Ion Ratio Lower Upper

139 100

109 64.4 44.9 84.9

65 91.8 74.7 114.7









#57
2,4-Dinitrotoluene
Concen: 60.964 ng
RT: 10.134 min Scan# 11
Delta R.T. 0.005 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:165 Resp: 35200
Ion Ratio Lower Upper
165 100
63 47.5 38.8 58.2
89 70.1 57.1 85.7
182 2.6 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#58
Fluorene
Concen: 54.754 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:166 Resp: 1348902
Ion Ratio Lower Upper
166 100
165 92.8 73.9 110.9
167 13.8 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 63.765 ng
RT: 10.269 min Scan# 1373
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060




Tgt Ion:232 Resp: 367314
Ion Ratio Lower Upper
232 100
131 44.2 35.9 53.9
130 2.2 1.8 2.6
166 29.4 24.0 36.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#60
Diethylphthalate
Concen: 57.717 ng
RT: 10.369 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Tgt Ion:149 Resp: 1382905
Ion Ratio Lower Upper
149 100
177 21.8 17.6 26.4
150 12.8 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 55.466 ng
RT: 10.492 min Scan# 1410
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060




Tgt Ion:204 Resp: 648920
Ion Ratio Lower Upper
204 100
206 33.2 25.6 38.4
141 58.4 48.9 73.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#62
4-Nitroaniline
Concen: 61.243 ng
RT: 10.516 min Scan# 1415
Delta R.T. 0.012 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Tgt Ion:138 Resp: 319292
Ion Ratio Lower Upper
138 100
92 48.4 29.2 69.2
108 68.2 49.1 89.1





#63
Azobenzene
Concen: 57.603 ng
RT: 10.651 min Scan# 1438
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion: 77 Resp: 1362900
Ion Ratio Lower Upper
77 100
182 26.7 7.3 47.3
105 22.5 2.5 42.5
51 33.9 14.4 54.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:188 Resp: 624729
Ion Ratio Lower Upper
188 100
94 10.4 8.2 12.2
80 11.0 8.6 12.8









#65
4,6-Dinitro-2-methylphenol
Concen: 60.646 ng
RT: 10.539 min Scan# 1418
Delta R.T. 0.005 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Instrument :
BNA_F
ClientSampleId :
SSTDICC060

Tgt Ion:198 Resp: 132189
Ion Ratio Lower Upper
198 100
51 57.5 40.6 80.6
105 45.8 29.2 69.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#66
n-Nitrosodiphenylamine
Concen: 57.351 ng
RT: 10.610 min Scan# 1431
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:169 Resp: 1224189
Ion Ratio Lower Upper
169 100
168 66.7 53.0 79.6
167 36.3 28.8 43.2





#67
4-Bromophenyl-phenylether
Concen: 59.371 ng
RT: 10.980 min Scan# 1494
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060




Tgt Ion:248 Resp: 413069
Ion Ratio Lower Upper
248 100
250 97.0 77.2 115.8
141 73.0 58.4 87.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#68
Hexachlorobenzene
Concen: 58.689 ng
RT: 11.045 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15



Tgt Ion:284 Resp: 457317
Ion Ratio Lower Upper
284 100
142 31.6 24.6 37.0
249 29.6 23.2 34.8





#69

Atrazine

Concen: 60.513 ng

RT: 11.133 min Scan# 1520

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:200 Resp: 34047

Ion Ratio Lower Upper

200 100

173 28.7 8.6 48.6





215 49.3 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#70

Pentachlorophenol

Concen: 65.710 ng

RT: 11.233 min Scan# 1537

Delta R.T. 0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion:266 Resp: 270497


Ion Ratio Lower Upper

266 100

268 64.1 50.0 75.0


264 63.1 49.3 73.9





#71
Phenanthrene
Concen: 55.249 ng
RT: 11.463 min Scan# 1576
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion:178 Resp: 186441
Ion Ratio Lower Upper
178 100
176 19.7 15.5 23.3
179 15.7 12.6 19.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#72
Anthracene
Concen: 55.673 ng
RT: 11.516 min Scan# 1585
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:178 Resp: 1920817
Ion Ratio Lower Upper
178 100
176 19.4 15.3 22.9
179 16.0 12.7 19.1





#73
Carbazole
Concen: 55.651 ng
RT: 11.663 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060




Tgt Ion:167 Resp: 1751538
Ion Ratio Lower Upper
167 100
166 21.5 17.3 25.9
139 12.9 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 58.511 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Tgt Ion:149 Resp: 1852105
Ion Ratio Lower Upper
149 100
150 9.5 7.5 11.3
104 4.6 3.6 5.4





#75
Fluoranthene
Concen: 53.793 ng
RT: 12.651 min Scan# 1778
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion:202 Resp: 1794425
Ion Ratio Lower Upper
202 100
101 11.7 0.0 32.4
203 18.0 0.0 37.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.080 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:240 Resp: 328990
Ion Ratio Lower Upper
240 100
120 11.4 10.1 15.1
236 25.2 20.5 30.7





#77
Benzidine
Concen: 63.206 ng
RT: 12.769 min Scan# 1798
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion:184 Resp: 883568
Ion Ratio Lower Upper
184 100
185 15.2 11.8 17.6
183 10.8 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#78
Pyrene
Concen: 59.627 ng
RT: 12.880 min Scan# 1817
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:202 Resp: 1809576
Ion Ratio Lower Upper
202 100
200 20.5 16.1 24.1
203 17.8 14.1 21.1





#79

Terphenyl-d14

Concen: 112.585 ng

RT: 13.022 min Scan# 1841

Delta R.T. -0.000 min

Lab File: BF142245.D


Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060



Tgt Ion:244 Resp: 2546694

Ion Ratio Lower Upper

244 100

212 6.8 5.3 7.9




122 11.5 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#80

Butylbenzylphthalate

Concen: 61.167 ng

RT: 13.498 min Scan# 1922

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion:149 Resp: 516723


Ion Ratio Lower Upper

149 100

91 68.3 54.4 81.6

206 20.6 15.9 23.9





#81

Benzo(a)anthracene

Concen: 58.829 ng

RT: 14.069 min Scan# 2018

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:228 Resp: 1299270

Ion Ratio Lower Upper

228 100

226 27.5 21.9 32.9





229 20.0 16.0 24.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#82

3,3'-Dichlorobenzidine

Concen: 68.582 ng

RT: 14.027 min Scan# 2012

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion:252 Resp: 444901


Ion Ratio Lower Upper

252 100

254 64.5 51.0 76.6


126 12.4 9.4 14.2





#83
Chrysene
Concen: 58.781 ng
RT: 14.104 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion:228 Resp: 1181729
Ion Ratio Lower Upper
228 100
226 30.6 24.0 36.0
229 19.7 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 61.002 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15


Tgt Ion:149 Resp: 712716
Ion Ratio Lower Upper
149 100
167 26.8 21.7 32.5
279 3.8 3.0 4.4





#85
Di-n-octyl phthalate
Concen: 61.204 ng
RT: 14.680 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion:149 Resp: 132973
Ion Ratio Lower Upper
149 100
167 1.5 1.1 1.7
43 11.3 9.3 13.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:264 Resp: 361305
Ion Ratio Lower Upper
264 100
260 23.6 18.4 27.6
265 21.9 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 61.012 ng

RT: 17.086 min Scan# 2199

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:276 Resp: 1614650

Ion Ratio Lower Upper

276 100

138 28.5 22.6 33.8





277 25.1 20.1 30.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#88

Benzo(b)fluoranthene

Concen: 56.836 ng

RT: 15.127 min Scan# 2199

Delta R.T. -0.000 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion:252 Resp: 1275552


Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 11.5 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 58.601 ng

RT: 15.163 min Scan# 21

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15

Instrument :

BNA_F

ClientSampleId :

SSTDICC060

Tgt Ion:252 Resp: 1202952

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.8





125 10.8 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#90

Benzo(a)pyrene

Concen: 59.982 ng

RT: 15.510 min Scan# 2264

Delta R.T. 0.006 min

Lab File: BF142245.D

Acq: 30 Apr 2025 14:15




Tgt Ion:252 Resp: 1232454


Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4


125 11.5 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 60.163 ng
RT: 17.110 min Scan# 21
Delta R.T. 0.012 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15




Instrument :
BNA_F
ClientSampleId :
SSTDICC060



Tgt Ion:278 Resp: 1292219
Ion Ratio Lower Upper
278 100
139 18.4 15.2 22.8
279 23.9 19.0 28.6




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 60.904 ng
RT: 17.545 min Scan# 2610
Delta R.T. 0.006 min
Lab File: BF142245.D
Acq: 30 Apr 2025 14:15

Tgt Ion:276 Resp: 1315066
Ion Ratio Lower Upper
276 100
277 23.2 18.7 28.1
138 24.7 19.4 29.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142246.D
 Acq On : 30 Apr 2025 14:43
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC080

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:46:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	192490	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	730774	20.000	ng	0.00
39) Acenaphthene-d10	9.951	164	381378	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	608907	20.000	ng	0.00
76) Chrysene-d12	14.080	240	323925	20.000	ng	0.00
86) Perylene-d12	15.568	264	374102	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.522	112	1741015	144.449	ng	0.00
7) Phenol-d6	6.534	99	2109624	144.946	ng	0.00
23) Nitrobenzene-d5	7.481	82	1929643	173.477	ng	0.00
42) 2,4,6-Tribromophenol	10.739	330	579431	166.657	ng	0.00
45) 2-Fluorobiphenyl	9.275	172	3399868	125.694	ng	0.00
79) Terphenyl-d14	13.027	244	2993632	134.413	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.699	88	408111	75.816	ng	99
3) Pyridine	3.469	79	1062667	75.959	ng	99
4) n-Nitrosodimethylamine	3.434	42	562539	77.702	ng	98
6) Aniline	6.575	93	1525082	74.030	ng	99
8) 2-Chlorophenol	6.692	128	942996	75.716	ng	99
9) Benzaldehyde	6.463	77	673391	66.197	ng	99
10) Phenol	6.551	94	1130209m	72.134	ng	
11) bis(2-Chloroethyl)ether	6.651	93	871905	71.808	ng	99
12) 1,3-Dichlorobenzene	6.851	146	1000058	70.979	ng	99
13) 1,4-Dichlorobenzene	6.928	146	997448	70.267	ng	99
14) 1,2-Dichlorobenzene	7.081	146	966587	71.326	ng	99
15) Benzyl Alcohol	7.051	79	787715	76.310	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.186	45	1573151	71.749	ng	99
17) 2-Methylphenol	7.157	107	757239	74.219	ng	100
18) Hexachloroethane	7.428	117	347325	74.757	ng	99
19) n-Nitroso-di-n-propyla...	7.334	70	650174	72.291	ng	99
20) 3+4-Methylphenols	7.316	107	897475	70.340	ng	# 87
22) Acetophenone	7.328	105	1180939	69.853	ng	97
24) Nitrobenzene	7.498	77	917286	84.046	ng	100
25) Isophorone	7.739	82	1758014	75.306	ng	99
26) 2-Nitrophenol	7.810	139	416387	79.254	ng	98
27) 2,4-Dimethylphenol	7.839	122	883922	75.098	ng	100
28) bis(2-Chloroethoxy)met...	7.945	93	1070383	72.702	ng	99
29) 2,4-Dichlorophenol	8.045	162	768570	78.103	ng	99
30) 1,2,4-Trichlorobenzene	8.133	180	797476	72.623	ng	100
31) Naphthalene	8.216	128	2508082	68.820	ng	99
32) Benzoic acid	7.969	122	523215m	83.328	ng	
33) 4-Chloroaniline	8.263	127	1084543	71.465	ng	99
34) Hexachlorobutadiene	8.333	225	471301	73.564	ng	100
35) Caprolactam	8.645	113	242414	81.958	ng	97
36) 4-Chloro-3-methylphenol	8.739	107	776483	76.844	ng	99
37) 2-Methylnaphthalene	8.910	142	1559858	69.615	ng	99
38) 1-Methylnaphthalene	9.010	142	1583955	68.237	ng	99
40) 1,2,4,5-Tetrachloroben...	9.075	216	757210	70.548	ng	100
41) Hexachlorocyclopentadiene	9.063	237	523186	85.886	ng	99
43) 2,4,6-Trichlorophenol	9.180	196	545358	81.996	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142246.D
 Acq On : 30 Apr 2025 14:43
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC080

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 15:46:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.216	196	557542	79.732	ng	99
46) 1,1'-Biphenyl	9.375	154	2006955	67.691	ng	99
47) 2-Chloronaphthalene	9.398	162	1547207	70.140	ng	99
48) 2-Nitroaniline	9.486	65	481954	79.390	ng	97
49) Acenaphthylene	9.816	152	2555385	68.650	ng	99
50) Dimethylphthalate	9.675	163	1757977	72.052	ng	99
51) 2,6-Dinitrotoluene	9.733	165	379127	78.246	ng	97
52) Acenaphthene	9.986	154	1517991	69.843	ng	99
53) 3-Nitroaniline	9.904	138	456571	78.061	ng	96
54) 2,4-Dinitrophenol	10.004	184	129817	79.532	ng	# 28
55) Dibenzofuran	10.157	168	2237085	68.664	ng	98
56) 4-Nitrophenol	10.045	139	335015	87.357	ng	97
57) 2,4-Dinitrotoluene	10.133	165	467545	78.915	ng	99
58) Fluorene	10.498	166	1634486	65.980	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.269	232	465050	80.286	ng	99
60) Diethylphthalate	10.375	149	1717661	71.292	ng	99
61) 4-Chlorophenyl-phenyle...	10.492	204	802224	68.190	ng	98
62) 4-Nitroaniline	10.522	138	410928	77.614	ng	98
63) Azobenzene	10.651	77	1674434	70.379	ng	99
65) 4,6-Dinitro-2-methylph...	10.545	198	183184	81.749	ng	96
66) n-Nitrosodiphenylamine	10.610	169	1518420	72.983	ng	99
67) 4-Bromophenyl-phenylether	10.980	248	514462	75.866	ng	97
68) Hexachlorobenzene	11.045	284	577203	75.999	ng	98
69) Atrazine	11.133	200	423886	77.296	ng	99
70) Pentachlorophenol	11.233	266	350009	87.235	ng	99
71) Phenanthrene	11.463	178	2288481	69.578	ng	99
72) Anthracene	11.516	178	2336551	69.482	ng	100
73) Carbazole	11.663	167	2087769	68.058	ng	99
74) Di-n-butylphthalate	11.998	149	2221417	72.002	ng	100
75) Fluoranthene	12.651	202	2137636	65.747	ng	98
77) Benzidine	12.768	184	1049467	76.247	ng	99
78) Pyrene	12.880	202	2118847	70.910	ng	99
80) Butylbenzylphthalate	13.498	149	659574	78.197	ng	98
81) Benzo(a)anthracene	14.068	228	1586778	72.970	ng	100
82) 3,3'-Dichlorobenzidine	14.033	252	563738	88.259	ng	99
83) Chrysene	14.104	228	1474094	74.470	ng	99
84) Bis(2-ethylhexyl)phtha...	14.057	149	915684	78.547	ng	100
85) Di-n-octyl phthalate	14.680	149	1722215	78.594	ng	100
87) Indeno(1,2,3-cd)pyrene	17.092	276	2128475	77.677	ng	99
88) Benzo(b)fluoranthene	15.133	252	1798083	77.378	ng	99
89) Benzo(k)fluoranthene	15.162	252	1439304	67.716	ng	100
90) Benzo(a)pyrene	15.509	252	1610193	75.685	ng	99
91) Dibenzo(a,h)anthracene	17.115	278	1712182	76.989	ng	99
92) Benzo(g,h,i)perylene	17.551	276	1738450	77.757	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142246.D
 Acq On : 30 Apr 2025 14:43
 Operator : RC/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC080

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025


Quant Time: Apr 30 15:46:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 15:28:40 2025
 Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080




Tgt Ion: 152 Resp: 192490
Ion Ratio Lower Upper
152 100
150 159.4 130.2 195.2
115 58.9 47.0 70.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#2
1,4-Dioxane
Concen: 75.816 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43



Tgt Ion: 88 Resp: 408111
Ion Ratio Lower Upper
88 100
58 90.8 72.2 108.2
43 34.2 27.0 40.6





#3

Pyridine

Concen: 75.959 ng

RT: 3.469 min Scan# 216

Delta R.T. 0.006 min

Lab File: BF142246.D


Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080



Tgt Ion: 79 Resp: 106266

Ion Ratio Lower Upper

79 100

52 88.8

51 42.4

71.5 107.3




34.5 51.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#4

n-Nitrosodimethylamine

Concen: 77.702 ng

RT: 3.434 min Scan# 211

Delta R.T. 0.023 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Tgt Ion: 42 Resp: 562539

Ion Ratio Lower Upper




42 100


74 113.7

44 7.5

89.5 134.3


6.6 9.8





#5
2-Fluorophenol
Concen: 144.449 ng
RT: 5.522 min Scan# 50
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion: 112 Resp: 174101
Ion Ratio Lower Upper
112 100
64 60.3 48.4 72.6
63 31.1 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#6
Aniline
Concen: 74.030 ng
RT: 6.575 min Scan# 745
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion: 93 Resp: 1525082
Ion Ratio Lower Upper
93 100
66 40.8 32.1 48.1
65 20.5 16.6 24.8





#7
Phenol-d6
Concen: 144.946 ng
RT: 6.534 min Scan# 738
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion: 99 Resp: 2109624
Ion Ratio Lower Upper
99 100
42 20.7 16.3 24.5
71 32.3 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#8
2-Chlorophenol
Concen: 75.716 ng
RT: 6.692 min Scan# 765
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:128 Resp: 942996
Ion Ratio Lower Upper
128 100
130 32.6 13.1 53.1
64 49.8 28.6 68.6





#9

Benzaldehyde

Concen: 66.197 ng

RT: 6.463 min Scan# 71

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion: 77 Resp: 67339

Ion Ratio Lower Upper

77 100

105 100.2 79.1 119.1





106 97.0 76.0 116.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#10

Phenol

Concen: 72.134 ng m

RT: 6.551 min Scan# 741

Delta R.T. 0.012 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion: 94 Resp: 1130209


Ion Ratio Lower Upper

94 100

65 30.1 8.9 48.9


66 41.8 18.7 58.7





#11
bis(2-Chloroethyl)ether
Concen: 71.808 ng
RT: 6.651 min Scan# 758
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion: 93 Resp: 871905
Ion Ratio Lower Upper
93 100
63 82.9 62.2 102.2
95 32.7 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#12
1,3-Dichlorobenzene
Concen: 70.979 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:146 Resp: 1000058
Ion Ratio Lower Upper
146 100
148 63.6 50.5 75.7
75 30.3 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 70.267 ng

RT: 6.928 min Scan# 805

Delta R.T. -0.000 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:146 Resp: 997448

Ion Ratio Lower Upper

146 100

148 63.7 50.6 76.0





111 42.7 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#14

1,2-Dichlorobenzene

Concen: 71.326 ng

RT: 7.081 min Scan# 831

Delta R.T. -0.000 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion:146 Resp: 966587


Ion Ratio Lower Upper

146 100

148 63.8 50.5 75.7

111 44.6 34.9 52.3





#15

Benzyl Alcohol

Concen: 76.310 ng

RT: 7.051 min Scan# 81

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion: 79 Resp: 787715

Ion Ratio Lower Upper

79 100

108 81.5 64.5 96.7





77 62.3 49.2 73.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 71.749 ng

RT: 7.186 min Scan# 849

Delta R.T. -0.000 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion: 45 Resp: 1573151


Ion Ratio Lower Upper

45 100

77 11.5 0.0 31.3


79 8.8 0.0 28.5





#17
2-Methylphenol
Concen: 74.219 ng
RT: 7.157 min Scan# 843
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:107 Resp: 757239
Ion Ratio Lower Upper
107 100
108 113.7 91.0 136.4
77 43.0 34.4 51.6
79 43.6 34.4 51.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#18
Hexachloroethane
Concen: 74.757 ng
RT: 7.428 min Scan# 890
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:117 Resp: 347325
Ion Ratio Lower Upper
117 100
119 96.3 77.1 115.7
201 97.4 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 72.291 ng

RT: 7.334 min Scan# 81

Delta R.T. 0.012 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion: 70 Resp: 650174

Ion Ratio Lower Upper

70 100

42 66.3 53.4 80.2

101 10.0 7.7 11.5





130 22.1 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#20

3+4-Methylphenols

Concen: 70.340 ng

RT: 7.316 min Scan# 871

Delta R.T. 0.012 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Tgt Ion:107 Resp: 897475




Ion Ratio Lower Upper


107 100

108 91.7 71.2 111.2

77 89.8 41.1 81.1#

79 28.5 8.1 48.1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.192 min Scan# 1020

Delta R.T. -0.001 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:136 Resp: 730774

Ion Ratio Lower Upper

136 100

137 10.8 8.7 13.1

54 10.0 7.7 11.5





68 6.4 5.0 7.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#22

Acetophenone

Concen: 69.853 ng

RT: 7.328 min Scan# 873

Delta R.T. 0.012 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Tgt Ion:105 Resp: 1180939




Ion Ratio Lower Upper


105 100

71 7.0 5.7 8.5

51 33.9 30.1 45.1

120 23.8 19.0 28.4





#23

Nitrobenzene-d5

Concen: 173.477 ng

RT: 7.481 min Scan# 899

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion: 82 Resp: 192964

Ion Ratio Lower Upper

82 100

128 47.0 37.0 55.6





54 59.7 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#24

Nitrobenzene

Concen: 84.046 ng

RT: 7.498 min Scan# 902

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion: 77 Resp: 917286


Ion Ratio Lower Upper

77 100

123 48.4 38.6 58.0


65 14.4 11.5 17.3





#25
Isophorone
Concen: 75.306 ng
RT: 7.739 min Scan# 941
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion: 82 Resp: 1758014
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 19.4 15.3 22.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#26
2-Nitrophenol
Concen: 79.254 ng
RT: 7.810 min Scan# 955
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:139 Resp: 416387
Ion Ratio Lower Upper
139 100
109 26.8 21.8 32.8
65 49.4 41.1 61.7





#27

2,4-Dimethylphenol

Concen: 75.098 ng

RT: 7.839 min Scan# 90

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :





SSTDICC080

Tgt Ion	Ratio	Lower	Upper
122	100		
107	108.1	86.0	129.0
121	57.8	46.3	69.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#28

bis(2-Chloroethoxy)methane

Concen: 72.702 ng




RT: 7.945 min Scan# 978


Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Tgt Ion	Ratio	Lower	Upper
93	100		
95	33.4	26.2	39.2
123	12.5	9.7	14.5





#29

2,4-Dichlorophenol

Concen: 78.103 ng

RT: 8.045 min Scan# 994

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:162 Resp: 768570

Ion Ratio Lower Upper

162 100

164 63.7 43.0 83.0





98 37.3 16.7 56.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#30

1,2,4-Trichlorobenzene

Concen: 72.623 ng

RT: 8.133 min Scan# 1010

Delta R.T. -0.000 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion:180 Resp: 797476


Ion Ratio Lower Upper

180 100

182 95.0 75.7 113.5


145 31.7 25.4 38.0





#31
Naphthalene
Concen: 68.820 ng
RT: 8.216 min Scan# 1024
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:128 Resp: 250808
Ion Ratio Lower Upper
128 100
129 11.6 9.0 13.4
127 13.7 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#32
Benzoic acid
Concen: 83.328 ng m
RT: 7.969 min Scan# 982
Delta R.T. 0.035 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:122 Resp: 523215
Ion Ratio Lower Upper
122 100
105 118.4 98.2 138.2
77 89.2 66.8 106.8





#33
4-Chloroaniline
Concen: 71.465 ng
RT: 8.263 min Scan# 1031
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:127 Resp: 108454
Ion Ratio Lower Upper
127 100
129 32.7 26.1 39.1
65 31.4 25.5 38.3
92 19.6 15.4 23.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#34
Hexachlorobutadiene
Concen: 73.564 ng
RT: 8.333 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:225 Resp: 471301
Ion Ratio Lower Upper
225 100
223 62.5 49.7 74.5
227 63.6 51.0 76.4





#35
Caprolactam
Concen: 81.958 ng
RT: 8.645 min Scan# 1097
Delta R.T. 0.023 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:113 Resp: 242414
Ion Ratio Lower Upper
113 100
55 194.1 177.7 217.7
56 142.5 127.1 167.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#36
4-Chloro-3-methylphenol
Concen: 76.844 ng
RT: 8.739 min Scan# 1113
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:107 Resp: 776483
Ion Ratio Lower Upper
107 100
144 26.9 21.0 31.4
142 82.9 66.1 99.1





#37
2-Methylnaphthalene
Concen: 69.615 ng
RT: 8.910 min Scan# 1141
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:142 Resp: 155985
Ion Ratio Lower Upper
142 100
141 87.0 70.5 105.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#38
1-Methylnaphthalene
Concen: 68.237 ng
RT: 9.010 min Scan# 1159
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:142 Resp: 1583955
Ion Ratio Lower Upper
142 100
141 91.8 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.951 min Scan# 1170

Delta R.T. 0.000 min

Lab File: BF142246.D


Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080



Tgt Ion:164 Resp: 381378

Ion Ratio Lower Upper

164 100

162 98.2 78.6 117.8




160 44.2 35.7 53.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#40

1,2,4,5-Tetrachlorobenzene


Concen: 70.548 ng

RT: 9.075 min Scan# 1170

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43



Tgt Ion:216 Resp: 757210



Ion Ratio Lower Upper


216 100

214 79.7 63.8 95.6

179 21.0 16.7 25.1

108 18.9 15.6 23.4





#41

Hexachlorocyclopentadiene

Concen: 85.886 ng

RT: 9.063 min Scan# 1168

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:237 Resp: 523180

Ion Ratio Lower Upper

237 100

235 64.0 43.5 83.5





272 13.5 0.0 32.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#42

2,4,6-Tribromophenol

Concen: 166.657 ng

RT: 10.739 min Scan# 1453

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion:330 Resp: 579431


Ion Ratio Lower Upper

330 100

332 97.1 76.5 114.7

141 28.6 24.3 36.5





#43

2,4,6-Trichlorophenol

Concen: 81.996 ng

RT: 9.180 min Scan# 1187

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:196 Resp: 545358

Ion Ratio Lower Upper

196 100

198 93.9 77.4 116.2





200 30.6 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#44

2,4,5-Trichlorophenol

Concen: 79.732 ng

RT: 9.216 min Scan# 1194

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Tgt Ion:196 Resp: 557542




Ion Ratio Lower Upper


196 100

198 95.4 75.2 112.8

97 50.7 39.8 59.8


132 29.8 23.8 35.8





#45
2-Fluorobiphenyl
Concen: 125.694 ng
RT: 9.275 min Scan# 1204
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080




Tgt Ion:172 Resp: 339986
Ion Ratio Lower Upper
172 100
171 36.9 29.0 43.4
170 24.6 19.4 29.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#46
1,1'-Biphenyl
Concen: 67.691 ng
RT: 9.375 min Scan# 1221
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43



Tgt Ion:154 Resp: 2006955
Ion Ratio Lower Upper
154 100
153 40.9 20.5 60.5
76 12.9 0.0 32.8





#47

2-Chloronaphthalene

Concen: 70.140 ng

RT: 9.398 min Scan# 1224

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:162 Resp: 154720

Ion Ratio Lower Upper

162 100

127 41.9 32.6 49.0





164 33.0 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#48

2-Nitroaniline

Concen: 79.390 ng

RT: 9.486 min Scan# 1240

Delta R.T. -0.001 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion: 65 Resp: 481954


Ion Ratio Lower Upper

65 100

92 67.1 54.3 81.5

138 106.4 88.6 133.0





#49

Acenaphthylene

Concen: 68.650 ng

RT: 9.816 min Scan# 1295

Delta R.T. 0.006 min

Lab File: BF142246.D


Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080



Tgt Ion:152 Resp: 255538

Ion Ratio Lower Upper

152 100

151 20.9 16.2 24.2




153 13.7 10.7 16.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#50

Dimethylphthalate


Concen: 72.052 ng

RT: 9.675 min Scan# 1272

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43





Tgt Ion:163 Resp: 1757977


Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8


164 10.5 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 78.246 ng
RT: 9.733 min Scan# 1281
Delta R.T. 0.005 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:165 Resp: 37912
Ion Ratio Lower Upper
165 100
63 58.4 48.5 72.7
89 52.7 43.9 65.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#52
Acenaphthene
Concen: 69.843 ng
RT: 9.986 min Scan# 1325
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:154 Resp: 1517991
Ion Ratio Lower Upper
154 100
153 110.2 88.6 133.0
152 53.1 42.1 63.1





#53
3-Nitroaniline
Concen: 78.061 ng
RT: 9.904 min Scan# 1311
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :





SSTDICC080

Tgt Ion:138 Resp: 45657
Ion Ratio Lower Upper
138 100
108 11.8 11.7 17.5
92 109.7 91.2 136.8

Manual Integrations




APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#54
2,4-Dinitrophenol
Concen: 79.532 ng
RT: 10.004 min Scan# 1328
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:184 Resp: 129817
Ion Ratio Lower Upper
184 100
63 73.9 69.2 103.8
154 153.5 271.4 407.2#





#55

Dibenzenofuran

Concen: 68.664 ng

RT: 10.157 min Scan# 11

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:168 Resp: 2237085

Ion Ratio Lower Upper

168 100

139 38.0 29.4 44.0





169 14.1 11.0 16.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#56

4-Nitrophenol

Concen: 87.357 ng

RT: 10.045 min Scan# 1335

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion:139 Resp: 335015


Ion Ratio Lower Upper

139 100

109 65.2 44.9 84.9


65 98.8 74.7 114.7





#57
2,4-Dinitrotoluene
Concen: 78.915 ng
RT: 10.133 min Scan# 11
Delta R.T. 0.005 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:165 Resp: 46754
Ion Ratio Lower Upper
165 100
63 49.4 38.8 58.2
89 72.5 57.1 85.7
182 2.6 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#58
Fluorene
Concen: 65.980 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:166 Resp: 1634486
Ion Ratio Lower Upper
166 100
165 93.0 73.9 110.9
167 13.9 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 80.286 ng
RT: 10.269 min Scan# 1373
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080




Tgt Ion:232 Resp: 465050
Ion Ratio Lower Upper
232 100
131 44.0 35.9 53.9
130 2.2 1.8 2.6
166 29.4 24.0 36.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#60
Diethylphthalate
Concen: 71.292 ng
RT: 10.375 min Scan# 1391
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Tgt Ion:149 Resp: 1717661
Ion Ratio Lower Upper
149 100
177 22.5 17.6 26.4
150 12.8 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 68.190 ng
RT: 10.492 min Scan# 1411
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080




Tgt Ion:204 Resp: 802224
Ion Ratio Lower Upper
204 100
206 32.4 25.6 38.4
141 58.7 48.9 73.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#62
4-Nitroaniline
Concen: 77.614 ng
RT: 10.522 min Scan# 1416
Delta R.T. 0.018 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Tgt Ion:138 Resp: 410928
Ion Ratio Lower Upper
138 100
92 48.3 29.2 69.2
108 67.6 49.1 89.1





#63
Azobenzene
Concen: 70.379 ng
RT: 10.651 min Scan# 1438
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080




Tgt Ion: 77 Resp: 1674434
Ion Ratio Lower Upper
77 100
182 26.0 7.3 47.3
105 22.6 2.5 42.5
51 34.6 14.4 54.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Tgt Ion:188 Resp: 608907
Ion Ratio Lower Upper
188 100
94 10.7 8.2 12.2
80 11.0 8.6 12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 81.749 ng
RT: 10.545 min Scan# 1418
Delta R.T. 0.011 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080




Tgt Ion:198 Resp: 183184
Ion Ratio Lower Upper
198 100
51 56.4 40.6 80.6
105 47.1 29.2 69.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#66
n-Nitrosodiphenylamine
Concen: 72.983 ng
RT: 10.610 min Scan# 1431
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Tgt Ion:169 Resp: 1518420
Ion Ratio Lower Upper
169 100
168 66.6 53.0 79.6
167 36.8 28.8 43.2





#67
4-Bromophenyl-phenylether
Concen: 75.866 ng
RT: 10.980 min Scan# 1494
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080




Tgt Ion:248 Resp: 514462
Ion Ratio Lower Upper
248 100
250 99.1 77.2 115.8
141 76.7 58.4 87.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#68
Hexachlorobenzene
Concen: 75.999 ng
RT: 11.045 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43



Tgt Ion:284 Resp: 577203
Ion Ratio Lower Upper
284 100
142 31.8 24.6 37.0
249 30.0 23.2 34.8





#69

Atrazine

Concen: 77.296 ng

RT: 11.133 min Scan# 1520

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:200 Resp: 42388

Ion Ratio Lower Upper

200 100

173 28.7 8.6 48.6





215 49.0 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#70

Pentachlorophenol

Concen: 87.235 ng

RT: 11.233 min Scan# 1537

Delta R.T. 0.000 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion:266 Resp: 350009


Ion Ratio Lower Upper

266 100

268 62.7 50.0 75.0


264 63.2 49.3 73.9





#71
Phenanthrene
Concen: 69.578 ng
RT: 11.463 min Scan# 1575
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:178 Resp: 228848
Ion Ratio Lower Upper
178 100
176 20.1 15.5 23.3
179 16.1 12.6 19.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#72
Anthracene
Concen: 69.482 ng
RT: 11.516 min Scan# 1585
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:178 Resp: 2336551
Ion Ratio Lower Upper
178 100
176 19.5 15.3 22.9
179 15.9 12.7 19.1





#73
Carbazole
Concen: 68.058 ng
RT: 11.663 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080




Tgt Ion:167 Resp: 2087769
Ion Ratio Lower Upper
167 100
166 21.8 17.3 25.9
139 13.4 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 72.002 ng
RT: 11.998 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Tgt Ion:149 Resp: 2221417
Ion Ratio Lower Upper
149 100
150 9.6 7.5 11.3
104 4.6 3.6 5.4





#75
Fluoranthene
Concen: 65.747 ng
RT: 12.651 min Scan# 1778
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:202 Resp: 2137630
Ion Ratio Lower Upper
202 100
101 11.8 0.0 32.4
203 18.2 0.0 37.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.080 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:240 Resp: 323925
Ion Ratio Lower Upper
240 100
120 12.2 10.1 15.1
236 25.4 20.5 30.7





#77
Benzidine
Concen: 76.247 ng
RT: 12.768 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:184 Resp: 104946
Ion Ratio Lower Upper
184 100
185 15.3 11.8 17.6
183 11.0 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#78
Pyrene
Concen: 70.910 ng
RT: 12.880 min Scan# 1817
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:202 Resp: 2118847
Ion Ratio Lower Upper
202 100
200 20.7 16.1 24.1
203 18.3 14.1 21.1





#79

Terphenyl-d14

Concen: 134.413 ng

RT: 13.027 min Scan# 1841

Delta R.T. 0.006 min

Lab File: BF142246.D


Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080



Tgt Ion:244 Resp: 299363

Ion Ratio Lower Upper

244 100

212 6.7 5.3 7.9




122 10.4 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#80

Butylbenzylphthalate

Concen: 78.197 ng

RT: 13.498 min Scan# 1922

Delta R.T. -0.000 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion:149 Resp: 659574


Ion Ratio Lower Upper

149 100

91 70.0 54.4 81.6

206 20.7 15.9 23.9





#81

Benzo(a)anthracene

Concen: 72.970 ng

RT: 14.068 min Scan# 2018

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:228 Resp: 1586778

Ion Ratio Lower Upper

228 100

226 27.5 21.9 32.9





229 20.2 16.0 24.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#82

3,3'-Dichlorobenzidine

Concen: 88.259 ng

RT: 14.033 min Scan# 2013

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion:252 Resp: 563738


Ion Ratio Lower Upper

252 100

254 64.9 51.0 76.6


126 11.5 9.4 14.2





#83
Chrysene
Concen: 74.470 ng
RT: 14.104 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:228 Resp: 1474094
Ion Ratio Lower Upper
228 100
226 30.5 24.0 36.0
229 20.2 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 78.547 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43


Tgt Ion:149 Resp: 915684
Ion Ratio Lower Upper
149 100
167 27.3 21.7 32.5
279 3.9 3.0 4.4





#85
Di-n-octyl phthalate
Concen: 78.594 ng
RT: 14.680 min Scan# 2123
Delta R.T. 0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:149 Resp: 1722215
Ion Ratio Lower Upper
149 100
167 1.5 1.1 1.7
43 11.4 9.3 13.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:264 Resp: 374102
Ion Ratio Lower Upper
264 100
260 23.6 18.4 27.6
265 21.8 17.3 25.9






#87
Indeno(1,2,3-cd)pyrene
Concen: 77.677 ng
RT: 17.092 min Scan# 211
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080






Tgt Ion:276 Resp: 2128475
Ion Ratio Lower Upper
276 100
138 28.4 22.6 33.8
277 25.4 20.1 30.1

Manual Integrations




APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#88
Benzo(b)fluoranthene
Concen: 77.378 ng
RT: 15.133 min Scan# 2200
Delta R.T. 0.006 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:252 Resp: 1798083
Ion Ratio Lower Upper
252 100
253 22.2 17.7 26.5
125 10.6 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 67.716 ng

RT: 15.162 min Scan# 2205

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43

Instrument :

BNA_F

ClientSampleId :

SSTDICC080

Tgt Ion:252 Resp: 1439304

Ion Ratio Lower Upper

252 100

253 22.2 17.8 26.8





125 11.2 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#90

Benzo(a)pyrene

Concen: 75.685 ng

RT: 15.509 min Scan# 2264

Delta R.T. 0.006 min

Lab File: BF142246.D

Acq: 30 Apr 2025 14:43




Tgt Ion:252 Resp: 1610193


Ion Ratio Lower Upper

252 100

253 22.4 17.6 26.4


125 12.2 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 76.989 ng
RT: 17.115 min Scan# 2611
Delta R.T. 0.017 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43




Instrument :
BNA_F
ClientSampleId :
SSTDICC080



Tgt Ion:278 Resp: 171218
Ion Ratio Lower Upper
278 100
139 17.9 15.2 22.8
279 24.1 19.0 28.6




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 77.757 ng
RT: 17.551 min Scan# 2611
Delta R.T. 0.012 min
Lab File: BF142246.D
Acq: 30 Apr 2025 14:43

Tgt Ion:276 Resp: 1738450
Ion Ratio Lower Upper
276 100
277 23.9 18.7 28.1
138 25.1 19.4 29.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142247.D
 Acq On : 30 Apr 2025 16:17
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF043025

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 16:36:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	185327	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	725482	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	381170	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	633040	20.000	ng	0.00
76) Chrysene-d12	14.074	240	345729	20.000	ng	0.00
86) Perylene-d12	15.562	264	342826	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	987704	85.116	ng	0.00
7) Phenol-d6	6.522	99	1201328	85.730	ng	0.00
23) Nitrobenzene-d5	7.469	82	1091522	98.845	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	323319	93.044	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2157499	79.807	ng	0.00
79) Terphenyl-d14	13.021	244	2037561	85.716	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.699	88	219939	42.438	ng	98
3) Pyridine	3.469	79	572034	42.469	ng	99
4) n-Nitrosodimethylamine	3.416	42	299316	42.942	ng	96
6) Aniline	6.569	93	844910	42.598	ng	99
8) 2-Chlorophenol	6.692	128	530021	44.202	ng	98
9) Benzaldehyde	6.457	77	420408	42.925	ng	99
10) Phenol	6.539	94	645155m	42.765	ng	
11) bis(2-Chloroethyl)ether	6.639	93	494843	42.329	ng	98
12) 1,3-Dichlorobenzene	6.851	146	567280	41.819	ng	99
13) 1,4-Dichlorobenzene	6.928	146	572974	41.924	ng	100
14) 1,2-Dichlorobenzene	7.081	146	551216	42.247	ng	99
15) Benzyl Alcohol	7.045	79	431108	43.378	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.181	45	901053	42.684	ng	99
17) 2-Methylphenol	7.151	107	415403	42.288	ng	98
18) Hexachloroethane	7.422	117	197959	44.255	ng	95
19) n-Nitroso-di-n-propyla...	7.322	70	363621	41.992	ng	99
20) 3+4-Methylphenols	7.304	107	539514	43.919	ng	98
22) Acetophenone	7.316	105	695539	41.441	ng	98
24) Nitrobenzene	7.492	77	513604	47.402	ng	99
25) Isophorone	7.728	82	968502	41.789	ng	100
26) 2-Nitrophenol	7.804	139	219881	48.175	ng	100
27) 2,4-Dimethylphenol	7.833	122	502502	43.004	ng	100
28) bis(2-Chloroethoxy)met...	7.939	93	610669	41.780	ng	99
29) 2,4-Dichlorophenol	8.039	162	429967	44.013	ng	99
30) 1,2,4-Trichlorobenzene	8.133	180	454848	41.724	ng	100
31) Naphthalene	8.216	128	1483568	41.005	ng	100
32) Benzoic acid	7.933	122	246172m	45.113	ng	
33) 4-Chloroaniline	8.257	127	622894	41.344	ng	100
34) Hexachlorobutadiene	8.333	225	267940	42.127	ng	99
35) Caprolactam	8.622	113	130473	44.434	ng	98
36) 4-Chloro-3-methylphenol	8.728	107	436215	43.485	ng	100
37) 2-Methylnaphthalene	8.904	142	915474	41.155	ng	100
38) 1-Methylnaphthalene	9.004	142	950163	41.232	ng	100
40) 1,2,4,5-Tetrachloroben...	9.069	216	443561	41.349	ng	100
41) Hexachlorocyclopentadiene	9.057	237	288102	47.321	ng	100
43) 2,4,6-Trichlorophenol	9.175	196	303872	45.713	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142247.D
 Acq On : 30 Apr 2025 16:17
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF043025

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/01/2025
 Supervised By :Jagrut Upadhyay 05/01/2025

Quant Time: Apr 30 16:36:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	312237	44.676	ng	99
46) 1,1'-Biphenyl	9.369	154	1213047	40.936	ng	100
47) 2-Chloronaphthalene	9.392	162	903974	41.003	ng	99
48) 2-Nitroaniline	9.480	65	269528	46.523	ng	95
49) Acenaphthylene	9.810	152	1533164	41.211	ng	100
50) Dimethylphthalate	9.669	163	1025498	42.054	ng	100
51) 2,6-Dinitrotoluene	9.727	165	211107	45.376	ng	96
52) Acenaphthene	9.980	154	904325	41.631	ng	99
53) 3-Nitroaniline	9.892	138	257926	45.509	ng	96
54) 2,4-Dinitrophenol	9.992	184	62568	49.090	ng	# 6
55) Dibenzofuran	10.151	168	1340934	41.181	ng	100
56) 4-Nitrophenol	10.033	139	191463	49.952	ng	99
57) 2,4-Dinitrotoluene	10.127	165	265412	46.994	ng	96
58) Fluorene	10.492	166	1008032	40.714	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	264709	45.724	ng	100
60) Diethylphthalate	10.369	149	1016722	42.222	ng	100
61) 4-Chlorophenyl-phenyle...	10.486	204	477688	40.626	ng	99
62) 4-Nitroaniline	10.504	138	232165	45.070	ng	98
63) Azobenzene	10.645	77	984443	41.400	ng	99
65) 4,6-Dinitro-2-methylph...	10.533	198	96341	46.598	ng	95
66) n-Nitrosodiphenylamine	10.604	169	904466	41.816	ng	100
67) 4-Bromophenyl-phenylether	10.980	248	299605	42.497	ng	98
68) Hexachlorobenzene	11.039	284	330048	41.800	ng	97
69) Atrazine	11.127	200	255030	44.732	ng	99
70) Pentachlorophenol	11.233	266	197476	47.342	ng	99
71) Phenanthrene	11.457	178	1400439	40.955	ng	100
72) Anthracene	11.510	178	1463247	41.854	ng	100
73) Carbazole	11.663	167	1332856	41.793	ng	99
74) Di-n-butylphthalate	11.992	149	1412919	44.051	ng	100
75) Fluoranthene	12.645	202	1409189	41.690	ng	99
77) Benzidine	12.768	184	682155	46.435	ng	99
78) Pyrene	12.874	202	1405875	44.082	ng	100
80) Butylbenzylphthalate	13.492	149	371146	42.984	ng	98
81) Benzo(a)anthracene	14.062	228	960484	41.383	ng	99
82) 3,3'-Dichlorobenzidine	14.027	252	310153	45.495	ng	98
83) Chrysene	14.098	228	890671	42.158	ng	100
84) Bis(2-ethylhexyl)phtha...	14.057	149	487122	40.882	ng	99
85) Di-n-octyl phthalate	14.674	149	889844	41.178	ng	100
87) Indeno(1,2,3-cd)pyrene	17.074	276	1098774	43.757	ng	99
88) Benzo(b)fluoranthene	15.121	252	860370	40.403	ng	100
89) Benzo(k)fluoranthene	15.157	252	811540	41.665	ng	99
90) Benzo(a)pyrene	15.498	252	826250	42.380	ng	99
91) Dibenzo(a,h)anthracene	17.098	278	887855	43.565	ng	99
92) Benzo(g,h,i)perylene	17.527	276	899474	43.902	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\  
Data File : BF142247.D  
Acq On    : 30 Apr 2025  16:17  
Operator  : RC/JU  
Sample    : SSTDICV040  
Misc      :  
ALS Vial  : 11    Sample Multiplier: 1
```


Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Manual Integrations

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025


Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:152 Resp: 185327
Ion Ratio Lower Upper
152 100
150 162.7 130.2 195.2
115 59.5 47.0 70.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#2
1,4-Dioxane
Concen: 42.438 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 88 Resp: 219939
Ion Ratio Lower Upper
88 100
58 91.6 72.2 108.2
43 34.8 27.0 40.6





#3

Pyridine

Concen: 42.469 ng

RT: 3.469 min Scan# 216

Delta R.T. 0.006 min

Lab File: BF142247.D


Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025



Tgt Ion: 79 Resp: 572034




Ion	Ratio	Lower	Upper
79	100		
52	88.3	71.5	107.3
51	41.9	34.5	51.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#4

n-Nitrosodimethylamine


Concen: 42.942 ng

RT: 3.416 min Scan# 208

Delta R.T. 0.006 min



Lab File: BF142247.D


Acq: 30 Apr 2025 16:17



Tgt Ion: 42 Resp: 299316


Ion	Ratio	Lower	Upper
42	100		
74	116.5	89.5	134.3
44	7.9	6.6	9.8





#5
2-Fluorophenol
Concen: 85.116 ng
RT: 5.516 min Scan# 565
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion: 112 Resp: 987704
Ion Ratio Lower Upper
112 100
64 60.5 48.4 72.6
63 31.2 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#6
Aniline
Concen: 42.598 ng
RT: 6.569 min Scan# 744
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion: 93 Resp: 844910
Ion Ratio Lower Upper
93 100
66 40.4 32.1 48.1
65 20.3 16.6 24.8





#7

Phenol-d6

Concen: 85.730 ng

RT: 6.522 min Scan# 736

Delta R.T. -0.006 min

Lab File: BF142247.D


Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025



Tgt Ion: 99 Resp: 1201328

Ion Ratio Lower Upper

99 100

42 21.4 16.3 24.5




71 32.2 25.0 37.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#8

2-Chlorophenol

Concen: 44.202 ng

RT: 6.692 min Scan# 765

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion:128 Resp: 530021


Ion Ratio Lower Upper

128 100

130 32.7 13.1 53.1

64 47.1 28.6 68.6





#9

Benzaldehyde

Concen: 42.925 ng

RT: 6.457 min Scan# 725

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025

Tgt Ion: 77 Resp: 420408

Ion Ratio Lower Upper

77 100

105 98.5 79.1 119.1





106 95.0 76.0 116.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#10

Phenol

Concen: 42.765 ng m

RT: 6.539 min Scan# 739

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion: 94 Resp: 645155


Ion Ratio Lower Upper

94 100

65 29.1 8.9 48.9

66 38.7 18.7 58.7









#11
bis(2-Chloroethyl)ether
Concen: 42.329 ng
RT: 6.639 min Scan# 756
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Tgt Ion: 93 Resp: 49484
Ion Ratio Lower Upper
93 100
63 84.5 62.2 102.2
95 32.4 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#12
1,3-Dichlorobenzene
Concen: 41.819 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:146 Resp: 567280
Ion Ratio Lower Upper
146 100
148 63.8 50.5 75.7
75 29.2 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 41.924 ng

RT: 6.928 min Scan# 805

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025

Tgt Ion:146 Resp: 572974

Ion Ratio Lower Upper

146 100

148 63.3 50.6 76.0





111 41.8 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#14

1,2-Dichlorobenzene

Concen: 42.247 ng

RT: 7.081 min Scan# 831

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion:146 Resp: 551216


Ion Ratio Lower Upper

146 100

148 64.4 50.5 75.7


111 43.8 34.9 52.3





#15
Benzyl Alcohol
Concen: 43.378 ng
RT: 7.045 min Scan# 811
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion: 79 Resp: 431108
Ion Ratio Lower Upper
79 100
108 81.5 64.5 96.7
77 62.1 49.2 73.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#16
2,2'-oxybis(1-chloropropane)
Concen: 42.684 ng
RT: 7.181 min Scan# 848
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion: 45 Resp: 901053
Ion Ratio Lower Upper
45 100
77 10.9 0.0 31.3
79 8.2 0.0 28.5





#17
2-Methylphenol
Concen: 42.288 ng
RT: 7.151 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:107 Resp: 41540
Ion Ratio Lower Upper
107 100
108 116.0 91.0 136.4
77 43.7 34.4 51.6
79 42.7 34.4 51.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#18
Hexachloroethane
Concen: 44.255 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:117 Resp: 197959
Ion Ratio Lower Upper
117 100
119 96.3 77.1 115.7
201 89.8 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 41.992 ng

RT: 7.322 min Scan# 872

Delta R.T. -0.000 min

Lab File: BF142247.D


Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025



Tgt Ion: 70 Resp: 36362

Ion Ratio Lower Upper

70 100

42 66.9 53.4 80.2

101 9.7 7.7 11.5




130 22.4 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#20

3+4-Methylphenols

Concen: 43.919 ng

RT: 7.304 min Scan# 869

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Tgt Ion:107 Resp: 539514




Ion Ratio Lower Upper


107 100

108 90.2 71.2 111.2

77 63.5 41.1 81.1


79 27.5 8.1 48.1





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1020
Delta R.T. -0.001 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:136 Resp: 72548
Ion Ratio Lower Upper
136 100
137 11.2 8.7 13.1
54 9.7 7.7 11.5
68 6.1 5.0 7.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#22
Acetophenone
Concen: 41.441 ng
RT: 7.316 min Scan# 871
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:105 Resp: 695539
Ion Ratio Lower Upper
105 100
71 7.3 5.7 8.5
51 35.9 30.1 45.1
120 23.5 19.0 28.4





#23

Nitrobenzene-d5

Concen: 98.845 ng

RT: 7.469 min Scan# 898

Delta R.T. -0.006 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025

Tgt Ion: 82 Resp: 109152

Ion Ratio Lower Upper

82 100

128 45.7 37.0 55.6





54 61.2 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#24

Nitrobenzene

Concen: 47.402 ng

RT: 7.492 min Scan# 901

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion: 77 Resp: 513604


Ion Ratio Lower Upper

77 100

123 48.8 38.6 58.0

65 14.5 11.5 17.3





#25

Isophorone

Concen: 41.789 ng

RT: 7.728 min Scan# 941

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :





ICVBF043025

Tgt Ion:	82	Resp:	968502
Ion Ratio	Lower	Upper	
82	100		
95	7.0	5.6	8.4
138	19.2	15.3	22.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#26

2-Nitrophenol

Concen: 48.175 ng




RT: 7.804 min Scan# 954


Delta R.T. 0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Tgt Ion:	139	Resp:	219881
Ion Ratio	Lower	Upper	
139	100		
109	27.6	21.8	32.8
65	51.2	41.1	61.7





#27

2,4-Dimethylphenol

Concen: 43.004 ng

RT: 7.833 min Scan# 91

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025





Tgt Ion	122	Resp	50250
Ion Ratio	100		
Lower			
Upper			
107	108.0	86.0	129.0
121	58.1	46.3	69.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#28

bis(2-Chloroethoxy)methane

Concen: 41.780 ng




RT: 7.939 min Scan# 977


Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Tgt Ion	93	Resp	610669
Ion Ratio <td>100</td> <td></td> <td></td>	100		
Lower			
Upper			
95	33.1	26.2	39.2
123	12.6	9.7	14.5





#29

2,4-Dichlorophenol

Concen: 44.013 ng

RT: 8.039 min Scan# 994

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :





ICVBF043025

Tgt Ion:162 Resp: 42996
Ion Ratio Lower Upper
162 100
164 63.9 43.0 83.0
98 37.3 16.7 56.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#30

1,2,4-Trichlorobenzene

Concen: 41.724 ng




RT: 8.133 min Scan# 1010


Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17


Tgt Ion:180 Resp: 454848
Ion Ratio Lower Upper
180 100
182 94.8 75.7 113.5
145 31.1 25.4 38.0





#31
Naphthalene
Concen: 41.005 ng
RT: 8.216 min Scan# 1024
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:128 Resp: 148356
Ion Ratio Lower Upper
128 100
129 11.3 9.0 13.4
127 13.3 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#32
Benzoic acid
Concen: 45.113 ng m
RT: 7.933 min Scan# 976
Delta R.T. -0.001 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:122 Resp: 246172
Ion Ratio Lower Upper
122 100
105 117.2 98.2 138.2
77 86.9 66.8 106.8





#33
4-Chloroaniline
Concen: 41.344 ng
RT: 8.257 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:127 Resp: 622894
Ion Ratio Lower Upper
127 100
129 32.7 26.1 39.1
65 32.0 25.5 38.3
92 19.0 15.4 23.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#34
Hexachlorobutadiene
Concen: 42.127 ng
RT: 8.333 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:225 Resp: 267940
Ion Ratio Lower Upper
225 100
223 63.0 49.7 74.5
227 63.9 51.0 76.4





#35
Caprolactam
Concen: 44.434 ng
RT: 8.622 min Scan# 1093
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:113 Resp: 13047
Ion Ratio Lower Upper
113 100
55 199.6 177.7 217.7
56 144.7 127.1 167.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#36
4-Chloro-3-methylphenol
Concen: 43.485 ng
RT: 8.728 min Scan# 1111
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:107 Resp: 436215
Ion Ratio Lower Upper
107 100
144 26.7 21.0 31.4
142 82.5 66.1 99.1





#37
2-Methylnaphthalene
Concen: 41.155 ng
RT: 8.904 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:142 Resp: 915474
Ion Ratio Lower Upper
142 100
141 87.6 70.5 105.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#38
1-Methylnaphthalene
Concen: 41.232 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:142 Resp: 950163
Ion Ratio Lower Upper
142 100
141 90.9 72.8 109.2





#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.945 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:164 Resp: 381170
Ion Ratio Lower Upper
164 100
162 98.0 78.6 117.8
160 44.5 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#40
1,2,4,5-Tetrachlorobenzene
Concen: 41.349 ng
RT: 9.069 min Scan# 1169
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:216 Resp: 443561
Ion Ratio Lower Upper
216 100
214 79.6 63.8 95.6
179 21.0 16.7 25.1
108 19.2 15.6 23.4





#41
Hexachlorocyclopentadiene
Concen: 47.321 ng
RT: 9.057 min Scan# 1167
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:237 Resp: 28810
Ion Ratio Lower Upper
237 100
235 63.3 43.5 83.5
272 13.2 0.0 32.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#42
2,4,6-Tribromophenol
Concen: 93.044 ng
RT: 10.733 min Scan# 1452
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:330 Resp: 323319
Ion Ratio Lower Upper
330 100
332 96.9 76.5 114.7
141 31.4 24.3 36.5





#43

2,4,6-Trichlorophenol

Concen: 45.713 ng

RT: 9.175 min Scan# 1187

Delta R.T. -0.000 min

Lab File: BF142247.D


Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025



Tgt Ion:196 Resp: 303872

Ion Ratio Lower Upper

196 100

198 95.5 77.4 116.2




200 30.1 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#44

2,4,5-Trichlorophenol


Concen: 44.676 ng

RT: 9.210 min Scan# 1193

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17



Tgt Ion:196 Resp: 312237



Ion Ratio Lower Upper


196 100

198 96.3 75.2 112.8

97 49.7 39.8 59.8


132 29.6 23.8 35.8





#45
2-Fluorobiphenyl
Concen: 79.807 ng
RT: 9.269 min Scan# 1203
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:172 Resp: 2157499
Ion Ratio Lower Upper
172 100
171 36.1 29.0 43.4
170 24.0 19.4 29.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#46
1,1'-Biphenyl
Concen: 40.936 ng
RT: 9.369 min Scan# 1220
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:154 Resp: 1213047
Ion Ratio Lower Upper
154 100
153 40.7 20.5 60.5
76 12.6 0.0 32.8





#47

2-Chloronaphthalene

Concen: 41.003 ng

RT: 9.392 min Scan# 1

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025

Tgt Ion:162 Resp: 903974

Ion Ratio Lower Upper

162 100

127 41.2 32.6 49.0





164 32.8 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#48

2-Nitroaniline

Concen: 46.523 ng

RT: 9.480 min Scan# 1239

Delta R.T. -0.007 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion: 65 Resp: 269528


Ion Ratio Lower Upper

65 100

92 65.5 54.3 81.5

138 104.3 88.6 133.0





#49

Acenaphthylene

Concen: 41.211 ng

RT: 9.810 min Scan# 1295

Delta R.T. -0.000 min

Lab File: BF142247.D


Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025



Tgt Ion:152 Resp: 153316

Ion Ratio Lower Upper

152 100

151 20.2 16.2 24.2




153 13.4 10.7 16.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#50

Dimethylphthalate

Concen: 42.054 ng

RT: 9.669 min Scan# 1271

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion:163 Resp: 1025498


Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8


164 10.1 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 45.376 ng
RT: 9.727 min Scan# 1281
Delta R.T. -0.001 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025




Tgt Ion:165 Resp: 21110
Ion Ratio Lower Upper
165 100
63 58.0 48.5 72.7
89 52.2 43.9 65.9

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#52
Acenaphthene
Concen: 41.631 ng
RT: 9.980 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17



Tgt Ion:154 Resp: 904325
Ion Ratio Lower Upper
154 100
153 110.3 88.6 133.0
152 53.0 42.1 63.1





#53

3-Nitroaniline

Concen: 45.509 ng

RT: 9.892 min Scan# 1309

Delta R.T. 0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025

Tgt Ion:138 Resp: 257920

Ion Ratio Lower Upper

138 100

108 11.7 11.7 17.5





92 109.7 91.2 136.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#54

2,4-Dinitrophenol

Concen: 49.090 ng

RT: 9.992 min Scan# 1326

Delta R.T. -0.006 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion:184 Resp: 62568


Ion Ratio Lower Upper

184 100

63 109.3 69.2 103.8#

154 577.6 271.4 407.2#





#55

Dibenzenofuran

Concen: 41.181 ng

RT: 10.151 min Scan# 1353

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025

Tgt Ion:168 Resp: 1340934

Ion Ratio Lower Upper

168 100

139 36.8 29.4 44.0





169 13.7 11.0 16.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#56

4-Nitrophenol

Concen: 49.952 ng

RT: 10.033 min Scan# 1333

Delta R.T. -0.006 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion:139 Resp: 191463


Ion Ratio Lower Upper

139 100

109 65.2 44.9 84.9

65 96.4 74.7 114.7






#57
2,4-Dinitrotoluene
Concen: 46.994 ng
RT: 10.127 min Scan# 11
Delta R.T. -0.001 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025






Tgt Ion:165 Resp: 265412
Ion Ratio Lower Upper
165 100
63 45.6 38.8 58.2
89 68.8 57.1 85.7
182 2.8 2.2 3.2

Manual Integrations




APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#58
Fluorene
Concen: 40.714 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:166 Resp: 1008032
Ion Ratio Lower Upper
166 100
165 93.0 73.9 110.9
167 13.7 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 45.724 ng
RT: 10.263 min Scan# 1372
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:232 Resp: 264709
Ion Ratio Lower Upper
232 100
131 44.8 35.9 53.9
130 2.3 1.8 2.6
166 29.7 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#60
Diethylphthalate
Concen: 42.222 ng
RT: 10.369 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:149 Resp: 1016722
Ion Ratio Lower Upper
149 100
177 22.1 17.6 26.4
150 12.7 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 40.626 ng
RT: 10.486 min Scan# 1410
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025




Tgt Ion:204 Resp: 477688
Ion Ratio Lower Upper
204 100
206 32.7 25.6 38.4
141 61.2 48.9 73.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#62
4-Nitroaniline
Concen: 45.070 ng
RT: 10.504 min Scan# 1413
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Tgt Ion:138 Resp: 232165
Ion Ratio Lower Upper
138 100
92 48.8 29.2 69.2
108 67.1 49.1 89.1





#63
Azobenzene
Concen: 41.400 ng
RT: 10.645 min Scan# 1437
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion: 77 Resp: 98444
Ion Ratio Lower Upper
77 100
182 25.8 7.3 47.3
105 22.4 2.5 42.5
51 34.9 14.4 54.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:188 Resp: 633040
Ion Ratio Lower Upper
188 100
94 10.0 8.2 12.2
80 10.8 8.6 12.8









#65
4,6-Dinitro-2-methylphenol
Concen: 46.598 ng
RT: 10.533 min Scan# 1418
Delta R.T. -0.001 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Tgt Ion:198 Resp: 9634
Ion Ratio Lower Upper
198 100
51 56.1 40.6 80.6
105 46.7 29.2 69.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#66
n-Nitrosodiphenylamine
Concen: 41.816 ng
RT: 10.604 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:169 Resp: 904466
Ion Ratio Lower Upper
169 100
168 66.4 53.0 79.6
167 36.3 28.8 43.2





#67
4-Bromophenyl-phenylether
Concen: 42.497 ng
RT: 10.980 min Scan# 1494
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:248 Resp: 299608
Ion Ratio Lower Upper
248 100
250 96.6 77.2 115.8
141 70.0 58.4 87.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#68
Hexachlorobenzene
Concen: 41.800 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:284 Resp: 330048
Ion Ratio Lower Upper
284 100
142 33.5 24.6 37.0
249 30.1 23.2 34.8





#69

Atrazine

Concen: 44.732 ng

RT: 11.127 min Scan# 1519

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025

Tgt Ion:200 Resp: 255030

Ion Ratio Lower Upper

200 100

173 27.9 8.6 48.6





215 48.4 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#70

Pentachlorophenol

Concen: 47.342 ng

RT: 11.233 min Scan# 1537

Delta R.T. 0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion:266 Resp: 197476


Ion Ratio Lower Upper

266 100

268 62.3 50.0 75.0


264 62.9 49.3 73.9





#71
Phenanthrene
Concen: 40.955 ng
RT: 11.457 min Scan# 1575
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:178 Resp: 1400439
Ion Ratio Lower Upper
178 100
176 19.2 15.5 23.3
179 15.6 12.6 19.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#72
Anthracene
Concen: 41.854 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:178 Resp: 1463247
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.8 12.7 19.1





#73
Carbazole
Concen: 41.793 ng
RT: 11.663 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025




Tgt Ion:167 Resp: 1332850
Ion Ratio Lower Upper
167 100
166 21.4 17.3 25.9
139 12.5 10.3 15.5

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#74
Di-n-butylphthalate
Concen: 44.051 ng
RT: 11.992 min Scan# 1666
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Tgt Ion:149 Resp: 1412919
Ion Ratio Lower Upper
149 100
150 9.4 7.5 11.3
104 4.6 3.6 5.4





#75
Fluoranthene
Concen: 41.690 ng
RT: 12.645 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:202 Resp: 1409189
Ion Ratio Lower Upper
202 100
101 12.0 0.0 32.4
203 17.8 0.0 37.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:240 Resp: 345729
Ion Ratio Lower Upper
240 100
120 12.1 10.1 15.1
236 25.9 20.5 30.7





#77
Benzidine
Concen: 46.435 ng
RT: 12.768 min Scan# 1798
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:184 Resp: 68215
Ion Ratio Lower Upper
184 100
185 15.3 11.8 17.6
183 10.9 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#78
Pyrene
Concen: 44.082 ng
RT: 12.874 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:202 Resp: 1405875
Ion Ratio Lower Upper
202 100
200 20.2 16.1 24.1
203 17.8 14.1 21.1





#79

Terphenyl-d14

Concen: 85.716 ng

RT: 13.021 min Scan# 1841

Delta R.T. -0.000 min

Lab File: BF142247.D


Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025



Tgt Ion:244 Resp: 203756

Ion Ratio Lower Upper

244 100

212 6.6 5.3 7.9




122 10.9 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#80

Butylbenzylphthalate


Concen: 42.984 ng

RT: 13.492 min Scan# 1921

Delta R.T. -0.006 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17





Tgt Ion:149 Resp: 371146


Ion Ratio Lower Upper

149 100

91 69.8 54.4 81.6


206 19.9 15.9 23.9





#81
Benzo(a)anthracene
Concen: 41.383 ng
RT: 14.062 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025




Tgt Ion:228 Resp: 960484
Ion Ratio Lower Upper
228 100
226 27.0 21.9 32.9
229 19.6 16.0 24.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025




#82
3,3'-Dichlorobenzidine
Concen: 45.495 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Tgt Ion:252 Resp: 310153
Ion Ratio Lower Upper
252 100
254 65.0 51.0 76.6
126 11.3 9.4 14.2





#83
Chrysene
Concen: 42.158 ng
RT: 14.098 min Scan# 2025
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:228 Resp: 89067
Ion Ratio Lower Upper
228 100
226 30.0 24.0 36.0
229 19.8 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 40.882 ng
RT: 14.057 min Scan# 2017
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17


Tgt Ion:149 Resp: 487122
Ion Ratio Lower Upper
149 100
167 26.8 21.7 32.5
279 3.9 3.0 4.4





#85
Di-n-octyl phthalate
Concen: 41.178 ng
RT: 14.674 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:149 Resp: 889844
Ion Ratio Lower Upper
149 100
167 1.5 1.1 1.7
43 11.4 9.3 13.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.562 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:264 Resp: 342826
Ion Ratio Lower Upper
264 100
260 23.6 18.4 27.6
265 22.0 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 43.757 ng

RT: 17.074 min Scan# 2198

Delta R.T. -0.006 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025

Tgt Ion:276 Resp: 1098774

Ion Ratio Lower Upper

276 100

138 28.7 22.6 33.8





277 25.1 20.1 30.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#88

Benzo(b)fluoranthene

Concen: 40.403 ng

RT: 15.121 min Scan# 2198

Delta R.T. -0.006 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion:252 Resp: 860370


Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 11.4 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 41.665 ng

RT: 15.157 min Scan# 2204

Delta R.T. -0.000 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17

Instrument :

BNA_F

ClientSampleId :

ICVBF043025

Tgt Ion:252 Resp: 811540

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.8





125 10.9 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/01/2025

Supervised By :Jagrut Upadhyay 05/01/2025



#90

Benzo(a)pyrene

Concen: 42.380 ng

RT: 15.498 min Scan# 2262

Delta R.T. -0.006 min

Lab File: BF142247.D

Acq: 30 Apr 2025 16:17




Tgt Ion:252 Resp: 826250


Ion Ratio Lower Upper

252 100

253 21.8 17.6 26.4


125 12.3 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 43.565 ng
RT: 17.098 min Scan# 21
Delta R.T. -0.000 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17




Instrument :
BNA_F
ClientSampleId :
ICVBF043025



Tgt Ion:278 Resp: 88785
Ion Ratio Lower Upper
278 100
139 18.2 15.2 22.8
279 23.7 19.0 28.6




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/01/2025
Supervised By :Jagrut Upadhyay 05/01/2025



#92
Benzo(g,h,i)perylene
Concen: 43.902 ng
RT: 17.527 min Scan# 2607
Delta R.T. -0.012 min
Lab File: BF142247.D
Acq: 30 Apr 2025 16:17

Tgt Ion:276 Resp: 899474
Ion Ratio Lower Upper
276 100
277 23.4 18.7 28.1
138 24.8 19.4 29.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142247.D
 Acq On : 30 Apr 2025 16:17
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF043025

Quant Time: Apr 30 16:36:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.559	0.593	-6.1	103	0.00
3	Pyridine	1.454	1.543	-6.1	104	0.00
4	n-Nitrosodimethylamine	0.752	0.808	-7.4	103	0.00
5 S	2-Fluorophenol	1.252	1.332	-6.4	104	0.00
6	Aniline	2.140	2.280	-6.5	105	0.00
7 S	Phenol-d6	1.512	1.621	-7.2	105	0.00
8	2-Chlorophenol	1.294	1.430	-10.5	107	0.00
9	Benzaldehyde	1.057	1.134	-7.3	105	0.00
10 C	Phenol	1.628	1.741	-6.9	105	0.00
11	bis(2-Chloroethyl)ether	1.262	1.335	-5.8	104	0.00
12	1,3-Dichlorobenzene	1.464	1.530	-4.5	103	0.00
13 C	1,4-Dichlorobenzene	1.475	1.546	-4.8	103	0.00
14	1,2-Dichlorobenzene	1.408	1.487	-5.6	104	0.00
15	Benzyl Alcohol	1.073	1.163	-8.4	104	0.00
16	2,2'-oxybis(1-Chloropropane	2.278	2.431	-6.7	105	0.00
17	2-Methylphenol	1.060	1.121	-5.8	103	0.00
18	Hexachloroethane	0.483	0.534	-10.6	108	0.00
19 P	n-Nitroso-di-n-propylamine	0.934	0.981	-5.0	104	0.00
20	3+4-Methylphenols	1.326	1.456	-9.8	107	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.463	0.479	-3.5	105	0.00
23 S	Nitrobenzene-d5	0.304	0.376	-23.7	114	0.00
24	Nitrobenzene	0.299	0.354	-18.4	111	0.00
25	Isophorone	0.639	0.667	-4.4	104	0.00
26 C	2-Nitrophenol	0.102	0.152	-49.0#	135	0.00
27	2,4-Dimethylphenol	0.322	0.346	-7.5	106	0.00
28	bis(2-Chloroethoxy)methane	0.403	0.421	-4.5	104	0.00
29 C	2,4-Dichlorophenol	0.269	0.296	-10.0	105	0.00
30	1,2,4-Trichlorobenzene	0.301	0.313	-4.0	103	0.00
31	Naphthalene	0.997	1.022	-2.5	103	0.00
32	Benzoic acid	0.125	0.170	-36.0#	134	0.00
33	4-Chloroaniline	0.415	0.429	-3.4	103	0.00
34 C	Hexachlorobutadiene	0.175	0.185	-5.7	103	0.00
35	Caprolactam	0.081	0.090	-11.1	106	0.00
36 C	4-Chloro-3-methylphenol	0.277	0.301	-8.7	105	0.00
37	2-Methylnaphthalene	0.613	0.631	-2.9	104	0.00
38	1-Methylnaphthalene	0.635	0.655	-3.1	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.563	0.582	-3.4	104	0.00
41 P	Hexachlorocyclopentadiene	0.319	0.378	-18.5	110	0.00
42 S	2,4,6-Tribromophenol	0.182	0.212	-16.5	107	0.00
43 C	2,4,6-Trichlorophenol	0.349	0.399	-14.3	111	0.00
44	2,4,5-Trichlorophenol	0.367	0.410	-11.7	104	0.00
45 S	2-Fluorobiphenyl	1.418	1.415	0.2	103	0.00
46	1,1'-Biphenyl	1.555	1.591	-2.3	103	0.00
47	2-Chloronaphthalene	1.157	1.186	-2.5	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142247.D
 Acq On : 30 Apr 2025 16:17
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF043025

Quant Time: Apr 30 16:36:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.249	0.354	-42.2#	123	0.00
49	Acenaphthylene	1.952	2.011	-3.0	103	0.00
50	Dimethylphthalate	1.280	1.345	-5.1	105	0.00
51	2,6-Dinitrotoluene	0.204	0.277	-35.8#	118	0.00
52 C	Acenaphthene	1.140	1.186	-4.0	104	0.00
53	3-Nitroaniline	0.257	0.338	-31.5#	117	0.00
54 P	2,4-Dinitrophenol	0.058	0.082	-41.4#	148	0.00
55	Dibenzofuran	1.709	1.759	-2.9	104	0.00
56 P	4-Nitrophenol	0.201	0.251	-24.9	120	0.00
57	2,4-Dinitrotoluene	0.239	0.348	-45.6#	126	0.00
58	Fluorene	1.299	1.322	-1.8	103	0.00
59	2,3,4,6-Tetrachlorophenol	0.304	0.347	-14.1	107	0.00
60	Diethylphthalate	1.263	1.334	-5.6	104	0.00
61	4-Chlorophenyl-phenylether	0.617	0.627	-1.6	102	0.00
62	4-Nitroaniline	0.238	0.305	-28.2#	115	0.00
63	Azobenzene	1.248	1.291	-3.4	104	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.054	0.076	-40.7#	142	0.00
66 c	n-Nitrosodiphenylamine	0.683	0.714	-4.5	103	0.00
67	4-Bromophenyl-phenylether	0.223	0.237	-6.3	103	0.00
68	Hexachlorobenzene	0.249	0.261	-4.8	103	0.00
69	Atrazine	0.180	0.201	-11.7	109	0.00
70 C	Pentachlorophenol	0.132	0.156	-18.2	111	0.00
71	Phenanthrene	1.080	1.106	-2.4	102	0.00
72	Anthracene	1.105	1.156	-4.6	105	0.00
73	Carbazole	1.008	1.053	-4.5	105	0.00
74	Di-n-butylphthalate	1.013	1.116	-10.2	107	0.00
75 C	Fluoranthene	1.068	1.113	-4.2	105	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
77	Benzydine	0.850	0.987	-16.1	107	0.00
78	Pyrene	1.845	2.033	-10.2	105	0.00
79 S	Terphenyl-d14	1.375	1.473	-7.1	104	0.00
80	Butylbenzylphthalate	0.425	0.537	-26.4#	111	0.00
81	Benzo(a)anthracene	1.343	1.389	-3.4	101	0.00
82	3,3'-Dichlorobenzidine	0.394	0.449	-14.0	106	0.00
83	Chrysene	1.222	1.288	-5.4	109	0.00
84	Bis(2-ethylhexyl)phthalate	0.604	0.704	-16.6	106	0.00
85 c	Di-n-octyl phthalate	1.133	1.287	-13.6	106	0.00
86 I	Perylene-d12	1.000	1.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	1.465	1.603	-9.4	104	0.00
88	Benzo(b)fluoranthene	1.242	1.255	-1.0	101	0.00
89	Benzo(k)fluoranthene	1.136	1.184	-4.2	108	0.00
90 C	Benzo(a)pyrene	1.137	1.205	-6.0	105	0.00
91	Dibenzo(a,h)anthracene	1.189	1.295	-8.9	103	0.00
92	Benzo(g,h,i)perylene	1.195	1.312	-9.8	105	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 1

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142247.D
 Acq On : 30 Apr 2025 16:17
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF043025

Quant Time: Apr 30 16:36:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	42.438	-6.1	103	0.00
3	Pyridine	40.000	42.469	-6.2	104	0.00
4	n-Nitrosodimethylamine	40.000	42.942	-7.4	103	0.00
5 S	2-Fluorophenol	80.000	85.116	-6.4	104	0.00
6	Aniline	40.000	42.598	-6.5	105	0.00
7 S	Phenol-d6	80.000	85.730	-7.2	105	0.00
8	2-Chlorophenol	40.000	44.202	-10.5	107	0.00
9	Benzaldehyde	40.000	42.925	-7.3	105	0.00
10 C	Phenol	40.000	42.765	-6.9	105	0.00
11	bis(2-Chloroethyl)ether	40.000	42.329	-5.8	104	0.00
12	1,3-Dichlorobenzene	40.000	41.819	-4.5	103	0.00
13 C	1,4-Dichlorobenzene	40.000	41.924	-4.8	103	0.00
14	1,2-Dichlorobenzene	40.000	42.247	-5.6	104	0.00
15	Benzyl Alcohol	40.000	43.378	-8.4	104	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	42.684	-6.7	105	0.00
17	2-Methylphenol	40.000	42.288	-5.7	103	0.00
18	Hexachloroethane	40.000	44.255	-10.6	108	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.992	-5.0	104	0.00
20	3+4-Methylphenols	40.000	43.919	-9.8	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	103	0.00
22	Acetophenone	40.000	41.441	-3.6	105	0.00
23 S	Nitrobenzene-d5	80.000	98.845	-23.6	114	0.00
24	Nitrobenzene	40.000	47.402	-18.5	111	0.00
25	Isophorone	40.000	41.789	-4.5	104	0.00
26 C	2-Nitrophenol	40.000	48.175	-20.4#	135	0.00
27	2,4-Dimethylphenol	40.000	43.004	-7.5	106	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.780	-4.5	104	0.00
29 C	2,4-Dichlorophenol	40.000	44.013	-10.0	105	0.00
30	1,2,4-Trichlorobenzene	40.000	41.724	-4.3	103	0.00
31	Naphthalene	40.000	41.005	-2.5	103	0.00
32	Benzoic acid	40.000	45.113	-12.8	134	0.00
33	4-Chloroaniline	40.000	41.344	-3.4	103	0.00
34 C	Hexachlorobutadiene	40.000	42.127	-5.3	103	0.00
35	Caprolactam	40.000	44.434	-11.1	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.485	-8.7	105	0.00
37	2-Methylnaphthalene	40.000	41.155	-2.9	104	0.00
38	1-Methylnaphthalene	40.000	41.232	-3.1	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.349	-3.4	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	47.321	-18.3	110	0.00
42 S	2,4,6-Tribromophenol	80.000	93.044	-16.3	107	0.00
43 C	2,4,6-Trichlorophenol	40.000	45.713	-14.3	111	0.00
44	2,4,5-Trichlorophenol	40.000	44.676	-11.7	104	0.00
45 S	2-Fluorobiphenyl	80.000	79.807	0.2	103	0.00
46	1,1'-Biphenyl	40.000	40.936	-2.3	103	0.00
47	2-Chloronaphthalene	40.000	41.003	-2.5	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
 Data File : BF142247.D
 Acq On : 30 Apr 2025 16:17
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF043025

Quant Time: Apr 30 16:36:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.523	-16.3	123	0.00
49	Acenaphthylene	40.000	41.211	-3.0	103	0.00
50	Dimethylphthalate	40.000	42.054	-5.1	105	0.00
51	2,6-Dinitrotoluene	40.000	45.376	-13.4	118	0.00
52 C	Acenaphthene	40.000	41.631	-4.1	104	0.00
53	3-Nitroaniline	40.000	45.509	-13.8	117	0.00
54 P	2,4-Dinitrophenol	40.000	49.090	-22.7	148	0.00
55	Dibenzofuran	40.000	41.181	-3.0	104	0.00
56 P	4-Nitrophenol	40.000	49.952	-24.9	120	0.00
57	2,4-Dinitrotoluene	40.000	46.994	-17.5	126	0.00
58	Fluorene	40.000	40.714	-1.8	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	45.724	-14.3	107	0.00
60	Diethylphthalate	40.000	42.222	-5.6	104	0.00
61	4-Chlorophenyl-phenylether	40.000	40.626	-1.6	102	0.00
62	4-Nitroaniline	40.000	45.070	-12.7	115	0.00
63	Azobenzene	40.000	41.400	-3.5	104	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.598	-16.5	142	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.816	-4.5	103	0.00
67	4-Bromophenyl-phenylether	40.000	42.497	-6.2	103	0.00
68	Hexachlorobenzene	40.000	41.800	-4.5	103	0.00
69	Atrazine	40.000	44.732	-11.8	109	0.00
70 C	Pentachlorophenol	40.000	47.342	-18.4	111	0.00
71	Phenanthrene	40.000	40.955	-2.4	102	0.00
72	Anthracene	40.000	41.854	-4.6	105	0.00
73	Carbazole	40.000	41.793	-4.5	105	0.00
74	Di-n-butylphthalate	40.000	44.051	-10.1	107	0.00
75 C	Fluoranthene	40.000	41.690	-4.2	105	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	46.435	-16.1	107	0.00
78	Pyrene	40.000	44.082	-10.2	105	0.00
79 S	Terphenyl-d14	80.000	85.716	-7.1	104	0.00
80	Butylbenzylphthalate	40.000	42.984	-7.5	111	0.00
81	Benzo(a)anthracene	40.000	41.383	-3.5	101	0.00
82	3,3'-Dichlorobenzidine	40.000	45.495	-13.7	106	0.00
83	Chrysene	40.000	42.158	-5.4	109	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.882	-2.2	106	0.00
85 c	Di-n-octyl phthalate	40.000	41.178	-2.9	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.757	-9.4	104	0.00
88	Benzo(b)fluoranthene	40.000	40.403	-1.0	101	0.00
89	Benzo(k)fluoranthene	40.000	41.665	-4.2	108	0.00
90 C	Benzo(a)pyrene	40.000	42.380	-6.0	105	0.00
91	Dibenzo(a,h)anthracene	40.000	43.565	-8.9	103	0.00
92	Benzo(g,h,i)perylene	40.000	43.902	-9.8	105	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142247.D
Acq On : 30 Apr 2025 16:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043025

Quant Time: Apr 30 16:36:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG No.: Q1915

Instrument ID: BNA_F Calibration Date/Time: 05/01/2025 10:17

Lab File ID: BF142250.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.532		5.4	
2-Fluorophenol	1.252	1.250		-0.2	
Phenol-d6	1.512	1.575		4.2	
1,4-Dichlorobenzene	1.475	1.526		3.5	20.0
2-Methylphenol	1.060	1.103		4.1	
3+4-Methylphenols	1.326	1.404		5.9	
Nitrobenzene-d5	0.304	0.333		9.5	
Hexachloroethane	0.483	0.505		4.6	
Nitrobenzene	0.299	0.320		7.0	
Hexachlorobutadiene	0.175	0.186		6.3	20.0
2,4,6-Trichlorophenol	0.349	0.354		1.4	20.0
2-Fluorobiphenyl	1.418	1.409		-0.6	
2,4,5-Trichlorophenol	0.367	0.386		5.2	
2,4-Dinitrotoluene	0.239	0.289		20.9	
2,4,6-Tribromophenol	0.182	0.196		7.7	
Hexachlorobenzene	0.249	0.257		3.2	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.422		3.4	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
 Data File : BF142250.D
 Acq On : 01 May 2025 10:17
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025

Quant Time: May 01 11:46:27 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Apr 30 16:00:01 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards

1) 1,4-Dichlorobenzene-d4	6.910	152	186578	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	719586	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	381198	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	647816	20.000	ng	0.00
76) Chrysene-d12	14.074	240	384442	20.000	ng	0.00
86) Perylene-d12	15.562	264	349272	20.000	ng	0.00

05/02/2025

Supervised By :Jagrut

Upadhyay

System Monitoring Compounds

5) 2-Fluorophenol	5.516	112	932561	79.825	ng	0.00
7) Phenol-d6	6.522	99	1175099	83.296	ng	0.00
23) Nitrobenzene-d5	7.475	82	958053	87.469	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	298805	85.983	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2147916	79.446	ng	0.00
79) Terphenyl-d14	13.021	244	2186580	82.722	ng	0.00

05/02/2025

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.699	88	216157	41.429	ng	98
3) Pyridine	3.463	79	571670	42.158	ng	98
4) n-Nitrosodimethylamine	3.410	42	297662	42.418	ng	97
6) Aniline	6.569	93	830257	41.579	ng	100
8) 2-Chlorophenol	6.687	128	475745	39.409	ng	98
9) Benzaldehyde	6.457	77	418699	42.464	ng	100
10) Phenol	6.540	94	617909m	40.684	ng	
11) bis(2-Chloroethyl)ether	6.640	93	494668	42.030	ng	99
12) 1,3-Dichlorobenzene	6.851	146	566821	41.505	ng	99
13) 1,4-Dichlorobenzene	6.928	146	569459	41.388	ng	100
14) 1,2-Dichlorobenzene	7.081	146	550181	41.885	ng	99
15) Benzyl Alcohol	7.045	79	411183	41.095	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.181	45	889834	41.870	ng	99
17) 2-Methylphenol	7.151	107	411493	41.610	ng	99
18) Hexachloroethane	7.422	117	188542	41.867	ng	94
19) n-Nitroso-di-n-propyla...	7.322	70	356048	40.842	ng	99
20) 3+4-Methylphenols	7.304	107	523765	42.351	ng	97
22) Acetophenone	7.316	105	690233	41.462	ng	98
24) Nitrobenzene	7.492	77	461069	42.902	ng	96
25) Isophorone	7.728	82	966817	42.059	ng	100
26) 2-Nitrophenol	7.804	139	140561	34.044	ng	99
27) 2,4-Dimethylphenol	7.834	122	485594	41.898	ng	100
28) bis(2-Chloroethoxy)met...	7.939	93	606286	41.820	ng	100
29) 2,4-Dichlorophenol	8.039	162	409683	42.280	ng	99
30) 1,2,4-Trichlorobenzene	8.134	180	453172	41.910	ng	99
31) Naphthalene	8.216	128	1482230	41.304	ng	100
32) Benzoic acid	7.928	122	168727m	34.514	ng	
33) 4-Chloroaniline	8.257	127	622889	41.683	ng	100
34) Hexachlorobutadiene	8.334	225	267370	42.382	ng	100
35) Caprolactam	8.622	113	129344	44.410	ng	98
36) 4-Chloro-3-methylphenol	8.728	107	431195	43.337	ng	99
37) 2-Methylnaphthalene	8.904	142	913767	41.414	ng	99
38) 1-Methylnaphthalene	9.004	142	950787	41.597	ng	99
40) 1,2,4,5-Tetrachloroben...	9.069	216	435999	40.641	ng	100
41) Hexachlorocyclopentadiene	9.057	237	241144	39.605	ng	99
43) 2,4,6-Trichlorophenol	9.175	196	269577	40.551	ng	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
 Data File : BF142250.D
 Acq On : 01 May 2025 10:17
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025

Quant Time: May 01 11:46:27 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
44) 2,4,5-Trichlorophenol	9.210	196	294308	42.108	ng	99	05/02/2025
46) 1,1'-Biphenyl	9.369	154	1212716	40.922	ng	99	Supervised By :Jagrut
47) 2-Chloronaphthalene	9.392	162	896541	40.663	ng	99	Upadhyay
48) 2-Nitroaniline	9.480	65	211351	37.508	ng	97	
49) Acenaphthylene	9.810	152	1521943	40.906	ng	100	
50) Dimethylphthalate	9.669	163	1026519	42.093	ng	100	
51) 2,6-Dinitrotoluene	9.728	165	176156	38.527	ng	98	05/02/2025
52) Acenaphthene	9.980	154	903201	41.576	ng	100	
53) 3-Nitroaniline	9.892	138	225294	40.152	ng	100	
54) 2,4-Dinitrophenol	9.992	184	42525	37.926	ng	# 1	
55) Dibenzofuran	10.151	168	1350916	41.484	ng	99	
56) 4-Nitrophenol	10.039	139	159868	41.706	ng	98	
57) 2,4-Dinitrotoluene	10.128	165	220248	39.850	ng	99	
58) Fluorene	10.498	166	1007989	40.709	ng	99	
59) 2,3,4,6-Tetrachlorophenol	10.263	232	245919	42.475	ng	99	
60) Diethylphthalate	10.369	149	1030685	42.799	ng	99	
61) 4-Chlorophenyl-phenyle...	10.486	204	486030	41.333	ng	99	
62) 4-Nitroaniline	10.504	138	215440	42.018	ng	98	
63) Azobenzene	10.645	77	996298	41.895	ng	98	
65) 4,6-Dinitro-2-methylph...	10.533	198	69294	35.905	ng	99	
66) n-Nitrosodiphenylamine	10.604	169	914630	41.322	ng	100	
67) 4-Bromophenyl-phenylether	10.980	248	304870	42.258	ng	99	
68) Hexachlorobenzene	11.039	284	332960	41.207	ng	94	
69) Atrazine	11.127	200	264074	45.262	ng	100	
70) Pentachlorophenol	11.233	266	177109	41.491	ng	99	
71) Phenanthrene	11.457	178	1446110	41.326	ng	100	
72) Anthracene	11.510	178	1479001	41.339	ng	100	
73) Carbazole	11.663	167	1352106	41.429	ng	100	
74) Di-n-butylphthalate	11.998	149	1472584	44.864	ng	100	
75) Fluoranthene	12.645	202	1485579	42.947	ng	100	
77) Benzidine	12.769	184	754042	46.160	ng	99	
78) Pyrene	12.874	202	1490017	42.016	ng	100	
80) Butylbenzylphthalate	13.498	149	385725	40.417	ng	98	
81) Benzo(a)anthracene	14.063	228	1059415	41.049	ng	99	
82) 3,3'-Dichlorobenzidine	14.027	252	323049	42.615	ng	100	
83) Chrysene	14.104	228	997135	42.445	ng	100	
84) Bis(2-ethylhexyl)phtha...	14.057	149	477774	36.467	ng	100	
85) Di-n-octyl phthalate	14.674	149	782817	33.845	ng	100	
87) Indeno(1,2,3-cd)pyrene	17.074	276	1071010	41.864	ng	100	
88) Benzo(b)fluoranthene	15.127	252	867766	39.998	ng	99	
89) Benzo(k)fluoranthene	15.157	252	866064	43.643	ng	99	
90) Benzo(a)pyrene	15.504	252	833612	41.969	ng	99	
91) Dibenzo(a,h)anthracene	17.098	278	878980	42.333	ng	99	
92) Benzo(g,h,i)perylene	17.533	276	873517	41.848	ng	100	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\  
Data File : BF142250.D  
Acq On    : 01 May 2025  10:17  
Operator  : RC/JU  
Sample    : SSTDCCC040  
Misc      :  
ALS Vial  : 2    Sample Multiplier: 1
```

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Manual Integrations

Reviewed By :Rahul
Chavli


05/02/2025


Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay

0510 Terphenyl-d14,S

05/02/2025






#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:152 Resp: 186578
Ion Ratio Lower Upper
152 100
150 159.1 130.2 195.2
115 58.3 47.0 70.4



Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli

05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#2
1,4-Dioxane
Concen: 41.429 ng
RT: 2.699 min Scan# 86
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17


Tgt Ion: 88 Resp: 216157
Ion Ratio Lower Upper
88 100
58 92.7 72.2 108.2
43 34.7 27.0 40.6





#3
Pyridine
Concen: 42.158 ng
RT: 3.463 min Scan# 216
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 571670
Ion Ratio Lower Upper
79 100
52 87.4 71.5 107.3
51 42.2 34.5 51.7




Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli

05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#4
n-Nitrosodimethylamine
Concen: 42.418 ng
RT: 3.410 min Scan# 207
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17


Tgt Ion: 42 Resp: 297662
Ion Ratio Lower Upper
42 100
74 108.6 89.5 134.3
44 8.5 6.6 9.8





#5
2-Fluorophenol
Concen: 79.825 ng
RT: 5.516 min Scan# 565
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040






Tgt Ion: 112 Resp: 93256
Ion Ratio Lower Upper
112 100
64 59.8 48.4 72.6
63 30.8 24.8 37.2

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#6
Aniline
Concen: 41.579 ng
RT: 6.569 min Scan# 744
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17


Tgt Ion: 93 Resp: 830257
Ion Ratio Lower Upper
93 100
66 40.1 32.1 48.1
65 20.7 16.6 24.8





#7
Phenol-d6
Concen: 83.296 ng
RT: 6.522 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040






Tgt Ion: 99 Resp: 1175099
Ion Ratio Lower Upper
99 100
42 21.0 16.3 24.5
71 31.9 25.0 37.4

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#8
2-Chlorophenol
Concen: 39.409 ng
RT: 6.687 min Scan# 764
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:128 Resp: 475745
Ion Ratio Lower Upper
128 100
130 32.3 13.1 53.1
64 49.9 28.6 68.6





#9

Benzaldehyde

Concen: 42.464 ng

RT: 6.457 min Scan# 725

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion: 77 Resp: 418699

Ion Ratio Lower Upper

77 100

105 98.6 79.1 119.1

106 95.6 76.0 116.0

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut




Upadhyay

05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay



#10

Phenol

Concen: 40.684 ng m

RT: 6.540 min Scan# 739

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion: 94 Resp: 617909


Ion Ratio Lower Upper

94 100

65 28.8 8.9 48.9


66 37.5 18.7 58.7





#11
bis(2-Chloroethyl)ether
Concen: 42.030 ng
RT: 6.640 min Scan# 757
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040






Tgt Ion: 93 Resp: 494668
Ion Ratio Lower Upper
93 100
63 83.6 62.2 102.2
95 32.2 12.7 52.7

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#12
1,3-Dichlorobenzene
Concen: 41.505 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:146 Resp: 566821
Ion Ratio Lower Upper
146 100
148 63.8 50.5 75.7
75 29.6 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 41.388 ng

RT: 6.928 min Scan# 805

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:146 Resp: 569459

Ion Ratio Lower Upper

146 100

148 63.1 50.6 76.0

111 42.0 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut



Upadhyay

05/02/2025


05/02/2025

Supervised By :Jagrut

Upadhyay



05/02/2025



#14

1,2-Dichlorobenzene

Concen: 41.885 ng

RT: 7.081 min Scan# 831

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:146 Resp: 550181


Ion Ratio Lower Upper

146 100

148 63.9 50.5 75.7

111 43.0 34.9 52.3





#15

Benzyl Alcohol

Concen: 41.095 ng

RT: 7.045 min Scan# 81

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 79 Resp: 41118

Ion Ratio Lower Upper

79 100

108 81.7 64.5 96.7

77 62.1 49.2 73.8

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay





05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 41.870 ng

RT: 7.181 min Scan# 848

Delta R.T. -0.006 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion: 45 Resp: 889834


Ion Ratio Lower Upper

45 100

77 10.8 0.0 31.3


79 8.4 0.0 28.5





#17
2-Methylphenol
Concen: 41.610 ng
RT: 7.151 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040






Tgt Ion:107 Resp: 41149
Ion Ratio Lower Upper
107 100
108 114.4 91.0 136.4
77 43.3 34.4 51.6
79 42.9 34.4 51.6

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli
05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025



05/02/2025
Supervised By :Jagrut
Upadhyay




#18
Hexachloroethane
Concen: 41.867 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17



Tgt Ion:117 Resp: 188542
Ion Ratio Lower Upper
117 100
119 94.5 77.1 115.7
201 90.2 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 40.842 ng

RT: 7.322 min Scan# 81

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion: 70 Resp: 35604

Ion Ratio Lower Upper

70 100

42 66.0 53.4 80.2

101 9.7 7.7 11.5

130 21.3 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut



Upadhyay

05/02/2025


05/02/2025

Supervised By :Jagrut

Upadhyay



05/02/2025



#20

3+4-Methylphenols

Concen: 42.351 ng

RT: 7.304 min Scan# 869

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Tgt Ion:107 Resp: 523765




Ion Ratio Lower Upper


107 100

108 90.6 71.2 111.2

77 65.8 41.1 81.1


79 27.6 8.1 48.1





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1020
Delta R.T. -0.001 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040






Tgt Ion:136 Resp: 719580
Ion Ratio Lower Upper
136 100
137 11.0 8.7 13.1
54 9.6 7.7 11.5
68 6.1 5.0 7.6

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#22
Acetophenone
Concen: 41.462 ng
RT: 7.316 min Scan# 871
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:105 Resp: 690233
Ion Ratio Lower Upper
105 100
71 7.2 5.7 8.5
51 36.0 30.1 45.1
120 23.7 19.0 28.4





#23

Nitrobenzene-d5

Concen: 87.469 ng

RT: 7.475 min Scan# 898

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 95805

Ion Ratio Lower Upper

82 100

128 42.5 37.0 55.6

54 59.1 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay





05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#24

Nitrobenzene

Concen: 42.902 ng

RT: 7.492 min Scan# 901

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion: 77 Resp: 461069


Ion Ratio Lower Upper

77 100

123 45.1 38.6 58.0

65 13.5 11.5 17.3





#25

Isophorone

Concen: 42.059 ng

RT: 7.728 min Scan# 941

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 96681

Ion Ratio Lower Upper

82 100

95 7.1 5.6 8.4

138 19.2 15.3 22.9

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay





05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#26

2-Nitrophenol

Concen: 34.044 ng

RT: 7.804 min Scan# 954

Delta R.T. 0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:139 Resp: 140561


Ion Ratio Lower Upper

139 100

109 27.7 21.8 32.8

65 52.3 41.1 61.7





#27

2,4-Dimethylphenol

Concen: 41.898 ng

RT: 7.834 min Scan# 91

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:122 Resp: 485594

Ion Ratio Lower Upper

122 100

107 107.9 86.0 129.0

121 57.8 46.3 69.5

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay





05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#28

bis(2-Chloroethoxy)methane

Concen: 41.820 ng

RT: 7.939 min Scan# 977

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion: 93 Resp: 606286


Ion Ratio Lower Upper

93 100

95 32.6 26.2 39.2

123 12.0 9.7 14.5





#29

2,4-Dichlorophenol

Concen: 42.280 ng

RT: 8.039 min Scan# 994

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:162 Resp: 40968

Ion Ratio Lower Upper

162 100

164 63.5 43.0 83.0

98 36.9 16.7 56.7

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut




Upadhyay

05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay



#30

1,2,4-Trichlorobenzene

Concen: 41.910 ng

RT: 8.134 min Scan# 1010

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:180 Resp: 453172


Ion Ratio Lower Upper

180 100

182 93.3 75.7 113.5

145 31.3 25.4 38.0





#31

Naphthalene

Concen: 41.304 ng

RT: 8.216 min Scan# 1024

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:128 Resp: 1482230

Ion Ratio Lower Upper

128 100

129 11.3 9.0 13.4

127 13.2 10.6 15.8

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut




Upadhyay

05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay



#32

Benzoic acid

Concen: 34.514 ng m

RT: 7.928 min Scan# 975

Delta R.T. -0.006 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:122 Resp: 168727


Ion Ratio Lower Upper

122 100

105 118.8 98.2 138.2

77 89.1 66.8 106.8






#33
4-Chloroaniline
Concen: 41.683 ng
RT: 8.257 min Scan# 1031
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :



SSTDCCC040



Tgt Ion:127 Resp: 622889
Ion Ratio Lower Upper
127 100
129 32.6 26.1 39.1
65 31.6 25.5 38.3
92 19.3 15.4 23.2




Manual Integrations
APPROVEDReviewed By :Rahul
Chavli


05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/202505/02/2025
Supervised By :Jagrut
Upadhyay

#34
Hexachlorobutadiene
Concen: 42.382 ng
RT: 8.334 min Scan# 1044
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:225 Resp: 267370
Ion Ratio Lower Upper
225 100
223 62.2 49.7 74.5
227 63.5 51.0 76.4





#35

Caprolactam

Concen: 44.410 ng

RT: 8.622 min Scan# 1093

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:113 Resp: 129344

Ion Ratio Lower Upper

113 100

55 195.1 177.7 217.7

56 143.4 127.1 167.1

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut




Upadhyay

05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay



#36

4-Chloro-3-methylphenol

Concen: 43.337 ng

RT: 8.728 min Scan# 1111

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:107 Resp: 431195


Ion Ratio Lower Upper

107 100

144 26.2 21.0 31.4


142 81.4 66.1 99.1





#37
2-Methylnaphthalene
Concen: 41.414 ng
RT: 8.904 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:142 Resp: 91376
Ion Ratio Lower Upper
142 100
141 87.6 70.5 105.7


Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli


05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#38
1-Methylnaphthalene
Concen: 41.597 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:142 Resp: 950787
Ion Ratio Lower Upper
142 100
141 90.5 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.945 min Scan# 1169

Delta R.T. -0.006 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:164 Resp: 381198

Ion Ratio Lower Upper

164 100

162 98.2 78.6 117.8

160 44.5 35.7 53.5

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut




Upadhyay

05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay



#40

1,2,4,5-Tetrachlorobenzene

Concen: 40.641 ng

RT: 9.069 min Scan# 1169

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Tgt Ion:216 Resp: 435999




Ion Ratio Lower Upper


216 100

214 79.7 63.8 95.6

179 21.1 16.7 25.1

108 19.0 15.6 23.4





#41

Hexachlorocyclopentadiene

Concen: 39.605 ng

RT: 9.057 min Scan# 1167

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:237 Resp: 241144

Ion Ratio Lower Upper

237 100

235 62.7 43.5 83.5

272 12.6 0.0 32.9

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay





05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#42

2,4,6-Tribromophenol

Concen: 85.983 ng

RT: 10.733 min Scan# 1452

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:330 Resp: 298805


Ion Ratio Lower Upper

330 100

332 96.5 76.5 114.7

141 31.9 24.3 36.5





#43

2,4,6-Trichlorophenol

Concen: 40.551 ng

RT: 9.175 min Scan# 1187

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:196 Resp: 26957

Ion Ratio Lower Upper

196 100

198 97.1 77.4 116.2

200 31.0 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut




Upadhyay

05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay



#44

2,4,5-Trichlorophenol

Concen: 42.108 ng

RT: 9.210 min Scan# 1193

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Tgt Ion:196 Resp: 294308




Ion Ratio Lower Upper


196 100

198 95.7 75.2 112.8

97 50.9 39.8 59.8

132 29.8 23.8 35.8





#45

2-Fluorobiphenyl

Concen: 79.446 ng

RT: 9.269 min Scan# 1203

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:172 Resp: 2147910

Ion Ratio Lower Upper

172 100

171 35.8 29.0 43.4

170 24.1 19.4 29.0

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay





05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#46

1,1'-Biphenyl

Concen: 40.922 ng

RT: 9.369 min Scan# 1220

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:154 Resp: 1212716


Ion Ratio Lower Upper

154 100

153 41.0 20.5 60.5

76 12.6 0.0 32.8





#47

2-Chloronaphthalene

Concen: 40.663 ng

RT: 9.392 min Scan# 11

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:162 Resp: 89654

Ion Ratio Lower Upper

162 100

127 41.5 32.6 49.0

164 33.3 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay





05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#48

2-Nitroaniline

Concen: 37.508 ng

RT: 9.480 min Scan# 1239

Delta R.T. -0.007 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion: 65 Resp: 211351


Ion Ratio Lower Upper

65 100

92 66.4 54.3 81.5

138 106.6 88.6 133.0





#49

Acenaphthylene

Concen: 40.906 ng

RT: 9.810 min Scan# 1295

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:152 Resp: 152194

Ion Ratio Lower Upper

152 100

151 20.2 16.2 24.2

153 13.5 10.7 16.1

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut




Upadhyay

05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay



#50

Dimethylphthalate

Concen: 42.093 ng

RT: 9.669 min Scan# 1271

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:163 Resp: 1026519


Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8


164 10.4 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 38.527 ng
RT: 9.728 min Scan# 1281
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 176150
Ion Ratio Lower Upper
165 100
63 61.3 48.5 72.7
89 56.7 43.9 65.9




Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli

05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#52
Acenaphthene
Concen: 41.576 ng
RT: 9.980 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17


Tgt Ion:154 Resp: 903201
Ion Ratio Lower Upper
154 100
153 110.3 88.6 133.0
152 52.6 42.1 63.1





#53
3-Nitroaniline
Concen: 40.152 ng
RT: 9.892 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:138 Resp: 225294
Ion Ratio Lower Upper
138 100
108 14.5 11.7 17.5
92 113.6 91.2 136.8




Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli

05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#54
2,4-Dinitrophenol
Concen: 37.926 ng
RT: 9.992 min Scan# 1326
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17


Tgt Ion:184 Resp: 42525
Ion Ratio Lower Upper
184 100
63 136.6 69.2 103.8#
154 883.4 271.4 407.2#





#55
Dibenzofuran
Concen: 41.484 ng
RT: 10.151 min Scan# 1353
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:168 Resp: 1350910
Ion Ratio Lower Upper
168 100
139 37.2 29.4 44.0
169 13.7 11.0 16.4

Manual Integrations
APPROVED



Reviewed By :Rahul
Chavli

05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025




05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025



#56
4-Nitrophenol
Concen: 41.706 ng
RT: 10.039 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:139 Resp: 159868
Ion Ratio Lower Upper
139 100
109 67.4 44.9 84.9
65 93.9 74.7 114.7






#57
2,4-Dinitrotoluene
Concen: 39.850 ng
RT: 10.128 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :




BNA_F

ClientSampleId :

SSTDCCC040







Tgt Ion:165 Resp: 220248
Ion Ratio Lower Upper
165 100
63 48.8 38.8 58.2
89 72.1 57.1 85.7
182 2.5 2.2 3.2

Manual Integrations
APPROVEDReviewed By :Rahul
Chavli
05/02/2025Supervised By :Jagrut
Upadhyay
05/02/202505/02/2025
Supervised By :Jagrut
Upadhyay

#58
Fluorene
Concen: 40.709 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:166 Resp: 1007989
Ion Ratio Lower Upper
166 100
165 93.4 73.9 110.9
167 13.7 10.8 16.2






#59
2,3,4,6-Tetrachlorophenol
Concen: 42.475 ng
RT: 10.263 min Scan# 1372
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :




SSTDCCC040




Tgt Ion:232 Resp: 245919
Ion Ratio Lower Upper
232 100
131 46.0 35.9 53.9
130 2.1 1.8 2.6
166 29.4 24.0 36.0

Manual Integrations
APPROVEDReviewed By :Rahul
Chavli



05/02/2025


Supervised By :Jagrut
Upadhyay
05/02/202505/02/2025
Supervised By :Jagrut
Upadhyay

#60
Diethylphthalate
Concen: 42.799 ng
RT: 10.369 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17




Tgt Ion:149 Resp: 1030685
Ion Ratio Lower Upper
149 100
177 21.7 17.6 26.4
150 12.5 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 41.333 ng
RT: 10.486 min Scan# 1410
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040





Tgt Ion:204 Resp: 486030
Ion Ratio Lower Upper
204 100
206 32.7 25.6 38.4
141 60.4 48.9 73.3

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#62
4-Nitroaniline
Concen: 42.018 ng
RT: 10.504 min Scan# 1413
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17


Tgt Ion:138 Resp: 215440
Ion Ratio Lower Upper
138 100
92 50.8 29.2 69.2
108 70.7 49.1 89.1





#63
Azobenzene
Concen: 41.895 ng
RT: 10.645 min Scan# 1438
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040






Tgt Ion: 77 Resp: 996298
Ion Ratio Lower Upper
77 100
182 25.2 7.3 47.3
105 22.3 2.5 42.5
51 35.1 14.4 54.4

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17


Tgt Ion:188 Resp: 647816
Ion Ratio Lower Upper
188 100
94 9.9 8.2 12.2
80 10.7 8.6 12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 35.905 ng
RT: 10.533 min Scan# 1418
Delta R.T. -0.001 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040






Tgt Ion:198 Resp: 69294
Ion Ratio Lower Upper
198 100
51 61.0 40.6 80.6
105 48.7 29.2 69.2

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli
05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#66
n-Nitrosodiphenylamine
Concen: 41.322 ng
RT: 10.604 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:169 Resp: 914630
Ion Ratio Lower Upper
169 100
168 66.2 53.0 79.6
167 36.5 28.8 43.2






#67
4-Bromophenyl-phenylether
Concen: 42.258 ng
RT: 10.980 min Scan# 1494
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:248 Resp: 304870
Ion Ratio Lower Upper
248 100
250 97.4 77.2 115.8
141 71.7 58.4 87.6




Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli


05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025



05/02/2025
Supervised By :Jagrut
Upadhyay




#68
Hexachlorobenzene
Concen: 41.207 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17



Tgt Ion:284 Resp: 332960
Ion Ratio Lower Upper
284 100
142 35.1 24.6 37.0
249 31.0 23.2 34.8





#69

Atrazine

Concen: 45.262 ng

RT: 11.127 min Scan# 1519

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :



SSTDCCC040



Tgt Ion:200	Resp: 264074
Ion Ratio	Lower Upper
200 100	
173 28.4	8.6 48.6
215 48.7	28.5 68.5

Manual Integrations
APPROVEDReviewed By :Rahul
Chavli

05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/202505/02/2025
Supervised By :Jagrut
Upadhyay

#70

Pentachlorophenol


Concen: 41.491 ng

RT: 11.233 min Scan# 1537



Delta R.T. 0.000 min


Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:266	Resp: 177109
Ion Ratio	Lower Upper
266 100	
268 63.2	50.0 75.0
264 61.4	49.3 73.9





#71
Phenanthrene
Concen: 41.326 ng
RT: 11.457 min Scan# 1575
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:178 Resp: 1446110
Ion Ratio Lower Upper
178 100
176 19.4 15.5 23.3
179 15.7 12.6 19.0

Manual Integrations
APPROVED



Reviewed By :Rahul
Chavli

05/02/2025

Supervised By :Jagrut
Upadhyay




05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#72
Anthracene
Concen: 41.339 ng
RT: 11.510 min Scan# 1584
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17


Tgt Ion:178 Resp: 1479001
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.8 12.7 19.1





#73
 Carbazole
 Concen: 41.429 ng
 RT: 11.663 min Scan# 1610
 Delta R.T. -0.000 min
 Lab File: BF142250.D
 Acq: 01 May 2025 10:17

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040



Tgt Ion:167 Resp: 1352100
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.3 25.9
 139 13.0 10.3 15.5




Manual Integrations
 APPROVED

Reviewed By :Rahul
 Chavli


05/02/2025

Supervised By :Jagrut
 Upadhyay
 05/02/2025



05/02/2025
 Supervised By :Jagrut
 Upadhyay




#74
 Di-n-butylphthalate
 Concen: 44.864 ng
 RT: 11.998 min Scan# 1667
 Delta R.T. -0.000 min
 Lab File: BF142250.D
 Acq: 01 May 2025 10:17



Tgt Ion:149 Resp: 1472584
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 4.6 3.6 5.4





#75

Fluoranthene

Concen: 42.947 ng

RT: 12.645 min Scan# 1777

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:202 Resp: 1485579

Ion Ratio Lower Upper

202 100

101 12.3 0.0 32.4

203 17.8 0.0 37.5

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut




Upadhyay

05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.074 min Scan# 2020

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:240 Resp: 384442


Ion Ratio Lower Upper

240 100

120 12.5 10.1 15.1


236 25.7 20.5 30.7





#77
Benzidine
Concen: 46.160 ng
RT: 12.769 min Scan# 1798
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:184 Resp: 75404
Ion Ratio Lower Upper
184 100
185 15.0 11.8 17.6
183 10.9 8.9 13.3



Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli

05/02/2025




Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



#78
Pyrene
Concen: 42.016 ng
RT: 12.874 min Scan# 1816
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:202 Resp: 1490017
Ion Ratio Lower Upper
202 100
200 20.1 16.1 24.1
203 18.0 14.1 21.1





#79

Terphenyl-d14

Concen: 82.722 ng

RT: 13.021 min Scan# 1841

Delta R.T. -0.000 min

Lab File: BF142250.D


Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:244 Resp: 2186580

Ion Ratio Lower Upper

244 100

212 6.7 5.3 7.9

122 11.3 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut




Upadhyay

05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay



#80

Butylbenzylphthalate

Concen: 40.417 ng

RT: 13.498 min Scan# 1922

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:149 Resp: 385725


Ion Ratio Lower Upper

149 100

91 66.6 54.4 81.6

206 20.5 15.9 23.9






#81
Benzo(a)anthracene
Concen: 41.049 ng
RT: 14.063 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:228 Resp: 105941

Ion	Ratio	Lower	Upper
228	100		
226	27.1	21.9	32.9
229	19.6	16.0	24.0


Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli


05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025


05/02/2025
Supervised By :Jagrut
Upadhyay



05/02/2025






#82
3,3'-Dichlorobenzidine
Concen: 42.615 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17



Tgt Ion:252 Resp: 323049

Ion	Ratio	Lower	Upper
252	100		
254	63.9	51.0	76.6
126	12.0	9.4	14.2





#83

Chrysene

Concen: 42.445 ng

RT: 14.104 min Scan# 2025

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:228 Resp: 997135

Ion Ratio Lower Upper

228 100

226 30.1 24.0 36.0

229 19.8 15.7 23.5

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay





05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#84

Bis(2-ethylhexyl)phthalate

Concen: 36.467 ng

RT: 14.057 min Scan# 2017

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:149 Resp: 477774


Ion Ratio Lower Upper

149 100

167 27.1 21.7 32.5


279 3.7 3.0 4.4





#85
Di-n-octyl phthalate
Concen: 33.845 ng
RT: 14.674 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:149 Resp: 78281
Ion Ratio Lower Upper
149 100
167 1.5 1.1 1.7
43 11.7 9.3 13.9

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli




05/02/2025

Supervised By :Jagrut
Upadhyay
05/02/2025

05/02/2025




Supervised By :Jagrut
Upadhyay


05/02/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.562 min Scan# 2273
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:264 Resp: 349272
Ion Ratio Lower Upper
264 100
260 23.5 18.4 27.6
265 21.2 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 41.864 ng

RT: 17.074 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:276 Resp: 1071010

Ion Ratio Lower Upper

276 100

138 28.3 22.6 33.8

277 25.2 20.1 30.1

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut

Upadhyay





05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#88

Benzo(b)fluoranthene

Concen: 39.998 ng

RT: 15.127 min Scan# 2199

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:252 Resp: 867766


Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 10.5 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 43.643 ng

RT: 15.157 min Scan# 21

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:252 Resp: 866064

Ion Ratio Lower Upper

252 100

253 21.6 17.8 26.8

125 10.9 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli

05/02/2025

Supervised By :Jagrut





Upadhyay

05/02/2025

05/02/2025

Supervised By :Jagrut

Upadhyay



#90

Benzo(a)pyrene

Concen: 41.969 ng

RT: 15.504 min Scan# 2263

Delta R.T. -0.000 min

Lab File: BF142250.D

Acq: 01 May 2025 10:17




Tgt Ion:252 Resp: 833612


Ion Ratio Lower Upper

252 100

253 21.8 17.6 26.4


125 11.7 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 42.333 ng
RT: 17.098 min Scan# 2534
Delta R.T. -0.000 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:278 Resp: 878980
Ion Ratio Lower Upper
278 100
139 18.3 15.2 22.8
279 23.5 19.0 28.6

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli




05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025




Supervised By :Jagrut
Upadhyay

05/02/2025



#92
Benzo(g,h,i)perylene
Concen: 41.848 ng
RT: 17.533 min Scan# 2608
Delta R.T. -0.006 min
Lab File: BF142250.D
Acq: 01 May 2025 10:17

Tgt Ion:276 Resp: 873517
Ion Ratio Lower Upper
276 100
277 23.4 18.7 28.1
138 24.3 19.4 29.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
 Data File : BF142250.D
 Acq On : 01 May 2025 10:17
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 01 11:46:27 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.559	0.579	-3.6	101	0.00
3	Pyridine	1.454	1.532	-5.4	104	0.00
4	n-Nitrosodimethylamine	0.752	0.798	-6.1	103	0.00
5 S	2-Fluorophenol	1.252	1.250	0.2	99	0.00
6	Aniline	2.140	2.225	-4.0	103	0.00
7 S	Phenol-d6	1.512	1.575	-4.2	102	0.00
8	2-Chlorophenol	1.294	1.275	1.5	96	0.00
9	Benzaldehyde	1.057	1.122	-6.1	104	0.00
10 C	Phenol	1.628	1.656	-1.7	100	0.00
11	bis(2-Chloroethyl)ether	1.262	1.326	-5.1	104	0.00
12	1,3-Dichlorobenzene	1.464	1.519	-3.8	103	0.00
13 C	1,4-Dichlorobenzene	1.475	1.526	-3.5	102	0.00
14	1,2-Dichlorobenzene	1.408	1.474	-4.7	104	0.00
15	Benzyl Alcohol	1.073	1.102	-2.7	99	0.00
16	2,2'-oxybis(1-Chloropropane	2.278	2.385	-4.7	104	0.00
17	2-Methylphenol	1.060	1.103	-4.1	102	0.00
18	Hexachloroethane	0.483	0.505	-4.6	103	0.00
19 P	n-Nitroso-di-n-propylamine	0.934	0.954	-2.1	102	0.00
20	3+4-Methylphenols	1.326	1.404	-5.9	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.463	0.480	-3.7	104	0.00
23 S	Nitrobenzene-d5	0.304	0.333	-9.5	100	0.00
24	Nitrobenzene	0.299	0.320	-7.0	100	0.00
25	Isophorone	0.639	0.672	-5.2	104	0.00
26 C	2-Nitrophenol	0.102	0.098	3.9	86	0.00
27	2,4-Dimethylphenol	0.322	0.337	-4.7	102	0.00
28	bis(2-Chloroethoxy)methane	0.403	0.421	-4.5	103	0.00
29 C	2,4-Dichlorophenol	0.269	0.285	-5.9	100	0.00
30	1,2,4-Trichlorobenzene	0.301	0.315	-4.7	103	0.00
31	Naphthalene	0.997	1.030	-3.3	103	0.00
32	Benzoic acid	0.125	0.117	6.4	92	0.00
33	4-Chloroaniline	0.415	0.433	-4.3	103	0.00
34 C	Hexachlorobutadiene	0.175	0.186	-6.3	103	0.00
35	Caprolactam	0.081	0.090	-11.1	106	0.00
36 C	4-Chloro-3-methylphenol	0.277	0.300	-8.3	103	0.00
37	2-Methylnaphthalene	0.613	0.635	-3.6	104	0.00
38	1-Methylnaphthalene	0.635	0.661	-4.1	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.563	0.572	-1.6	102	0.00
41 P	Hexachlorocyclopentadiene	0.319	0.316	0.9	92	0.00
42 S	2,4,6-Tribromophenol	0.182	0.196	-7.7	99	0.00
43 C	2,4,6-Trichlorophenol	0.349	0.354	-1.4	98	0.00
44	2,4,5-Trichlorophenol	0.367	0.386	-5.2	98	0.00
45 S	2-Fluorobiphenyl	1.418	1.409	0.6	103	0.00
46	1,1'-Biphenyl	1.555	1.591	-2.3	103	0.00
47	2-Chloronaphthalene	1.157	1.176	-1.6	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
 Data File : BF142250.D
 Acq On : 01 May 2025 10:17
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 01 11:46:27 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.249	0.277	-11.2	97	0.00
49	Acenaphthylene	1.952	1.996	-2.3	103	0.00
50	Dimethylphthalate	1.280	1.346	-5.2	105	0.00
51	2,6-Dinitrotoluene	0.204	0.231	-13.2	98	0.00
52 C	Acenaphthene	1.140	1.185	-3.9	104	0.00
53	3-Nitroaniline	0.257	0.296	-15.2	102	0.00
54 P	2,4-Dinitrophenol	0.058	0.056	3.4	101	0.00
55	Dibenzofuran	1.709	1.772	-3.7	104	0.00
56 P	4-Nitrophenol	0.201	0.210	-4.5	100	0.00
57	2,4-Dinitrotoluene	0.239	0.289	-20.9	104	0.00
58	Fluorene	1.299	1.322	-1.8	103	0.00
59	2,3,4,6-Tetrachlorophenol	0.304	0.323	-6.3	99	0.00
60	Diethylphthalate	1.263	1.352	-7.0	106	0.00
61	4-Chlorophenyl-phenylether	0.617	0.638	-3.4	104	0.00
62	4-Nitroaniline	0.238	0.283	-18.9	107	0.00
63	Azobenzene	1.248	1.307	-4.7	105	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	0.054	0.053	1.9	102	0.00
66 c	n-Nitrosodiphenylamine	0.683	0.706	-3.4	104	0.00
67	4-Bromophenyl-phenylether	0.223	0.235	-5.4	105	0.00
68	Hexachlorobenzene	0.249	0.257	-3.2	104	0.00
69	Atrazine	0.180	0.204	-13.3	112	0.00
70 C	Pentachlorophenol	0.132	0.137	-3.8	100	0.00
71	Phenanthrene	1.080	1.116	-3.3	105	0.00
72	Anthracene	1.105	1.142	-3.3	106	0.00
73	Carbazole	1.008	1.044	-3.6	106	0.00
74	Di-n-butylphthalate	1.013	1.137	-12.2	112	0.00
75 C	Fluoranthene	1.068	1.147	-7.4	111	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
77	Benzidine	0.850	0.981	-15.4	118	0.00
78	Pyrene	1.845	1.938	-5.0	111	0.00
79 S	Terphenyl-d14	1.375	1.422	-3.4	111	0.00
80	Butylbenzylphthalate	0.425	0.502	-18.1	115	0.00
81	Benzo(a)anthracene	1.343	1.378	-2.6	111	0.00
82	3,3'-Dichlorobenzidine	0.394	0.420	-6.6	110	0.00
83	Chrysene	1.222	1.297	-6.1	122	0.00
84	Bis(2-ethylhexyl)phthalate	0.604	0.621	-2.8	104	0.00
85 c	Di-n-octyl phthalate	1.133	1.018	10.2	94	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	1.465	1.533	-4.6	101	0.00
88	Benzo(b)fluoranthene	1.242	1.242	0.0	101	0.00
89	Benzo(k)fluoranthene	1.136	1.240	-9.2	115	0.00
90 C	Benzo(a)pyrene	1.137	1.193	-4.9	105	0.00
91	Dibenzo(a,h)anthracene	1.189	1.258	-5.8	102	0.00
92	Benzo(g,h,i)perylene	1.195	1.250	-4.6	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
 Data File : BF142250.D
 Acq On : 01 May 2025 10:17
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 01 11:46:27 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	41.429	-3.6	101	0.00
3	Pyridine	40.000	42.158	-5.4	104	0.00
4	n-Nitrosodimethylamine	40.000	42.418	-6.0	103	0.00
5 S	2-Fluorophenol	80.000	79.825	0.2	99	0.00
6	Aniline	40.000	41.579	-3.9	103	0.00
7 S	Phenol-d6	80.000	83.296	-4.1	102	0.00
8	2-Chlorophenol	40.000	39.409	1.5	96	0.00
9	Benzaldehyde	40.000	42.464	-6.2	104	0.00
10 C	Phenol	40.000	40.684	-1.7	100	0.00
11	bis(2-Chloroethyl)ether	40.000	42.030	-5.1	104	0.00
12	1,3-Dichlorobenzene	40.000	41.505	-3.8	103	0.00
13 C	1,4-Dichlorobenzene	40.000	41.388	-3.5	102	0.00
14	1,2-Dichlorobenzene	40.000	41.885	-4.7	104	0.00
15	Benzyl Alcohol	40.000	41.095	-2.7	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.870	-4.7	104	0.00
17	2-Methylphenol	40.000	41.610	-4.0	102	0.00
18	Hexachloroethane	40.000	41.867	-4.7	103	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.842	-2.1	102	0.00
20	3+4-Methylphenols	40.000	42.351	-5.9	104	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	41.462	-3.7	104	0.00
23 S	Nitrobenzene-d5	80.000	87.469	-9.3	100	0.00
24	Nitrobenzene	40.000	42.902	-7.3	100	0.00
25	Isophorone	40.000	42.059	-5.1	104	0.00
26 C	2-Nitrophenol	40.000	34.044	14.9	86	0.00
27	2,4-Dimethylphenol	40.000	41.898	-4.7	102	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.820	-4.6	103	0.00
29 C	2,4-Dichlorophenol	40.000	42.280	-5.7	100	0.00
30	1,2,4-Trichlorobenzene	40.000	41.910	-4.8	103	0.00
31	Naphthalene	40.000	41.304	-3.3	103	0.00
32	Benzoic acid	40.000	34.514	13.7	92	0.00
33	4-Chloroaniline	40.000	41.683	-4.2	103	0.00
34 C	Hexachlorobutadiene	40.000	42.382	-6.0	103	0.00
35	Caprolactam	40.000	44.410	-11.0	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.337	-8.3	103	0.00
37	2-Methylnaphthalene	40.000	41.414	-3.5	104	0.00
38	1-Methylnaphthalene	40.000	41.597	-4.0	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.641	-1.6	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.605	1.0	92	0.00
42 S	2,4,6-Tribromophenol	80.000	85.983	-7.5	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.551	-1.4	98	0.00
44	2,4,5-Trichlorophenol	40.000	42.108	-5.3	98	0.00
45 S	2-Fluorobiphenyl	80.000	79.446	0.7	103	0.00
46	1,1'-Biphenyl	40.000	40.922	-2.3	103	0.00
47	2-Chloronaphthalene	40.000	40.663	-1.7	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
 Data File : BF142250.D
 Acq On : 01 May 2025 10:17
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 01 11:46:27 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.508	6.2	97	0.00
49	Acenaphthylene	40.000	40.906	-2.3	103	0.00
50	Dimethylphthalate	40.000	42.093	-5.2	105	0.00
51	2,6-Dinitrotoluene	40.000	38.527	3.7	98	0.00
52 C	Acenaphthene	40.000	41.576	-3.9	104	0.00
53	3-Nitroaniline	40.000	40.152	-0.4	102	0.00
54 P	2,4-Dinitrophenol	40.000	37.926	5.2	101	0.00
55	Dibenzofuran	40.000	41.484	-3.7	104	0.00
56 P	4-Nitrophenol	40.000	41.706	-4.3	100	0.00
57	2,4-Dinitrotoluene	40.000	39.850	0.4	104	0.00
58	Fluorene	40.000	40.709	-1.8	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.475	-6.2	99	0.00
60	Diethylphthalate	40.000	42.799	-7.0	106	0.00
61	4-Chlorophenyl-phenylether	40.000	41.333	-3.3	104	0.00
62	4-Nitroaniline	40.000	42.018	-5.0	107	0.00
63	Azobenzene	40.000	41.895	-4.7	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	35.905	10.2	102	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.322	-3.3	104	0.00
67	4-Bromophenyl-phenylether	40.000	42.258	-5.6	105	0.00
68	Hexachlorobenzene	40.000	41.207	-3.0	104	0.00
69	Atrazine	40.000	45.262	-13.2	112	0.00
70 C	Pentachlorophenol	40.000	41.491	-3.7	100	0.00
71	Phenanthrene	40.000	41.326	-3.3	105	0.00
72	Anthracene	40.000	41.339	-3.3	106	0.00
73	Carbazole	40.000	41.429	-3.6	106	0.00
74	Di-n-butylphthalate	40.000	44.864	-12.2	112	0.00
75 C	Fluoranthene	40.000	42.947	-7.4	111	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	116	0.00
77	Benzydine	40.000	46.160	-15.4	118	0.00
78	Pyrene	40.000	42.016	-5.0	111	0.00
79 S	Terphenyl-d14	80.000	82.722	-3.4	111	0.00
80	Butylbenzylphthalate	40.000	40.417	-1.0	115	0.00
81	Benzo(a)anthracene	40.000	41.049	-2.6	111	0.00
82	3,3'-Dichlorobenzidine	40.000	42.615	-6.5	110	0.00
83	Chrysene	40.000	42.445	-6.1	122	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.467	8.8	104	0.00
85 c	Di-n-octyl phthalate	40.000	33.845	15.4	94	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.864	-4.7	101	0.00
88	Benzo(b)fluoranthene	40.000	39.998	0.0	101	0.00
89	Benzo(k)fluoranthene	40.000	43.643	-9.1	115	0.00
90 C	Benzo(a)pyrene	40.000	41.969	-4.9	105	0.00
91	Dibenzo(a,h)anthracene	40.000	42.333	-5.8	102	0.00
92	Benzo(g,h,i)perylene	40.000	41.848	-4.6	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142250.D
Acq On : 01 May 2025 10:17
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 01 11:46:27 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG No.: Q1915

Instrument ID: BNA_F Calibration Date/Time: 05/02/2025 10:39

Lab File ID: BF142274.D Init. Calib. Date(s): 04/30/2025 04/30/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:24 14:43

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Pyridine	1.454	1.384		-4.8	
2-Fluorophenol	1.252	1.195		-4.6	
Phenol-d6	1.512	1.457		-3.6	
1,4-Dichlorobenzene	1.475	1.364		-7.5	20.0
2-Methylphenol	1.060	1.005		-5.2	
3+4-Methylphenols	1.326	1.295		-2.3	
Nitrobenzene-d5	0.304	0.332		9.2	
Hexachloroethane	0.483	0.473		-2.1	
Nitrobenzene	0.299	0.315		5.4	
Hexachlorobutadiene	0.175	0.164		-6.3	20.0
2,4,6-Trichlorophenol	0.349	0.360		3.2	20.0
2-Fluorobiphenyl	1.418	1.293		-8.8	
2,4,5-Trichlorophenol	0.367	0.363		-1.1	
2,4-Dinitrotoluene	0.239	0.314		31.4	
2,4,6-Tribromophenol	0.182	0.190		4.4	
Hexachlorobenzene	0.249	0.232		-6.8	
Pentachlorophenol	0.132	0.137		3.8	20.0
Terphenyl-d14	1.375	1.344		-2.3	

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142274.D
 Acq On : 02 May 2025 10:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025

Quant Time: May 02 11:50:23 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Apr 30 16:00:01 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	189457	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	747235	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	389025	20.000	ng	0.00
64) Phenanthrene-d10	11.427	188	652685	20.000	ng	0.00
76) Chrysene-d12	14.068	240	354752	20.000	ng	0.00
86) Perylene-d12	15.557	264	327562	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	905917	76.366	ng	0.00
7) Phenol-d6	6.522	99	1104282	77.087	ng	0.00
23) Nitrobenzene-d5	7.469	82	992335	87.247	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	295525	83.329	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2012359	72.935	ng	0.00
79) Terphenyl-d14	13.016	244	1907599	78.207	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.687	88	199149	37.589	ng	97
3) Pyridine	3.457	79	524389	38.083	ng	99
4) n-Nitrosodimethylamine	3.404	42	276556	38.811	ng	99
6) Aniline	6.569	93	768076	37.880	ng	99
8) 2-Chlorophenol	6.687	128	480593	39.206	ng	98
9) Benzaldehyde	6.457	77	378814	37.835	ng	99
10) Phenol	6.540	94	585536m	37.967	ng	
11) bis(2-Chloroethyl)ether	6.640	93	452236	37.841	ng	99
12) 1,3-Dichlorobenzene	6.851	146	524346	37.811	ng	100
13) 1,4-Dichlorobenzene	6.928	146	516930	36.999	ng	99
14) 1,2-Dichlorobenzene	7.081	146	505354	37.888	ng	99
15) Benzyl Alcohol	7.045	79	393158	38.697	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.181	45	817791	37.895	ng	99
17) 2-Methylphenol	7.151	107	380905	37.931	ng	99
18) Hexachloroethane	7.422	117	179209	39.190	ng	96
19) n-Nitroso-di-n-propyla...	7.316	70	330837	37.373	ng	97
20) 3+4-Methylphenols	7.304	107	490537	39.062	ng	94
22) Acetophenone	7.316	105	628918	36.381	ng	98
24) Nitrobenzene	7.487	77	470320	42.143	ng	99
25) Isophorone	7.728	82	878449	36.800	ng	99
26) 2-Nitrophenol	7.804	139	201999	43.995	ng	97
27) 2,4-Dimethylphenol	7.834	122	458385	38.087	ng	100
28) bis(2-Chloroethoxy)met...	7.934	93	552958	36.730	ng	100
29) 2,4-Dichlorophenol	8.039	162	389539	38.714	ng	99
30) 1,2,4-Trichlorobenzene	8.128	180	414934	36.954	ng	100
31) Naphthalene	8.210	128	1363971	36.602	ng	100
32) Benzoic acid	7.928	122	233233m	42.364	ng	
33) 4-Chloroaniline	8.257	127	579952	37.373	ng	99
34) Hexachlorobutadiene	8.328	225	245743	37.512	ng	98
35) Caprolactam	8.622	113	120939	39.988	ng	98
36) 4-Chloro-3-methylphenol	8.728	107	397404	38.463	ng	98
37) 2-Methylnaphthalene	8.904	142	837448	36.551	ng	100
38) 1-Methylnaphthalene	9.004	142	871131	36.702	ng	100
40) 1,2,4,5-Tetrachloroben...	9.069	216	405798	37.065	ng	99
41) Hexachlorocyclopentadiene	9.057	237	254671	40.985	ng	99
43) 2,4,6-Trichlorophenol	9.175	196	280415	41.332	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142274.D
 Acq On : 02 May 2025 10:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025

Quant Time: May 02 11:50:23 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Apr 30 16:00:01 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	282545	39.611	ng	98
46) 1,1'-Biphenyl	9.363	154	1118632	36.988	ng	99
47) 2-Chloronaphthalene	9.392	162	833002	37.021	ng	100
48) 2-Nitroaniline	9.481	65	245382	42.014	ng	97
49) Acenaphthylene	9.810	152	1407244	37.062	ng	99
50) Dimethylphthalate	9.663	163	931138	37.413	ng	100
51) 2,6-Dinitrotoluene	9.722	165	191021	40.686	ng	98
52) Acenaphthene	9.981	154	832658	37.558	ng	99
53) 3-Nitroaniline	9.892	138	232079	40.499	ng	97
54) 2,4-Dinitrophenol	9.992	184	66816	50.599	ng	91
55) Dibenzofuran	10.151	168	1230121	37.015	ng	100
56) 4-Nitrophenol	10.033	139	176430	45.101	ng	98
57) 2,4-Dinitrotoluene	10.122	165	244640	42.928	ng	100
58) Fluorene	10.492	166	931222	36.852	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	240492	40.702	ng	99
60) Diethylphthalate	10.363	149	912963	37.148	ng	100
61) 4-Chlorophenyl-phenyle...	10.486	204	441273	36.772	ng	97
62) 4-Nitroaniline	10.504	138	220163	42.072	ng	98
63) Azobenzene	10.645	77	895687	36.907	ng	99
65) 4,6-Dinitro-2-methylph...	10.528	198	97288	45.858	ng	97
66) n-Nitrosodiphenylamine	10.604	169	829293	37.187	ng	100
67) 4-Bromophenyl-phenylether	10.975	248	277516	38.179	ng	100
68) Hexachlorobenzene	11.039	284	302639	37.175	ng	98
69) Atrazine	11.128	200	228706	38.907	ng	99
70) Pentachlorophenol	11.227	266	178739	41.560	ng	98
71) Phenanthrene	11.457	178	1313668	37.261	ng	100
72) Anthracene	11.510	178	1349761	37.446	ng	99
73) Carbazole	11.657	167	1245984	37.893	ng	99
74) Di-n-butylphthalate	11.992	149	1235690	37.366	ng	99
75) Fluoranthene	12.645	202	1316477	37.775	ng	99
77) Benzidine	12.763	184	629233	41.743	ng	100
78) Pyrene	12.874	202	1327219	40.557	ng	100
80) Butylbenzylphthalate	13.492	149	318246	36.531	ng	98
81) Benzo(a)anthracene	14.057	228	917653	38.532	ng	99
82) 3,3'-Dichlorobenzidine	14.021	252	264696	37.840	ng	99
83) Chrysene	14.098	228	786998	36.304	ng	99
84) Bis(2-ethylhexyl)phtha...	14.051	149	344949	29.284	ng	100
85) Di-n-octyl phthalate	14.668	149	576205	28.225	ng	100
87) Indeno(1,2,3-cd)pyrene	17.068	276	957230	39.896	ng	100
88) Benzo(b)fluoranthene	15.121	252	702872	34.545	ng	99
89) Benzo(k)fluoranthene	15.151	252	719163	38.642	ng	99
90) Benzo(a)pyrene	15.498	252	703822	37.783	ng	100
91) Dibenzo(a,h)anthracene	17.086	278	781717	40.144	ng	99
92) Benzo(g,h,i)perylene	17.521	276	796796	40.703	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142274.D
 Acq On : 02 May 2025 10:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/05/2025
 Supervised By :Jagrut Upadhyay 05/05/2025


Quant Time: May 02 11:50:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:152 Resp: 18945
Ion Ratio Lower Upper
152 100
150 162.5 130.2 195.2
115 59.4 47.0 70.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#2
1,4-Dioxane
Concen: 37.589 ng
RT: 2.687 min Scan# 84
Delta R.T. -0.012 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Tgt Ion: 88 Resp: 199149
Ion Ratio Lower Upper
88 100
58 92.8 72.2 108.2
43 35.2 27.0 40.6





#3
Pyridine
Concen: 38.083 ng
RT: 3.457 min Scan# 216
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 524389
Ion Ratio Lower Upper
79 100
52 88.7 71.5 107.3
51 43.0 34.5 51.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#4
n-Nitrosodimethylamine
Concen: 38.811 ng
RT: 3.404 min Scan# 206
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion: 42 Resp: 276556
Ion Ratio Lower Upper
42 100
74 112.7 89.5 134.3
44 7.7 6.6 9.8






#5
2-Fluorophenol
Concen: 76.366 ng
RT: 5.516 min Scan# 50
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:112 Resp: 905917
Ion Ratio Lower Upper
112 100
64 59.5 48.4 72.6
63 30.8 24.8 37.2

Manual Integrations




APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#6
Aniline
Concen: 37.880 ng
RT: 6.569 min Scan# 744
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion: 93 Resp: 768076
Ion Ratio Lower Upper
93 100
66 39.8 32.1 48.1
65 20.4 16.6 24.8





#7
Phenol-d6
Concen: 77.087 ng
RT: 6.522 min Scan# 714
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 99 Resp: 110428
Ion Ratio Lower Upper
99 100
42 21.5 16.3 24.5
71 32.0 25.0 37.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#8
2-Chlorophenol
Concen: 39.206 ng
RT: 6.687 min Scan# 764
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Tgt Ion:128 Resp: 480593
Ion Ratio Lower Upper
128 100
130 32.5 13.1 53.1
64 49.8 28.6 68.6





#9
Benzaldehyde
Concen: 37.835 ng
RT: 6.457 min Scan# 725
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 378814
Ion Ratio Lower Upper
77 100
105 99.9 79.1 119.1
106 95.6 76.0 116.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#10
Phenol
Concen: 37.967 ng m
RT: 6.540 min Scan# 739
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Tgt Ion: 94 Resp: 585536
Ion Ratio Lower Upper
94 100
65 28.8 8.9 48.9
66 38.2 18.7 58.7





#11
bis(2-Chloroethyl)ether
Concen: 37.841 ng
RT: 6.640 min Scan# 756
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion: 93 Resp: 452236
Ion Ratio Lower Upper
93 100
63 83.2 62.2 102.2
95 32.1 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#12
1,3-Dichlorobenzene
Concen: 37.811 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:146 Resp: 524346
Ion Ratio Lower Upper
146 100
148 63.1 50.5 75.7
75 29.1 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 36.999 ng

RT: 6.928 min Scan# 805

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:146 Resp: 516930

Ion Ratio Lower Upper

146 100

148 64.5 50.6 76.0





111 42.4 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#14

1,2-Dichlorobenzene

Concen: 37.888 ng

RT: 7.081 min Scan# 831

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:146 Resp: 505354


Ion Ratio Lower Upper

146 100

148 63.9 50.5 75.7

111 42.9 34.9 52.3





#15

Benzyl Alcohol

Concen: 38.697 ng

RT: 7.045 min Scan# 81

Delta R.T. -0.000 min

Lab File: BF142274.D


Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion: 79	Resp: 393158
Ion Ratio	Lower Upper
79	100
108	81.8 64.5 96.7
77	61.8 49.2 73.8

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#16

2,2'-oxybis(1-Chloropropane)


Concen: 37.895 ng

RT: 7.181 min Scan# 848



Delta R.T. -0.006 min


Lab File: BF142274.D

Acq: 02 May 2025 10:39



Tgt Ion: 45	Resp: 817791
Ion Ratio	Lower Upper
45	100
77	10.6 0.0 31.3
79	8.3 0.0 28.5






#17
2-Methylphenol
Concen: 37.931 ng
RT: 7.151 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:107 Resp: 38090
Ion Ratio Lower Upper
107 100
108 114.8 91.0 136.4
77 43.4 34.4 51.6
79 42.5 34.4 51.6


Manual Integrations

APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#18
Hexachloroethane
Concen: 39.190 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39



Tgt Ion:117 Resp: 179209
Ion Ratio Lower Upper
117 100
119 96.8 77.1 115.7
201 92.0 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 37.373 ng

RT: 7.316 min Scan# 81

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 70 Resp: 33083

Ion Ratio Lower Upper

70 100

42 70.2 53.4 80.2

101 9.7 7.7 11.5





130 21.2 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#20

3+4-Methylphenols

Concen: 39.062 ng

RT: 7.304 min Scan# 869

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Tgt Ion:107 Resp: 490537




Ion Ratio Lower Upper


107 100

108 90.4 71.2 111.2

77 73.0 41.1 81.1

79 27.9 8.1 48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1020
Delta R.T. -0.001 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:136 Resp: 747235
Ion Ratio Lower Upper
136 100
137 10.7 8.7 13.1
54 9.1 7.7 11.5
68 6.0 5.0 7.6

Manual Integrations

APPROVED




Reviewed By :Rahul Chavli 05/05/2025


Supervised By :Jagrut Upadhyay 05/05/2025



#22
Acetophenone
Concen: 36.381 ng
RT: 7.316 min Scan# 871
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:105 Resp: 628918
Ion Ratio Lower Upper
105 100
71 7.8 5.7 8.5
51 35.6 30.1 45.1
120 24.0 19.0 28.4





#23

Nitrobenzene-d5

Concen: 87.247 ng

RT: 7.469 min Scan# 89

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 99233

Ion Ratio Lower Upper

82 100

128 46.4 37.0 55.6





54 60.8 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#24

Nitrobenzene

Concen: 42.143 ng

RT: 7.487 min Scan# 900

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion: 77 Resp: 470320


Ion Ratio Lower Upper

77 100

123 47.0 38.6 58.0


65 14.5 11.5 17.3





#25
Isophorone
Concen: 36.800 ng
RT: 7.728 min Scan# 941
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion: 82 Resp: 878449
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 19.7 15.3 22.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#26
2-Nitrophenol
Concen: 43.995 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:139 Resp: 201999
Ion Ratio Lower Upper
139 100
109 26.2 21.8 32.8
65 48.9 41.1 61.7





#27

2,4-Dimethylphenol

Concen: 38.087 ng

RT: 7.834 min Scan# 91

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:122 Resp: 45838

Ion Ratio Lower Upper

122 100

107 107.6 86.0 129.0





121 58.0 46.3 69.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#28

bis(2-Chloroethoxy)methane

Concen: 36.730 ng

RT: 7.934 min Scan# 976

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion: 93 Resp: 552958


Ion Ratio Lower Upper

93 100

95 32.7 26.2 39.2

123 12.4 9.7 14.5





#29

2,4-Dichlorophenol

Concen: 38.714 ng

RT: 8.039 min Scan# 994

Delta R.T. -0.000 min

Lab File: BF142274.D


Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:162 Resp: 389539

Ion Ratio Lower Upper

162 100

164 64.3 43.0 83.0




98 36.3 16.7 56.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#30

1,2,4-Trichlorobenzene

Concen: 36.954 ng

RT: 8.128 min Scan# 1009

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:180 Resp: 414934


Ion Ratio Lower Upper

180 100

182 94.7 75.7 113.5


145 32.0 25.4 38.0





#31
Naphthalene
Concen: 36.602 ng
RT: 8.210 min Scan# 1024
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:128 Resp: 136397
Ion Ratio Lower Upper
128 100
129 11.3 9.0 13.4
127 13.0 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#32
Benzoic acid
Concen: 42.364 ng m
RT: 7.928 min Scan# 975
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:122 Resp: 233233
Ion Ratio Lower Upper
122 100
105 118.1 98.2 138.2
77 89.8 66.8 106.8





#33

4-Chloroaniline

Concen: 37.373 ng

RT: 8.257 min Scan# 1031

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:127 Resp: 579952

Ion Ratio Lower Upper

127 100

129 32.3 26.1 39.1

65 31.1 25.5 38.3





92 19.1 15.4 23.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#34

Hexachlorobutadiene

Concen: 37.512 ng

RT: 8.328 min Scan# 1043

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:225 Resp: 245743


Ion Ratio Lower Upper

225 100

223 62.8 49.7 74.5

227 62.0 51.0 76.4





#35

Caprolactam

Concen: 39.988 ng

RT: 8.622 min Scan# 1093

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:113 Resp: 120939

Ion Ratio Lower Upper

113 100

55 194.6 177.7 217.7





56 143.4 127.1 167.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#36

4-Chloro-3-methylphenol

Concen: 38.463 ng

RT: 8.728 min Scan# 1111

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:107 Resp: 397404


Ion Ratio Lower Upper

107 100

144 26.9 21.0 31.4


142 84.0 66.1 99.1





#37
2-Methylnaphthalene
Concen: 36.551 ng
RT: 8.904 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:142 Resp: 837448
Ion Ratio Lower Upper
142 100
141 88.5 70.5 105.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#38
1-Methylnaphthalene
Concen: 36.702 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:142 Resp: 871131
Ion Ratio Lower Upper
142 100
141 91.0 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.945 min Scan# 1169

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:164 Resp: 389025

Ion Ratio Lower Upper

164 100

162 98.2 78.6 117.8





160 44.4 35.7 53.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#40

1,2,4,5-Tetrachlorobenzene

Concen: 37.065 ng

RT: 9.069 min Scan# 1169

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Tgt Ion:216 Resp: 405798




Ion Ratio Lower Upper


216 100

214 79.9 63.8 95.6

179 21.4 16.7 25.1

108 18.8 15.6 23.4





#41

Hexachlorocyclopentadiene

Concen: 40.985 ng

RT: 9.057 min Scan# 1167

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:237 Resp: 25467

Ion Ratio Lower Upper

237 100

235 62.7 43.5 83.5





272 13.3 0.0 32.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#42

2,4,6-Tribromophenol

Concen: 83.329 ng

RT: 10.733 min Scan# 1452

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:330 Resp: 295525


Ion Ratio Lower Upper

330 100

332 96.1 76.5 114.7


141 31.6 24.3 36.5





#43
2,4,6-Trichlorophenol
Concen: 41.332 ng
RT: 9.175 min Scan# 1187
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:196 Resp: 280415
Ion Ratio Lower Upper
196 100
198 96.2 77.4 116.2
200 31.1 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#44
2,4,5-Trichlorophenol
Concen: 39.611 ng
RT: 9.210 min Scan# 1193
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:196 Resp: 282545
Ion Ratio Lower Upper
196 100
198 95.4 75.2 112.8
97 48.3 39.8 59.8
132 28.9 23.8 35.8





#45

2-Fluorobiphenyl

Concen: 72.935 ng

RT: 9.263 min Scan# 1202

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:172 Resp: 201235

Ion Ratio Lower Upper

172 100

171 35.8 29.0 43.4





170 23.7 19.4 29.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#46

1,1'-Biphenyl

Concen: 36.988 ng

RT: 9.363 min Scan# 1219

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:154 Resp: 1118632


Ion Ratio Lower Upper

154 100

153 39.7 20.5 60.5


76 13.1 0.0 32.8





#47
2-Chloronaphthalene
Concen: 37.021 ng
RT: 9.392 min Scan# 1224
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:162 Resp: 833002
Ion Ratio Lower Upper
162 100
127 40.6 32.6 49.0
164 32.9 26.1 39.1

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#48
2-Nitroaniline
Concen: 42.014 ng
RT: 9.481 min Scan# 1239
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39



Tgt Ion: 65 Resp: 245382
Ion Ratio Lower Upper
65 100
92 66.8 54.3 81.5
138 107.4 88.6 133.0





#49

Acenaphthylene

Concen: 37.062 ng

RT: 9.810 min Scan# 1295

Delta R.T. -0.000 min

Lab File: BF142274.D


Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:152 Resp: 140724

Ion Ratio Lower Upper

152 100

151 19.8 16.2 24.2




153 13.3 10.7 16.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#50

Dimethylphthalate


Concen: 37.413 ng

RT: 9.663 min Scan# 1270

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39





Tgt Ion:163 Resp: 931138


Ion Ratio Lower Upper

163 100

194 4.0 3.2 4.8


164 10.3 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 40.686 ng
RT: 9.722 min Scan# 1281
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:165 Resp: 19102
Ion Ratio Lower Upper
165 100
63 63.3 48.5 72.7
89 55.3 43.9 65.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#52
Acenaphthene
Concen: 37.558 ng
RT: 9.981 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:154 Resp: 832658
Ion Ratio Lower Upper
154 100
153 110.1 88.6 133.0
152 52.1 42.1 63.1





#53

3-Nitroaniline

Concen: 40.499 ng

RT: 9.892 min Scan# 1309

Delta R.T. 0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:138 Resp: 232079

Ion Ratio Lower Upper

138 100

108 11.7 11.7 17.5





92 111.1 91.2 136.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#54

2,4-Dinitrophenol

Concen: 50.599 ng

RT: 9.992 min Scan# 1326

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:184 Resp: 66816


Ion Ratio Lower Upper

184 100

63 90.7 69.2 103.8


154 361.3 271.4 407.2





#55
Dibenzenofuran
Concen: 37.015 ng
RT: 10.151 min Scan# 1353
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:168 Resp: 123012
Ion Ratio Lower Upper
168 100
139 36.5 29.4 44.0
169 13.4 11.0 16.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#56
4-Nitrophenol
Concen: 45.101 ng
RT: 10.033 min Scan# 1333
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:139 Resp: 176430
Ion Ratio Lower Upper
139 100
109 66.5 44.9 84.9
65 97.0 74.7 114.7






#57
2,4-Dinitrotoluene
Concen: 42.928 ng
RT: 10.122 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:165 Resp: 244640
Ion Ratio Lower Upper
165 100
63 48.6 38.8 58.2
89 70.9 57.1 85.7
182 2.7 2.2 3.2


Manual Integrations

APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#58
Fluorene
Concen: 36.852 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Tgt Ion:166 Resp: 931222
Ion Ratio Lower Upper
166 100
165 93.3 73.9 110.9
167 13.5 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 40.702 ng
RT: 10.263 min Scan# 1372
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:232 Resp: 240492
Ion Ratio Lower Upper
232 100
131 45.3 35.9 53.9
130 2.2 1.8 2.6
166 29.5 24.0 36.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#60
Diethylphthalate
Concen: 37.148 ng
RT: 10.363 min Scan# 1389
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Tgt Ion:149 Resp: 912963
Ion Ratio Lower Upper
149 100
177 21.7 17.6 26.4
150 12.6 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 36.772 ng
RT: 10.486 min Scan# 1410
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:204 Resp: 44127
Ion Ratio Lower Upper
204 100
206 32.9 25.6 38.4
141 58.7 48.9 73.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#62
4-Nitroaniline
Concen: 42.072 ng
RT: 10.504 min Scan# 1413
Delta R.T. 0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Tgt Ion:138 Resp: 220163
Ion Ratio Lower Upper
138 100
92 48.3 29.2 69.2
108 67.7 49.1 89.1





#63
Azobenzene
Concen: 36.907 ng
RT: 10.645 min Scan# 1437
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 895685
Ion Ratio Lower Upper
77 100
182 26.3 7.3 47.3
105 22.2 2.5 42.5
51 35.0 14.4 54.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.427 min Scan# 1570
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Tgt Ion:188 Resp: 652685
Ion Ratio Lower Upper
188 100
94 10.5 8.2 12.2
80 11.1 8.6 12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 45.858 ng
RT: 10.528 min Scan# 1418
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:198 Resp: 9728
Ion Ratio Lower Upper
198 100
51 63.1 40.6 80.6
105 47.7 29.2 69.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#66
n-Nitrosodiphenylamine
Concen: 37.187 ng
RT: 10.604 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion:169 Resp: 829293
Ion Ratio Lower Upper
169 100
168 66.1 53.0 79.6
167 36.0 28.8 43.2





#67
4-Bromophenyl-phenylether
Concen: 38.179 ng
RT: 10.975 min Scan# 1493
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:248 Resp: 277510
Ion Ratio Lower Upper
248 100
250 96.1 77.2 115.8
141 73.4 58.4 87.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#68
Hexachlorobenzene
Concen: 37.175 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39



Tgt Ion:284 Resp: 302639
Ion Ratio Lower Upper
284 100
142 32.8 24.6 37.0
249 29.7 23.2 34.8





#69

Atrazine

Concen: 38.907 ng

RT: 11.128 min Scan# 1519

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:200 Resp: 228700

Ion Ratio Lower Upper

200 100

173 27.5 8.6 48.6





215 49.1 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#70

Pentachlorophenol

Concen: 41.560 ng

RT: 11.227 min Scan# 1536

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:266 Resp: 178739


Ion Ratio Lower Upper

266 100

268 63.4 50.0 75.0

264 63.0 49.3 73.9





#71

Phenanthrene

Concen: 37.261 ng

RT: 11.457 min Scan# 1575

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:178 Resp: 131366

Ion Ratio Lower Upper

178 100

176 19.3 15.5 23.3





179 15.6 12.6 19.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#72

Anthracene

Concen: 37.446 ng

RT: 11.510 min Scan# 1584

Delta R.T. -0.000 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:178 Resp: 1349761


Ion Ratio Lower Upper

178 100

176 18.8 15.3 22.9


179 15.5 12.7 19.1





#73
 Carbazole
 Concen: 37.893 ng
 RT: 11.657 min Scan# 1609
 Delta R.T. -0.006 min
 Lab File: BF142274.D
 Acq: 02 May 2025 10:39




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040




Tgt Ion:167 Resp: 1245984
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.3 25.9
 139 12.8 10.3 15.5

Manual Integrations
 APPROVED



Reviewed By :Rahul Chavli 05/05/2025
 Supervised By :Jagrut Upadhyay 05/05/2025




#74
 Di-n-butylphthalate
 Concen: 37.366 ng
 RT: 11.992 min Scan# 1666
 Delta R.T. -0.006 min
 Lab File: BF142274.D
 Acq: 02 May 2025 10:39




Tgt Ion:149 Resp: 1235690
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.5 11.3
 104 4.5 3.6 5.4





#75
Fluoranthene
Concen: 37.775 ng
RT: 12.645 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:202 Resp: 131647
Ion Ratio Lower Upper
202 100
101 11.2 0.0 32.4
203 17.6 0.0 37.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.068 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:240 Resp: 354752
Ion Ratio Lower Upper
240 100
120 12.8 10.1 15.1
236 25.7 20.5 30.7





#77

Benzidine

Concen: 41.743 ng

RT: 12.763 min Scan# 1797

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:184 Resp: 629233

Ion Ratio Lower Upper

184 100

185 14.6 11.8 17.6





183 11.1 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#78

Pyrene

Concen: 40.557 ng

RT: 12.874 min Scan# 1816

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:202 Resp: 1327219


Ion Ratio Lower Upper

202 100

200 20.0 16.1 24.1


203 17.6 14.1 21.1





#79
Terphenyl-d14
Concen: 78.207 ng
RT: 13.016 min Scan# 1840
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:244 Resp: 1907599
Ion Ratio Lower Upper
244 100
212 6.7 5.3 7.9
122 11.4 9.1 13.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#80
Butylbenzylphthalate
Concen: 36.531 ng
RT: 13.492 min Scan# 1921
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39


Tgt Ion:149 Resp: 318246
Ion Ratio Lower Upper
149 100
91 66.3 54.4 81.6
206 20.4 15.9 23.9





#81
Benzo(a)anthracene
Concen: 38.532 ng
RT: 14.057 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:228 Resp: 91765
Ion Ratio Lower Upper
228 100
226 26.8 21.9 32.9
229 19.7 16.0 24.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 37.840 ng
RT: 14.021 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Tgt Ion:252 Resp: 264696
Ion Ratio Lower Upper
252 100
254 63.1 51.0 76.6
126 12.1 9.4 14.2





#83
Chrysene
Concen: 36.304 ng
RT: 14.098 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 786998
Ion Ratio Lower Upper
228 100
226 30.3 24.0 36.0
229 19.9 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 29.284 ng
RT: 14.051 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:149 Resp: 344949
Ion Ratio Lower Upper
149 100
167 27.0 21.7 32.5
279 3.8 3.0 4.4





#85

Di-n-octyl phthalate

Concen: 28.225 ng

RT: 14.668 min Scan# 2121

Delta R.T. -0.012 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:149 Resp: 576205

Ion Ratio Lower Upper

149 100

167 1.5 1.1 1.7





43 11.6 9.3 13.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.557 min Scan# 2272

Delta R.T. -0.012 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:264 Resp: 327562


Ion Ratio Lower Upper

264 100

260 23.5 18.4 27.6

265 21.8 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 39.896 ng

RT: 17.068 min Scan# 2198

Delta R.T. -0.012 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:276 Resp: 957230

Ion Ratio Lower Upper

276 100

138 28.0 22.6 33.8





277 24.9 20.1 30.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#88

Benzo(b)fluoranthene

Concen: 34.545 ng

RT: 15.121 min Scan# 2198

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:252 Resp: 702872


Ion Ratio Lower Upper

252 100

253 21.6 17.7 26.5

125 10.5 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 38.642 ng

RT: 15.151 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:252 Resp: 71916

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.8





125 10.5 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#90

Benzo(a)pyrene

Concen: 37.783 ng

RT: 15.498 min Scan# 2262

Delta R.T. -0.006 min

Lab File: BF142274.D

Acq: 02 May 2025 10:39




Tgt Ion:252 Resp: 703822


Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4


125 11.5 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 40.144 ng
RT: 17.086 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:278 Resp: 78171
Ion Ratio Lower Upper
278 100
139 18.2 15.2 22.8
279 23.9 19.0 28.6




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#92
Benzo(g,h,i)perylene
Concen: 40.703 ng
RT: 17.521 min Scan# 2606
Delta R.T. -0.018 min
Lab File: BF142274.D
Acq: 02 May 2025 10:39

Tgt Ion:276 Resp: 796796
Ion Ratio Lower Upper
276 100
277 23.5 18.7 28.1
138 24.5 19.4 29.0



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142274.D
 Acq On : 02 May 2025 10:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 02 11:50:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.559	0.526	5.9	93	-0.01
3	Pyridine	1.454	1.384	4.8	96	0.00
4	n-Nitrosodimethylamine	0.752	0.730	2.9	95	0.00
5 S	2-Fluorophenol	1.252	1.195	4.6	96	0.00
6	Aniline	2.140	2.027	5.3	95	0.00
7 S	Phenol-d6	1.512	1.457	3.6	96	0.00
8	2-Chlorophenol	1.294	1.268	2.0	97	0.00
9	Benzaldehyde	1.057	1.000	5.4	94	0.00
10 C	Phenol	1.628	1.545	5.1	95	0.00
11	bis(2-Chloroethyl)ether	1.262	1.194	5.4	95	0.00
12	1,3-Dichlorobenzene	1.464	1.384	5.5	95	0.00
13 C	1,4-Dichlorobenzene	1.475	1.364	7.5	93	0.00
14	1,2-Dichlorobenzene	1.408	1.334	5.3	95	0.00
15	Benzyl Alcohol	1.073	1.038	3.3	95	0.00
16	2,2'-oxybis(1-Chloropropane	2.278	2.158	5.3	95	0.00
17	2-Methylphenol	1.060	1.005	5.2	94	0.00
18	Hexachloroethane	0.483	0.473	2.1	97	0.00
19 P	n-Nitroso-di-n-propylamine	0.934	0.873	6.5	95	0.00
20	3+4-Methylphenols	1.326	1.295	2.3	97	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
22	Acetophenone	0.463	0.421	9.1	95	0.00
23 S	Nitrobenzene-d5	0.304	0.332	-9.2	103	0.00
24	Nitrobenzene	0.299	0.315	-5.4	102	0.00
25	Isophorone	0.639	0.588	8.0	94	0.00
26 C	2-Nitrophenol	0.102	0.135	-32.4#	124	0.00
27	2,4-Dimethylphenol	0.322	0.307	4.7	96	0.00
28	bis(2-Chloroethoxy)methane	0.403	0.370	8.2	94	0.00
29 C	2,4-Dichlorophenol	0.269	0.261	3.0	95	0.00
30	1,2,4-Trichlorobenzene	0.301	0.278	7.6	94	0.00
31	Naphthalene	0.997	0.913	8.4	95	0.00
32	Benzoic acid	0.125	0.156	-24.8	127	0.00
33	4-Chloroaniline	0.415	0.388	6.5	96	0.00
34 C	Hexachlorobutadiene	0.175	0.164	6.3	94	0.00
35	Caprolactam	0.081	0.081	0.0	99	0.00
36 C	4-Chloro-3-methylphenol	0.277	0.266	4.0	95	0.00
37	2-Methylnaphthalene	0.613	0.560	8.6	95	0.00
38	1-Methylnaphthalene	0.635	0.583	8.2	95	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	0.563	0.522	7.3	95	0.00
41 P	Hexachlorocyclopentadiene	0.319	0.327	-2.5	98	0.00
42 S	2,4,6-Tribromophenol	0.182	0.190	-4.4	98	0.00
43 C	2,4,6-Trichlorophenol	0.349	0.360	-3.2	102	0.00
44	2,4,5-Trichlorophenol	0.367	0.363	1.1	94	0.00
45 S	2-Fluorobiphenyl	1.418	1.293	8.8	96	0.00
46	1,1'-Biphenyl	1.555	1.438	7.5	95	0.00
47	2-Chloronaphthalene	1.157	1.071	7.4	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142274.D
 Acq On : 02 May 2025 10:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 02 11:50:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.249	0.315	-26.5#	112	0.00
49	Acenaphthylene	1.952	1.809	7.3	95	0.00
50	Dimethylphthalate	1.280	1.197	6.5	95	0.00
51	2,6-Dinitrotoluene	0.204	0.246	-20.6	107	0.00
52 C	Acenaphthene	1.140	1.070	6.1	96	0.00
53	3-Nitroaniline	0.257	0.298	-16.0	105	0.00
54 P	2,4-Dinitrophenol	0.058	0.086	-48.3#	158#	0.00
55	Dibenzofuran	1.709	1.581	7.5	95	0.00
56 P	4-Nitrophenol	0.201	0.227	-12.9	110	0.00
57	2,4-Dinitrotoluene	0.239	0.314	-31.4#	116	0.00
58	Fluorene	1.299	1.197	7.9	95	0.00
59	2,3,4,6-Tetrachlorophenol	0.304	0.309	-1.6	97	0.00
60	Diethylphthalate	1.263	1.173	7.1	94	0.00
61	4-Chlorophenyl-phenylether	0.617	0.567	8.1	94	0.00
62	4-Nitroaniline	0.238	0.283	-18.9	109	0.00
63	Azobenzene	1.248	1.151	7.8	94	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	0.054	0.075	-38.9#	144	0.00
66 c	n-Nitrosodiphenylamine	0.683	0.635	7.0	95	0.00
67	4-Bromophenyl-phenylether	0.223	0.213	4.5	95	0.00
68	Hexachlorobenzene	0.249	0.232	6.8	94	0.00
69	Atrazine	0.180	0.175	2.8	97	0.00
70 C	Pentachlorophenol	0.132	0.137	-3.8	101	0.00
71	Phenanthrene	1.080	1.006	6.9	96	0.00
72	Anthracene	1.105	1.034	6.4	97	0.00
73	Carbazole	1.008	0.955	5.3	98	0.00
74	Di-n-butylphthalate	1.013	0.947	6.5	94	0.00
75 C	Fluoranthene	1.068	1.009	5.5	98	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
77	Benzydine	0.850	0.887	-4.4	99	0.00
78	Pyrene	1.845	1.871	-1.4	99	0.00
79 S	Terphenyl-d14	1.375	1.344	2.3	97	0.00
80	Butylbenzylphthalate	0.425	0.449	-5.6	95	0.00
81	Benzo(a)anthracene	1.343	1.293	3.7	96	0.00
82	3,3'-Dichlorobenzidine	0.394	0.373	5.3	90	0.00
83	Chrysene	1.222	1.109	9.2	96	0.00
84	Bis(2-ethylhexyl)phthalate	0.604	0.486	19.5	75	0.00
85 c	Di-n-octyl phthalate	1.133	0.812	28.3#	69	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	100	-0.01
87	Indeno(1,2,3-cd)pyrene	1.465	1.461	0.3	90	-0.01
88	Benzo(b)fluoranthene	1.242	1.073	13.6	82	0.00
89	Benzo(k)fluoranthene	1.136	1.098	3.3	96	0.00
90 C	Benzo(a)pyrene	1.137	1.074	5.5	89	0.00
91	Dibenzo(a,h)anthracene	1.189	1.193	-0.3	91	-0.01
92	Benzo(g,h,i)perylene	1.195	1.216	-1.8	93	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

(#) = Out of Range

SPCC's out = 0 CCC's out = 2

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142274.D
 Acq On : 02 May 2025 10:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 02 11:50:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	37.589	6.0	93	-0.01
3	Pyridine	40.000	38.083	4.8	96	0.00
4	n-Nitrosodimethylamine	40.000	38.811	3.0	95	0.00
5 S	2-Fluorophenol	80.000	76.366	4.5	96	0.00
6	Aniline	40.000	37.880	5.3	95	0.00
7 S	Phenol-d6	80.000	77.087	3.6	96	0.00
8	2-Chlorophenol	40.000	39.206	2.0	97	0.00
9	Benzaldehyde	40.000	37.835	5.4	94	0.00
10 C	Phenol	40.000	37.967	5.1	95	0.00
11	bis(2-Chloroethyl)ether	40.000	37.841	5.4	95	0.00
12	1,3-Dichlorobenzene	40.000	37.811	5.5	95	0.00
13 C	1,4-Dichlorobenzene	40.000	36.999	7.5	93	0.00
14	1,2-Dichlorobenzene	40.000	37.888	5.3	95	0.00
15	Benzyl Alcohol	40.000	38.697	3.3	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.895	5.3	95	0.00
17	2-Methylphenol	40.000	37.931	5.2	94	0.00
18	Hexachloroethane	40.000	39.190	2.0	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.373	6.6	95	0.00
20	3+4-Methylphenols	40.000	39.062	2.3	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	36.381	9.0	95	0.00
23 S	Nitrobenzene-d5	80.000	87.247	-9.1	103	0.00
24	Nitrobenzene	40.000	42.143	-5.4	102	0.00
25	Isophorone	40.000	36.800	8.0	94	0.00
26 C	2-Nitrophenol	40.000	43.995	-10.0	124	0.00
27	2,4-Dimethylphenol	40.000	38.087	4.8	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.730	8.2	94	0.00
29 C	2,4-Dichlorophenol	40.000	38.714	3.2	95	0.00
30	1,2,4-Trichlorobenzene	40.000	36.954	7.6	94	0.00
31	Naphthalene	40.000	36.602	8.5	95	0.00
32	Benzoic acid	40.000	42.364	-5.9	127	0.00
33	4-Chloroaniline	40.000	37.373	6.6	96	0.00
34 C	Hexachlorobutadiene	40.000	37.512	6.2	94	0.00
35	Caprolactam	40.000	39.988	0.0	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	38.463	3.8	95	0.00
37	2-Methylnaphthalene	40.000	36.551	8.6	95	0.00
38	1-Methylnaphthalene	40.000	36.702	8.2	95	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.065	7.3	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.985	-2.5	98	0.00
42 S	2,4,6-Tribromophenol	80.000	83.329	-4.2	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.332	-3.3	102	0.00
44	2,4,5-Trichlorophenol	40.000	39.611	1.0	94	0.00
45 S	2-Fluorobiphenyl	80.000	72.935	8.8	96	0.00
46	1,1'-Biphenyl	40.000	36.988	7.5	95	0.00
47	2-Chloronaphthalene	40.000	37.021	7.4	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142274.D
 Acq On : 02 May 2025 10:39
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 02 11:50:23 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.014	-5.0	112	0.00
49	Acenaphthylene	40.000	37.062	7.3	95	0.00
50	Dimethylphthalate	40.000	37.413	6.5	95	0.00
51	2,6-Dinitrotoluene	40.000	40.686	-1.7	107	0.00
52 C	Acenaphthene	40.000	37.558	6.1	96	0.00
53	3-Nitroaniline	40.000	40.499	-1.2	105	0.00
54 P	2,4-Dinitrophenol	40.000	50.599	-26.5#	158	0.00
55	Dibenzofuran	40.000	37.015	7.5	95	0.00
56 P	4-Nitrophenol	40.000	45.101	-12.8	110	0.00
57	2,4-Dinitrotoluene	40.000	42.928	-7.3	116	0.00
58	Fluorene	40.000	36.852	7.9	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.702	-1.8	97	0.00
60	Diethylphthalate	40.000	37.148	7.1	94	0.00
61	4-Chlorophenyl-phenylether	40.000	36.772	8.1	94	0.00
62	4-Nitroaniline	40.000	42.072	-5.2	109	0.00
63	Azobenzene	40.000	36.907	7.7	94	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.858	-14.6	144	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.187	7.0	95	0.00
67	4-Bromophenyl-phenylether	40.000	38.179	4.6	95	0.00
68	Hexachlorobenzene	40.000	37.175	7.1	94	0.00
69	Atrazine	40.000	38.907	2.7	97	0.00
70 C	Pentachlorophenol	40.000	41.560	-3.9	101	0.00
71	Phenanthrene	40.000	37.261	6.8	96	0.00
72	Anthracene	40.000	37.446	6.4	97	0.00
73	Carbazole	40.000	37.893	5.3	98	0.00
74	Di-n-butylphthalate	40.000	37.366	6.6	94	0.00
75 C	Fluoranthene	40.000	37.775	5.6	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	41.743	-4.4	99	0.00
78	Pyrene	40.000	40.557	-1.4	99	0.00
79 S	Terphenyl-d14	80.000	78.207	2.2	97	0.00
80	Butylbenzylphthalate	40.000	36.531	8.7	95	0.00
81	Benzo(a)anthracene	40.000	38.532	3.7	96	0.00
82	3,3'-Dichlorobenzidine	40.000	37.840	5.4	90	0.00
83	Chrysene	40.000	36.304	9.2	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	29.284	26.8#	75	0.00
85 c	Di-n-octyl phthalate	40.000	28.225	29.4#	69	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	100	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	39.896	0.3	90	-0.01
88	Benzo(b)fluoranthene	40.000	34.545	13.6	82	0.00
89	Benzo(k)fluoranthene	40.000	38.642	3.4	96	0.00
90 C	Benzo(a)pyrene	40.000	37.783	5.5	89	0.00
91	Dibenzo(a,h)anthracene	40.000	40.144	-0.4	91	-0.01
92	Benzo(g,h,i)perylene	40.000	40.703	-1.8	93	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142274.D
Acq On : 02 May 2025 10:39
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 02 11:50:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range SPCC's out = 0 CCC's out = 1





QC SAMPLE DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142238.D
Acq On : 30 Apr 2025 10:55
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Apr 30 16:00:01 2025






AutoFind: Scans 1614, 1615, 1616; Background Corrected with Scan 1607

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.6	86007	PASS
68	69	0.00	2	1.8	1316	PASS
69	198	0.00	100	33.1	71903	PASS
70	69	0.00	2	0.5	377	PASS
127	198	10	80	45.1	97928	PASS
197	198	0.00	2	0.8	1760	PASS
198	198	100	100	100.0	217131	PASS
199	198	5	9	6.5	14095	PASS
275	198	10	60	27.6	60019	PASS
365	198	1	100	3.7	7976	PASS
441	198	0.01	100	19.0	41171	PASS
442	442	50	100	100.0	263808	PASS
443	442	15	24	19.9	52467	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142238.D
Acq On : 30 Apr 2025 10:55
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Apr 30 16:04:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration



TIC: BF142238.D\data.ms

(70) Pentachlorophenol (C)

11.233min (+ 0.000) 69575.24 ng




response 180630

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.50	63.39
264.00	61.60	62.57
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043025\
Data File : BF142238.D
Acq On : 30 Apr 2025 10:55
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Apr 30 16:04:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration



TIC: BF142238.D\data.ms

(77) Benzidine

12.774min (+ 0.006) 97416.22 ng

response 1440496

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.25
183.00	11.10	11.10
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/30/2025	BNA_F	<u>BF142238.D</u>
Compound Name	Response	Retention Time
DDT	378546	13.598
DDD	2841	13.333
DDE	0	12.904
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
2841	381387	0.74



Instrument :
BNA_F
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142249.D
Acq On : 01 May 2025 09:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Apr 30 16:00:01 2025



AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1607

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.2	103579	PASS
68	69	0.00	2	1.6	1339	PASS
69	198	0.00	100	35.5	83059	PASS
70	69	0.00	2	0.5	413	PASS
127	198	10	80	47.7	111579	PASS
197	198	0.00	2	0.7	1704	PASS
198	198	100	100	100.0	234155	PASS
199	198	5	9	6.7	15632	PASS
275	198	10	60	26.7	62555	PASS
365	198	1	100	3.4	8074	PASS
441	198	0.01	100	17.1	40008	PASS
442	442	50	100	100.0	263019	PASS
443	442	15	24	19.1	50208	PASS

DDT Breakdown




Date	Instrument Name	DFTPP Data File
5/1/2025	BNA_F	<u>BF142249.D</u>
Compound Name	Response	Retention Time
DDT	427368	13.598
DDD	6172	13.327
DDE	0	12.904
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
6172	433540	1.42

Instrument :
BNA_F
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142249.D
Acq On : 01 May 2025 09:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 01 12:00:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration



TIC: BF142249.D\data.ms

(70) Pentachlorophenol (C)

11.233min (+ 0.000) 113470.87 ng




response 170474

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.50	65.06
264.00	61.60	64.75
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142249.D
Acq On : 01 May 2025 09:48
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 01 12:00:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration



TIC: BF142249.D\data.ms

(77) Benzidine

12.774min (+ 0.006) 156734.40 ng

response 1611689



Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.68
183.00	11.10	11.08
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142273.D
Acq On : 02 May 2025 10:11
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Apr 30 16:00:01 2025



AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1606

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	42.1	103072	PASS
68	69	0.00	2	1.6	1298	PASS
69	198	0.00	100	33.5	82011	PASS
70	69	0.00	2	0.6	473	PASS
127	198	10	80	45.6	111691	PASS
197	198	0.00	2	0.8	2015	PASS
198	198	100	100	100.0	244928	PASS
199	198	5	9	6.8	16632	PASS
275	198	10	60	28.7	70309	PASS
365	198	1	100	3.7	9025	PASS
441	198	0.01	100	18.7	45787	PASS
442	442	50	100	100.0	296981	PASS
443	442	15	24	19.3	57432	PASS

DDT Breakdown




Date	Instrument Name	DFTPP Data File
5/2/2025	BNA_F	<u>BF142273.D</u>
Compound Name	Response	Retention Time
DDT	646019	13.592
DDD	7521	13.322
DDE	403	12.951
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
7924	653943	1.21

Instrument :
BNA_F
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142273.D
Acq On : 02 May 2025 10:11
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 02 11:49:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration



TIC: BF142273.D\data.ms

(70) Pentachlorophenol (C)

11.228min (-0.005) 200113.05 ng




response 342837

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.50	63.76
264.00	61.60	64.00
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142273.D
Acq On : 02 May 2025 10:11
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: May 02 11:49:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration



TIC: BF142273.D\data.ms

(77) Benzidine

12.769min (-0.000) 120826.54 ng

response 1673715

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.73
183.00	11.10	11.31
0.00	0.00	0.00

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BL	SDG No.:	Q1915
Lab Sample ID:	PB167810BL	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142254.D	1	04/30/25 13:15	05/01/25 12:10	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	5.00	U	1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	5.00	U	0.53	5.00	ug/L
95-48-7	2-Methylphenol	5.00	U	1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	10.0	U	1.10	10.0	ug/L
67-72-1	Hexachloroethane	5.00	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	5.00	U	0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	5.00	U	0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	5.00	U	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	5.00	U	0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	5.00	U	1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	5.00	U	0.52	5.00	ug/L
87-86-5	Pentachlorophenol	10.0	U	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	110		10 - 139	73%	SPK: 150
13127-88-3	Phenol-d6	109		10 - 134	73%	SPK: 150
4165-60-0	Nitrobenzene-d5	83.3		49 - 133	83%	SPK: 100
321-60-8	2-Fluorobiphenyl	69.1		52 - 132	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	124		44 - 137	83%	SPK: 150
1718-51-0	Terphenyl-d14	63.3		48 - 125	63%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	224000	6.904			
1146-65-2	Naphthalene-d8	858000	8.187			
15067-26-2	Acenaphthene-d10	450000	9.945			
1517-22-2	Phenanthrene-d10	795000	11.428			
1719-03-5	Chrysene-d12	568000	14.069			
1520-96-3	Perylene-d12	405000	15.563			

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BL	SDG No.:	Q1915
Lab Sample ID:	PB167810BL	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142254.D	1	04/30/25 13:15	05/01/25 12:10	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142254.D
Acq On : 01 May 2025 12:10
Operator : RC/JU
Sample : PB167810BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB167810BL

Quant Time: May 01 12:47:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.904	152	223997	20.000	ng	0.00
21) Naphthalene-d8	8.187	136	857825	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	449926	20.000	ng	0.00
64) Phenanthrene-d10	11.428	188	794711	20.000	ng	0.00
76) Chrysene-d12	14.069	240	568207	20.000	ng	0.00
86) Perylene-d12	15.563	264	404816	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	1546121	110.236	ng	0.01
7) Phenol-d6	6.522	99	1853180	109.417	ng	0.00
23) Nitrobenzene-d5	7.469	82	1087156	83.261	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	508946	124.082	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2205083	69.102	ng	0.00
79) Terphenyl-d14	13.022	244	2472921	63.298	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142254.D
Acq On : 01 May 2025 12:10
Operator : RC/JU
Sample : PB167810BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB167810BL


Quant Time: May 01 12:47:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.904 min Scan# 801
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Instrument :
BNA_F
ClientSampleId :
PB167810BL







Tgt Ion:152 Resp: 223997
Ion Ratio Lower Upper
152 100
150 156.1 130.2 195.2
115 59.2 47.0 70.4



#5
2-Fluorophenol
Concen: 110.236 ng
RT: 5.528 min Scan# 567
Delta R.T. 0.012 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Tgt Ion:112 Resp: 1546121
Ion Ratio Lower Upper
112 100
64 60.1 48.4 72.6
63 31.1 24.8 37.2









#7
Phenol-d6
Concen: 109.417 ng
RT: 6.522 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10




Instrument :
BNA_F
ClientSampleId :
PB167810BL


Tgt Ion: 99 Resp: 1853180
Ion Ratio Lower Upper
99 100
42 21.1 16.3 24.5
71 32.5 25.0 37.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.187 min Scan# 1019
Delta R.T. -0.006 min
Lab File: BF142254.D
Acq: 01 May 2025 12:10

Tgt Ion: 136 Resp: 857825
Ion Ratio Lower Upper
136 100
137 11.0 8.7 13.1
54 9.6 7.7 11.5
68 6.4 5.0 7.6





#23

Nitrobenzene-d5

Concen: 83.261 ng

RT: 7.469 min Scan# 89

Delta R.T. -0.006 min

Lab File: BF142254.D

Acq: 01 May 2025 12:10

Instrument :





BNA_F

ClientSampleId :

PB167810BL

Tgt Ion: 82 Resp: 1087156

Ion	Ratio	Lower	Upper
82	100		
128	46.4	37.0	55.6
54	59.8	48.0	72.0



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.945 min Scan# 1318




Delta R.T. -0.006 min


Lab File: BF142254.D

Acq: 01 May 2025 12:10

Tgt Ion:164 Resp: 449926

Ion	Ratio	Lower	Upper
164	100		
162	100.6	78.6	117.8
160	45.1	35.7	53.5





#42

2,4,6-Tribromophenol

Concen: 124.082 ng

RT: 10.733 min Scan# 1452

Delta R.T. -0.000 min

Lab File: BF142254.D


Acq: 01 May 2025 12:10

Instrument :

BNA_F

ClientSampleId :

PB167810BL





Tgt Ion:330 Resp: 508946

Ion Ratio Lower Upper

330 100

332 96.7 76.5 114.7

141 29.6 24.3 36.5



#45

2-Fluorobiphenyl

Concen: 69.102 ng

RT: 9.263 min Scan# 1202

Delta R.T. -0.006 min

Lab File: BF142254.D

Acq: 01 May 2025 12:10




Tgt Ion:172 Resp: 2205083


Ion Ratio Lower Upper

172 100

171 36.1 29.0 43.4

170 23.9 19.4 29.0





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.428 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF142254.D


Acq: 01 May 2025 12:10

Instrument :

BNA_F

ClientSampleId :

PB167810BL






Tgt Ion:188 Resp: 794711

Ion Ratio Lower Upper

188 100

94 10.8 8.2 12.2

80 11.1 8.6 12.8



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.069 min Scan# 2019

Delta R.T. -0.006 min

Lab File: BF142254.D

Acq: 01 May 2025 12:10




Tgt Ion:240 Resp: 568207


Ion Ratio Lower Upper

240 100

120 12.6 10.1 15.1

236 25.7 20.5 30.7









#79
 Terphenyl-d14
 Concen: 63.298 ng
 RT: 13.022 min Scan# 1841
 Delta R.T. -0.000 min
 Lab File: BF142254.D
 Acq: 01 May 2025 12:10




Instrument :
 BNA_F
 ClientSampleId :
 PB167810BL

Tgt Ion:244 Resp: 2472921
 Ion Ratio Lower Upper
 244 100
 212 6.7 5.3 7.9
 122 10.9 9.1 13.7



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.563 min Scan# 2273
 Delta R.T. -0.006 min
 Lab File: BF142254.D
 Acq: 01 May 2025 12:10

Tgt Ion:264 Resp: 404816
 Ion Ratio Lower Upper
 264 100
 260 22.8 18.4 27.6
 265 21.1 17.3 25.9



Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BS	SDG No.:	Q1915
Lab Sample ID:	PB167810BS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142255.D	1	04/30/25 13:15	05/01/25 12:39	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	37.2		1.30	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	39.6		0.53	5.00	ug/L
95-48-7	2-Methylphenol	41.0		1.10	5.00	ug/L
65794-96-9	3+4-Methylphenols	40.8		1.10	10.0	ug/L
67-72-1	Hexachloroethane	41.8		0.65	5.00	ug/L
98-95-3	Nitrobenzene	45.8		0.76	5.00	ug/L
87-68-3	Hexachlorobutadiene	41.5		0.54	5.00	ug/L
88-06-2	2,4,6-Trichlorophenol	43.4		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	44.9		0.62	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	49.8		1.20	5.00	ug/L
118-74-1	Hexachlorobenzene	41.6		0.52	5.00	ug/L
87-86-5	Pentachlorophenol	93.4	E	1.60	10.0	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	111		10 - 139	74%	SPK: 150
13127-88-3	Phenol-d6	113		10 - 134	75%	SPK: 150
4165-60-0	Nitrobenzene-d5	82.4		49 - 133	82%	SPK: 100
321-60-8	2-Fluorobiphenyl	68.5		52 - 132	69%	SPK: 100
118-79-6	2,4,6-Tribromophenol	132		44 - 137	88%	SPK: 150
1718-51-0	Terphenyl-d14	74.1		48 - 125	74%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	229000		6.91		
1146-65-2	Naphthalene-d8	893000		8.192		
15067-26-2	Acenaphthene-d10	472000		9.945		
1517-22-2	Phenanthrene-d10	805000		11.433		
1719-03-5	Chrysene-d12	464000		14.074		
1520-96-3	Perylene-d12	447000		15.568		

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB167810BS	SDG No.:	Q1915
Lab Sample ID:	PB167810BS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142255.D	1	04/30/25 13:15	05/01/25 12:39	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
 Data File : BF142255.D
 Acq On : 01 May 2025 12:39
 Operator : RC/JU
 Sample : PB167810BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB167810BS

Manual Integrations
 APPROVED

Reviewed By :Rahul
 Chavli

Quant Time: May 01 13:05:39 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

05/02/2025
 Supervised By :Jagrut
 Upadhyay

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards

1) 1,4-Dichlorobenzene-d4	6.910	152	228833	20.000	ng	0.00
21) Naphthalene-d8	8.192	136	892782	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	472284	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	804874	20.000	ng	0.00
76) Chrysene-d12	14.074	240	464382	20.000	ng	0.00
86) Perylene-d12	15.568	264	446826	20.000	ng	0.00

05/02/2025
 Supervised By :Jagrut
 Upadhyay

System Monitoring Compounds

5) 2-Fluorophenol	5.534	112	1593182	111.191	ng	0.02
7) Phenol-d6	6.528	99	1946714	112.510	ng	0.00
23) Nitrobenzene-d5	7.469	82	1119930	82.412	ng	0.00
42) 2,4,6-Tribromophenol	10.739	330	568456	132.029	ng	0.00
45) 2-Fluorobiphenyl	9.269	172	2294519	68.501	ng	0.00
79) Terphenyl-d14	13.021	244	2366082	74.104	ng	0.00

Target Compounds

Compound	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.798	88	217494	33.987	ng	98
3) Pyridine	3.551	79	618200	37.171	ng	100
4) n-Nitrosodimethylamine	3.504	42	358234	41.623	ng	99
6) Aniline	6.569	93	471322	19.245	ng	100
8) 2-Chlorophenol	6.692	128	631635	42.661	ng	99
9) Benzaldehyde	6.457	77	305773	25.285	ng	98
10) Phenol	6.545	94	755191m	40.542	ng	
11) bis(2-Chloroethyl)ether	6.639	93	588988	40.804	ng	99
12) 1,3-Dichlorobenzene	6.851	146	659968	39.402	ng	98
13) 1,4-Dichlorobenzene	6.928	146	667827	39.574	ng	99
14) 1,2-Dichlorobenzene	7.075	146	650967	40.407	ng	99
15) Benzyl Alcohol	7.045	79	515024	41.969	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.181	45	1072194	41.135	ng	100
17) 2-Methylphenol	7.151	107	497856	41.046	ng	99
18) Hexachloroethane	7.422	117	231050	41.832	ng	97
19) n-Nitroso-di-n-propyla...	7.322	70	427592	39.992	ng	99
20) 3+4-Methylphenols	7.304	107	618253	40.760	ng	92
22) Acetophenone	7.316	105	826671	40.025	ng	99
24) Nitrobenzene	7.492	77	610784	45.807	ng	99
25) Isophorone	7.728	82	1177064	41.271	ng	99
26) 2-Nitrophenol	7.804	139	256947	46.237	ng	99
27) 2,4-Dimethylphenol	7.839	122	602302	41.886	ng	100
28) bis(2-Chloroethoxy)met...	7.939	93	734241	40.821	ng	99
29) 2,4-Dichlorophenol	8.039	162	518664	43.143	ng	99
30) 1,2,4-Trichlorobenzene	8.128	180	544203	40.565	ng	99
31) Naphthalene	8.216	128	1769061	39.733	ng	100
32) Benzoic acid	7.945	122	321202m	47.181	ng	
33) 4-Chloroaniline	8.251	127	82158	4.431	ng	99
34) Hexachlorobutadiene	8.328	225	324756	41.492	ng	99
35) Caprolactam	8.622	113	163644m	45.287	ng	
36) 4-Chloro-3-methylphenol	8.733	107	523494	42.406	ng	99
37) 2-Methylnaphthalene	8.904	142	1090096	39.822	ng	99
38) 1-Methylnaphthalene	9.004	142	1143793	40.333	ng	100
40) 1,2,4,5-Tetrachloroben...	9.069	216	524634	39.471	ng	99
41) Hexachlorocyclopentadiene	9.057	237	668416	88.607	ng	100
43) 2,4,6-Trichlorophenol	9.175	196	357651	43.423	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
 Data File : BF142255.D
 Acq On : 01 May 2025 12:39
 Operator : RC/JU
 Sample : PB167810BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Manual Integrations

APPROVED

Reviewed By :Rahul
 Chavli

05/02/2025

Supervised By :Jagrut
 Upadhyay

05/02/2025

Supervised By :Jagrut
 Upadhyay

Quant Time: May 01 13:05:39 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.216	196	388570	44.872	ng	98
46) 1,1'-Biphenyl	9.369	154	1458088	39.713	ng	100
47) 2-Chloronaphthalene	9.392	162	1093139	40.017	ng	100
48) 2-Nitroaniline	9.486	65	321433	44.957	ng	99
49) Acenaphthylene	9.810	152	1847757	40.085	ng	100
50) Dimethylphthalate	9.669	163	1268421	41.981	ng	100
51) 2,6-Dinitrotoluene	9.727	165	263569	45.692	ng	98
52) Acenaphthene	9.980	154	1164359	43.261	ng	100
53) 3-Nitroaniline	9.892	138	98911	16.287	ng	95
54) 2,4-Dinitrophenol	9.998	184	196654	90.602	ng	# 18
55) Dibenzofuran	10.157	168	1606504	39.818	ng	100
56) 4-Nitrophenol	10.045	139	484293	101.975	ng	98
57) 2,4-Dinitrotoluene	10.133	165	350987	49.818	ng	98
58) Fluorene	10.498	166	1221943	39.832	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.269	232	330098	46.019	ng	99
60) Diethylphthalate	10.369	149	1267747	42.490	ng	100
61) 4-Chlorophenyl-phenyle...	10.486	204	585254	40.172	ng	99
62) 4-Nitroaniline	10.510	138	286425	44.888	ng	99
63) Azobenzene	10.651	77	1208231	41.009	ng	99
65) 4,6-Dinitro-2-methylph...	10.539	198	126239	47.699	ng	95
66) n-Nitrosodiphenylamine	10.610	169	1104532	40.164	ng	99
67) 4-Bromophenyl-phenylether	10.980	248	373384	41.655	ng	98
68) Hexachlorobenzene	11.045	284	417991	41.636	ng	100
69) Atrazine	11.133	200	332195	45.827	ng	98
70) Pentachlorophenol	11.233	266	495448	93.418	ng	98
71) Phenanthrene	11.463	178	1740817	40.041	ng	100
72) Anthracene	11.510	178	1786049	40.180	ng	100
73) Carbazole	11.663	167	1677441	41.368	ng	100
74) Di-n-butylphthalate	11.998	149	1839790	45.113	ng	100
75) Fluoranthene	12.645	202	1819275	42.331	ng	99
77) Benzidine	12.763	184	347980	17.635	ng	98
78) Pyrene	12.880	202	1803569	42.103	ng	99
80) Butylbenzylphthalate	13.498	149	532587	45.667	ng	99
81) Benzo(a)anthracene	14.062	228	1330282	42.672	ng	99
82) 3,3'-Dichlorobenzidine	14.027	252	74664	8.154	ng	98
83) Chrysene	14.104	228	1152993	40.631	ng	100
84) Bis(2-ethylhexyl)phtha...	14.057	149	632694	39.646	ng	99
85) Di-n-octyl phthalate	14.674	149	1144596	39.690	ng	100
87) Indeno(1,2,3-cd)pyrene	17.080	276	1374940	42.010	ng	100
88) Benzo(b)fluoranthene	15.127	252	1207610	43.510	ng	99
89) Benzo(k)fluoranthene	15.156	252	1003805	39.540	ng	99
90) Benzo(a)pyrene	15.504	252	1073003	42.227	ng	100
91) Dibenzo(a,h)anthracene	17.098	278	1112811	41.894	ng	99
92) Benzo(g,h,i)perylene	17.539	276	1122234	42.026	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050125\
Data File : BF142255.D
Acq On : 01 May 2025 12:39
Operator : RC/JU
Sample : PB167810BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB167810BS


Manual Integrations


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 80
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion: 152 Resp: 22883
Ion Ratio Lower Upper
152 100
150 162.7 130.2 195.2
115 58.2 47.0 70.4

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#2
1,4-Dioxane
Concen: 33.987 ng
RT: 2.798 min Scan# 103
Delta R.T. 0.100 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion: 88 Resp: 217494
Ion Ratio Lower Upper
88 100
58 92.0 72.2 108.2
43 35.3 27.0 40.6





#3
Pyridine
Concen: 37.171 ng
RT: 3.551 min Scan# 216
Delta R.T. 0.088 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion: 79 Resp: 618200
Ion Ratio Lower Upper
79 100
52 88.9 71.5 107.3
51 42.9 34.5 51.7

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#4
n-Nitrosodimethylamine
Concen: 41.623 ng
RT: 3.504 min Scan# 223
Delta R.T. 0.094 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion: 42 Resp: 358234
Ion Ratio Lower Upper
42 100
74 111.4 89.5 134.3
44 7.3 6.6 9.8





#5
2-Fluorophenol
Concen: 111.191 ng
RT: 5.534 min Scan# 50
Delta R.T. 0.017 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:112 Resp: 159318
Ion Ratio Lower Upper
112 100
64 59.3 48.4 72.6
63 30.6 24.8 37.2

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#6
Aniline
Concen: 19.245 ng
RT: 6.569 min Scan# 744
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion: 93 Resp: 471322
Ion Ratio Lower Upper
93 100
66 39.9 32.1 48.1
65 20.4 16.6 24.8





#7
Phenol-d6
Concen: 112.510 ng
RT: 6.528 min Scan# 737
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion: 99 Resp: 1946714
Ion Ratio Lower Upper
99 100
42 21.0 16.3 24.5
71 32.4 25.0 37.4

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#8
2-Chlorophenol
Concen: 42.661 ng
RT: 6.692 min Scan# 765
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:128 Resp: 631635
Ion Ratio Lower Upper
128 100
130 32.9 13.1 53.1
64 47.6 28.6 68.6





#9

Benzaldehyde

Concen: 25.285 ng

RT: 6.457 min Scan# 71

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion: 77 Resp: 30577

Ion Ratio Lower Upper

77 100

105 101.1 79.1 119.1

106 97.8 76.0 116.0

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#10

Phenol

Concen: 40.542 ng m

RT: 6.545 min Scan# 740

Delta R.T. 0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion: 94 Resp: 755191


Ion Ratio Lower Upper

94 100

65 29.0 8.9 48.9


66 39.0 18.7 58.7





#11
bis(2-Chloroethyl)ether
Concen: 40.804 ng
RT: 6.639 min Scan# 757
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion: 93 Resp: 588988
Ion Ratio Lower Upper
93 100
63 83.8 62.2 102.2
95 33.0 12.7 52.7

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#12
1,3-Dichlorobenzene
Concen: 39.402 ng
RT: 6.851 min Scan# 792
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:146 Resp: 659968
Ion Ratio Lower Upper
146 100
148 64.6 50.5 75.7
75 29.9 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 39.574 ng

RT: 6.928 min Scan# 805

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:146 Resp: 66782

Ion Ratio Lower Upper

146 100

148 64.2 50.6 76.0

111 41.9 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#14

1,2-Dichlorobenzene

Concen: 40.407 ng

RT: 7.075 min Scan# 830

Delta R.T. -0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:146 Resp: 650967


Ion Ratio Lower Upper

146 100

148 62.9 50.5 75.7

111 44.3 34.9 52.3





#15

Benzyl Alcohol

Concen: 41.969 ng

RT: 7.045 min Scan# 81

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion: 79 Resp: 515024

Ion Ratio Lower Upper

79 100

108 81.8 64.5 96.7

77 62.7 49.2 73.8

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 41.135 ng

RT: 7.181 min Scan# 848

Delta R.T. -0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion: 45 Resp: 1072194


Ion Ratio Lower Upper

45 100

77 11.1 0.0 31.3


79 8.4 0.0 28.5





#17
2-Methylphenol
Concen: 41.046 ng
RT: 7.151 min Scan# 843
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:107 Resp: 497850
Ion Ratio Lower Upper
107 100
108 113.7 91.0 136.4
77 43.7 34.4 51.6
79 43.8 34.4 51.6

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#18
Hexachloroethane
Concen: 41.832 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:117 Resp: 231050
Ion Ratio Lower Upper
117 100
119 96.8 77.1 115.7
201 93.2 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 39.992 ng

RT: 7.322 min Scan# 81

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion: 70 Resp: 427592

Ion Ratio Lower Upper

70 100

42 67.8 53.4 80.2

101 9.9 7.7 11.5

130 22.1 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#20

3+4-Methylphenols

Concen: 40.760 ng

RT: 7.304 min Scan# 869

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Tgt Ion:107 Resp: 618253




Ion Ratio Lower Upper


107 100

108 91.9 71.2 111.2

77 77.5 41.1 81.1

79 28.2 8.1 48.1






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.192 min Scan# 1020
Delta R.T. -0.001 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS



Tgt Ion:136 Resp: 892782
Ion Ratio Lower Upper
136 100
137 11.0 8.7 13.1
54 9.4 7.7 11.5
68 6.1 5.0 7.6

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli




05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025




Supervised By :Jagrut
Upadhyay


05/02/2025



#22
Acetophenone
Concen: 40.025 ng
RT: 7.316 min Scan# 871
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:105 Resp: 826671
Ion Ratio Lower Upper
105 100
71 7.5 5.7 8.5
51 37.2 30.1 45.1
120 23.5 19.0 28.4





#23

Nitrobenzene-d5

Concen: 82.412 ng

RT: 7.469 min Scan# 898

Delta R.T. -0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion: 82 Resp: 1119930

Ion Ratio Lower Upper

82 100

128 45.8 37.0 55.6

54 60.7 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#24

Nitrobenzene

Concen: 45.807 ng

RT: 7.492 min Scan# 901

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion: 77 Resp: 610784


Ion Ratio Lower Upper

77 100

123 49.2 38.6 58.0


65 14.4 11.5 17.3





#25
Isophorone
Concen: 41.271 ng
RT: 7.728 min Scan# 941
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion: 82 Resp: 1177064
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 19.4 15.3 22.9

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#26
2-Nitrophenol
Concen: 46.237 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:139 Resp: 256947
Ion Ratio Lower Upper
139 100
109 27.6 21.8 32.8
65 50.4 41.1 61.7






#27
2,4-Dimethylphenol
Concen: 41.886 ng
RT: 7.839 min Scan# 959
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS



Tgt Ion:122 Resp: 602302
Ion Ratio Lower Upper
122 100
107 107.6 86.0 129.0
121 57.3 46.3 69.5

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli




05/02/2025

Supervised By :Jagrut
Upadhyay


05/02/2025

Supervised By :Jagrut
Upadhyay



05/02/2025




#28
bis(2-Chloroethoxy)methane
Concen: 40.821 ng
RT: 7.939 min Scan# 977
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39



Tgt Ion: 93 Resp: 734241
Ion Ratio Lower Upper
93 100
95 32.8 26.2 39.2
123 12.9 9.7 14.5





#29

2,4-Dichlorophenol

Concen: 43.143 ng

RT: 8.039 min Scan# 994

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:162 Resp: 51864

Ion Ratio Lower Upper

162 100

164 63.7 43.0 83.0

98 37.4 16.7 56.7

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#30

1,2,4-Trichlorobenzene

Concen: 40.565 ng

RT: 8.128 min Scan# 1009

Delta R.T. -0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:180 Resp: 544203


Ion Ratio Lower Upper

180 100

182 94.4 75.7 113.5


145 32.5 25.4 38.0





#31
Naphthalene
Concen: 39.733 ng
RT: 8.216 min Scan# 1024
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:128 Resp: 176906
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.4
127 13.2 10.6 15.8

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#32
Benzoic acid
Concen: 47.181 ng m
RT: 7.945 min Scan# 978
Delta R.T. 0.011 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:122 Resp: 321202
Ion Ratio Lower Upper
122 100
105 121.8 98.2 138.2
77 90.5 66.8 106.8





#33

4-Chloroaniline

Concen: 4.431 ng

RT: 8.251 min Scan# 1030

Delta R.T. -0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:127 Resp: 82158

Ion Ratio Lower Upper

127 100

129 32.0 26.1 39.1

65 32.6 25.5 38.3

92 18.3 15.4 23.2

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#34

Hexachlorobutadiene

Concen: 41.492 ng

RT: 8.328 min Scan# 1043

Delta R.T. -0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:225 Resp: 324756


Ion Ratio Lower Upper

225 100

223 62.7 49.7 74.5

227 63.2 51.0 76.4





#35

Caprolactam

Concen: 45.287 ng m

RT: 8.622 min Scan# 1093

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:113 Resp: 163644

Ion Ratio Lower Upper

113 100

55 204.8 177.7 217.7

56 148.3 127.1 167.1

Manual Integrations





APPROVED

Reviewed By :Rahul
Chavli

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

#36

4-Chloro-3-methylphenol

Concen: 42.406 ng

RT: 8.733 min Scan# 1112

Delta R.T. 0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:107 Resp: 523494


Ion Ratio Lower Upper

107 100

144 26.8 21.0 31.4


142 83.5 66.1 99.1





#37
2-Methylnaphthalene
Concen: 39.822 ng
RT: 8.904 min Scan# 1141
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:142 Resp: 1090090
Ion Ratio Lower Upper
142 100
141 88.9 70.5 105.7

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#38
1-Methylnaphthalene
Concen: 40.333 ng
RT: 9.004 min Scan# 1158
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:142 Resp: 1143793
Ion Ratio Lower Upper
142 100
141 91.4 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.945 min Scan# 1169

Delta R.T. -0.006 min

Lab File: BF142255.D


Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS



Tgt Ion:164 Resp: 472284

Ion Ratio Lower Upper

164 100

162 100.2 78.6 117.8

160 44.6 35.7 53.5

Manual Integrations



APPROVED

Reviewed By :Rahul
Chavli


05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#40

1,2,4,5-Tetrachlorobenzene

Concen: 39.471 ng

RT: 9.069 min Scan# 1169

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Tgt Ion:216 Resp: 524634




Ion Ratio Lower Upper


216 100

214 79.4 63.8 95.6

179 20.8 16.7 25.1

108 20.9 15.6 23.4





#41

Hexachlorocyclopentadiene

Concen: 88.607 ng

RT: 9.057 min Scan# 1167

Delta R.T. -0.000 min

Lab File: BF142255.D


Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS



Tgt Ion:237 Resp: 668410

Ion Ratio Lower Upper

237 100

235 63.6 43.5 83.5

272 13.1 0.0 32.9

Manual Integrations




APPROVED

Reviewed By :Rahul
Chavli

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

#42

2,4,6-Tribromophenol


Concen: 132.029 ng

RT: 10.739 min Scan# 1453

Delta R.T. 0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39





Tgt Ion:330 Resp: 568456


Ion Ratio Lower Upper

330 100

332 96.0 76.5 114.7

141 29.1 24.3 36.5





#43

2,4,6-Trichlorophenol

Concen: 43.423 ng

RT: 9.175 min Scan# 1187

Delta R.T. -0.000 min

Lab File: BF142255.D


Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS



Tgt Ion:196 Resp: 35765

Ion Ratio Lower Upper

196 100

198 95.6 77.4 116.2

200 30.7 24.6 37.0

Manual Integrations



APPROVED

Reviewed By :Rahul
Chavli


05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#44

2,4,5-Trichlorophenol


Concen: 44.872 ng

RT: 9.216 min Scan# 1194

Delta R.T. 0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39



Tgt Ion:196 Resp: 388570



Ion Ratio Lower Upper


196 100

198 96.0 75.2 112.8

97 48.5 39.8 59.8

132 28.8 23.8 35.8





#45

2-Fluorobiphenyl

Concen: 68.501 ng

RT: 9.269 min Scan# 1203

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:172 Resp: 2294519

Ion Ratio Lower Upper

172 100

171 35.8 29.0 43.4

170 24.0 19.4 29.0

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#46

1,1'-Biphenyl

Concen: 39.713 ng

RT: 9.369 min Scan# 1220

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:154 Resp: 1458088


Ion Ratio Lower Upper

154 100

153 40.6 20.5 60.5

76 12.8 0.0 32.8





#47

2-Chloronaphthalene

Concen: 40.017 ng

RT: 9.392 min Scan# 1224

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:162 Resp: 1093139

Ion Ratio Lower Upper

162 100

127 40.9 32.6 49.0

164 32.7 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#48

2-Nitroaniline

Concen: 44.957 ng

RT: 9.486 min Scan# 1240

Delta R.T. -0.001 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion: 65 Resp: 321433


Ion Ratio Lower Upper

65 100

92 67.2 54.3 81.5

138 109.3 88.6 133.0





#49

Acenaphthylene

Concen: 40.085 ng

RT: 9.810 min Scan# 1295

Delta R.T. -0.000 min

Lab File: BF142255.D


Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS



Tgt Ion:152 Resp: 184775

Ion Ratio Lower Upper

152 100

151 20.1 16.2 24.2

153 13.6 10.7 16.1

Manual Integrations




APPROVED

Reviewed By :Rahul
Chavli

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

#50

Dimethylphthalate

Concen: 41.981 ng

RT: 9.669 min Scan# 1271

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:163 Resp: 1268421


Ion Ratio Lower Upper

163 100

194 4.0 3.2 4.8

164 10.3 8.2 12.2






#51
2,6-Dinitrotoluene
Concen: 45.692 ng
RT: 9.727 min Scan# 1281
Delta R.T. -0.001 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS



Tgt Ion:165 Resp: 263569
Ion Ratio Lower Upper
165 100
63 59.7 48.5 72.7
89 53.2 43.9 65.9

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli




05/02/2025

Supervised By :Jagrut
Upadhyay


05/02/2025

Supervised By :Jagrut
Upadhyay



05/02/2025




#52
Acenaphthene
Concen: 43.261 ng
RT: 9.980 min Scan# 1324
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39



Tgt Ion:154 Resp: 1164359
Ion Ratio Lower Upper
154 100
153 110.8 88.6 133.0
152 52.8 42.1 63.1





#53

3-Nitroaniline

Concen: 16.287 ng

RT: 9.892 min Scan# 1309

Delta R.T. 0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:138	Resp: 9891
Ion Ratio	Lower Upper
138 100	
108 12.4	11.7 17.5
92 108.3	91.2 136.8

Manual Integrations
APPROVEDReviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#54

2,4-Dinitrophenol

Concen: 90.602 ng




RT: 9.998 min Scan# 1327


Delta R.T. 0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39


Tgt Ion:184	Resp: 196654
Ion Ratio	Lower Upper
184 100	
63 76.2	69.2 103.8
154 127.3	271.4 407.2#





#55
Dibenzofuran
Concen: 39.818 ng
RT: 10.157 min Scan# 1353
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:168 Resp: 1606504
Ion Ratio Lower Upper
168 100
139 36.8 29.4 44.0
169 13.8 11.0 16.4

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#56
4-Nitrophenol
Concen: 101.975 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:139 Resp: 484293
Ion Ratio Lower Upper
139 100
109 66.2 44.9 84.9
65 97.6 74.7 114.7





#57
2,4-Dinitrotoluene
Concen: 49.818 ng
RT: 10.133 min Scan# 11
Delta R.T. 0.005 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:165 Resp: 35098
Ion Ratio Lower Upper
165 100
63 48.1 38.8 58.2
89 69.0 57.1 85.7
182 2.7 2.2 3.2

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#58
Fluorene
Concen: 39.832 ng
RT: 10.498 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:166 Resp: 1221943
Ion Ratio Lower Upper
166 100
165 92.5 73.9 110.9
167 13.5 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 46.019 ng
RT: 10.269 min Scan# 1373
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:232 Resp: 330098
Ion Ratio Lower Upper
232 100
131 44.4 35.9 53.9
130 2.1 1.8 2.6
166 28.8 24.0 36.0

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#60
Diethylphthalate
Concen: 42.490 ng
RT: 10.369 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:149 Resp: 1267747
Ion Ratio Lower Upper
149 100
177 21.8 17.6 26.4
150 12.8 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 40.172 ng
RT: 10.486 min Scan# 1410
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:204 Resp: 585254
Ion Ratio Lower Upper
204 100
206 32.3 25.6 38.4
141 59.8 48.9 73.3

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025
Supervised By :Jagrut
Upadhyay





05/02/2025




#62
4-Nitroaniline
Concen: 44.888 ng
RT: 10.510 min Scan# 1414
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39




Tgt Ion:138 Resp: 286425
Ion Ratio Lower Upper
138 100
92 49.2 29.2 69.2
108 67.6 49.1 89.1





#63
Azobenzene
Concen: 41.009 ng
RT: 10.651 min Scan# 1438
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS






Tgt Ion: 77 Resp: 120823
Ion Ratio Lower Upper
77 100
182 27.8 7.3 47.3
105 22.7 2.5 42.5
51 34.8 14.4 54.4

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli




05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025
Supervised By :Jagrut
Upadhyay



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39


Tgt Ion:188 Resp: 804874
Ion Ratio Lower Upper
188 100
94 10.3 8.2 12.2
80 10.9 8.6 12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 47.699 ng
RT: 10.539 min Scan# 1418
Delta R.T. 0.005 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:198 Resp: 126239
Ion Ratio Lower Upper
198 100
51 55.6 40.6 80.6
105 46.4 29.2 69.2

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025
Supervised By :Jagrut
Upadhyay





05/02/2025




#66
n-Nitrosodiphenylamine
Concen: 40.164 ng
RT: 10.610 min Scan# 1431
Delta R.T. 0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39




Tgt Ion:169 Resp: 1104532
Ion Ratio Lower Upper
169 100
168 65.8 53.0 79.6
167 36.5 28.8 43.2





#67
4-Bromophenyl-phenylether
Concen: 41.655 ng
RT: 10.980 min Scan# 1494
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS






Tgt Ion:248 Resp: 373384
Ion Ratio Lower Upper
248 100
250 97.9 77.2 115.8
141 71.4 58.4 87.6

Manual Integrations
APPROVED

Reviewed By :Rahul
Chavli




05/02/2025
Supervised By :Jagrut
Upadhyay


05/02/2025
Supervised By :Jagrut
Upadhyay



#68
Hexachlorobenzene
Concen: 41.636 ng
RT: 11.045 min Scan# 1505
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:284 Resp: 417991
Ion Ratio Lower Upper
284 100
142 30.8 24.6 37.0
249 29.2 23.2 34.8





#69

Atrazine

Concen: 45.827 ng

RT: 11.133 min Scan# 1519

Delta R.T. 0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:200 Resp: 33219

Ion Ratio Lower Upper

200 100

173 28.5 8.6 48.6

215 50.7 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#70

Pentachlorophenol

Concen: 93.418 ng

RT: 11.233 min Scan# 1537

Delta R.T. 0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:266 Resp: 495448


Ion Ratio Lower Upper

266 100

268 63.8 50.0 75.0

264 63.3 49.3 73.9





#71

Phenanthrene

Concen: 40.041 ng

RT: 11.463 min Scan# 1575

Delta R.T. 0.006 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:178 Resp: 174081

Ion Ratio Lower Upper

178 100

176 19.6 15.5 23.3

179 15.5 12.6 19.0

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#72

Anthracene

Concen: 40.180 ng

RT: 11.510 min Scan# 1584

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:178 Resp: 1786049


Ion Ratio Lower Upper

178 100

176 19.3 15.3 22.9


179 15.9 12.7 19.1





#73
 Carbazole
 Concen: 41.368 ng
 RT: 11.663 min Scan# 1610
 Delta R.T. -0.000 min
 Lab File: BF142255.D
 Acq: 01 May 2025 12:39

Instrument :
 BNA_F
 ClientSampleId :
 PB167810BS





Tgt Ion:167 Resp: 167744
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.3 25.9
 139 12.9 10.3 15.5

Manual Integrations
 APPROVED


Reviewed By :Rahul
 Chavli

05/02/2025
 Supervised By :Jagrut
 Upadhyay


05/02/2025
 Supervised By :Jagrut
 Upadhyay





05/02/2025




#74
 Di-n-butylphthalate
 Concen: 45.113 ng
 RT: 11.998 min Scan# 1667
 Delta R.T. -0.000 min
 Lab File: BF142255.D
 Acq: 01 May 2025 12:39




Tgt Ion:149 Resp: 1839790
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.5 11.3
 104 4.5 3.6 5.4





#75
Fluoranthene
Concen: 42.331 ng
RT: 12.645 min Scan# 1777
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:202 Resp: 181927
Ion Ratio Lower Upper
202 100
101 12.3 0.0 32.4
203 17.9 0.0 37.5

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.074 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:240 Resp: 464382
Ion Ratio Lower Upper
240 100
120 12.4 10.1 15.1
236 25.9 20.5 30.7





#77

Benzidine

Concen: 17.635 ng

RT: 12.763 min Scan# 1797

Delta R.T. -0.006 min

Lab File: BF142255.D


Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS



Tgt Ion:184 Resp: 347980

Ion Ratio Lower Upper

184 100

185 15.8 11.8 17.6

183 10.5 8.9 13.3

Manual Integrations




APPROVED

Reviewed By :Rahul
Chavli

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

#78

Pyrene

Concen: 42.103 ng

RT: 12.880 min Scan# 1817

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:202 Resp: 1803569


Ion Ratio Lower Upper

202 100

200 20.3 16.1 24.1

203 17.9 14.1 21.1





#79

Terphenyl-d14

Concen: 74.104 ng

RT: 13.021 min Scan# 1841

Delta R.T. -0.000 min

Lab File: BF142255.D


Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS



Tgt Ion:244 Resp: 236608

Ion Ratio Lower Upper

244 100

212 6.7 5.3 7.9

122 11.1 9.1 13.7

Manual Integrations



APPROVED

Reviewed By :Rahul
Chavli


05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#80

Butylbenzylphthalate


Concen: 45.667 ng

RT: 13.498 min Scan# 1922

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39





Tgt Ion:149 Resp: 532587


Ion Ratio Lower Upper

149 100

91 68.7 54.4 81.6


206 20.9 15.9 23.9





#81
Benzo(a)anthracene
Concen: 42.672 ng
RT: 14.062 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:228 Resp: 133028
Ion Ratio Lower Upper
228 100
226 27.0 21.9 32.9
229 20.2 16.0 24.0

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#82
3,3'-Dichlorobenzidine
Concen: 8.154 ng
RT: 14.027 min Scan# 2012
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:252 Resp: 74664
Ion Ratio Lower Upper
252 100
254 65.3 51.0 76.6
126 11.3 9.4 14.2





#83

Chrysene

Concen: 40.631 ng

RT: 14.104 min Scan# 2025

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:228 Resp: 115299

Ion Ratio Lower Upper

228 100

226 29.9 24.0 36.0

229 19.5 15.7 23.5

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#84

Bis(2-ethylhexyl)phthalate

Concen: 39.646 ng

RT: 14.057 min Scan# 2017

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:149 Resp: 632694


Ion Ratio Lower Upper

149 100

167 26.8 21.7 32.5


279 4.0 3.0 4.4





#85
Di-n-octyl phthalate
Concen: 39.690 ng
RT: 14.674 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:149 Resp: 1144590
Ion Ratio Lower Upper
149 100
167 1.6 1.1 1.7
43 11.5 9.3 13.9

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay







05/02/2025



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.568 min Scan# 2274
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:264 Resp: 446826
Ion Ratio Lower Upper
264 100
260 23.6 18.4 27.6
265 21.3 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 42.010 ng

RT: 17.080 min Scan# 2199

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS

Tgt Ion:276 Resp: 1374940

Ion Ratio Lower Upper

276 100

138 28.3 22.6 33.8

277 25.1 20.1 30.1

Manual Integrations

APPROVED

Reviewed By :Rahul
Chavli





05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#88

Benzo(b)fluoranthene

Concen: 43.510 ng

RT: 15.127 min Scan# 2199

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39




Tgt Ion:252 Resp: 1207610


Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 10.7 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 39.540 ng

RT: 15.156 min Scan# 21

Delta R.T. -0.000 min

Lab File: BF142255.D


Acq: 01 May 2025 12:39

Instrument :

BNA_F

ClientSampleId :

PB167810BS



Tgt Ion:252 Resp: 100380

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.8

125 11.0 9.0 13.6

Manual Integrations



APPROVED

Reviewed By :Rahul
Chavli


05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025

Supervised By :Jagrut
Upadhyay

05/02/2025



#90

Benzo(a)pyrene


Concen: 42.227 ng

RT: 15.504 min Scan# 2263

Delta R.T. -0.000 min

Lab File: BF142255.D

Acq: 01 May 2025 12:39





Tgt Ion:252 Resp: 1073003


Ion Ratio Lower Upper

252 100

253 21.8 17.6 26.4


125 11.8 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 41.894 ng
RT: 17.098 min Scan# 2534
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Instrument :
BNA_F
ClientSampleId :
PB167810BS





Tgt Ion:278 Resp: 111281
Ion Ratio Lower Upper
278 100
139 19.3 15.2 22.8
279 24.1 19.0 28.6

Manual Integrations
APPROVED


Reviewed By :Rahul
Chavli

05/02/2025
Supervised By :Jagrut
Upadhyay

05/02/2025
Supervised By :Jagrut
Upadhyay






05/02/2025



#92
Benzo(g,h,i)perylene
Concen: 42.026 ng
RT: 17.539 min Scan# 2609
Delta R.T. -0.000 min
Lab File: BF142255.D
Acq: 01 May 2025 12:39

Tgt Ion:276 Resp: 1122234
Ion Ratio Lower Upper
276 100
277 23.8 18.7 28.1
138 24.4 19.4 29.0



Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142277.D	1	04/30/25 13:15	05/02/25 12:27	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	270		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	340		5.30	50.0	ug/L
95-48-7	2-Methylphenol	360		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	360		11.0	100	ug/L
67-72-1	Hexachloroethane	350		6.50	50.0	ug/L
98-95-3	Nitrobenzene	440		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	380		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	470		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	460		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	500		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	440		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	940	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	96.6		10 - 139	64%	SPK: 150
13127-88-3	Phenol-d6	89.8		10 - 134	60%	SPK: 150
4165-60-0	Nitrobenzene-d5	84.0		49 - 133	84%	SPK: 100
321-60-8	2-Fluorobiphenyl	70.7		52 - 132	71%	SPK: 100
118-79-6	2,4,6-Tribromophenol	126		44 - 137	84%	SPK: 150
1718-51-0	Terphenyl-d14	70.4		48 - 125	70%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	213000	6.91			
1146-65-2	Naphthalene-d8	808000	8.193			
15067-26-2	Acenaphthene-d10	402000	9.945			
1517-22-2	Phenanthrene-d10	630000	11.433			
1719-03-5	Chrysene-d12	370000	14.069			
1520-96-3	Perylene-d12	379000	15.557			

Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MS	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MS	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142277.D	1	04/30/25 13:15	05/02/25 12:27	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142277.D
 Acq On : 02 May 2025 12:27
 Operator : RC/JU
 Sample : Q1901-08MS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-167-SB01MS

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/05/2025
 Supervised By :Jagrut Upadhyay 05/05/2025

Quant Time: May 02 13:05:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	212656	20.000	ng	0.00
21) Naphthalene-d8	8.193	136	808361	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	402059	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	629997	20.000	ng	0.00
76) Chrysene-d12	14.069	240	370460	20.000	ng	0.00
86) Perylene-d12	15.557	264	379220	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	1285671	96.555	ng	0.01
7) Phenol-d6	6.528	99	1443545	89.776	ng	0.00
23) Nitrobenzene-d5	7.469	82	1033716	84.012	ng	0.00
42) 2,4,6-Tribromophenol	10.734	330	462379	126.149	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2016863	70.728	ng	0.00
79) Terphenyl-d14	13.016	244	1792164	70.359	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.822	88	169976	28.583	ng	# 84
3) Pyridine	3.546	79	415695	26.896	ng	99
4) n-Nitrosodimethylamine	3.481	42	278483	34.818	ng	99
6) Aniline	6.569	93	185208	8.138	ng	100
8) 2-Chlorophenol	6.693	128	531955	38.662	ng	99
9) Benzaldehyde	6.457	77	175465	15.613	ng	99
10) Phenol	6.540	94	537381m	31.043	ng	
11) bis(2-Chloroethyl)ether	6.646	93	502092	37.430	ng	99
12) 1,3-Dichlorobenzene	6.851	146	511679	32.873	ng	99
13) 1,4-Dichlorobenzene	6.928	146	527665	33.647	ng	99
14) 1,2-Dichlorobenzene	7.081	146	509008	33.999	ng	99
15) Benzyl Alcohol	7.046	79	445258	39.044	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.181	45	916352	37.830	ng	99
17) 2-Methylphenol	7.151	107	403260	35.777	ng	100
18) Hexachloroethane	7.422	117	180036	35.075	ng	97
19) n-Nitroso-di-n-propyla...	7.316	70	379242	38.168	ng	98
20) 3+4-Methylphenols	7.304	107	508536	36.077	ng	91
22) Acetophenone	7.316	105	714285	38.195	ng	97
24) Nitrobenzene	7.487	77	532026	44.068	ng	100
25) Isophorone	7.728	82	1017548	39.404	ng	99
26) 2-Nitrophenol	7.804	139	249952	48.946	ng	98
27) 2,4-Dimethylphenol	7.834	122	465938	35.787	ng	100
28) bis(2-Chloroethoxy)met...	7.934	93	653302	40.114	ng	100
29) 2,4-Dichlorophenol	8.040	162	450542	41.390	ng	99
30) 1,2,4-Trichlorobenzene	8.128	180	460388	37.902	ng	100
31) Naphthalene	8.210	128	1515067	37.582	ng	100
32) Benzoic acid	7.922	122	217267	37.981	ng	97
33) 4-Chloroaniline	8.251	127	79798	4.754	ng	99
34) Hexachlorobutadiene	8.328	225	267842	37.794	ng	99
35) Caprolactam	8.616	113	114486m	34.992	ng	
36) 4-Chloro-3-methylphenol	8.728	107	450003	40.260	ng	99
37) 2-Methylnaphthalene	8.898	142	965025	38.934	ng	99
38) 1-Methylnaphthalene	8.998	142	1001608	39.008	ng	99
40) 1,2,4,5-Tetrachloroben...	9.069	216	466320	41.212	ng	99
41) Hexachlorocyclopentadiene	9.057	237	570339	88.811	ng	100
43) 2,4,6-Trichlorophenol	9.175	196	326161	46.517	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142277.D
 Acq On : 02 May 2025 12:27
 Operator : RC/JU
 Sample : Q1901-08MS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-167-SB01MS

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/05/2025
 Supervised By :Jagrut Upadhyay 05/05/2025

Quant Time: May 02 13:05:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	336279	45.616	ng	99
46) 1,1'-Biphenyl	9.363	154	1292053	41.337	ng	99
47) 2-Chloronaphthalene	9.392	162	960343	41.296	ng	99
48) 2-Nitroaniline	9.481	65	294596	48.035	ng	98
49) Acenaphthylene	9.810	152	1605889	40.923	ng	100
50) Dimethylphthalate	9.663	163	1104923	42.957	ng	100
51) 2,6-Dinitrotoluene	9.722	165	230556	46.839	ng	99
52) Acenaphthene	9.981	154	1000886	43.682	ng	99
53) 3-Nitroaniline	9.886	138	93454	17.726	ng	98
54) 2,4-Dinitrophenol	9.992	184	156192	86.643	ng	# 52
55) Dibenzofuran	10.151	168	1420255	41.350	ng	100
56) 4-Nitrophenol	10.039	139	369705	91.443	ng	99
57) 2,4-Dinitrotoluene	10.128	165	301103	50.164	ng	96
58) Fluorene	10.492	166	1070926	41.007	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	273427	44.776	ng	99
60) Diethylphthalate	10.363	149	1082961	42.636	ng	100
61) 4-Chlorophenyl-phenyle...	10.486	204	522609	42.138	ng	96
62) 4-Nitroaniline	10.504	138	228504	42.238	ng	97
63) Azobenzene	10.645	77	1037185	41.352	ng	99
65) 4,6-Dinitro-2-methylph...	10.534	198	106866	50.723	ng	90
66) n-Nitrosodiphenylamine	10.604	169	918104	42.652	ng	99
67) 4-Bromophenyl-phenylether	10.975	248	319645	45.559	ng	98
68) Hexachlorobenzene	11.039	284	349088	44.425	ng	98
69) Atrazine	11.128	200	274854	48.442	ng	99
70) Pentachlorophenol	11.233	266	390510	94.071	ng	98
71) Phenanthrene	11.457	178	1459064	42.876	ng	100
72) Anthracene	11.510	178	1476623	42.440	ng	99
73) Carbazole	11.657	167	1322084	41.655	ng	100
74) Di-n-butylphthalate	11.992	149	1508741	47.265	ng	100
75) Fluoranthene	12.645	202	1362829	40.513	ng	98
77) Benzidine	12.763	184	40553	2.576	ng	97
78) Pyrene	12.875	202	1353270	39.600	ng	100
80) Butylbenzylphthalate	13.492	149	486120	51.715	ng	99
81) Benzo(a)anthracene	14.057	228	1077111	43.310	ng	100
82) 3,3'-Dichlorobenzidine	14.022	252	61349	8.398	ng	98
83) Chrysene	14.098	228	946310	41.802	ng	100
84) Bis(2-ethylhexyl)phtha...	14.051	149	628324	48.509	ng	99
85) Di-n-octyl phthalate	14.669	149	1009277	43.232	ng	99
87) Indeno(1,2,3-cd)pyrene	17.068	276	1154061	41.548	ng	100
88) Benzo(b)fluoranthene	15.121	252	1089899	46.269	ng	99
89) Benzo(k)fluoranthene	15.151	252	857957	39.820	ng	99
90) Benzo(a)pyrene	15.498	252	936845	43.441	ng	99
91) Dibenzo(a,h)anthracene	17.086	278	933265	41.398	ng	100
92) Benzo(g,h,i)perylene	17.521	276	936160	41.307	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
Data File : BF142277.D
Acq On : 02 May 2025 12:27
Operator : RC/JU
Sample : Q1901-08MS
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jaagrut Upadhyay 05/05/2025


Quant Time: May 02 13:05:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS




Tgt Ion:152 Resp: 212650
Ion Ratio Lower Upper
152 100
150 159.0 130.2 195.2
115 58.1 47.0 70.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#2
1,4-Dioxane
Concen: 28.583 ng
RT: 2.822 min Scan# 107
Delta R.T. 0.124 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27



Tgt Ion: 88 Resp: 169976
Ion Ratio Lower Upper
88 100
58 90.1 72.2 108.2
43 0.0 27.0 40.6





#3

Pyridine

Concen: 26.896 ng

RT: 3.546 min Scan# 216

Delta R.T. 0.082 min

Lab File: BF142277.D


Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS



Tgt Ion: 79 Resp: 41569

Ion Ratio Lower Upper

79 100

52 88.4 71.5 107.3




51 43.7 34.5 51.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#4

n-Nitrosodimethylamine

Concen: 34.818 ng

RT: 3.481 min Scan# 219

Delta R.T. 0.071 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion: 42 Resp: 278483


Ion Ratio Lower Upper

42 100

74 111.1 89.5 134.3


44 7.8 6.6 9.8





#5
2-Fluorophenol
Concen: 96.555 ng
RT: 5.528 min Scan# 50
Delta R.T. 0.012 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion: 112 Resp: 128567
Ion Ratio Lower Upper
112 100
64 60.6 48.4 72.6
63 31.2 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#6
Aniline
Concen: 8.138 ng
RT: 6.569 min Scan# 744
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion: 93 Resp: 185208
Ion Ratio Lower Upper
93 100
66 40.4 32.1 48.1
65 20.7 16.6 24.8





#7
Phenol-d6
Concen: 89.776 ng
RT: 6.528 min Scan# 737
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion: 99 Resp: 144354
Ion Ratio Lower Upper
99 100
42 20.9 16.3 24.5
71 31.8 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#8
2-Chlorophenol
Concen: 38.662 ng
RT: 6.693 min Scan# 765
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:128 Resp: 531955
Ion Ratio Lower Upper
128 100
130 32.7 13.1 53.1
64 47.9 28.6 68.6





#9
Benzaldehyde
Concen: 15.613 ng
RT: 6.457 min Scan# 725
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS




Tgt Ion: 77 Resp: 17546
Ion Ratio Lower Upper
77 100
105 100.9 79.1 119.1
106 96.2 76.0 116.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#10
Phenol
Concen: 31.043 ng m
RT: 6.540 min Scan# 739
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Tgt Ion: 94 Resp: 537381
Ion Ratio Lower Upper
94 100
65 29.4 8.9 48.9
66 40.3 18.7 58.7





#11
bis(2-Chloroethyl)ether
Concen: 37.430 ng
RT: 6.646 min Scan# 757
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion: 93 Resp: 502092
Ion Ratio Lower Upper
93 100
63 81.5 62.2 102.2
95 32.2 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#12
1,3-Dichlorobenzene
Concen: 32.873 ng
RT: 6.851 min Scan# 792
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:146 Resp: 511679
Ion Ratio Lower Upper
146 100
148 64.1 50.5 75.7
75 28.9 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 33.647 ng

RT: 6.928 min Scan# 805

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:146 Resp: 52766

Ion Ratio Lower Upper

146 100

148 63.0 50.6 76.0





111 41.0 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#14

1,2-Dichlorobenzene

Concen: 33.999 ng

RT: 7.081 min Scan# 831

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:146 Resp: 509008


Ion Ratio Lower Upper

146 100

148 64.1 50.5 75.7

111 43.0 34.9 52.3





#15

Benzyl Alcohol

Concen: 39.044 ng

RT: 7.046 min Scan# 81

Delta R.T. 0.000 min

Lab File: BF142277.D


Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS



Tgt Ion: 79 Resp: 445258

Ion Ratio Lower Upper

79 100

108 81.1 64.5 96.7




77 62.0 49.2 73.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 37.830 ng

RT: 7.181 min Scan# 848

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion: 45 Resp: 916352


Ion Ratio Lower Upper

45 100

77 10.8 0.0 31.3


79 8.3 0.0 28.5





#17
2-Methylphenol
Concen: 35.777 ng
RT: 7.151 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion:107 Resp: 403260
Ion Ratio Lower Upper
107 100
108 113.7 91.0 136.4
77 43.0 34.4 51.6
79 43.0 34.4 51.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#18
Hexachloroethane
Concen: 35.075 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:117 Resp: 180036
Ion Ratio Lower Upper
117 100
119 96.1 77.1 115.7
201 94.4 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 38.168 ng

RT: 7.316 min Scan# 871

Delta R.T. -0.006 min

Lab File: BF142277.D


Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS



Tgt Ion: 70 Resp: 37924

Ion Ratio Lower Upper

70 100

42 69.0 53.4 80.2

101 9.9 7.7 11.5




130 21.4 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#20

3+4-Methylphenols

Concen: 36.077 ng

RT: 7.304 min Scan# 869

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Tgt Ion:107 Resp: 508536




Ion Ratio Lower Upper


107 100

108 90.6 71.2 111.2

77 80.7 41.1 81.1

79 28.2 8.1 48.1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.193 min Scan# 1020

Delta R.T. -0.000 min

Lab File: BF142277.D


Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS



Tgt Ion:136 Resp: 80836

Ion Ratio Lower Upper

136 100

137 10.9 8.7 13.1

54 9.2 7.7 11.5




68 6.2 5.0 7.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#22

Acetophenone

Concen: 38.195 ng

RT: 7.316 min Scan# 871

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Tgt Ion:105 Resp: 714285




Ion Ratio Lower Upper


105 100

71 8.5 5.7 8.5

51 35.0 30.1 45.1

120 23.7 19.0 28.4





#23

Nitrobenzene-d5

Concen: 84.012 ng

RT: 7.469 min Scan# 89

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion: 82 Resp: 1033710

Ion Ratio Lower Upper

82 100

128 46.1 37.0 55.6





54 60.7 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#24

Nitrobenzene

Concen: 44.068 ng

RT: 7.487 min Scan# 900

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion: 77 Resp: 532026


Ion Ratio Lower Upper

77 100

123 48.3 38.6 58.0

65 14.6 11.5 17.3









#25
Isophorone
Concen: 39.404 ng
RT: 7.728 min Scan# 941
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Tgt Ion: 82 Resp: 101754
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 19.9 15.3 22.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#26
2-Nitrophenol
Concen: 48.946 ng
RT: 7.804 min Scan# 954
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:139 Resp: 249952
Ion Ratio Lower Upper
139 100
109 27.3 21.8 32.8
65 49.3 41.1 61.7





#27

2,4-Dimethylphenol

Concen: 35.787 ng

RT: 7.834 min Scan# 91

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:122 Resp: 465938

Ion Ratio Lower Upper

122 100

107 107.3 86.0 129.0





121 57.6 46.3 69.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#28

bis(2-Chloroethoxy)methane

Concen: 40.114 ng

RT: 7.934 min Scan# 976

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion: 93 Resp: 653302


Ion Ratio Lower Upper

93 100

95 32.7 26.2 39.2

123 12.5 9.7 14.5





#29

2,4-Dichlorophenol

Concen: 41.390 ng

RT: 8.040 min Scan# 994

Delta R.T. 0.000 min

Lab File: BF142277.D


Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS



Tgt Ion:162 Resp: 450542

Ion Ratio Lower Upper

162 100

164 63.6 43.0 83.0




98 36.1 16.7 56.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#30

1,2,4-Trichlorobenzene

Concen: 37.902 ng

RT: 8.128 min Scan# 1009

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:180 Resp: 460388


Ion Ratio Lower Upper

180 100

182 94.7 75.7 113.5


145 31.6 25.4 38.0





#31
Naphthalene
Concen: 37.582 ng
RT: 8.210 min Scan# 1023
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS




Tgt Ion:128 Resp: 151506
Ion Ratio Lower Upper
128 100
129 11.3 9.0 13.4
127 13.2 10.6 15.8

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#32
Benzoic acid
Concen: 37.981 ng
RT: 7.922 min Scan# 974
Delta R.T. -0.012 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27



Tgt Ion:122 Resp: 217267
Ion Ratio Lower Upper
122 100
105 116.5 98.2 138.2
77 91.2 66.8 106.8





#33

4-Chloroaniline

Concen: 4.754 ng

RT: 8.251 min Scan# 1030

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:127 Resp: 79798

Ion Ratio Lower Upper

127 100

129 31.7 26.1 39.1

65 31.0 25.5 38.3





92 19.6 15.4 23.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#34

Hexachlorobutadiene

Concen: 37.794 ng

RT: 8.328 min Scan# 1043

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:225 Resp: 267842


Ion Ratio Lower Upper

225 100

223 63.2 49.7 74.5

227 63.5 51.0 76.4





#35

Caprolactam

Concen: 34.992 ng m

RT: 8.616 min Scan# 1092

Delta R.T. -0.006 min

Lab File: BF142277.D


Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS



Tgt Ion:113 Resp: 114480

Ion Ratio Lower Upper

113 100

55 188.8 177.7 217.7




56 140.2 127.1 167.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#36

4-Chloro-3-methylphenol

Concen: 40.260 ng

RT: 8.728 min Scan# 1111

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:107 Resp: 450003


Ion Ratio Lower Upper

107 100

144 26.8 21.0 31.4


142 83.8 66.1 99.1





#37
2-Methylnaphthalene
Concen: 38.934 ng
RT: 8.898 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion:142 Resp: 965025
Ion Ratio Lower Upper
142 100
141 87.2 70.5 105.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#38
1-Methylnaphthalene
Concen: 39.008 ng
RT: 8.998 min Scan# 1157
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:142 Resp: 1001608
Ion Ratio Lower Upper
142 100
141 89.9 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.945 min Scan# 1169

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:164 Resp: 402059

Ion Ratio Lower Upper

164 100

162 99.8 78.6 117.8





160 44.9 35.7 53.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#40

1,2,4,5-Tetrachlorobenzene

Concen: 41.212 ng

RT: 9.069 min Scan# 1169

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Tgt Ion:216 Resp: 466320




Ion Ratio Lower Upper


216 100

214 78.7 63.8 95.6

179 20.9 16.7 25.1

108 20.5 15.6 23.4





#41

Hexachlorocyclopentadiene

Concen: 88.811 ng

RT: 9.057 min Scan# 1167

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:237 Resp: 570339

Ion Ratio Lower Upper

237 100

235 63.2 43.5 83.5





272 13.2 0.0 32.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#42

2,4,6-Tribromophenol

Concen: 126.149 ng

RT: 10.734 min Scan# 1452

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:330 Resp: 462379


Ion Ratio Lower Upper

330 100

332 96.5 76.5 114.7

141 29.6 24.3 36.5





#43

2,4,6-Trichlorophenol

Concen: 46.517 ng

RT: 9.175 min Scan# 1187

Delta R.T. 0.000 min

Lab File: BF142277.D


Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS



Tgt Ion:196 Resp: 32616

Ion Ratio Lower Upper

196 100

198 94.4 77.4 116.2




200 30.6 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#44

2,4,5-Trichlorophenol

Concen: 45.616 ng

RT: 9.210 min Scan# 1193

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Tgt Ion:196 Resp: 336279




Ion Ratio Lower Upper


196 100

198 95.4 75.2 112.8

97 49.4 39.8 59.8

132 29.9 23.8 35.8





#45

2-Fluorobiphenyl

Concen: 70.728 ng

RT: 9.263 min Scan# 1202

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:172 Resp: 201686

Ion Ratio Lower Upper

172 100

171 35.9 29.0 43.4





170 23.8 19.4 29.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#46

1,1'-Biphenyl

Concen: 41.337 ng

RT: 9.363 min Scan# 1219

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:154 Resp: 1292053


Ion Ratio Lower Upper

154 100

153 39.8 20.5 60.5

76 13.1 0.0 32.8





#47

2-Chloronaphthalene

Concen: 41.296 ng

RT: 9.392 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:162 Resp: 96034

Ion Ratio Lower Upper

162 100

127 40.9 32.6 49.0





164 33.3 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#48

2-Nitroaniline

Concen: 48.035 ng

RT: 9.481 min Scan# 1239

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion: 65 Resp: 294596


Ion Ratio Lower Upper

65 100

92 65.4 54.3 81.5

138 108.7 88.6 133.0





#49

Acenaphthylene

Concen: 40.923 ng

RT: 9.810 min Scan# 1295

Delta R.T. 0.000 min

Lab File: BF142277.D


Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS



Tgt Ion:152 Resp: 1605889

Ion Ratio Lower Upper

152 100

151 20.3 16.2 24.2




153 13.3 10.7 16.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#50

Dimethylphthalate


Concen: 42.957 ng

RT: 9.663 min Scan# 1270

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27





Tgt Ion:163 Resp: 1104923


Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8


164 10.0 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 46.839 ng
RT: 9.722 min Scan# 1280
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion:165 Resp: 230550
Ion Ratio Lower Upper
165 100
63 62.0 48.5 72.7
89 54.3 43.9 65.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#52
Acenaphthene
Concen: 43.682 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:154 Resp: 1000886
Ion Ratio Lower Upper
154 100
153 109.9 88.6 133.0
152 52.4 42.1 63.1





#53

3-Nitroaniline

Concen: 17.726 ng

RT: 9.886 min Scan# 11

Delta R.T. -0.005 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS





Tgt Ion:	138	Resp:	93454
Ion Ratio	Lower	Upper	
138	100		
108	11.7	11.7	17.5
92	113.0	91.2	136.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#54

2,4-Dinitrophenol

Concen: 86.643 ng




RT: 9.992 min Scan# 1326


Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27


Tgt Ion:	184	Resp:	156192
Ion Ratio	Lower	Upper	
184	100		
63	80.1	69.2	103.8
154	214.0	271.4	407.2#





#55
Dibenzenofuran
Concen: 41.350 ng
RT: 10.151 min Scan# 1353
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion:168 Resp: 142025
Ion Ratio Lower Upper
168 100
139 36.7 29.4 44.0
169 13.5 11.0 16.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#56
4-Nitrophenol
Concen: 91.443 ng
RT: 10.039 min Scan# 1334
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:139 Resp: 369705
Ion Ratio Lower Upper
139 100
109 65.6 44.9 84.9
65 96.2 74.7 114.7





#57
2,4-Dinitrotoluene
Concen: 50.164 ng
RT: 10.128 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS




Tgt Ion:165 Resp: 30110
Ion Ratio Lower Upper
165 100
63 46.0 38.8 58.2
89 67.2 57.1 85.7
182 2.5 2.2 3.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#58
Fluorene
Concen: 41.007 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Tgt Ion:166 Resp: 1070926
Ion Ratio Lower Upper
166 100
165 93.0 73.9 110.9
167 13.7 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 44.776 ng
RT: 10.263 min Scan# 1372
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion:232 Resp: 273421
Ion Ratio Lower Upper
232 100
131 45.5 35.9 53.9
130 2.1 1.8 2.6
166 29.5 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#60
Diethylphthalate
Concen: 42.636 ng
RT: 10.363 min Scan# 1389
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:149 Resp: 1082961
Ion Ratio Lower Upper
149 100
177 21.9 17.6 26.4
150 12.7 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 42.138 ng
RT: 10.486 min Scan# 1410
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS




Tgt Ion:204 Resp: 522609
Ion Ratio Lower Upper
204 100
206 32.1 25.6 38.4
141 56.9 48.9 73.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#62
4-Nitroaniline
Concen: 42.238 ng
RT: 10.504 min Scan# 1413
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Tgt Ion:138 Resp: 228504
Ion Ratio Lower Upper
138 100
92 47.9 29.2 69.2
108 66.2 49.1 89.1





#63
Azobenzene
Concen: 41.352 ng
RT: 10.645 min Scan# 1437
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion: 77 Resp: 103718
Ion Ratio Lower Upper
77 100
182 26.8 7.3 47.3
105 22.5 2.5 42.5
51 35.0 14.4 54.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:188 Resp: 629997
Ion Ratio Lower Upper
188 100
94 9.8 8.2 12.2
80 10.1 8.6 12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 50.723 ng
RT: 10.534 min Scan# 1418
Delta R.T. -0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS




Tgt Ion:198 Resp: 106860
Ion Ratio Lower Upper
198 100
51 52.8 40.6 80.6
105 42.9 29.2 69.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#66
n-Nitrosodiphenylamine
Concen: 42.652 ng
RT: 10.604 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Tgt Ion:169 Resp: 918104
Ion Ratio Lower Upper
169 100
168 65.8 53.0 79.6
167 35.6 28.8 43.2





#67
4-Bromophenyl-phenylether
Concen: 45.559 ng
RT: 10.975 min Scan# 1493
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS




Tgt Ion:248 Resp: 31964
Ion Ratio Lower Upper
248 100
250 98.6 77.2 115.8
141 73.8 58.4 87.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#68
Hexachlorobenzene
Concen: 44.425 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27



Tgt Ion:284 Resp: 349088
Ion Ratio Lower Upper
284 100
142 32.4 24.6 37.0
249 29.3 23.2 34.8





#69

Atrazine

Concen: 48.442 ng

RT: 11.128 min Scan# 1519

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:200 Resp: 274854

Ion Ratio Lower Upper

200 100

173 28.4 8.6 48.6





215 49.6 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#70

Pentachlorophenol

Concen: 94.071 ng

RT: 11.233 min Scan# 1537

Delta R.T. 0.000 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:266 Resp: 390510


Ion Ratio Lower Upper

266 100

268 63.6 50.0 75.0

264 62.9 49.3 73.9









#71
Phenanthrene
Concen: 42.876 ng
RT: 11.457 min Scan# 1575
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS

Tgt Ion:178 Resp: 1459064
Ion Ratio Lower Upper
178 100
176 19.4 15.5 23.3
179 15.9 12.6 19.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#72
Anthracene
Concen: 42.440 ng
RT: 11.510 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27


Tgt Ion:178 Resp: 1476623
Ion Ratio Lower Upper
178 100
176 18.9 15.3 22.9
179 15.6 12.7 19.1





#73
 Carbazole
 Concen: 41.655 ng
 RT: 11.657 min Scan# 1609
 Delta R.T. -0.006 min
 Lab File: BF142277.D
 Acq: 02 May 2025 12:27




Instrument :
 BNA_F
 ClientSampleId :
 B-167-SB01MS




Tgt Ion:167 Resp: 1322084
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.3 25.9
 139 12.8 10.3 15.5

Manual Integrations
 APPROVED



Reviewed By :Rahul Chavli 05/05/2025
 Supervised By :Jagrut Upadhyay 05/05/2025




#74
 Di-n-butylphthalate
 Concen: 47.265 ng
 RT: 11.992 min Scan# 1666
 Delta R.T. -0.006 min
 Lab File: BF142277.D
 Acq: 02 May 2025 12:27



Tgt Ion:149 Resp: 1508741
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.5 11.3
 104 4.5 3.6 5.4





#75

Fluoranthene

Concen: 40.513 ng

RT: 12.645 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF142277.D


Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS



Tgt Ion:202 Resp: 1362829

Ion Ratio Lower Upper

202 100

101 11.3 0.0 32.4




203 17.8 0.0 37.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#76

Chrysene-d12


Concen: 20.000 ng

RT: 14.069 min Scan# 2019

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27





Tgt Ion:240 Resp: 370460


Ion Ratio Lower Upper

240 100

120 12.3 10.1 15.1

236 25.2 20.5 30.7





#77

Benzidine

Concen: 2.576 ng

RT: 12.763 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:184 Resp: 4055

Ion Ratio Lower Upper

184 100

185 15.8 11.8 17.6





183 10.0 8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#78

Pyrene

Concen: 39.600 ng

RT: 12.875 min Scan# 1816

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:202 Resp: 1353270


Ion Ratio Lower Upper

202 100

200 20.0 16.1 24.1

203 17.7 14.1 21.1





#79

Terphenyl-d14

Concen: 70.359 ng

RT: 13.016 min Scan# 1840

Delta R.T. -0.006 min

Lab File: BF142277.D


Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS



Tgt Ion:244 Resp: 179216

Ion Ratio Lower Upper

244 100

212 6.5 5.3 7.9




122 11.2 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#80

Butylbenzylphthalate

Concen: 51.715 ng

RT: 13.492 min Scan# 1921

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:149 Resp: 486120


Ion Ratio Lower Upper

149 100

91 68.1 54.4 81.6

206 20.8 15.9 23.9





#81

Benzo(a)anthracene

Concen: 43.310 ng

RT: 14.057 min Scan# 2011

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:228 Resp: 107711

Ion Ratio Lower Upper

228 100

226 27.5 21.9 32.9





229 20.1 16.0 24.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#82

3,3'-Dichlorobenzidine

Concen: 8.398 ng

RT: 14.022 min Scan# 2011

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:252 Resp: 61349


Ion Ratio Lower Upper

252 100

254 65.6 51.0 76.6

126 12.2 9.4 14.2





#83

Chrysene

Concen: 41.802 ng

RT: 14.098 min Scan# 2016

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:228 Resp: 946310

Ion Ratio Lower Upper

228 100

226 30.1 24.0 36.0





229 19.7 15.7 23.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#84

Bis(2-ethylhexyl)phthalate

Concen: 48.509 ng

RT: 14.051 min Scan# 2016

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:149 Resp: 628324


Ion Ratio Lower Upper

149 100

167 27.7 21.7 32.5

279 4.1 3.0 4.4





#85

Di-n-octyl phthalate

Concen: 43.232 ng

RT: 14.669 min Scan# 2121

Delta R.T. -0.011 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:149 Resp: 100927

Ion Ratio Lower Upper

149 100

167 1.5 1.1 1.7





43 11.3 9.3 13.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.557 min Scan# 2272

Delta R.T. -0.012 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:264 Resp: 379220


Ion Ratio Lower Upper

264 100

260 23.2 18.4 27.6


265 21.4 17.3 25.9





#87
Indeno(1,2,3-cd)pyrene
Concen: 41.548 ng
RT: 17.068 min Scan# 2198
Delta R.T. -0.012 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion:276 Resp: 115406
Ion Ratio Lower Upper
276 100
138 28.1 22.6 33.8
277 25.0 20.1 30.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#88
Benzo(b)fluoranthene
Concen: 46.269 ng
RT: 15.121 min Scan# 2198
Delta R.T. -0.006 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:252 Resp: 1089899
Ion Ratio Lower Upper
252 100
253 21.7 17.7 26.5
125 10.3 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 39.820 ng

RT: 15.151 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MS

Tgt Ion:252 Resp: 85795

Ion Ratio Lower Upper

252 100

253 22.2 17.8 26.8





125 10.7 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#90

Benzo(a)pyrene

Concen: 43.441 ng

RT: 15.498 min Scan# 2262

Delta R.T. -0.006 min

Lab File: BF142277.D

Acq: 02 May 2025 12:27




Tgt Ion:252 Resp: 936845


Ion Ratio Lower Upper

252 100

253 22.4 17.6 26.4


125 11.4 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 41.398 ng
RT: 17.086 min Scan# 21
Delta R.T. -0.012 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MS



Tgt Ion:278 Resp: 93326
Ion Ratio Lower Upper
278 100
139 18.6 15.2 22.8
279 23.8 19.0 28.6




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#92
Benzo(g,h,i)perylene
Concen: 41.307 ng
RT: 17.521 min Scan# 2606
Delta R.T. -0.018 min
Lab File: BF142277.D
Acq: 02 May 2025 12:27

Tgt Ion:276 Resp: 936160
Ion Ratio Lower Upper
276 100
277 24.1 18.7 28.1
138 24.6 19.4 29.0



Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MSD	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142278.D	1	04/30/25 13:15	05/02/25 12:55	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	280		12.8	50.0	ug/L
106-46-7	1,4-Dichlorobenzene	340		5.30	50.0	ug/L
95-48-7	2-Methylphenol	370		11.2	50.0	ug/L
65794-96-9	3+4-Methylphenols	370		11.0	100	ug/L
67-72-1	Hexachloroethane	370		6.50	50.0	ug/L
98-95-3	Nitrobenzene	460		7.60	50.0	ug/L
87-68-3	Hexachlorobutadiene	390		5.40	50.0	ug/L
88-06-2	2,4,6-Trichlorophenol	480		5.10	50.0	ug/L
95-95-4	2,4,5-Trichlorophenol	470		6.20	50.0	ug/L
121-14-2	2,4-Dinitrotoluene	530		12.2	50.0	ug/L
118-74-1	Hexachlorobenzene	450		5.20	50.0	ug/L
87-86-5	Pentachlorophenol	970	E	15.8	100	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	99.4		10 - 139	66%	SPK: 150
13127-88-3	Phenol-d6	92.8		10 - 134	62%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.8		49 - 133	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	72.3		52 - 132	72%	SPK: 100
118-79-6	2,4,6-Tribromophenol	133		44 - 137	88%	SPK: 150
1718-51-0	Terphenyl-d14	72.8		48 - 125	73%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	217000	6.91			
1146-65-2	Naphthalene-d8	833000	8.193			
15067-26-2	Acenaphthene-d10	419000	9.945			
1517-22-2	Phenanthrene-d10	684000	11.433			
1719-03-5	Chrysene-d12	400000	14.069			
1520-96-3	Perylene-d12	401000	15.563			

Report of Analysis

Client:	CDM Smith	Date Collected:	04/26/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	04/28/25
Client Sample ID:	B-167-SB01MSD	SDG No.:	Q1915
Lab Sample ID:	Q1901-08MSD	Matrix:	TCLP
Analytical Method:	SW8270	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142278.D	1	04/30/25 13:15	05/02/25 12:55	PB167810

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142278.D
 Acq On : 02 May 2025 12:55
 Operator : RC/JU
 Sample : Q1901-08MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-167-SB01MSD

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 05/05/2025
 Supervised By :Jagrut Upadhyay 05/05/2025

Quant Time: May 02 13:41:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 30 16:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.910	152	217028	20.000	ng	0.00
21) Naphthalene-d8	8.193	136	832936	20.000	ng	0.00
39) Acenaphthene-d10	9.945	164	419407	20.000	ng	0.00
64) Phenanthrene-d10	11.433	188	684254	20.000	ng	0.00
76) Chrysene-d12	14.069	240	400079	20.000	ng	0.00
86) Perylene-d12	15.563	264	401381	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	1351308	99.440	ng	0.01
7) Phenol-d6	6.528	99	1523396	92.834	ng	0.00
23) Nitrobenzene-d5	7.469	82	1112775	87.769	ng	0.00
42) 2,4,6-Tribromophenol	10.734	330	506639	132.507	ng	0.00
45) 2-Fluorobiphenyl	9.263	172	2151303	72.323	ng	0.00
79) Terphenyl-d14	13.016	244	2003750	72.842	ng	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	2.828	88	174625	28.773	ng	# 82
3) Pyridine	3.546	79	441222	27.973	ng	100
4) n-Nitrosodimethylamine	3.487	42	294114	36.032	ng	99
6) Aniline	6.569	93	197398	8.499	ng	98
8) 2-Chlorophenol	6.693	128	562936	40.089	ng	99
9) Benzaldehyde	6.457	77	183950	16.038	ng	99
10) Phenol	6.540	94	571858m	32.370	ng	
11) bis(2-Chloroethyl)ether	6.646	93	532809	38.919	ng	99
12) 1,3-Dichlorobenzene	6.851	146	542334	34.140	ng	99
13) 1,4-Dichlorobenzene	6.928	146	548211	34.253	ng	99
14) 1,2-Dichlorobenzene	7.081	146	540561	35.379	ng	99
15) Benzyl Alcohol	7.046	79	478246	41.092	ng	100
16) 2,2'-oxybis(1-Chloropr...	7.181	45	960997	38.874	ng	100
17) 2-Methylphenol	7.151	107	427556	37.168	ng	99
18) Hexachloroethane	7.422	117	192593	36.766	ng	96
19) n-Nitroso-di-n-propyla...	7.322	70	400784	39.523	ng	99
20) 3+4-Methylphenols	7.304	107	538223	37.414	ng	92
22) Acetophenone	7.316	105	748030	38.819	ng	97
24) Nitrobenzene	7.487	77	571114	45.910	ng	100
25) Isophorone	7.728	82	1075681	40.426	ng	99
26) 2-Nitrophenol	7.804	139	277477	51.929	ng	98
27) 2,4-Dimethylphenol	7.834	122	495291	36.919	ng	100
28) bis(2-Chloroethoxy)met...	7.934	93	685392	40.843	ng	99
29) 2,4-Dichlorophenol	8.040	162	483547	43.112	ng	99
30) 1,2,4-Trichlorobenzene	8.128	180	482951	38.586	ng	99
31) Naphthalene	8.210	128	1609412	38.745	ng	100
32) Benzoic acid	7.928	122	263320	42.769	ng	97
33) 4-Chloroaniline	8.251	127	72531	4.193	ng	99
34) Hexachlorobutadiene	8.328	225	287976	39.436	ng	99
35) Caprolactam	8.616	113	125300m	37.167	ng	
36) 4-Chloro-3-methylphenol	8.728	107	484815	42.095	ng	99
37) 2-Methylnaphthalene	8.904	142	1012785	39.656	ng	100
38) 1-Methylnaphthalene	8.998	142	1065609	40.276	ng	99
40) 1,2,4,5-Tetrachloroben...	9.069	216	493589	41.817	ng	99
41) Hexachlorocyclopentadiene	9.057	237	611613	91.299	ng	99
43) 2,4,6-Trichlorophenol	9.175	196	351147	48.009	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\
 Data File : BF142278.D
 Acq On : 02 May 2025 12:55
 Operator : RC/JU
 Sample : Q1901-08MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025

Quant Time: May 02 13:41:32 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Apr 30 16:00:01 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.210	196	361688	47.034	ng	99
46) 1,1'-Biphenyl	9.363	154	1368430	41.970	ng	100
47) 2-Chloronaphthalene	9.392	162	1020867	42.083	ng	99
48) 2-Nitroaniline	9.481	65	319718	49.782	ng	98
49) Acenaphthylene	9.810	152	1710079	41.776	ng	100
50) Dimethylphthalate	9.663	163	1199810	44.716	ng	100
51) 2,6-Dinitrotoluene	9.722	165	253926	49.228	ng	99
52) Acenaphthene	9.981	154	1087078	45.482	ng	98
53) 3-Nitroaniline	9.886	138	99043	17.958	ng	98
54) 2,4-Dinitrophenol	9.998	184	184644	93.898	ng	# 1
55) Dibenzofuran	10.151	168	1522541	42.495	ng	100
56) 4-Nitrophenol	10.045	139	407185	96.548	ng	99
57) 2,4-Dinitrotoluene	10.128	165	336232	53.345	ng	96
58) Fluorene	10.492	166	1144424	42.008	ng	99
59) 2,3,4,6-Tetrachlorophenol	10.263	232	303708	47.678	ng	99
60) Diethylphthalate	10.369	149	1182771	44.640	ng	99
61) 4-Chlorophenyl-phenyle...	10.486	204	556055	42.980	ng	97
62) 4-Nitroaniline	10.504	138	257712	45.444	ng	98
63) Azobenzene	10.645	77	1119990	42.806	ng	99
65) 4,6-Dinitro-2-methylph...	10.534	198	125145	53.859	ng	94
66) n-Nitrosodiphenylamine	10.604	169	983989	42.088	ng	99
67) 4-Bromophenyl-phenylether	10.975	248	347963	45.662	ng	99
68) Hexachlorobenzene	11.039	284	382262	44.789	ng	97
69) Atrazine	11.128	200	312642	50.733	ng	99
70) Pentachlorophenol	11.233	266	437319	96.993	ng	99
71) Phenanthrene	11.457	178	1591587	43.062	ng	100
72) Anthracene	11.510	178	1599092	42.316	ng	100
73) Carbazole	11.657	167	1465976	42.526	ng	100
74) Di-n-butylphthalate	11.992	149	1677931	48.397	ng	100
75) Fluoranthene	12.645	202	1520599	41.619	ng	98
77) Benzidine	12.763	184	44894	2.641	ng	97
78) Pyrene	12.875	202	1524967	41.321	ng	100
80) Butylbenzylphthalate	13.492	149	558469	54.776	ng	99
81) Benzo(a)anthracene	14.063	228	1194007	44.456	ng	100
82) 3,3'-Dichlorobenzidine	14.022	252	70246	8.904	ng	99
83) Chrysene	14.098	228	1076915	44.049	ng	99
84) Bis(2-ethylhexyl)phtha...	14.051	149	722253	51.410	ng	99
85) Di-n-octyl phthalate	14.669	149	1156306	45.494	ng	99
87) Indeno(1,2,3-cd)pyrene	17.068	276	1219041	41.464	ng	100
88) Benzo(b)fluoranthene	15.121	252	1168207	46.856	ng	99
89) Benzo(k)fluoranthene	15.151	252	987950	43.322	ng	100
90) Benzo(a)pyrene	15.498	252	1023759	44.850	ng	100
91) Dibenzo(a,h)anthracene	17.092	278	988821	41.441	ng	99
92) Benzo(g,h,i)perylene	17.527	276	978093	40.775	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050225\  
Data File : BF142278.D  
Acq On    : 02 May 2025  12:55  
Operator  : RC/JU  
Sample    : Q1901-08MSD  
Misc      :  
ALS Vial  : 6    Sample Multiplier: 1
```


Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD

Manual Integrations

Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025


Quant Time: May 02 13:41:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 30 16:00:01 2025
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.910 min Scan# 802
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD




Tgt Ion: 152 Resp: 217028
Ion Ratio Lower Upper
152 100
150 159.7 130.2 195.2
115 57.6 47.0 70.4

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#2
1,4-Dioxane
Concen: 28.773 ng
RT: 2.828 min Scan# 108
Delta R.T. 0.129 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Tgt Ion: 88 Resp: 174625
Ion Ratio Lower Upper
88 100
58 92.4 72.2 108.2
43 0.0 27.0 40.6





#3
Pyridine
Concen: 27.973 ng
RT: 3.546 min Scan# 216
Delta R.T. 0.082 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion: 79 Resp: 441222
Ion Ratio Lower Upper
79 100
52 89.1 71.5 107.3
51 42.8 34.5 51.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#4
n-Nitrosodimethylamine
Concen: 36.032 ng
RT: 3.487 min Scan# 220
Delta R.T. 0.076 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion: 42 Resp: 294114
Ion Ratio Lower Upper
42 100
74 112.6 89.5 134.3
44 7.7 6.6 9.8





#5
2-Fluorophenol
Concen: 99.440 ng
RT: 5.528 min Scan# 50
Delta R.T. 0.012 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion: 112 Resp: 1351308
Ion Ratio Lower Upper
112 100
64 61.0 48.4 72.6
63 31.8 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#6
Aniline
Concen: 8.499 ng
RT: 6.569 min Scan# 744
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion: 93 Resp: 197398
Ion Ratio Lower Upper
93 100
66 41.3 32.1 48.1
65 21.0 16.6 24.8





#7

Phenol-d6

Concen: 92.834 ng

RT: 6.528 min Scan# 71

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion: 99 Resp: 1523390

Ion Ratio Lower Upper

99 100

42 21.0 16.3 24.5





71 32.1 25.0 37.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#8

2-Chlorophenol

Concen: 40.089 ng

RT: 6.693 min Scan# 765

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:128 Resp: 562936


Ion Ratio Lower Upper

128 100

130 32.6 13.1 53.1

64 48.3 28.6 68.6





#9

Benzaldehyde

Concen: 16.038 ng

RT: 6.457 min Scan# 725

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion: 77 Resp: 183950

Ion Ratio Lower Upper

77 100

105 99.6 79.1 119.1





106 96.6 76.0 116.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#10

Phenol

Concen: 32.370 ng m

RT: 6.540 min Scan# 739

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion: 94 Resp: 571858


Ion Ratio Lower Upper

94 100

65 29.7 8.9 48.9


66 41.7 18.7 58.7





#11
bis(2-Chloroethyl)ether
Concen: 38.919 ng
RT: 6.646 min Scan# 757
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion: 93 Resp: 532809
Ion Ratio Lower Upper
93 100
63 81.3 62.2 102.2
95 31.8 12.7 52.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#12
1,3-Dichlorobenzene
Concen: 34.140 ng
RT: 6.851 min Scan# 792
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:146 Resp: 542334
Ion Ratio Lower Upper
146 100
148 63.9 50.5 75.7
75 28.8 23.2 34.8





#13

1,4-Dichlorobenzene

Concen: 34.253 ng

RT: 6.928 min Scan# 805

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:146 Resp: 54821

Ion Ratio Lower Upper

146 100

148 64.0 50.6 76.0





111 42.1 33.4 50.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#14

1,2-Dichlorobenzene

Concen: 35.379 ng

RT: 7.081 min Scan# 831

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:146 Resp: 540561


Ion Ratio Lower Upper

146 100

148 63.6 50.5 75.7

111 42.7 34.9 52.3





#15

Benzyl Alcohol

Concen: 41.092 ng

RT: 7.046 min Scan# 81

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion: 79 Resp: 478240

Ion Ratio Lower Upper

79 100

108 80.7 64.5 96.7





77 62.2 49.2 73.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#16

2,2'-oxybis(1-Chloropropane)

Concen: 38.874 ng

RT: 7.181 min Scan# 848

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion: 45 Resp: 960997


Ion Ratio Lower Upper

45 100

77 11.0 0.0 31.3


79 8.4 0.0 28.5





#17
2-Methylphenol
Concen: 37.168 ng
RT: 7.151 min Scan# 843
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:107 Resp: 427550
Ion Ratio Lower Upper
107 100
108 114.4 91.0 136.4
77 43.5 34.4 51.6
79 42.8 34.4 51.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#18
Hexachloroethane
Concen: 36.766 ng
RT: 7.422 min Scan# 889
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:117 Resp: 192593
Ion Ratio Lower Upper
117 100
119 94.7 77.1 115.7
201 93.5 79.4 119.2





#19

n-Nitroso-di-n-propylamine

Concen: 39.523 ng

RT: 7.322 min Scan# 81

Delta R.T. 0.000 min

Lab File: BF142278.D


Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD



Tgt Ion: 70 Resp: 400784

Ion Ratio Lower Upper

70 100

42 67.2 53.4 80.2

101 9.9 7.7 11.5




130 23.0 17.2 25.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#20

3+4-Methylphenols

Concen: 37.414 ng

RT: 7.304 min Scan# 869

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Tgt Ion:107 Resp: 538223




Ion Ratio Lower Upper


107 100

108 91.5 71.2 111.2

77 78.2 41.1 81.1


79 28.1 8.1 48.1





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.193 min Scan# 1020
Delta R.T. -0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:136 Resp: 832930
Ion Ratio Lower Upper
136 100
137 10.8 8.7 13.1
54 9.3 7.7 11.5
68 6.0 5.0 7.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#22
Acetophenone
Concen: 38.819 ng
RT: 7.316 min Scan# 871
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:105 Resp: 748030
Ion Ratio Lower Upper
105 100
71 8.1 5.7 8.5
51 35.4 30.1 45.1
120 24.0 19.0 28.4





#23

Nitrobenzene-d5

Concen: 87.769 ng

RT: 7.469 min Scan# 89

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion: 82 Resp: 111277

Ion Ratio Lower Upper

82 100

128 46.9 37.0 55.6





54 60.6 48.0 72.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#24

Nitrobenzene

Concen: 45.910 ng

RT: 7.487 min Scan# 900

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion: 77 Resp: 571114


Ion Ratio Lower Upper

77 100

123 48.0 38.6 58.0

65 14.6 11.5 17.3





#25

Isophorone

Concen: 40.426 ng

RT: 7.728 min Scan# 941

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion: 82 Resp: 107568

Ion Ratio Lower Upper

82 100

95 7.2 5.6 8.4





138 19.8 15.3 22.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#26

2-Nitrophenol

Concen: 51.929 ng

RT: 7.804 min Scan# 954

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:139 Resp: 277477


Ion Ratio Lower Upper

139 100

109 27.3 21.8 32.8

65 48.8 41.1 61.7





#27

2,4-Dimethylphenol

Concen: 36.919 ng

RT: 7.834 min Scan# 91

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:122 Resp: 49529

Ion Ratio Lower Upper

122 100

107 107.9 86.0 129.0





121 58.0 46.3 69.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#28

bis(2-Chloroethoxy)methane

Concen: 40.843 ng

RT: 7.934 min Scan# 976

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion: 93 Resp: 685392


Ion Ratio Lower Upper

93 100

95 33.0 26.2 39.2


123 13.0 9.7 14.5





#29
2,4-Dichlorophenol
Concen: 43.112 ng
RT: 8.040 min Scan# 994
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:162 Resp: 48354
Ion Ratio Lower Upper
162 100
164 64.0 43.0 83.0
98 35.6 16.7 56.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#30
1,2,4-Trichlorobenzene
Concen: 38.586 ng
RT: 8.128 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:180 Resp: 482951
Ion Ratio Lower Upper
180 100
182 95.4 75.7 113.5
145 32.3 25.4 38.0





#31
Naphthalene
Concen: 38.745 ng
RT: 8.210 min Scan# 1024
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:128 Resp: 160941
Ion Ratio Lower Upper
128 100
129 11.5 9.0 13.4
127 13.3 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#32
Benzoic acid
Concen: 42.769 ng
RT: 7.928 min Scan# 975
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:122 Resp: 263320
Ion Ratio Lower Upper
122 100
105 119.9 98.2 138.2
77 91.8 66.8 106.8





#33

4-Chloroaniline

Concen: 4.193 ng

RT: 8.251 min Scan# 1030

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:127 Resp: 7253

Ion Ratio Lower Upper

127 100

129 31.9 26.1 39.1

65 31.0 25.5 38.3





92 19.2 15.4 23.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#34

Hexachlorobutadiene

Concen: 39.436 ng

RT: 8.328 min Scan# 1043

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:225 Resp: 287976


Ion Ratio Lower Upper

225 100

223 62.7 49.7 74.5

227 62.9 51.0 76.4





#35

Caprolactam

Concen: 37.167 ng m

RT: 8.616 min Scan# 1092

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:113 Resp: 125300

Ion Ratio Lower Upper

113 100

55 197.1 177.7 217.7





56 142.9 127.1 167.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#36

4-Chloro-3-methylphenol

Concen: 42.095 ng

RT: 8.728 min Scan# 1111

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:107 Resp: 484815


Ion Ratio Lower Upper

107 100

144 26.8 21.0 31.4


142 83.9 66.1 99.1





#37
2-Methylnaphthalene
Concen: 39.656 ng
RT: 8.904 min Scan# 1141
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:142 Resp: 101278
Ion Ratio Lower Upper
142 100
141 88.1 70.5 105.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#38
1-Methylnaphthalene
Concen: 40.276 ng
RT: 8.998 min Scan# 1157
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:142 Resp: 1065609
Ion Ratio Lower Upper
142 100
141 90.2 72.8 109.2





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.945 min Scan# 11

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:164 Resp: 41940

Ion Ratio Lower Upper

164 100

162 98.5 78.6 117.8





160 44.5 35.7 53.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#40

1,2,4,5-Tetrachlorobenzene

Concen: 41.817 ng

RT: 9.069 min Scan# 1169

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Tgt Ion:216 Resp: 493589




Ion Ratio Lower Upper


216 100

214 79.0 63.8 95.6

179 21.2 16.7 25.1

108 20.6 15.6 23.4





#41

Hexachlorocyclopentadiene

Concen: 91.299 ng

RT: 9.057 min Scan# 1167

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:237 Resp: 61161

Ion Ratio Lower Upper

237 100

235 63.0 43.5 83.5





272 13.1 0.0 32.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#42

2,4,6-Tribromophenol

Concen: 132.507 ng

RT: 10.734 min Scan# 1452

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:330 Resp: 506639


Ion Ratio Lower Upper

330 100

332 97.3 76.5 114.7

141 29.5 24.3 36.5





#43

2,4,6-Trichlorophenol

Concen: 48.009 ng

RT: 9.175 min Scan# 1187

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:196 Resp: 35114

Ion Ratio Lower Upper

196 100

198 95.1 77.4 116.2





200 30.8 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#44

2,4,5-Trichlorophenol

Concen: 47.034 ng

RT: 9.210 min Scan# 1193

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Tgt Ion:196 Resp: 361688




Ion Ratio Lower Upper


196 100

198 94.8 75.2 112.8

97 50.1 39.8 59.8

132 30.5 23.8 35.8





#45

2-Fluorobiphenyl

Concen: 72.323 ng

RT: 9.263 min Scan# 1202

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:172 Resp: 215130

Ion Ratio Lower Upper

172 100

171 35.5 29.0 43.4





170 23.8 19.4 29.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#46

1,1'-Biphenyl

Concen: 41.970 ng

RT: 9.363 min Scan# 1219

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:154 Resp: 1368430


Ion Ratio Lower Upper

154 100

153 40.4 20.5 60.5

76 12.9 0.0 32.8





#47

2-Chloronaphthalene

Concen: 42.083 ng

RT: 9.392 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:162 Resp: 102086

Ion Ratio Lower Upper

162 100

127 40.5 32.6 49.0





164 33.1 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#48

2-Nitroaniline

Concen: 49.782 ng

RT: 9.481 min Scan# 1239

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion: 65 Resp: 319718


Ion Ratio Lower Upper

65 100

92 67.2 54.3 81.5

138 108.2 88.6 133.0





#49

Acenaphthylene

Concen: 41.776 ng

RT: 9.810 min Scan# 1295

Delta R.T. 0.000 min

Lab File: BF142278.D


Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD



Tgt Ion:152 Resp: 1710079

Ion Ratio Lower Upper

152 100

151 20.4 16.2 24.2




153 13.3 10.7 16.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#50

Dimethylphthalate


Concen: 44.716 ng

RT: 9.663 min Scan# 1270

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55





Tgt Ion:163 Resp: 1199810


Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8


164 10.4 8.2 12.2





#51
2,6-Dinitrotoluene
Concen: 49.228 ng
RT: 9.722 min Scan# 1280
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:165 Resp: 253920
Ion Ratio Lower Upper
165 100
63 60.1 48.5 72.7
89 53.9 43.9 65.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#52
Acenaphthene
Concen: 45.482 ng
RT: 9.981 min Scan# 1324
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:154 Resp: 1087078
Ion Ratio Lower Upper
154 100
153 108.5 88.6 133.0
152 52.1 42.1 63.1





#53

3-Nitroaniline

Concen: 17.958 ng

RT: 9.886 min Scan# 1308

Delta R.T. -0.005 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:138 Resp: 9904

Ion Ratio Lower Upper

138 100

108 12.6 11.7 17.5





92 112.3 91.2 136.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#54

2,4-Dinitrophenol

Concen: 93.898 ng

RT: 9.998 min Scan# 1327

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:184 Resp: 184644


Ion Ratio Lower Upper

184 100

63 65.1 69.2 103.8#


154 79.6 271.4 407.2#





#55
Dibenzenofuran
Concen: 42.495 ng
RT: 10.151 min Scan# 1353
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:168 Resp: 152254
Ion Ratio Lower Upper
168 100
139 36.5 29.4 44.0
169 13.6 11.0 16.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#56
4-Nitrophenol
Concen: 96.548 ng
RT: 10.045 min Scan# 1335
Delta R.T. 0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:139 Resp: 407185
Ion Ratio Lower Upper
139 100
109 64.3 44.9 84.9
65 93.1 74.7 114.7





#57
2,4-Dinitrotoluene
Concen: 53.345 ng
RT: 10.128 min Scan# 11
Delta R.T. -0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:165 Resp: 336232
Ion Ratio Lower Upper
165 100
63 46.4 38.8 58.2
89 66.8 57.1 85.7
182 2.8 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#58
Fluorene
Concen: 42.008 ng
RT: 10.492 min Scan# 1411
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:166 Resp: 1144424
Ion Ratio Lower Upper
166 100
165 93.0 73.9 110.9
167 13.8 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 47.678 ng
RT: 10.263 min Scan# 1372
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:232 Resp: 303708
Ion Ratio Lower Upper
232 100
131 44.5 35.9 53.9
130 2.1 1.8 2.6
166 29.4 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#60
Diethylphthalate
Concen: 44.640 ng
RT: 10.369 min Scan# 1390
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:149 Resp: 1182771
Ion Ratio Lower Upper
149 100
177 22.3 17.6 26.4
150 12.5 10.2 15.2





#61
4-Chlorophenyl-phenylether
Concen: 42.980 ng
RT: 10.486 min Scan# 1410
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD




Tgt Ion:204 Resp: 55605
Ion Ratio Lower Upper
204 100
206 32.9 25.6 38.4
141 58.0 48.9 73.3

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#62
4-Nitroaniline
Concen: 45.444 ng
RT: 10.504 min Scan# 1413
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Tgt Ion:138 Resp: 257712
Ion Ratio Lower Upper
138 100
92 48.3 29.2 69.2
108 66.7 49.1 89.1





#63
Azobenzene
Concen: 42.806 ng
RT: 10.645 min Scan# 1438
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion: 77 Resp: 1119990
Ion Ratio Lower Upper
77 100
182 27.0 7.3 47.3
105 22.6 2.5 42.5
51 34.9 14.4 54.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.433 min Scan# 1571
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:188 Resp: 684254
Ion Ratio Lower Upper
188 100
94 9.7 8.2 12.2
80 10.1 8.6 12.8





#65
4,6-Dinitro-2-methylphenol
Concen: 53.859 ng
RT: 10.534 min Scan# 1418
Delta R.T. -0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:198 Resp: 125145
Ion Ratio Lower Upper
198 100
51 55.4 40.6 80.6
105 45.3 29.2 69.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#66
n-Nitrosodiphenylamine
Concen: 42.088 ng
RT: 10.604 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55


Tgt Ion:169 Resp: 983989
Ion Ratio Lower Upper
169 100
168 66.7 53.0 79.6
167 36.4 28.8 43.2





#67
4-Bromophenyl-phenylether
Concen: 45.662 ng
RT: 10.975 min Scan# 1493
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD




Tgt Ion:248 Resp: 347963
Ion Ratio Lower Upper
248 100
250 98.4 77.2 115.8
141 73.4 58.4 87.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#68
Hexachlorobenzene
Concen: 44.789 ng
RT: 11.039 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55



Tgt Ion:284 Resp: 382262
Ion Ratio Lower Upper
284 100
142 32.4 24.6 37.0
249 30.4 23.2 34.8





#69

Atrazine

Concen: 50.733 ng

RT: 11.128 min Scan# 1519

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:200 Resp: 31264

Ion Ratio Lower Upper

200 100

173 28.0 8.6 48.6





215 48.8 28.5 68.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#70

Pentachlorophenol

Concen: 96.993 ng

RT: 11.233 min Scan# 1537

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:266 Resp: 437319


Ion Ratio Lower Upper

266 100

268 63.5 50.0 75.0

264 62.5 49.3 73.9





#71

Phenanthrene

Concen: 43.062 ng

RT: 11.457 min Scan# 1575

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:178 Resp: 159158

Ion Ratio Lower Upper

178 100

176 19.5 15.5 23.3





179 15.7 12.6 19.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#72

Anthracene

Concen: 42.316 ng

RT: 11.510 min Scan# 1584

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:178 Resp: 1599092


Ion Ratio Lower Upper

178 100

176 18.9 15.3 22.9

179 15.7 12.7 19.1





#73

Carbazole

Concen: 42.526 ng

RT: 11.657 min Scan# 1609

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:167 Resp: 1465970

Ion Ratio Lower Upper

167 100

166 21.5 17.3 25.9





139 13.1 10.3 15.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#74

Di-n-butylphthalate

Concen: 48.397 ng

RT: 11.992 min Scan# 1666

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:149 Resp: 1677931


Ion Ratio Lower Upper

149 100

150 9.4 7.5 11.3

104 4.5 3.6 5.4





#75

Fluoranthene

Concen: 41.619 ng

RT: 12.645 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:202 Resp: 152059

Ion Ratio Lower Upper

202 100

101 11.2 0.0 32.4





203 18.0 0.0 37.5

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.069 min Scan# 2019

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:240 Resp: 400079


Ion Ratio Lower Upper

240 100

120 12.6 10.1 15.1

236 25.2 20.5 30.7





#77

Benzidine

Concen: 2.641 ng

RT: 12.763 min Scan# 1797

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD





Tgt Ion:184	Resp:	44894
Ion Ratio	Lower	Upper
184	100	
185	16.7	11.8 17.6
183	11.2	8.9 13.3

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#78

Pyrene

Concen: 41.321 ng




RT: 12.875 min Scan# 1816


Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Tgt Ion:202	Resp:	1524967
Ion Ratio	Lower	Upper
202	100	
200	20.2	16.1 24.1
203	17.6	14.1 21.1





#79

Terphenyl-d14

Concen: 72.842 ng

RT: 13.016 min Scan# 1840

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:244 Resp: 2003750

Ion Ratio Lower Upper

244 100

212 6.6 5.3 7.9





122 11.4 9.1 13.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#80

Butylbenzylphthalate

Concen: 54.776 ng

RT: 13.492 min Scan# 1921

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:149 Resp: 558469


Ion Ratio Lower Upper

149 100

91 68.6 54.4 81.6


206 20.7 15.9 23.9





#81
Benzo(a)anthracene
Concen: 44.456 ng
RT: 14.063 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD




Tgt Ion:228 Resp: 119400
Ion Ratio Lower Upper
228 100
226 27.2 21.9 32.9
229 19.9 16.0 24.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025




#82
3,3'-Dichlorobenzidine
Concen: 8.904 ng
RT: 14.022 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Tgt Ion:252 Resp: 70246
Ion Ratio Lower Upper
252 100
254 64.2 51.0 76.6
126 12.1 9.4 14.2





#83
Chrysene
Concen: 44.049 ng
RT: 14.098 min Scan# 2025
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD



Tgt Ion:228 Resp: 107691
Ion Ratio Lower Upper
228 100
226 30.4 24.0 36.0
229 20.0 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#84
Bis(2-ethylhexyl)phthalate
Concen: 51.410 ng
RT: 14.051 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55

Tgt Ion:149 Resp: 722253
Ion Ratio Lower Upper
149 100
167 27.6 21.7 32.5
279 4.1 3.0 4.4





#85

Di-n-octyl phthalate

Concen: 45.494 ng

RT: 14.669 min Scan# 2121

Delta R.T. -0.011 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:149 Resp: 1156300

Ion Ratio Lower Upper

149 100

167 1.5 1.1 1.7





43 11.3 9.3 13.9

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.563 min Scan# 2273

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:264 Resp: 401381


Ion Ratio Lower Upper

264 100

260 23.1 18.4 27.6

265 21.7 17.3 25.9





#87

Indeno(1,2,3-cd)pyrene

Concen: 41.464 ng

RT: 17.068 min Scan# 21904

Delta R.T. -0.012 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:276 Resp: 121904

Ion Ratio Lower Upper

276 100

138 27.9 22.6 33.8





277 25.0 20.1 30.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#88

Benzo(b)fluoranthene

Concen: 46.856 ng

RT: 15.121 min Scan# 2198

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:252 Resp: 1168207


Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 10.5 9.0 13.4





#89

Benzo(k)fluoranthene

Concen: 43.322 ng

RT: 15.151 min Scan# 21

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55

Instrument :

BNA_F

ClientSampleId :

B-167-SB01MSD

Tgt Ion:252 Resp: 987950

Ion Ratio Lower Upper

252 100

253 22.2 17.8 26.8





125 11.1 9.0 13.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 05/05/2025

Supervised By :Jagrut Upadhyay 05/05/2025



#90

Benzo(a)pyrene

Concen: 44.850 ng

RT: 15.498 min Scan# 2262

Delta R.T. -0.006 min

Lab File: BF142278.D

Acq: 02 May 2025 12:55




Tgt Ion:252 Resp: 1023759


Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4


125 11.7 9.5 14.3





#91
Dibenzo(a,h)anthracene
Concen: 41.441 ng
RT: 17.092 min Scan# 21
Delta R.T. -0.006 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55




Instrument :
BNA_F
ClientSampleId :
B-167-SB01MSD




Tgt Ion:278 Resp: 98882
Ion Ratio Lower Upper
278 100
139 18.0 15.2 22.8
279 24.0 19.0 28.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 05/05/2025
Supervised By :Jagrut Upadhyay 05/05/2025



#92
Benzo(g,h,i)perylene
Concen: 40.775 ng
RT: 17.527 min Scan# 2607
Delta R.T. -0.012 min
Lab File: BF142278.D
Acq: 02 May 2025 12:55



Tgt Ion:276 Resp: 978093
Ion Ratio Lower Upper
276 100
277 23.0 18.7 28.1
138 23.9 19.4 29.0



Manual Integration Report

Sequence:	BF043025	Instrument	BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDICC005	BF142240.D	2,3,4,6-Tetrachlorophen ol	Rahul	5/1/2025 11:26:17 AM	Jagrut	5/1/2025 1:17:43 PM	Peak Integrated by Software
SSTDICC020	BF142242.D	Benzoic acid	Rahul	5/1/2025 11:26:23 AM	Jagrut	5/1/2025 1:17:45 PM	Peak Integrated by Software
SSTDICCC040	BF142243.D	Benzoic acid	Rahul	5/1/2025 11:26:25 AM	Jagrut	5/1/2025 1:17:48 PM	Peak Integrated by Software
SSTDICC050	BF142244.D	Phenol	Rahul	5/1/2025 11:26:27 AM	Jagrut	5/1/2025 1:17:50 PM	Peak Integrated by Software
SSTDICC060	BF142245.D	Phenol	Rahul	5/1/2025 11:26:30 AM	Jagrut	5/1/2025 1:17:52 PM	Peak Integrated by Software
SSTDICC080	BF142246.D	Benzoic acid	Rahul	5/1/2025 11:26:33 AM	Jagrut	5/1/2025 1:17:55 PM	Peak Integrated by Software
SSTDICC080	BF142246.D	Phenol	Rahul	5/1/2025 11:26:33 AM	Jagrut	5/1/2025 1:17:55 PM	Peak Integrated by Software
SSTDICV040	BF142247.D	Benzoic acid	Rahul	5/1/2025 11:26:36 AM	Jagrut	5/1/2025 1:17:57 PM	Peak Integrated by Software
SSTDICV040	BF142247.D	Phenol	Rahul	5/1/2025 11:26:36 AM	Jagrut	5/1/2025 1:17:57 PM	Peak Integrated by Software

Manual Integration Report

Sequence:

bf050125

Instrument

BNA_f

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC040	BF142250.D	Benzoic acid	Rahul	5/2/2025 10:45:14 AM	Jagrut	5/2/2025 4:47:49 PM	Peak Integrated by Software
SSTDCCC040	BF142250.D	Phenol	Rahul	5/2/2025 10:45:14 AM	Jagrut	5/2/2025 4:47:49 PM	Peak Integrated by Software
PB167810BS	BF142255.D	Benzoic acid	Rahul	5/2/2025 10:45:25 AM	Jagrut	5/2/2025 4:47:56 PM	Peak Integrated by Software
PB167810BS	BF142255.D	Caprolactam	Rahul	5/2/2025 10:45:25 AM	Jagrut	5/2/2025 4:47:56 PM	Peak Integrated by Software
PB167810BS	BF142255.D	Phenol	Rahul	5/2/2025 10:45:25 AM	Jagrut	5/2/2025 4:47:56 PM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

bf050225

Instrument

BNA_f

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF043025

Review By	Rahul	Review On	5/1/2025 1:20:20 PM
Supervise By	Jagrut	Supervise On	5/1/2025 1:23:08 PM
SubDirectory	BF043025	HP Acquire Method	BNA_F
		HP Processing Method	BF043025
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729		
CCC	SP6725		
Internal Standard/PEM	S12661,10ul/1000ul sample		
ICV/I.BLK	SP6770		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF142238.D	30 Apr 2025 10:55	RC/JU	Ok
2	SSTDICC2.5	BF142239.D	30 Apr 2025 11:24	RC/JU	Ok
3	SSTDICC005	BF142240.D	30 Apr 2025 11:52	RC/JU	Ok,M
4	SSTDICC010	BF142241.D	30 Apr 2025 12:20	RC/JU	Ok
5	SSTDICC020	BF142242.D	30 Apr 2025 12:49	RC/JU	Ok,M
6	SSTDICCC040	BF142243.D	30 Apr 2025 13:17	RC/JU	Ok,M
7	SSTDICC050	BF142244.D	30 Apr 2025 13:46	RC/JU	Ok,M
8	SSTDICC060	BF142245.D	30 Apr 2025 14:15	RC/JU	Ok,M
9	SSTDICC080	BF142246.D	30 Apr 2025 14:43	RC/JU	Ok,M
10	SSTDICV040	BF142247.D	30 Apr 2025 16:17	RC/JU	Ok,M
11	PB167726BL	BF142248.D	30 Apr 2025 17:15	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF050125

Review By	Rahul	Review On	5/2/2025 10:46:38 AM
Supervise By	Jagrut	Supervise On	5/2/2025 4:49:12 PM
SubDirectory	BF050125	HP Acquire Method	BNA_F
		HP Processing Method	BF043025
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729		
CCC	SP6725		
Internal Standard/PEM	S12661,10ul/1000ul sample		
ICV/I.BLK	SP6770		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF142249.D	01 May 2025 09:48	RC/JU	Ok
2	SSTDCCC040	BF142250.D	01 May 2025 10:17	RC/JU	Ok,M
3	PB167798BL	BF142251.D	01 May 2025 10:45	RC/JU	Ok
4	PB167798BS	BF142252.D	01 May 2025 11:13	RC/JU	Ok,M
5	PB167798BSD	BF142253.D	01 May 2025 11:42	RC/JU	Ok,M
6	PB167810BL	BF142254.D	01 May 2025 12:10	RC/JU	Ok
7	PB167810BS	BF142255.D	01 May 2025 12:39	RC/JU	Ok,M
8	PB167774TB	BF142256.D	01 May 2025 13:07	RC/JU	Ok
9	Q1890-01	BF142257.D	01 May 2025 13:40	RC/JU	Ok
10	Q1920-01	BF142258.D	01 May 2025 14:09	RC/JU	Ok,M
11	Q1914-14	BF142259.D	01 May 2025 14:37	RC/JU	Ok
12	Q1913-04	BF142260.D	01 May 2025 15:06	RC/JU	Ok
13	Q1915-01	BF142261.D	01 May 2025 15:34	RC/JU	Ok
14	Q1903-01	BF142262.D	01 May 2025 16:03	RC/JU	Ok
15	Q1903-02	BF142263.D	01 May 2025 16:31	RC/JU	Ok
16	Q1903-03	BF142264.D	01 May 2025 17:00	RC/JU	Ok
17	Q1906-01	BF142265.D	01 May 2025 17:29	RC/JU	Ok
18	Q1906-13	BF142266.D	01 May 2025 17:57	RC/JU	Ok
19	Q1901-06	BF142267.D	01 May 2025 18:26	RC/JU	Ok
20	Q1901-02	BF142268.D	01 May 2025 18:55	RC/JU	Ok,M
21	Q1901-02MS	BF142269.D	01 May 2025 19:24	RC/JU	Ok,M

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF050125

Review By	Rahul	Review On	5/2/2025 10:46:38 AM		
Supervise By	Jagrut	Supervise On	5/2/2025 4:49:12 PM		
SubDirectory	BF050125	HP Acquire Method	BNA_F	HP Processing Method	BF043025
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729			
CCC		SP6725			
Internal Standard/PEM		S12661,10ul/1000ul sample			
ICV/I.BLK		SP6770			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q1901-02MSD	BF142270.D	01 May 2025 19:52	RC/JU	Ok,M
23	Q1901-04	BF142271.D	01 May 2025 20:20	RC/JU	Ok
24	Q1901-05	BF142272.D	01 May 2025 20:49	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF050225

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF050225	HP Acquire Method	BNA_F	HP Processing Method	BF043025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729				
CCC	SP6725				
Internal Standard/PEM	S12662,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF142273.D	02 May 2025 10:11	RC/JU	Ok
2	SSTDCCC040	BF142274.D	02 May 2025 10:39	RC/JU	Ok,NR
3	PB167774TB	BF142275.D	02 May 2025 11:07	RC/JU	Not Ok
4	Q1901-08	BF142276.D	02 May 2025 11:58	RC/JU	Ok
5	Q1901-08MS	BF142277.D	02 May 2025 12:27	RC/JU	Ok,NR
6	Q1901-08MSD	BF142278.D	02 May 2025 12:55	RC/JU	Ok,NR
7	Q1901-09	BF142279.D	02 May 2025 13:24	RC/JU	Ok
8	Q1901-10	BF142280.D	02 May 2025 13:53	RC/JU	ReRun
9	Q1901-11	BF142281.D	02 May 2025 14:22	RC/JU	Ok
10	Q1907-02	BF142282.D	02 May 2025 14:51	RC/JU	ReRun
11	Q1916-04	BF142283.D	02 May 2025 16:20	RC/JU	Ok
12	Q1916-04MS	BF142284.D	02 May 2025 16:49	RC/JU	Not Ok
13	Q1916-04MSD	BF142285.D	02 May 2025 17:18	RC/JU	Not Ok
14	Q1932-02	BF142286.D	02 May 2025 17:47	RC/JU	Ok
15	Q1932-04	BF142287.D	02 May 2025 18:16	RC/JU	Ok
16	Q1932-06	BF142288.D	02 May 2025 18:46	RC/JU	ReRun
17	Q1932-08	BF142289.D	02 May 2025 19:15	RC/JU	Ok
18	Q1917-04	BF142290.D	02 May 2025 19:44	RC/JU	Ok
19	Q1922-04	BF142291.D	02 May 2025 20:13	RC/JU	Ok
20	Q1922-08	BF142292.D	02 May 2025 20:42	RC/JU	Ok
21	Q1929-03	BF142293.D	02 May 2025 21:11	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF043025

Review By	Rahul	Review On	5/1/2025 1:20:20 PM
Supervise By	Jagrut	Supervise On	5/1/2025 1:23:08 PM
SubDirectory	BF043025	HP Acquire Method	BNA_F
		HP Processing Method	BF043025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142238.D	30 Apr 2025 10:55		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF142239.D	30 Apr 2025 11:24		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF142240.D	30 Apr 2025 11:52	Compound#32,54,56,65,70,85 removed from 5 ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF142241.D	30 Apr 2025 12:20		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF142242.D	30 Apr 2025 12:49		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF142243.D	30 Apr 2025 13:17	Compound#32,48,51,53,57,62,65,80,84,85 Kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF142244.D	30 Apr 2025 13:46	Compound#26,54 Kept on QR	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF142245.D	30 Apr 2025 14:15		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF142246.D	30 Apr 2025 14:43		RC/JU	Ok,M
10	SSTDICV040	ICVBF043025	BF142247.D	30 Apr 2025 16:17		RC/JU	Ok,M
11	PB167726BL	PB167726BL	BF142248.D	30 Apr 2025 17:15		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF050125

Review By	Rahul	Review On	5/2/2025 10:46:38 AM
Supervise By	Jagrut	Supervise On	5/2/2025 4:49:12 PM
SubDirectory	BF050125	HP Acquire Method	BNA_F
		HP Processing Method	BF043025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12661,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142249.D	01 May 2025 09:48		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142250.D	01 May 2025 10:17		RC/JU	Ok,M
3	PB167798BL	PB167798BL	BF142251.D	01 May 2025 10:45		RC/JU	Ok
4	PB167798BS	PB167798BS	BF142252.D	01 May 2025 11:13		RC/JU	Ok,M
5	PB167798BSD	PB167798BSD	BF142253.D	01 May 2025 11:42		RC/JU	Ok,M
6	PB167810BL	PB167810BL	BF142254.D	01 May 2025 12:10		RC/JU	Ok
7	PB167810BS	PB167810BS	BF142255.D	01 May 2025 12:39		RC/JU	Ok,M
8	PB167774TB	PB167774TB	BF142256.D	01 May 2025 13:07		RC/JU	Ok
9	Q1890-01	TW-WTS-07	BF142257.D	01 May 2025 13:40		RC/JU	Ok
10	Q1920-01	291	BF142258.D	01 May 2025 14:09		RC/JU	Ok,M
11	Q1914-14	FB04282025	BF142259.D	01 May 2025 14:37		RC/JU	Ok
12	Q1913-04	WC-13-A-202504	BF142260.D	01 May 2025 15:06		RC/JU	Ok
13	Q1915-01	WC-04282025	BF142261.D	01 May 2025 15:34		RC/JU	Ok
14	Q1903-01	COMP-4	BF142262.D	01 May 2025 16:03		RC/JU	Ok
15	Q1903-02	COMP-5	BF142263.D	01 May 2025 16:31		RC/JU	Ok
16	Q1903-03	COMP-6	BF142264.D	01 May 2025 17:00		RC/JU	Ok
17	Q1906-01	WC-4	BF142265.D	01 May 2025 17:29		RC/JU	Ok
18	Q1906-13	WC-7	BF142266.D	01 May 2025 17:57		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF050125

Review By	Rahul	Review On	5/2/2025 10:46:38 AM		
Supervise By	Jagrut	Supervise On	5/2/2025 4:49:12 PM		
SubDirectory	BF050125	HP Acquire Method	BNA_F	HP Processing Method	BF043025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729				
CCC	SP6725				
Internal Standard/PEM	S12661,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1901-06	FB04262025	BF142267.D	01 May 2025 18:26		RC/JU	Ok
20	Q1901-02	B-167-SB01	BF142268.D	01 May 2025 18:55		RC/JU	Ok,M
21	Q1901-02MS	B-167-SB01MS	BF142269.D	01 May 2025 19:24		RC/JU	Ok,M
22	Q1901-02MSD	B-167-SB01MSD	BF142270.D	01 May 2025 19:52		RC/JU	Ok,M
23	Q1901-04	B-167-SB02	BF142271.D	01 May 2025 20:20		RC/JU	Ok
24	Q1901-05	B-170-SB02	BF142272.D	01 May 2025 20:49		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF050225

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF050225	HP Acquire Method	BNA_F	HP Processing Method	BF043025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729				
CCC	SP6725				
Internal Standard/PEM	S12662,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142273.D	02 May 2025 10:11		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142274.D	02 May 2025 10:39		RC/JU	Ok,NR
3	PB167774TB	PB167774TB	BF142275.D	02 May 2025 11:07	Analyzed for contamination check	RC/JU	Not Ok
4	Q1901-08	B-167-SB01	BF142276.D	02 May 2025 11:58		RC/JU	Ok
5	Q1901-08MS	B-167-SB01MS	BF142277.D	02 May 2025 12:27		RC/JU	Ok,NR
6	Q1901-08MSD	B-167-SB01MSD	BF142278.D	02 May 2025 12:55		RC/JU	Ok,NR
7	Q1901-09	B-170-SB01	BF142279.D	02 May 2025 13:24		RC/JU	Ok
8	Q1901-10	B-167-SB02	BF142280.D	02 May 2025 13:53	Internal Standard Fail	RC/JU	ReRun
9	Q1901-11	B-170-SB02	BF142281.D	02 May 2025 14:22		RC/JU	Ok
10	Q1907-02	CO-8R-WC	BF142282.D	02 May 2025 14:51	Internal Standard and surrogates Fail	RC/JU	ReRun
11	Q1916-04	WC-12	BF142283.D	02 May 2025 16:20		RC/JU	Ok
12	Q1916-04MS	WC-12MS	BF142284.D	02 May 2025 16:49	Recovery is zero for compound #3	RC/JU	Not Ok
13	Q1916-04MSD	WC-12MSD	BF142285.D	02 May 2025 17:18	Recovery is zero for compound #3	RC/JU	Not Ok
14	Q1932-02	COMP-4	BF142286.D	02 May 2025 17:47		RC/JU	Ok
15	Q1932-04	COMP-5	BF142287.D	02 May 2025 18:16		RC/JU	Ok
16	Q1932-06	COMP-6	BF142288.D	02 May 2025 18:46	Surrogate Fail	RC/JU	ReRun
17	Q1932-08	COMP-7	BF142289.D	02 May 2025 19:15		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF050225

Review By	Review On
Supervise By	Supervise On
SubDirectory	BF050225
HP Acquire Method	BNA_F
HP Processing Method	BF043025
STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12662,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q1917-04	MH-JJ	BF142290.D	02 May 2025 19:44		RC/JU	Ok
19	Q1922-04	MH-R	BF142291.D	02 May 2025 20:13		RC/JU	Ok
20	Q1922-08	MH-S	BF142292.D	02 May 2025 20:42		RC/JU	Ok
21	Q1929-03	WC-A4-02-C	BF142293.D	02 May 2025 21:11		RC/JU	Ok

M : Manual Integration

SOP ID : M1311-TCLP-15

SDG No : N/A

Weight By : JP

Balance ID : WC SC-7

pH Meter ID : WC PH METER-1

Extraction By : JP

Filter By : JP

Pipette ID : WC

Tumbler ID : T-1 / T-2

TCLP Filter ID : 115525

Start Prep Date : 04/29/2025 Time : 17:20

End Prep Date : 04/30/2025 Time : 11:30

Combination Ratio : 20

ZHE Cleaning Batch : ¹⁰ N/A

Initial Room Temperature: 23 °C

Final Room Temperature: 22 °C

TCLP Technician Signature : *[Signature]*

Supervisor By : *12*

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP112795
HCL-TCLP,1N	N/A	WP112797
HNO3-TCLP,1N	N/A	WP112799
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1940,W1941,W1942	W3166,W1938,W1939,
1 Liter Amber	N/A	90924-08
120ml Plastic bottle	N/A	2738
1:1 HNO3	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-1 / T-2 checked,30 rpm. Particle size reduction is not required. q1894-01,02 and 03 we receive limited volume so no fluid determination. Q1901-08 IS USED FOR MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
04/30/25 13:00	<i>JP</i> Prep Room	<i>[Signature]</i> RJE-1
	Preparation Group	Analysis Group <i>[Signature]</i>

TCLP EXTRACTION LOGPAGE

PB167774

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB167774TB	LEB774	21	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-2
Q1894-01	PAINTED-GROUND-PIPE--1	01	25.00	500	N/A	N/A	N/A	5.5	1.0	T-1
Q1894-02	PAINTED-GROUND-PIPE-2	02	35.00	700	N/A	N/A	N/A	5.6	1.5	T-1
Q1894-03	PAINTED-GROUND-PIPE-3	03	34.00	680	N/A	N/A	N/A	5.6	1.5	T-1
Q1901-08	B-167-SB01	04	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1901-09	B-170-SB01	05	100.03	2000	N/A	N/A	N/A	7.0	1.5	T-1
Q1901-10	B-167-SB02	06	100.04	2000	N/A	N/A	N/A	5.8	1.0	T-1
Q1901-11	B-170-SB02	07	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q1904-02	VNJ-210	08	100.03	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1905-04	MH-G	09	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q1905-08	MH-H	10	100.01	2000	N/A	N/A	N/A	3.0	1.5	T-1
Q1906-04	WC-4	11	100.02	2000	N/A	N/A	N/A	5.5	1.0	T-2
Q1906-08	WC-5	12	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-2
Q1906-12	WC-6	13	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-2
Q1906-16	WC-7	14	100.03	2000	N/A	N/A	N/A	5.0	1.5	T-2
Q1907-02	CO-8R-WC	15	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-2
Q1912-04	MH-E	16	100.03	2000	N/A	N/A	N/A	4.5	1.0	T-2
Q1912-08	MH-F	17	100.02	2000	N/A	N/A	N/A	4.0	1.5	T-2
Q1913-01	WC-12-S-202504	18	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2
Q1913-03	WC-13-S-202504	19	100.03	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q1915-01	WC-04282025	20	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB167774TB	LEB774	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-01	PAINTED-GROUND-PIPE--1	N/A	N/A	N/A	N/A	100	N/A
Q1894-02	PAINTED-GROUND-PIPE-2	N/A	N/A	N/A	N/A	100	N/A
Q1894-03	PAINTED-GROUND-PIPE-3	N/A	N/A	N/A	N/A	100	N/A
Q1901-08	B-167-SB01	N/A	N/A	N/A	N/A	100	N/A
Q1901-09	B-170-SB01	N/A	N/A	N/A	N/A	100	N/A
Q1901-10	B-167-SB02	N/A	N/A	N/A	N/A	100	N/A
Q1901-11	B-170-SB02	N/A	N/A	N/A	N/A	100	N/A
Q1904-02	VNJ-210	N/A	N/A	N/A	N/A	100	N/A
Q1905-04	MH-G	N/A	N/A	N/A	N/A	100	N/A
Q1905-08	MH-H	N/A	N/A	N/A	N/A	100	N/A
Q1906-04	WC-4	N/A	N/A	N/A	N/A	100	N/A
Q1906-08	WC-5	N/A	N/A	N/A	N/A	100	N/A
Q1906-12	WC-6	N/A	N/A	N/A	N/A	100	N/A
Q1906-16	WC-7	N/A	N/A	N/A	N/A	100	N/A
Q1907-02	CO-8R-WC	N/A	N/A	N/A	N/A	100	N/A
Q1912-04	MH-E	N/A	N/A	N/A	N/A	100	N/A
Q1912-08	MH-F	N/A	N/A	N/A	N/A	100	N/A
Q1913-01	WC-12-S-202504	N/A	N/A	N/A	N/A	100	N/A
Q1913-03	WC-13-S-202504	N/A	N/A	N/A	N/A	100	N/A
Q1915-01	WC-04282025	N/A	N/A	N/A	N/A	100	N/A

Hot Block ID : WC S-1 / WC S-2

 Thermometer ID : FLASHPOINT

SampleID	ClientID	Sample Weight (g)	Volume DI Water (mL)	PH after 5 min stir	PH after 10 min stir	Extraction Fluid 1 or 2	pH Extraction Fluid
PB167774TB	LEB774	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-01	PAINTED-GROUND-PIPE--1	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-02	PAINTED-GROUND-PIPE-2	N/A	N/A	N/A	N/A	N/A	N/A
Q1894-03	PAINTED-GROUND-PIPE-3	N/A	N/A	N/A	N/A	N/A	N/A
Q1901-08	B-167-SB01	5.02	96.5	8.4	3.0	#1	4.93
Q1901-09	B-170-SB01	5.01	96.5	9.0	4.0	#1	4.93
Q1901-10	B-167-SB02	5.02	96.5	7.0	3.5	#1	4.93
Q1901-11	B-170-SB02	5.03	96.5	8.2	3.5	#1	4.93
Q1904-02	VNJ-210	5.04	96.5	8.4	3.0	#1	4.93
Q1905-04	MH-G	5.03	96.5	5.6	1.5	#1	4.93
Q1905-08	MH-H	5.02	96.5	6.0	2.0	#1	4.93
Q1906-04	WC-4	5.03	96.5	7.6	2.5	#1	4.93
Q1906-08	WC-5	5.02	96.5	7.0	2.5	#1	4.93
Q1906-12	WC-6	5.01	96.5	6.6	2.0	#1	4.93
Q1906-16	WC-7	5.02	96.5	6.6	2.0	#1	4.93
Q1907-02	CO-8R-WC	5.02	96.5	8.4	3.0	#1	4.93
Q1912-04	MH-E	5.01	96.5	6.6	2.0	#1	4.93
Q1912-08	MH-F	5.02	96.5	6.4	2.0	#1	4.93
Q1913-01	WC-12-S-202504	5.02	96.5	8.2	3.5	#1	4.93
Q1913-03	WC-13-S-202504	5.01	96.5	8.0	3.0	#1	4.93
Q1915-01	WC-04282025	5.02	96.5	8.4	3.5	#1	4.93

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TCLP Q1907 WorkList ID : 189188 Department : TCLP Extraction Date : 04-29-2025 07:58:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1913-01	WC-12-S-202504	Solid	TCLP Extraction	Cool 4 deg C	AECO02	L21	04/29/2025	1311
Q1913-03	WC-13-S-202504	Solid	TCLP Extraction	Cool 4 deg C	AECO02	L21	04/29/2025	1311
Q1915-01	WC-04282025	Solid	TCLP Extraction	Cool 4 deg C	CAMP02	L41	04/28/2025	1311
Q1901-08	B-167-SB01	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1901-09	B-170-SB01	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1901-10	B-167-SB02	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1901-11	B-170-SB02	Solid	TCLP Extraction	Cool 4 deg C	PORT06	L51	04/26/2025	1311
Q1894-01	PAINTED-GROUND-PIPE--1	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L31	04/28/2025	1311
Q1894-02	PAINTED-GROUND-PIPE-2	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L31	04/28/2025	1311
Q1894-03	PAINTED-GROUND-PIPE-3	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L31	04/28/2025	1311
Q1904-02	VNJ-210	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1905-04	MH-G	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/28/2025	1311
Q1905-08	MH-H	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/28/2025	1311
Q1906-04	WC-4	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1906-08	WC-5	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1906-12	WC-6	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1906-16	WC-7	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L41	04/28/2025	1311
Q1912-04	MH-E	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/29/2025	1311
Q1912-08	MH-F	Solid	TCLP Extraction	Cool 4 deg C	PSEG03	L51	04/29/2025	1311
Q1907-02	CO-8R-WC	Solid	TCLP Extraction	Cool 4 deg C	WALS01		04/28/2025	1311

Date/Time 04/29/25 17:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 04/29/25 19:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

SOP ID:	M3510C,3580A-Extraction SVOC-20		
Clean Up SOP #:	N/A	Extraction Start Date :	04/30/2025
Matrix :	Water	Extraction Start Time :	13:15
Weigh By:	N/A	Extraction By:	RJ
Balance check:	N/A	Filter By:	RJ
Balance ID:	N/A	pH Meter ID:	N/A
pH Strip Lot#:	E3880	Hood ID:	4,5,6,7
Extraction Method:	☒ Separatory Funnel	☐ Continuous Liquid/Liquid	☐ Sonication
		☐ Waste Dilution	☐ Soxhlet
		Extraction End Date :	04/30/2025
		Extraction End Time :	18:25
		Concentration By:	EH
		Supervisor By :	RUPESH

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	50/100 PPM	SP6752
Surrogate	1.0ML	100/150 PPM	SP6754
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3926
Baked Na2SO4	N/A	EP2607
10N NaOH	N/A	EP2559
H2SO4 1:1	N/A	EP2865
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5 ML Vial lot# 2210443. pH Adjusted<2 with 1:1 H2SO4 &>11 with 10 N NaOH.

KD Bath ID:	WATER BATH-1,2	Envap ID:	NEVAP-02
KD Bath Temperature:	60 °C	Envap Temperature:	40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
4/30/25	RS (Eut-06)	Rc/svr
18:30	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction SVOC-20

Concentration Date: 04/30/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167774TB	PB167774TB	TCLP BNA	100	6	RUPESH	ritesh	1			SEP-1
PB167805TB	PB167805TB	TCLP BNA	100	6	RUPESH	ritesh	1			2
PB167810BL	PB167810BL	TCLP BNA	1000	6	RUPESH	ritesh	1			3
PB167810BS	PB167810BS	TCLP BNA	1000	6	RUPESH	ritesh	1			4
Q1901-08	B-167-SB01	TCLP BNA	100	6	RUPESH	ritesh	1	A		5
Q1901-08MS	B-167-SB01MS	TCLP BNA	100	6	RUPESH	ritesh	1	A		6
Q1901-08MS D	B-167-SB01MSD	TCLP BNA	100	6	RUPESH	ritesh	1	A		7
Q1901-09	B-170-SB01	TCLP BNA	100	6	RUPESH	ritesh	1	A		8
Q1901-10	B-167-SB02	TCLP BNA	100	6	RUPESH	ritesh	1	A		9
Q1901-11	B-170-SB02	TCLP BNA	100	6	RUPESH	ritesh	1	A		10
Q1904-02	VNJ-210	TCLP BNA	100	6	RUPESH	ritesh	1	A		11
Q1905-04	MH-G	TCLP BNA	100	6	RUPESH	ritesh	1	A		12
Q1905-08	MH-H	TCLP BNA	100	6	RUPESH	ritesh	1	A		13
Q1906-04	WC-4	TCLP BNA	100	6	RUPESH	ritesh	1	A		14
Q1906-08	WC-5	TCLP BNA	100	6	RUPESH	ritesh	1	A		15
Q1906-12	WC-6	TCLP BNA	100	6	RUPESH	ritesh	1	A		16
Q1906-16	WC-7	TCLP BNA	100	6	RUPESH	ritesh	1	A		SEP-1
Q1907-02	CO-8R-WC	TCLP BNA	100	6	RUPESH	ritesh	1	A		2
Q1912-04	MH-E	TCLP BNA	100	6	RUPESH	ritesh	1	A		3
Q1912-08	MH-F	TCLP BNA	100	6	RUPESH	ritesh	1	A		4
Q1913-01	WC-12-S-202504	TCLP BNA	100	6	RUPESH	ritesh	1	A		5
Q1913-02	WC-12-A-202504	TCLP BNA	100	6	RUPESH	ritesh	1	A		6
Q1913-03	WC-13-S-202504	TCLP BNA	100	6	RUPESH	ritesh	1	A		7
Q1913-04	WC-13-A-202504	TCLP BNA	100	6	RUPESH	ritesh	1	A		8
Q1915-01	WC-04282025	TCLP BNA	100	6	RUPESH	ritesh	1	A		9

TCLP EXTRACTION LOGPAGE

PB167774

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prej Pos
PB167774TB	LEB774	21	N/A	2000	N/A	N/A	N/A	4.93	1.0	T-2
Q1894-01	PAINTED-GROUND-PIPE--1	01	25.00	500	N/A	N/A	N/A	5.5	1.0	T-1
Q1894-02	PAINTED-GROUND-PIPE-2	02	35.00	700	N/A	N/A	N/A	5.6	1.5	T-1
Q1894-03	PAINTED-GROUND-PIPE-3	03	34.00	680	N/A	N/A	N/A	5.6	1.5	T-1
Q1901-08	B-167-SB01	04	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1901-09	B-170-SB01	05	100.03	2000	N/A	N/A	N/A	7.0	1.5	T-1
Q1901-10	B-167-SB02	06	100.04	2000	N/A	N/A	N/A	5.8	1.0	T-1
Q1901-11	B-170-SB02	07	100.02	2000	N/A	N/A	N/A	5.6	1.5	T-1
Q1904-02	VNJ-210	08	100.03	2000	N/A	N/A	N/A	6.0	1.0	T-1
Q1905-04	MH-G	09	100.02	2000	N/A	N/A	N/A	3.0	1.0	T-1
Q1905-08	MH-H	10	100.01	2000	N/A	N/A	N/A	3.0	1.5	T-1
Q1906-04	WC-4	11	100.02	2000	N/A	N/A	N/A	5.5	1.0	T-2
Q1906-08	WC-5	12	100.03	2000	N/A	N/A	N/A	5.6	1.5	T-2
Q1906-12	WC-6	13	100.02	2000	N/A	N/A	N/A	5.0	1.0	T-2
Q1906-16	WC-7	14	100.03	2000	N/A	N/A	N/A	5.0	1.5	T-2
Q1907-02	CO-8R-WC	15	100.02	2000	N/A	N/A	N/A	5.8	1.5	T-2
Q1912-04	MH-E	16	100.03	2000	N/A	N/A	N/A	4.5	1.0	T-2
Q1912-08	MH-F	17	100.02	2000	N/A	N/A	N/A	4.0	1.5	T-2
Q1913-01	WC-12-S-202504	18	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2
Q1913-03	WC-13-S-202504	19	100.03	2000	N/A	N/A	N/A	6.2	1.5	T-2
Q1915-01	WC-04282025	20	100.02	2000	N/A	N/A	N/A	6.0	1.0	T-2

04/30/25
13 LOC

TCLP EXTRACTION LOGPAGE

PB167805

Sample ID	ClientID	TCLP Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB167805TB	LEB805	N/A	N/A	N/A	N/A	N/A	N/A	4.93	1.0	N/A
Q1913-02	WC-12-A-202504	N/A	N/A	N/A	N/A	N/A	N/A	12.0	1.5	N/A
Q1913-04	WC-13-A-202504	N/A	N/A	N/A	N/A	N/A	N/A	12.0	1.5	N/A

04/30/25
131-00



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: CSM Smith
ADDRESS: 110 Fieldcrest Ave #8 6th Flr
CITY Edison STATE: NJ ZIP: 08837
ATTENTION: M Encinas
PHONE: 732 590 4697 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: UTEN VST
PROJECT NO.: LOCATION: MT Vernon NJ
PROJECT MANAGER: M. Encinas
e-mail: Encinas M@cdm.com
PHONE: 732 590 4697 FAX:

CLIENT BILLING INFORMATION

BILL TO: CSM Smith PO#:
ADDRESS: 110 Fieldcrest Ave #8 6th Flr
CITY Edison STATE: NJ ZIP: 08837
ATTENTION: M. Encinas PHONE: 732 590 4697

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT

1. TCLP VOC
2. TCLP SVOC
3. TCLP metals
4. liquids
5. compositing
6. leachability
7. leachability
8. leachability
9. leachability

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS
			COMP	GRAB	DATE	TIME		E	E	E	E	E	E				
1.	<u>WC 04282025</u>	<u>Soil</u>	☒	<u>only</u>	<u>4/28/25</u>	<u>1415</u>	<u>4</u>	☒	☒	☒	☒	☒	☒				
2.																	
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP
1. <u>[Signature]</u>	<u>4/24/25 1345</u>	<u>[Signature]</u>	<u>2.5</u> °C
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Comments: <u>Temp 2.5°C Adjustment factor + 1 IR Gun #1</u>
2. <u>[Signature]</u>			
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3. <u>[Signature]</u>	<u>4-29-25</u>		

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete

☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488