

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915
 Lab File ID: VX045985.D Date Analyzed: 04/30/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:01
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	83019	5.55	141256	6.75	122754	10.05
UPPER LIMIT	166038	6.049	282512	7.25	245508	10.549
LOWER LIMIT	41509.5	5.049	70628	6.25	61377	9.549
EPA SAMPLE NO.						
WC-04282025	56553	5.55	111268	6.76	107559	10.06
VX0430WBL01	62274	5.54	125239	6.76	116463	10.05
VX0430WBS01	79591	5.54	138531	6.76	123154	10.05
VX0430WBSD01	77091	5.54	132529	6.76	119517	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.