

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q1915 SAS No.: Q1915 SDG NO.: Q1915
 Lab File ID: VX045985.D Date Analyzed: 04/30/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:01
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	60558	12.018				
UPPER LIMIT	121116	12.518				
LOWER LIMIT	30279	11.518				
EPA SAMPLE NO.						
WC-04282025	44562	12.02				
VX0430WBL01	49781	12.02				
VX0430WBS01	58135	12.02				
VX0430WBSD01	57970	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.