

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR16
 Lab Code: CHEM Case No.: Q1927 SAS No.: Q1927 SDG No.: Q1927
 Instrument ID: MSVOA_L Calibration Date(s): 05/01/2025 05/01/2025
 Heated Purge: (Y/N) N Calibration Time(s): 09:03 14:10
 GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042458.D		RRF002 = VL042459.D		RRF001 = VL042460.D		
		RRF0.5 = VL042461.D		RRF0.1 = VL042462.D		RRF015 = VL042465.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF015	RRF	% RSD
Dichlorodifluoromethane	1.470	1.629	1.650	1.887		1.639	1.655	9
Chloromethane	0.624	0.734	0.696	0.738		0.666	0.692	6.9
Vinyl Chloride	0.567	0.623	0.625	0.708	0.964	0.631	0.696	19.2
Bromomethane	0.286	0.327	0.359	0.349		0.324	0.329	8.6
Chloroethane	0.231	0.271	0.272	0.250		0.261	0.257	6.5
Tetrahydrofuran	0.326	0.387	0.354	0.329		0.322	0.344	8
Trichlorofluoromethane	1.376	1.528	1.489	1.660		1.582	1.527	7
1,1,2-Trichlorotrifluoroethane	1.099	1.245	1.234	1.318		1.201	1.219	6.5
Dichlorotetrafluoroethane	1.235	1.402	1.371	1.445		1.368	1.364	5.8
tert-Butyl alcohol	1.489	1.620	1.615	1.816		1.502	1.608	8.1
Heptane	1.445	1.593	1.634	1.276		1.474	1.484	9.5
1,1-Dichloroethene	0.495	0.567	0.587	0.650		0.547	0.569	10
Acetone	1.318	2.022	2.094	2.229		1.465	1.826	22.3
Carbon Disulfide	1.418	1.499	1.446	1.509		1.558	1.486	3.7
Methyl tert-Butyl Ether	0.671	0.759	0.853	0.802		0.740	0.765	8.9
Methylene Chloride	0.449	0.518	0.547	0.753		0.477	0.549	21.9
trans-1,2-Dichloroethene	0.552	0.645	0.742	0.741		0.644	0.665	12
1,1-Dichloroethane	1.196	1.368	1.465	1.432		1.322	1.357	7.8
Cyclohexane	1.045	1.096	0.998	1.135		1.048	1.065	4.9
2-Butanone	0.659	0.774	0.732	0.802		0.650	0.723	9.3
Carbon Tetrachloride	0.827	0.875	0.935	0.856	0.799	0.902	0.908	13.4
cis-1,2-Dichloroethene	1.342	1.545	1.605	1.551		1.386	1.486	7.7
Chloroform	2.198	2.399	2.482	2.551		2.301	2.386	5.9
1,1,1-Trichloroethane	2.288	2.511	2.501	2.567	2.842	2.433	2.589	9.3
2,2,4-Trimethylpentane	1.437	1.633	1.590	1.490		1.498	1.530	5.2
Benzene	0.892	1.036	0.974	0.924		0.892	0.944	6.5
1,2-Dichloroethane	0.666	0.769	0.755	0.738		0.709	0.727	5.7
Trichloroethene	0.347	0.404	0.364	0.389	0.419	0.355	0.360	16.3
1,2-Dichloropropane	0.317	0.362	0.405	0.368		0.315	0.353	10.7
Bromodichloromethane	0.839	0.931	0.921	0.848		0.904	0.889	4.8

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR16
 Lab Code: CHEM Case No.: Q1927 SAS No.: Q1927 SDG No.: Q1927
 Instrument ID: MSVOA_L Calibration Date(s): 05/01/2025 05/01/2025
 Heated Purge: (Y/N) N Calibration Time(s): 09:03 14:10
 GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:		RRF010 = VL042458.D		RRF002 = VL042459.D		RRF001 = VL042460.D			
		RRF0.5 = VL042461.D		RRF0.1 = VL042462.D		RRF015 = VL042465.D			
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF015	RRF	% RSD	
4-Methyl-2-Pentanone	0.870	0.985	0.980	0.852		0.887	0.915	6.9	
Toluene	1.016	1.025	0.969	0.935		1.050	0.999	4.6	
t-1,3-Dichloropropene	0.480	0.460	0.423	0.438		0.510	0.462	7.5	
cis-1,3-Dichloropropene	0.555	0.549	0.539	0.526		0.577	0.549	3.5	
1,1,2-Trichloroethane	0.357	0.428	0.409	0.372		0.371	0.387	7.7	
Dibromochloromethane	0.630	0.673	0.643	0.651		0.674	0.654	3	
1,2-Dibromoethane	0.554	0.610	0.599	0.589	0.625	0.577	0.592	4.3	
Tetrachloroethene	0.305	0.338	0.339	0.325	0.380	0.318	0.359	19.5	
Chlorobenzene	0.904	1.036	1.060	1.101		0.923	1.005	8.6	
Ethyl Benzene	1.725	1.752	1.634	1.492		1.754	1.671	6.7	
m/p-Xylene	1.429	1.506	1.378	1.212		1.479	1.401	8.3	
o-Xylene	1.423	1.519	1.393	1.197		1.481	1.403	8.9	
Styrene	0.583	0.538	0.500	0.415		0.582	0.523	13.3	
Bromoform	0.527	0.573	0.535	0.540		0.568	0.549	3.7	
1,1,2,2-Tetrachloroethane	0.846	0.990	0.985	0.976	0.945	0.887	0.967	9.7	
2-Chlorotoluene	1.597	1.697	1.569	1.405		1.667	1.587	7.2	
1,3,5-Trimethylbenzene	1.475	1.618	1.467	1.177		1.535	1.455	11.4	
1,2,4-Trimethylbenzene	1.646	2.075	1.788	1.568		1.723	1.760	11.1	
1,3-Dichlorobenzene	0.893	1.056	1.022	0.975		0.952	0.980	6.4	
1,4-Dichlorobenzene	0.907	0.966	0.945	0.869		0.949	0.927	4.2	
1,2-Dichlorobenzene	0.864	1.033	1.007	0.977		0.914	0.959	7.2	
1,2,4-Trichlorobenzene	0.778	0.947	0.798	0.599		0.828	0.790	15.8	
Hexachloro-1,3-Butadiene	0.772	1.070	1.081	1.028		0.838	0.958	14.9	
1,3-Butadiene	0.681	0.836	0.874	0.883		0.751	0.805	10.8	
Naphthalene	1.527	2.155	1.657	1.168	0.720	1.661	1.482	33	
4-Ethyltoluene	1.722	1.828	1.647	1.281		1.792	1.654	13.3	
1-Bromo-4-Fluorobenzene	0.951	0.915	0.928	0.902	0.911	0.963	0.927	2.4	
Hexane	1.215	1.345	1.290	1.278		1.239	1.273	3.9	
Allyl Chloride	1.076	1.164	1.181	1.283		1.166	1.174	6.3	
1,4-Dioxane	150.544	172.245	185.047	175.632		143.678	165.429	10.6	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR16
 Lab Code: CHEM Case No.: Q1927 SAS No.: Q1927 SDG No.: Q1927
 Instrument ID: MSVOA_L Calibration Date(s): 05/01/2025 05/01/2025
 Heated Purge: (Y/N) N Calibration Time(s): 09:03 14:10
 GC Column: RTX-1 ID: 0.32 (mm)

LAB FILE ID:								
RRF010 = VL042458.D			RRF002 = VL042459.D			RRF001 = VL042460.D		
RRF0.5 = VL042461.D			RRF0.1 = VL042462.D			RRF015 = VL042465.D		
COMPOUND	RRF010	RRF002	RRF001	RRF0.5	RRF0.1	RRF015	RRF	% RSD
Methyl Methacrylate	0.363	0.362	0.341	0.328		0.366	0.352	4.7

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.