

Duplicate Sample Summary

Lab Sample Id :	Q1927-01DUP	Q1927-01
Client Id :	EBR-P06-SC-IA01-202504	EBR-P06-SC-IA01-202504
DF :	1	1
Datafile :	VL042477.D	VL042476.D
Anal Date & Time :	05/05/2025 15:53	05/05/2025 15:18

Parameter	Result	Result	RPD
1,1,1-Trichloroethane	0	0	0
1,1,2,2-Tetrachloroethane	0	0	0
1,1,2-Trichloroethane	0	0	0
1,1,2-Trichlorotrifluoroethane	0	0	0
1,1-Dichloroethane	0	0	0
1,1-Dichloroethene	0	0	0
1,2,4-Trichlorobenzene	0	0	0
1,2,4-Trimethylbenzene	0	0	0
1,2-Dibromoethane	0	0	0
1,2-Dichlorobenzene	0	0	0
1,2-Dichloroethane	0	0	0
1,2-Dichloropropane	0	0	0
1,3,5-Trimethylbenzene	0	0	0
1,3-Butadiene	0	0	0
1,3-Dichlorobenzene	0	0	0
1,4-Dichlorobenzene	0	0	0
1,4-Dioxane	0	0	0
2,2,4-Trimethylpentane	0	0	0
2-Butanone	0.59	0.61	3.3
2-Chlorotoluene	0	0	0
4-Ethyltoluene	0	0	0
4-Methyl-2-Pentanone	0	0	0
Acetone	12.7	12.8	0.78
Allyl Chloride	0	0	0
Benzene	0	0	0
Bromodichloromethane	0	0	0
Bromoform	0	0	0
Bromomethane	0	0	0
Carbon Disulfide	0	0	0
Carbon Tetrachloride	0.07	0.08	13.3

Duplicate Sample Summary

Lab Sample Id :	Q1927-01DUP	Q1927-01
Client Id :	EBR-P06-SC-IA01-202504	EBR-P06-SC-IA01-2025C
DF :	1	1
Datafile :	VL042477.D	VL042476.D
Anal Date & Time :	05/05/2025 15:53	05/05/2025 15:18

Parameter	Result	Result	RPD
Chlorobenzene	0	0	0
Chloroethane	0	0	0
Chloroform	0	0	0
Chloromethane	0.45	0.59	26.9 *
cis-1,2-Dichloroethene	0	0	0
cis-1,3-Dichloropropene	0	0	0
Cyclohexane	0	0	0
Dibromochloromethane	0	0	0
Dichlorodifluoromethane	0.47	0.47	0
Dichlorotetrafluoroethane	0	0	0
Ethyl Benzene	0.14	0.14	0
Heptane	0	0	0
Hexachloro-1,3-Butadiene	0	0	0
Hexane	0.25	0.24	4.1
m/p-Xylene	0.42	0.42	0
Methyl Methacrylate	0	0	0
Methyl tert-Butyl Ether	0	0	0
Methylene Chloride	1.7	1.3	26.7 *
Naphthalene	0	0	0
o-Xylene	0.16	0.16	0
Styrene	0.14	0.15	6.9
t-1,3-Dichloropropene	0	0	0
tert-Butyl alcohol	0	0	0
Tetrachloroethene	0.08	0.07	13.3
Tetrahydrofuran	0	0	0
Toluene	0.4	0.36	10.5
trans-1,2-Dichloroethene	0	0	0
Trichloroethene	0	0	0
Trichlorofluoromethane	0.27	0.25	7.7
Vinyl Chloride	0	0	0